



INSTRUMENT PROCESSING SHEET

Agency Florida Department of Law EnforcementS/N 80-000229Florida Department of
Law EnforcementDate In N/ADI Completion Date 03/11/2025☒ Ship☐ P/U☐ H/D☐ CMI☐ EE

Intake	By TDG	Date	Quality Checks	By TDG	Date	Flow Calibration	By	Date																														
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: <u>Evidential Loaner.</u>		<u>03/11/2025</u>	☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>147</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP101</u> 32 mm <u>0.156</u> (.139 - .169) 36 mm <u>0.167</u> (.156 - .190) 53 mm <u>0.238</u> (.228 - .278) 103 mm <u>0.507</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>26932</u> ☒ Stability Checks		<u>03/11/2025</u>	Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																
			<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6286</td><td>202303K 03/29/2025</td></tr><tr><td>0.080</td><td>MP6287</td><td>202303L 03/29/2025</td></tr><tr><td>0.200</td><td>MP6288</td><td>202304C 04/05/2025</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG429602 10/22/2026</td></tr></tbody></table>	Simulator	Serial #	Lot #/Exp	0.050	MP6286	202303K 03/29/2025	0.080	MP6287	202303L 03/29/2025	0.200	MP6288	202304C 04/05/2025	0.080 DGS	N/A	AG429602 10/22/2026																				
Simulator	Serial #	Lot #/Exp																																				
0.050	MP6286	202303K 03/29/2025																																				
0.080	MP6287	202303L 03/29/2025																																				
0.200	MP6288	202304C 04/05/2025																																				
0.080 DGS	N/A	AG429602 10/22/2026																																				
			<table border="1"><thead><tr><th colspan="2">Maintenance By</th><th>Date</th></tr></thead><tbody><tr><td colspan="2">☐ Battery Replacement</td><td></td></tr><tr><td colspan="2">☐ Dry Gas Regulator Replacement</td><td></td></tr><tr><td colspan="2">☐ Breath Tube Replacement</td><td></td></tr><tr><td colspan="2">☐ Other _____</td><td></td></tr></tbody></table>						Maintenance By		Date	☐ Battery Replacement			☐ Dry Gas Regulator Replacement			☐ Breath Tube Replacement			☐ Other _____																	
Maintenance By		Date																																				
☐ Battery Replacement																																						
☐ Dry Gas Regulator Replacement																																						
☐ Breath Tube Replacement																																						
☐ Other _____																																						
Calibration Adjustment			Department Inspection																																			
By _____			By <u>TDG</u>																																			
Barometric Pressure Gauge _____ ID # _____			Barometric Pressure ID# <u>26932</u>																																			
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>			Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Gauge <u>1017</u> Instrument <u>1015</u> Mouth Alcohol Solution Lot # <u>2024-A</u> Acetone Stock Solution Lot # <u>2023-B</u>							
Simulator	Serial #	Lot #	Expiration																																			
0.000		N/A	N/A																																			
0.040																																						
0.100																																						
0.200																																						
0.300																																						
0.080 DGS	N/A																																					
☐ Post Calibration Adjustment Stability Checks			<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6284</td></tr><tr><td>Interferent</td><td>MP6285</td></tr><tr><td>0.050</td><td>MP6286</td></tr><tr><td>0.080</td><td>MP6287</td></tr><tr><td>0.200</td><td>MP6288</td></tr></tbody></table>						Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288																		
Simulator	Serial Number																																					
0.000	MP6284																																					
Interferent	MP6285																																					
0.050	MP6286																																					
0.080	MP6287																																					
0.200	MP6288																																					
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			<table border="1"><thead><tr><th colspan="2">Attachments</th></tr></thead><tbody><tr><td>☒ Form 41</td><td>☐ Post-Stability Checks</td></tr><tr><td>☒ Stability Checks</td><td>☐ Flow Calibration</td></tr><tr><td>☒ Calibration Certificate</td><td>☒ Form 40</td></tr><tr><td>☐ Calibration Adjustment</td><td>☒ Other <u>Extra Stabilities</u></td></tr></tbody></table>						Attachments		☒ Form 41	☐ Post-Stability Checks	☒ Stability Checks	☐ Flow Calibration	☒ Calibration Certificate	☒ Form 40	☐ Calibration Adjustment	☒ Other <u>Extra Stabilities</u>
Simulator	Serial #	Lot #	Expiration																																			
0.050																																						
0.080																																						
0.200																																						
0.080 DGS	N/A																																					
Attachments																																						
☒ Form 41	☐ Post-Stability Checks																																					
☒ Stability Checks	☐ Flow Calibration																																					
☒ Calibration Certificate	☒ Form 40																																					
☐ Calibration Adjustment	☒ Other <u>Extra Stabilities</u>																																					
Notes/Suggested Service: <u>Performed extra 0.20 Stabilities after an Ambient Fail message was received during that part of the inspection. The extra stabilities passed with no further Ambient Fail messages generated. (TDG)</u>			☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																																			
<u>Loaner was requested by Everglades National Park and will be sent there after completion of paperwork. (TDG)</u>			Digitally signed by Shayla Platt Destinee Armstrong Date: 2025.03.12 12:13:25 -0400 Date: 2025.03.13 10:01:09 -0400																																			
			Tech Review / Date _____ Admin Review / Date _____																																			

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FDLE

Time of Inspection: 10:48

Date of Inspection: 03/11/2025

Serial Number: 80-000229

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____

Remarks:

AI NOT CONDUCTED. BYPASSED TO BRING OUT OF DISABLED MODE.

Not determined *3/11/25*
ML

The above instrument complies (☒) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Taylor D Gutschow

TAYLOR D GUTSCHOW

Signature and Printed Name

03/11/2025
Date

FDLE

Intoxilyzer - Alcohol Analyzer

Model 8000

SN 80-000229

03/11/2025

Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:00
Control Test	0.049	11:00
Air Blank	0.000	11:01
Control Test	0.048	11:02
Air Blank	0.000	11:02
Control Test	0.048	11:03
Air Blank	0.000	11:03
Control Test Stats		
Average	0.0483	
Std Dev	0.0006	
Rel Std Dev(%)	1.1945	

ML

Operator's Signature

FDLE

Intoxilyzer - Alcohol Analyzer

Model 8000

SN 80-000229

03/11/2025

Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:07
Control Test	0.079	11:08
Air Blank	0.000	11:08
Control Test	0.078	11:09
Air Blank	0.000	11:09
Control Test	0.078	11:10
Air Blank	0.000	11:11
Control Test Stats		
Average	0.0783	
Std Dev	0.0006	
Rel Std Dev(%)	0.7370	

Wet

ML

Operator's Signature

FDLE
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000229
03/11/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:22
Control Test	0.196	11:23
Air Blank	0.000	11:23
Control Test	0.197	11:24
Air Blank	0.000	11:24
Control Test	0.195	11:25
Air Blank	0.000	11:26
Control Test Stats		
Average	0.1960	
Std Dev	0.0010	
Rel Std Dev(%)	0.5102	

MC

Operator's Signature

FDLE

Intoxilyzer - Alcohol Analyzer

Model 8000

SN 80-000229

03/11/2025

Software: 8100.27

DLG

Test	g/210L	Time
Air Blank	0.000	10:51
Control Test	0.078	10:52
Air Blank	0.000	10:52
Control Test	0.078	10:53
Air Blank	0.000	10:53
Control Test	0.080	10:53
Air Blank	0.000	10:54
Control Test Stats		
Average	0.0787	
Std Dev	0.0012	
Rel Std Dev(%)	1.4678	

ML

Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FDLE

Time of Inspection: 13:50

Date of Inspection: 03/11/2025

Serial Number: 80-000229

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG429602 Exp: 10/22/2026
0.000	0.048	0.078	0.196 / 0.197	0.078
0.000	0.048	0.077	0.194 / 0.198	0.079
0.000	0.048	0.077	/ 0.197	0.079
0.000	0.049	0.077	/ 0.199	0.079
0.000	0.049	0.078	/ 0.199	0.079
0.000	0.049	0.078	/ 0.200	0.079
0.000	0.049	0.078	/ 0.200	0.079
0.000	0.049	0.078	/ 0.199	0.079
0.000	0.049	0.079	/ 0.199	0.079
0.000	0.049	0.078	/ 0.199	0.079

Standard Deviations	0.0004	0.0006	/ 0.0010	0.0003
---------------------	--------	--------	----------	--------


Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

20: Ambient Fail. AIRED OUT ROOM AND REPEATED.

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

03/11/2025
Date

Extra Stabilities (0.20)

FOLE
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000229
03/11/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:22
Control Test	0.196	14:22
Air Blank	0.000	14:23
Control Test	0.196	14:24
Air Blank	0.000	14:24
Control Test	0.196	14:25
Air Blank	0.000	14:26
Control Test	0.196	14:26
Air Blank	0.000	14:27
Control Test	0.196	14:28
Air Blank	0.000	14:28
Control Test	0.195	14:29
Air Blank	0.000	14:29
Control Test	0.196	14:30
Air Blank	0.000	14:31
Control Test	0.195	14:31
Air Blank	0.000	14:32
Control Test	0.195	14:33
Air Blank	0.000	14:33
Control Test	0.195	14:34
Air Blank	0.000	14:34
Control Test Stats		
Average	0.1956	
Std Dev	0.0005	
Rel Std Dev(%)	0.2640	


Operator's Signature



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000229, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000229</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>FDLE</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>03/11/2025</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>13:50</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.



03/11/2025

Date

TAYLOR D GUTSCHOW,
Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality