



INSTRUMENT PROCESSING SHEET

Agency ALACHUA COUNTY SOS/N 80-000737Florida Department of
Law EnforcementDate In 06/23/2025 DI Completion Date N/A☐ Ship ☐ P/U ☐ H/D ☒ CMI ☐ EE

Intake	By <u>WKP</u>	Date <u>06/23/2025</u>	Quality Checks	By <u>KTS</u>	Date <u>6/24/25</u>	Flow Calibration	By _____	Date _____																											
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: Two bottles of mouth alcohol solution and acetone were sent inside of the box. WKP No leakage observed of these bottles inside the box, but will discard when shipping to CMI. SLH 7/1/2025			<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>191</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP105 / 103*</u> 32 mm <u>0.000 / 0.000*</u> (.139 - .169) 36 mm <u>0.000 / 0.000*</u> (.156 - .190) 53 mm <u>0.066</u> (.228 - .278) 103 mm <u>0.242</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28662</u> <input checked="" type="checkbox"/> Stability Checks			Flow Column # _____ <input type="checkbox"/> 5L/min – 17mm <input type="checkbox"/> 15L/min – 53mm <input type="checkbox"/> 30L/min – 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																													
			<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP5088</td><td>202406K 6/19/26</td></tr><tr><td>0.080</td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG429602 * 10/22/26 *</td></tr></tbody></table>		Simulator	Serial #	Lot #/Exp	0.050	MP5088	202406K 6/19/26	0.080			0.200			0.080 DGS	N/A	AG429602 * 10/22/26 *			Maintenance By _____ Date _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____													
Simulator	Serial #	Lot #/Exp																																	
0.050	MP5088	202406K 6/19/26																																	
0.080																																			
0.200																																			
0.080 DGS	N/A	AG429602 * 10/22/26 *																																	
Calibration Adjustment By _____			Department Inspection By _____																																
Barometric Pressure Gauge _____ ID # _____			Barometric Pressure ID# _____																																
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>			Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Gauge _____ Instrument _____ Mouth Alcohol Solution Lot # _____ Acetone Stock Solution Lot # _____				
Simulator	Serial #	Lot #	Expiration																																
0.000		N/A	N/A																																
0.040																																			
0.100																																			
0.200																																			
0.300																																			
0.080 DGS	N/A																																		
<input type="checkbox"/> Post Calibration Adjustment Stability Checks			<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td></td></tr><tr><td>Interferent</td><td></td></tr><tr><td>0.050</td><td></td></tr><tr><td>0.080</td><td></td></tr><tr><td>0.200</td><td></td></tr></tbody></table>					Simulator	Serial Number	0.000		Interferent		0.050		0.080		0.200																	
Simulator	Serial Number																																		
0.000																																			
Interferent																																			
0.050																																			
0.080																																			
0.200																																			
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Attachments <input type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks (from 2 dates) <input type="checkbox"/> Calibration Certificate <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Form 40 <input checked="" type="checkbox"/> Other <u>Form 51</u>												
Simulator	Serial #	Lot #	Expiration																																
0.050																																			
0.080																																			
0.200																																			
0.080 DGS	N/A																																		
Notes/Suggested Service: Root cause analyses (RCA) performed during flow verification. No user or equipment error determined. RCA performed on DGS. cleared area near instrument. Second DGS run and the 0.050g/210L read 0.000 with tight connections. KTS 6/24/25 *Evaluated flow with different flowmeter, but still zero flows at low pressures. Reassessed DGS stability which yielded similar results. Return back to CMI. SLH 7/1/25 Tech: CMI repair for zero flow and stability readings and exception messages occurring without an environmental reason. SLH 7/2/2025			<input type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input type="checkbox"/> Return to/Place into Evidentiary Use <input checked="" type="checkbox"/> Remain Out of Evidentiary Use <input type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use																																
Taylor Gutschow <small>Digitally signed by Taylor Gutschow Date: 2025.07.02 13:40:49 +04'00'</small>			Phil Nicodemo <small>Digitally signed by Phil Nicodemo Date: 2025.07.10 13:29:24 -04'00'</small>																																
Tech Review / Date _____			Admin Review / Date _____																																

ALACHUA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000737
06/24/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:34
Control Test	0.083	08:34
Air Blank	PUR*	08:35
Air Blank	PUR*	08:36

*Purge Fail



Operator's Signature

DGS

ALACHUA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000737
06/24/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:38
Control Test	0.000	08:38
Air Blank	0.000	08:38
Control Test	0.000	08:39
Air Blank	0.000	08:39
Control Test	0.000	08:39
Air Blank	0.000	08:40
Control Test Stats		
Average	0.0000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	




Operator's Signature

DGS

ALACHUA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000737
06/24/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:49
Control Test	0.000	08:49
Air Blank	0.000	08:50
Control Test	0.000	08:51
Air Blank	0.000	08:51
Control Test	0.000	08:52
Air Blank	0.000	08:52
Control Test Stats		
Average	0.0000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	



Operator's Signature

0.050

ALACHUA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000737
→ 07/01/2025
Software: 8100.27

#1

Test	g/210L	Time
Air Blank	0.000	14:52
Control Test	0.082	14:52
Air Blank	PUR*	14:53
Air Blank	PUR*	14:54

*Purge Fail


Operator's Signature

Same results as ran 6/24/25

Sub 7/1/25

Faint smell of cleaner in lab
due to janitorial service; used
fan in area to clear.


Sub 7/1/25

ALACHUA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000737
07/01/2025
Software: 8100.27

#2

Test	g/210L	Time
Air Blank	0.000	15:02
Control Test	INT*	15:02
Air Blank	0.000	15:03
Control Test	0.000	15:03
Air Blank	0.000	15:04
Control Test	0.000	15:04
Air Blank	0.000	15:04
Control Test Stats		
Average	0.0097	
Std Dev	0.0167	
Rel Std Dev(%)	173.2051	

*Interferent Detect


Operator's Signature

Used fan to ^{SUT}circulate/remove
residual cleaner smell. SUT
7/1/25


ALACHUA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000737
→ 07/01/2025
Software: 8100.27

#3

Test	g/210L	Time
Air Blank	0.000	15:15
Control Test	INT*	15:15
Air Blank	0.000	15:16
Control Test	0.000	15:17
Air Blank	0.000	15:17
Control Test	0.000	15:17
Air Blank	0.000	15:18
Control Test Stats		
Average	0.0097	
Std Dev	0.0167	
Rel Std Dev(%)	173.2051	

*Interferent Detect

- result
reading
on instru-
ment
0.029 g/210L
prior to
exception
sett
7/1/25


Operator's Signature

Post _{sett} V stability check _{sett}
use of fan

Exception not due to environment
in lab. _{sett} 7/1/25

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: ALACHUA COUNTY SO
Time of Inspection: 14:28

Date of Inspection: 07/01/2025

Serial Number: 80-000737
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08g/210L Test (g/210L) Lot#:_____ Exp:_____	0.20g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:_____ Exp:_____

Number of Simulators Used: _____

Remarks:

BYPASS AI FOR OPERATION, COMPLIANCE UNDETERMINED

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Mejla

LEANDRA HIGGINBOTHAM

Signature and Printed Name

07/01/2025
Date

Return Material Authorization

Ship to:

☒ CMI, Inc.

☐ Enforcement Electronics

Shipment to repair facility authorized by: Toni Fulton on 07/02/2025

Items Returned: Instrument ☒ Supplies ☐ Other ☐ Describe: _____

Instrument Model: Intoxilyzer 8000 Serial Number: 80-000737

Bill To Address:

Alachua County SO

Attn Toni Fulton

Ship to Address:

FDLE Tallahassee

Reason for Return:

Previously authorized 3/11/25 and was sent to CMI. It was 'repaired' and returned to FDLE for inspection, arriving 6/23/2025. The instrument has multiple issues with purge fail and interferent exceptions when all precautions performed. In addition, zero flow into the instrument at the lower levels tested (32 and 36mm). Stabilities with dry gas and simulators are not able to be performed due to exceptions or 0.000 result. SLH 7/2/2025

Please choose one of the following options:

- ☐ 1. I _____, authorize all repairs.
- ☐ 2. I _____, authorize repairs up to \$_____.
- ☒ 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Toni Fulton

Phone #: 352-538-7970 Email: tfulton@alachuasheriff.org

ATP Contact Name: LeAndra Higginbotham ATP Email: leandrahigginbotham@fdle.state.fl.us