



INSTRUMENT PROCESSING SHEET

Agency Haines City PDS/N 80-000796Florida Department of
Law EnforcementDate In 03/21/2025 DI Completion Date 04/02/2025☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake By <u>ALL</u> Date <u>03/21/2025</u>		Quality Checks By <u>DA</u> Date <u>03/31/2025</u>		Flow Calibration By _____ Date _____																													
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>203</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP103</u> 32 mm <u>0.148</u> (.139 - .169) 36 mm <u>0.164</u> (.156 - .190) 53 mm <u>0.234</u> (.228 - .278) 103 mm <u>0.511</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>30793</u> ☒ Stability Checks		☐ Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																													
		<table border="1" style="width:100%"><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr><tr><td>0.050</td><td>MP5088</td><td>202406K 06/19/2026</td></tr><tr><td>0.080</td><td>MP5089</td><td>202406L 06/19/2026</td></tr><tr><td>0.200</td><td>MP5090</td><td>202406N 06/20/2026</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG429602 10/22/2026</td></tr></table>		Simulator	Serial #	Lot #/Exp	0.050	MP5088	202406K 06/19/2026	0.080	MP5089	202406L 06/19/2026	0.200	MP5090	202406N 06/20/2026	0.080 DGS	N/A	AG429602 10/22/2026	Maintenance By <u>DA</u> Date <u>03/31/2025</u> ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☒ Other <u>Replaced internal printer paper prior to completing any quality checks DA 03/31/2025</u> _____ _____ _____														
Simulator	Serial #	Lot #/Exp																															
0.050	MP5088	202406K 06/19/2026																															
0.080	MP5089	202406L 06/19/2026																															
0.200	MP5090	202406N 06/20/2026																															
0.080 DGS	N/A	AG429602 10/22/2026																															
Calibration Adjustment By <u>DA</u>				Department Inspection By <u>DA</u>																													
Barometric Pressure Gauge <u>1014</u> ID # <u>28427</u>				Barometric Pressure ID# <u>28662</u>																													
<table border="1" style="width:100%"><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr><tr><td>0.000</td><td>MP6294</td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td>MP6295</td><td>24330</td><td>09/10/2026</td></tr><tr><td>0.100</td><td>MP6296</td><td>24390</td><td>10/29/2026</td></tr><tr><td>0.200</td><td>MP6297</td><td>24080</td><td>02/13/2026</td></tr><tr><td>0.300</td><td>MP6298</td><td>24430</td><td>12/10/2026</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>06723080A5</td><td>04/05/2025</td></tr></table>				Simulator	Serial #	Lot #	Expiration	0.000	MP6294	N/A	N/A	0.040	MP6295	24330	09/10/2026	0.100	MP6296	24390	10/29/2026	0.200	MP6297	24080	02/13/2026	0.300	MP6298	24430	12/10/2026	0.080 DGS	N/A	06723080A5	04/05/2025	Gauge <u>1014</u> Instrument <u>1013</u> Mouth Alcohol Solution Lot # <u>2024-A</u> Acetone Stock Solution Lot # <u>2024-B</u>	
Simulator	Serial #	Lot #	Expiration																														
0.000	MP6294	N/A	N/A																														
0.040	MP6295	24330	09/10/2026																														
0.100	MP6296	24390	10/29/2026																														
0.200	MP6297	24080	02/13/2026																														
0.300	MP6298	24430	12/10/2026																														
0.080 DGS	N/A	06723080A5	04/05/2025																														
☒ Post Calibration Adjustment Stability Checks				<table border="1" style="width:100%"><tr><th>Simulator</th><th>Serial Number</th></tr><tr><td>0.000</td><td>MP5086</td></tr><tr><td>Interferent</td><td>MP5087</td></tr><tr><td>0.050</td><td>MP6291</td></tr><tr><td>0.080</td><td>MP6292</td></tr><tr><td>0.200</td><td>MP6293</td></tr></table>		Simulator	Serial Number	0.000	MP5086	Interferent	MP5087	0.050	MP6291	0.080	MP6292	0.200	MP6293																
Simulator	Serial Number																																
0.000	MP5086																																
Interferent	MP5087																																
0.050	MP6291																																
0.080	MP6292																																
0.200	MP6293																																
<table border="1" style="width:100%"><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr><tr><td>0.050</td><td>MP6291</td><td>202406K</td><td>06/19/2026</td></tr><tr><td>0.080</td><td>MP6292</td><td>202406L</td><td>06/19/2026</td></tr><tr><td>0.200</td><td>MP6293</td><td>202406N</td><td>06/20/2026</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG429602</td><td>10/22/2026</td></tr></table>				Simulator	Serial #	Lot #	Expiration	0.050	MP6291	202406K	06/19/2026	0.080	MP6292	202406L	06/19/2026	0.200	MP6293	202406N	06/20/2026	0.080 DGS	N/A	AG429602	10/22/2026	Attachments ☒ Form 41 ☒ Post-Stability Checks ☒ Stability Checks ☐ Flow Calibration ☒ Calibration Certificate ☐ Form 40 ☒ Calibration Adjustment ☐ Other _____									
Simulator	Serial #	Lot #	Expiration																														
0.050	MP6291	202406K	06/19/2026																														
0.080	MP6292	202406L	06/19/2026																														
0.200	MP6293	202406N	06/20/2026																														
0.080 DGS	N/A	AG429602	10/22/2026																														
Notes/Suggested Service: <u>I performed root-cause analysis per the TSOP for out-of-range stability value. No user or equipment error was determined. DA 03/31/2025</u> _____ _____ _____ _____ _____				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																													
Phil Nicodemo Digitally signed by Phil Nicodemo Date: 2025.04.03 11:31:31 -0400				Shayla Platt Digitally signed by Shayla Platt Date: 2025.04.03 14:26:24 -0400																													
Tech Review / Date _____				Admin Review / Date _____																													

Stability Checks

80-000796

DA

3/31/25

HAINES CITY PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000796
03/31/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:32
Control Test	0.049	12:33
Air Blank	0.000	12:33
Control Test	0.047	12:34
Air Blank	0.000	12:35
Control Test	0.047	12:35
Air Blank	0.000	12:36
Control Test Stats		
Average	0.0473	
Std Dev	0.0006	
Rel Std Dev(%)	1.2198	

DA

Operator's Signature

HAINES CITY PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000796
03/31/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:24
Control Test	0.077	12:24
Air Blank	0.000	12:25
Control Test	0.077	12:26
Air Blank	0.000	12:26
Control Test	0.077	12:27
Air Blank	0.000	12:27
Control Test Stats		
Average	0.0770	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Wet

DA

Operator's Signature

HAINES CITY PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000796
03/31/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:11
Control Test	0.193	12:12
Air Blank	0.000	12:12
Control Test	0.194	12:13
Air Blank	0.000	12:13
Control Test	0.194	12:14
Air Blank	0.000	12:14
Control Test Stats		
Average	0.1937	
Std Dev	0.0006	
Rel Std Dev(%)	0.2981	

DA

Operator's Signature

HAINES CITY PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000796
03/31/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:17
Control Test	0.080	12:17
Air Blank	0.000	12:17
Control Test	0.079	12:18
Air Blank	0.000	12:18
Control Test	0.080	12:19
Air Blank	0.000	12:19
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

DGS

DA

Operator's Signature

HAINES CITY PD
Intoxilyzer - Alcohol Analyzer
Model: 8000 SN 60-000736
04/02/2025 12:27:00

Auto Calibration
Max Power Res Value = 13
Auto Range Res Value = 4

Sol Value = 0.000 g/210L ***
Fit Value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12543, Sum Io = 14028

***** CHANNEL 1 *****
Sample % Abs (% Abs Ref)
Sample #1 = 0.0480 (-0.0200)
Sample #2 = 0.0780 (-0.0310)
Sample #3 = 0.0810 (-0.0260)
Sample #4 = 0.0850 (-0.0180)
Avg % Abs = 0.0813 (-0.0130)
STD DEV = 0.0035 (0.0270)
REL STD DEV = 4.318 (207.407)

***** CHANNEL 2 *****
Sample % Abs (% Abs Ref)
Sample #1 = 0.1080 (-0.0090)
Sample #2 = 0.1050 (-0.0160)
Sample #3 = 0.1040 (-0.0150)
Sample #4 = 0.0920 (-0.0160)
Avg % Abs = 0.1003 (-0.0050)
STD DEV = 0.0072 (0.0182)
REL STD DEV = 7.210 (363.668)

Sol Value = 0.040 g/210L ***
Fit Value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12537, Sum Io = 14026

***** CHANNEL 1 *****
Sample % Abs (% Abs Ref)
Sample #1 = 0.7270 (-0.0180)
Sample #2 = 0.7360 (0.0050)
Sample #3 = 0.7340 (-0.0090)
Sample #4 = 0.7590 (0.0030)
Avg % Abs = 0.7430 (-0.0003)
STD DEV = 0.0139 (0.0076)
REL STD DEV = 1.870 (2271.564)

***** CHANNEL 2 *****
Sample % Abs (% Abs Ref)
Sample #1 = 1.5110 (0.0040)
Sample #2 = 1.5180 (0.0070)
Sample #3 = 1.5460 (-0.0070)
Sample #4 = 1.5350 (0.0110)
Avg % Abs = 1.5330 (0.0037)
STD DEV = 0.0141 (0.0095)
REL STD DEV = 0.920 (257.772)

Sol Value = 0.100 g/210L ***
Fit Value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12532, Sum Io = 14021
***** CHANNEL 1 *****
Sample % Abs (% Abs Ref)
Sample #1 = 1.7270 (-0.0280)
Sample #2 = 1.7060 (0.0010)
Sample #3 = 1.7370 (0.0040)
Sample #4 = 1.7400 (0.0120)
Avg % Abs = 1.7277 (0.0057)
STD DEV = 0.0188 (0.0057)
REL STD DEV = 1.090 (100.345)

***** CHANNEL 2 *****
Sample % Abs (% Abs Ref)
Sample #1 = 3.6030 (-0.0210)
Sample #2 = 3.6080 (-0.0140)
Sample #3 = 3.6130 (-0.0050)
Sample #4 = 3.6050 (-0.0070)
Avg % Abs = 3.6087 (-0.0087)
STD DEV = 0.0040 (0.0047)
REL STD DEV = 0.112 (54.523)

Sol Value = 0.200 g/210L ***
Fit Value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12528, Sum Io = 14021
***** CHANNEL 1 *****
Sample % Abs (% Abs Ref)
Sample #1 = 3.3510 (-0.0100)
Sample #2 = 3.3600 (-0.0240)
Sample #3 = 3.3540 (0.0060)
Sample #4 = 3.3580 (-0.0140)
Avg % Abs = 3.3573 (-0.0107)
STD DEV = 0.0031 (0.0153)
REL STD DEV = 0.091 (143.205)

***** CHANNEL 2 *****
Sample % Abs (% Abs Ref)
Sample #1 = 6.9390 (-0.0100)
Sample #2 = 6.9370 (-0.0080)
Sample #3 = 6.9250 (0.0070)
Sample #4 = 6.9130 (0.0050)
Avg % Abs = 6.9250 (0.0013)
STD DEV = 0.0120 (0.0081)
REL STD DEV = 0.173 (610.840)

Sol Value = 0.300 g/210L ***
Fit Value = 1.4286 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12529, Sum Io = 14020
***** CHANNEL 1 *****
Sample % Abs (% Abs Ref)
Sample #1 = 4.9390 (0.0030)
Sample #2 = 4.9270 (-0.0030)
Sample #3 = 4.9330 (0.0310)
Sample #4 = 4.9050 (0.0350)
Avg % Abs = 4.9217 (0.0210)
STD DEV = 0.0147 (0.0209)
REL STD DEV = 0.300 (99.431)

***** CHANNEL 2 *****
Sample % Abs (% Abs Ref)
Sample #1 = 10.0650 (0.0080)
Sample #2 = 10.0390 (0.0120)
Sample #3 = 10.0750 (0.0030)
Sample #4 = 10.0660 (0.0140)
Avg % Abs = 10.0600 (0.0097)
STD DEV = 0.0187 (0.0059)
REL STD DEV = 0.186 (60.615)

***** AUTO CAL DATA *****
***** CHANNEL 1 *****
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.081
Std Dev = 0.00 Rel Std Dev = 4.32
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.743
Std Dev = 0.01 Rel Std Dev = 1.87
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.728
Std Dev = 0.02 Rel Std Dev = 1.09
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.357
Std Dev = 0.00 Rel Std Dev = 0.09
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 4.922
Std Dev = 0.01 Rel Std Dev = 0.30
Zero Order Coef = -218.99
First Order Coef = 2831.55
Second Order Coef = 23.12
Standard Deviation = 17.632050

***** CHANNEL 2 *****
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.100
Std Dev = 0.01 Rel Std Dev = 7.21
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.533
Std Dev = 0.01 Rel Std Dev = 0.92
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.609
Std Dev = 0.00 Rel Std Dev = 0.11
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.925
Std Dev = 0.01 Rel Std Dev = 0.17
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 10.060
Std Dev = 0.02 Rel Std Dev = 0.19
Zero Order Coef = -131.27
First Order Coef = 1310.69
Second Order Coef = 12.15
Standard Deviation = 3.517509

Solution Stats Quadratic Fit Chan 2		
% Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.000	-0.0000
0.040	0.040	-0.0000
0.100	0.100	0.0001
0.200	0.200	-0.0001
0.300	0.300	0.0000

Sol Value = 0.080 g/210L ***
Fit Value = 0.3810 mg/l %%%
Samples Taken = 4, Discarded = 1

***** CHANNEL 1 *****
Sample #1 = 3279.00
Sample #2 = 3294.00
Sample #3 = 3209.00
Sample #4 = 3247.00
Average Result = 3250.0000
STD DEV = 42.5793
REL STD DEV = 1.310

***** CHANNEL 2 *****
Sample #1 = 3378.00
Sample #2 = 3348.00
Sample #3 = 3347.00
Sample #4 = 3356.00
Average Result = 3353.6667
STD DEV = 10.6927
REL STD DEV = 0.319

Dry Gas H2O Adjust Results *****
Barometric Pressure = 1013
3 um H2O Adjust (mg/l*10,000) = 559
9 um H2O Adjust (mg/l*10,000) = 456
**** AUTO CAL PASS

Solution Stats Quadratic Fit Chan 1		
% Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.000	-0.0002
0.040	0.040	0.0002
0.100	0.100	0.0004
0.200	0.200	-0.0005
0.300	0.300	0.0002

Post-Stability Checks 80-000796 DA 4/2/25

HAINES CITY PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000796
04/02/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:16
Control Test	0.049	13:17
Air Blank	0.000	13:17
Control Test	0.049	13:18
Air Blank	0.000	13:18
Control Test	0.050	13:19
Air Blank	0.000	13:20
Control Test Stats		
Average	0.0493	
Std Dev	0.0016	
Rel Std Dev(%)	1.1703	

DA

Operator's Signature

HAINES CITY PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000796
04/02/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:25
Control Test	0.079	13:26
Air Blank	0.000	13:26
Control Test	0.079	13:27
Air Blank	0.000	13:27
Control Test	0.079	13:28
Air Blank	0.000	13:29
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

wet

DA

Operator's Signature

HAINES CITY PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000796
04/02/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:21
Control Test	0.198	13:21
Air Blank	0.000	13:22
Control Test	0.199	13:23
Air Blank	0.000	13:23
Control Test	0.199	13:24
Air Blank	0.000	13:24
Control Test Stats		
Average	0.1967	
Std Dev	0.0016	
Rel Std Dev(%)	0.2906	

DA

Operator's Signature

HAINES CITY PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000796
04/02/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:30
Control Test	0.080	13:30
Air Blank	0.000	13:31
Control Test	0.077	13:31
Air Blank	0.000	13:32
Control Test	0.077	13:32
Air Blank	0.000	13:32
Control Test Stats		
Average	0.0780	
Std Dev	0.0017	
Rel Std Dev(%)	2.2206	

DGS

DA

Operator's Signature

Florida Department of Law Enforcement

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: HAINES CITY PD

Serial Number: 80-000796

Time of Inspection: 16:19

Date of Inspection: 04/02/2025

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#: 202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#: 202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: AG429602 Exp: 10/22/2026
0.000	0.050	0.079	0.199	0.079
0.000	0.049	0.079	0.199	0.079
0.000	0.050	0.079	0.199	0.079
0.000	0.049	0.079	0.199	0.079
0.000	0.049	0.080	0.200	0.079
0.000	0.050	0.080	0.199	0.079
0.000	0.049	0.079	0.200	0.079
0.000	0.050	0.080	0.199	0.078
0.000	0.049	0.079	0.199	0.079
0.000	0.050	0.079	0.199	0.078

Standard Deviations	0.0005	0.0004	0.0004	0.0004
---------------------	--------	--------	--------	--------


Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004

Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.




DESTINEE N ARMSTRONG

Signature and Printed Name

04/02/2025

Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000796, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000796</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>HAINES CITY PD</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>04/02/2025</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>16:19</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

04/02/2025

Date


DESTINEE N ARMSTRONG,
Department Inspector

FDLE/ATP Form 69 October 2024

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality