

Stability Checks

80-000810

DA

2/26/25

DIXIE COUNTY SO

Intoxilyzer - Alcohol Analyzer

Model 8000

SN 80-000810

02/26/2025

Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:47
Control Test	0.050	12:48
Air Blank	0.000	12:48
Control Test	0.049	12:49
Air Blank	0.000	12:49
Control Test	0.049	12:50
Air Blank	0.000	12:51
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel Std Dev(%)	1.1703	



Operator's Signature

DIXIE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000810
02/26/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:01
Control Test	0.080	13:02
Air Blank	0.000	13:02
Control Test	0.079	13:03
Air Blank	0.000	13:04
Control Test	0.079	13:04
Air Blank	0.000	13:05
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

Wet

LA

Operator's Signature

DIXIE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000810
02/26/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:53
Control Test	0.203	12:53
Air Blank	0.000	12:54
Control Test	0.201	12:55
Air Blank	0.000	12:55
Control Test	0.202	12:56
Air Blank	0.000	12:56
Control Test Stats		
Average	0.2020	
Std Dev	0.0010	
Rel Std Dev(%)	0.4950	



Operator's Signature

DIXIE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000810
02/26/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:58
Control Test	0.078	12:58
Air Blank	0.000	12:59
Control Test	0.078	12:59
Air Blank	0.000	12:59
Control Test	0.078	13:00
Air Blank	0.000	13:00
Control Test Stats		
Average	0.0780	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS



Operator's Signature

Florida Department of Law Enforcement

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: DIXIE COUNTY SO
Time of Inspection: 15:03

Date of Inspection: 02/26/2025

Serial Number: 80-000810
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG429602 Exp: 10/22/2026
0.000	0.049	0.080	0.203	0.078
0.000	0.049	0.080	0.204	0.078
0.000	0.050	0.080	0.203	0.077
0.000	0.049	0.081	0.204	0.077
0.000	0.050	0.080	0.203	0.078
0.000	0.050	0.080	0.204	0.078
0.000	0.050	0.080	0.204	0.078
0.000	0.050	0.080	0.204	0.077
0.000	0.050	0.080	0.204	0.077
0.000	0.050	0.080	0.204	0.077
0.000	0.049	0.081	0.204	0.077

Standard Deviations	0.0005	0.0004	0.0004	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

INTERNAL PRINTER FAIL DURING DIAGNOSTIC CHECKNon-compliance:FAILED POST DIAGNOSTIC CHECK.

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

DA DESTINEE N ARMSTRONG
Signature and Printed Name

02/26/2025
Date

DIXIE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000810
12/27/2025
Software: 8100.27

***** Printer Test *****

40 characters per line
abcdefghijklmnopqrstuvwxyz1234567890-
ABCDEFGHIJKLMNOPQRSTUVWXYZ1234567890-?

Double Char/line
abcdefghijklmnopqrstuvwxyz
1234567890-!@#%&^*
ABCDEFGHIJKLMNOPQRSTUVWXYZ

WIDE Char/line
abcdefghijklmnopqrstuvwxyz
1234567890-!@#%&^*
ABCDEFGHIJKLMNOPQRSTUVWXYZ

TALL Char/line
abcdefghijklmnopqrstuvwxyz1234567890-
ABCDEFGHIJKLMNOPQRSTUVWXYZ1234567890-?

***** Test End *****

DIXIE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000810
12/27/2025
Software: 8100.27

***** Printer Test *****

40 characters per line
abcdefghijklmnopqrstuvwxyz1234567890-
ABCDEFGHIJKLMNOPQRSTUVWXYZ1234567890-?

Double Char/line
abcdefghijklmnopqrstuvwxyz
1234567890-!@#%&^*
ABCDEFGHIJKLMNOPQRSTUVWXYZ

WIDE Char/line
abcdefghijklmnopqrstuvwxyz
1234567890-!@#%&^*
ABCDEFGHIJKLMNOPQRSTUVWXYZ

TALL Char/line
abcdefghijklmnopqrstuvwxyz1234567890-
ABCDEFGHIJKLMNOPQRSTUVWXYZ1234567890-?

***** Test End *****

DIXIE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000810
02/27/2025
Software: 8100.27

***** Printer Test *****

Normal 10.0 Characters per inch.
12345678901234567890123456789012345678901234567890123456
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%^&*()_+?

Arial
12345678901234567890123456789012345678901234567890123456
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%^&*()_+?

Elite 12.0 Char/inch.
123456789012345678901234567890123456789012345678901234567
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%^&*()_+?

Compressed 16.5 Char/inch.
1234567890123456789012345678901234567890123456789012345678901234567890123
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%^&*()_+?

***** Test End *****

Florida Department of Law Enforcement

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: DIXIE COUNTY SO
Time of Inspection: 12:13

Date of Inspection: 02/27/2025

Serial Number: 80-000810
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG429602 Exp: 10/22/2026
0.000	0.048	0.079	0.203	0.077
0.000	0.048	0.079	0.203	0.077
0.000	0.049	0.079	0.203	0.077
0.000	0.049	0.079	0.203	0.076
0.000	0.049	0.078	0.203	0.076
0.000	0.049	0.079	0.203	0.077
0.000	0.049	0.079	0.204	0.076
0.000	0.048	0.079	0.203	0.076
0.000	0.049	0.079	0.203	0.076
0.000	0.048	0.079	0.203	0.076

Standard Deviations	0.0005	0.0003	0.0003	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

DESTINEE N ARMSTRONG
Signature and Printed Name

02/27/2025
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000810, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000810</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>DIXIE COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>02/27/2025</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>12:13</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

02/27/2025

Date

DESTINEE N ARMSTRONG,
Department Inspector

FDLE/ATP Form 69 October 2024
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality