



INSTRUMENT PROCESSING SHEET

Agency MARION COUNTY SOS/N 80-000813Florida Department of
Law EnforcementDate In 06/30/2025 DI Completion Date 06/30/2025☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake By <u>WKP</u> Date <u>6/30/2025</u> ☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: <u>Received with one foot unscrewed and brass anchor detached. Reinforced anchor and tightened feet.</u> <u>WKP 6/30/2025</u>	Quality Checks By <u>SLH</u> Date <u>06/30/2025</u> ☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>121</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP102</u> 32 mm <u>0.140</u> (.139 - .169) 36 mm <u>0.152</u> (.156 - .190) 53 mm <u>0.222</u> (.228 - .278) 103 mm <u>0.500</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28662</u> ☒ Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP5088</td><td>202406K 06/19/2026</td></tr><tr><td>0.080</td><td>MP5089</td><td>202406L 06/19/2026</td></tr><tr><td>0.200</td><td>MP5090</td><td>202406N 06/20/2026</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG429602 10/22/2026</td></tr></tbody></table>	Simulator	Serial #	Lot #/Exp	0.050	MP5088	202406K 06/19/2026	0.080	MP5089	202406L 06/19/2026	0.200	MP5090	202406N 06/20/2026	0.080 DGS	N/A	AG429602 10/22/2026	Flow Calibration By <u>SLH</u> Date <u>6/30/2025</u> Flow Column # <u>ATP103</u> ☒ 5L/min – 17mm ☒ 15L/min – 53mm ☒ 30L/min – 103mm ☒ R-Value <u>125</u> ☒ Post Calibration Verification (L/s) Flow Column # <u>ATP102</u> 32 mm <u>0.152</u> (.139 - .169) 36 mm <u>0.167</u> (.156 - .190) 53 mm <u>0.234</u> (.228 - .278) 103 mm <u>0.511</u> (.447 - .547) Maintenance By <u>SLH</u> Date <u>06/30/2025</u> ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☒ Other <u>Replaced o-ring on simulator.SLH</u>																																																				
Simulator	Serial #	Lot #/Exp																																																																			
0.050	MP5088	202406K 06/19/2026																																																																			
0.080	MP5089	202406L 06/19/2026																																																																			
0.200	MP5090	202406N 06/20/2026																																																																			
0.080 DGS	N/A	AG429602 10/22/2026																																																																			
Calibration Adjustment By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table> ☐ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table> Notes/Suggested Service: _____		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Department Inspection By <u>SLH</u> Barometric Pressure ID# <u>28421</u> Gauge <u>1014</u> Instrument <u>1010</u> Mouth Alcohol Solution Lot # <u>2025-A</u> Acetone Stock Solution Lot # <u>2024-B</u> <table border="1" style="width:100%; border-collapse: collapse;"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP5086</td></tr><tr><td>Interferent</td><td>MP5087</td></tr><tr><td>0.050</td><td>MP5088</td></tr><tr><td>0.080</td><td>MP5089</td></tr><tr><td>0.200</td><td>MP5090</td></tr></tbody></table> Attachments <table border="1" style="width:100%; border-collapse: collapse;"><tr><td>☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment</td><td>☐ Post-Stability Checks ☒ Flow Calibration ☐ Form 40 ☐ Other _____</td></tr></table> ☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use <table style="width:100%;"><tr><td style="width:50%; text-align: center;">Phil Nicodemo Digitally signed by Phil Nicodemo Date: 2025.07.01 12:33:35 +04'00'</td><td style="width:50%; text-align: center;">Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2025.07.02 12:16:53 -04'00'</td></tr><tr><td style="text-align: center;">Tech Review / Date</td><td style="text-align: center;">Admin Review / Date</td></tr></table>		Simulator	Serial Number	0.000	MP5086	Interferent	MP5087	0.050	MP5088	0.080	MP5089	0.200	MP5090	☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment	☐ Post-Stability Checks ☒ Flow Calibration ☐ Form 40 ☐ Other _____	Phil Nicodemo Digitally signed by Phil Nicodemo Date: 2025.07.01 12:33:35 +04'00'	Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2025.07.02 12:16:53 -04'00'	Tech Review / Date	Admin Review / Date
Simulator	Serial #	Lot #	Expiration																																																																		
0.000		N/A	N/A																																																																		
0.040																																																																					
0.100																																																																					
0.200																																																																					
0.300																																																																					
0.080 DGS	N/A																																																																				
Simulator	Serial #	Lot #	Expiration																																																																		
0.050																																																																					
0.080																																																																					
0.200																																																																					
0.080 DGS	N/A																																																																				
Simulator	Serial Number																																																																				
0.000	MP5086																																																																				
Interferent	MP5087																																																																				
0.050	MP5088																																																																				
0.080	MP5089																																																																				
0.200	MP5090																																																																				
☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment	☐ Post-Stability Checks ☒ Flow Calibration ☐ Form 40 ☐ Other _____																																																																				
Phil Nicodemo Digitally signed by Phil Nicodemo Date: 2025.07.01 12:33:35 +04'00'	Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2025.07.02 12:16:53 -04'00'																																																																				
Tech Review / Date	Admin Review / Date																																																																				

80-000813

Sut 6/30/2025

MARION COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000813
06/30/2025
Software: 0100.27

Flow Rate Calibration*****

1: Rate (Liters/min) = 5

SQRT(Diff) = 6.324

2: Rate (Liters/min) = 15

SQRT(Diff) = 11.043

3: Rate (Liters/min) = 30

SQRT(Diff) = 20.121

Dependent Data Scale Factor = 100000 L/min

Independent Data Scale Factor = 256

Rounded Slope = 699

Rounded Intercept = -570864



Correlation = 0.99786

Root cause Analysis -
no equipment or user error -
performed Flow Calibration.
Sut

Stability Checks

6/30/2025 Sat
80-000813


* passed on rerun

0.050 g/210L	0.080 g/210L	0.200 g/210L	DGS 0.080 g/210L																																																																																																																																																
0.047 to 0.053 g/210L ☒	0.077 to 0.083 g/210L ☒	0.194 to 0.206 g/210L * ☒	0.077 to 0.083 g/210L ☒ ≤0.003 g/210L of Wet ☒																																																																																																																																																
Performed Root Case Analysis ☐	Performed Root Case Analysis ☐	Performed Root Case Analysis ☒	Performed Root Case Analysis ☐																																																																																																																																																
MARION COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000813 06/30/2025 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>16:36</td></tr> <tr><td>Control Test</td><td>0.048</td><td>16:37</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:37</td></tr> <tr><td>Control Test</td><td>0.047</td><td>16:38</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:38</td></tr> <tr><td>Control Test</td><td>0.048</td><td>16:39</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:40</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0477</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.2112</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	16:36	Control Test	0.048	16:37	Air Blank	0.000	16:37	Control Test	0.047	16:38	Air Blank	0.000	16:38	Control Test	0.048	16:39	Air Blank	0.000	16:40	Control Test Stats			Average	0.0477		Std Dev	0.0006		Rel Std Dev(%)	1.2112		MARION COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000813 06/30/2025 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>16:31</td></tr> <tr><td>Control Test</td><td>0.077</td><td>16:31</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:32</td></tr> <tr><td>Control Test</td><td>0.077</td><td>16:32</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:33</td></tr> <tr><td>Control Test</td><td>0.077</td><td>16:34</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:34</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0770</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	16:31	Control Test	0.077	16:31	Air Blank	0.000	16:32	Control Test	0.077	16:32	Air Blank	0.000	16:33	Control Test	0.077	16:34	Air Blank	0.000	16:34	Control Test Stats			Average	0.0770		Std Dev	0.0000		Rel Std Dev(%)	0.0000		MARION COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000813 06/30/2025 Software: 8100.27 #1 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>16:46</td></tr> <tr><td>Control Test</td><td>0.195</td><td>16:47</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:47</td></tr> <tr><td>Control Test</td><td>0.195</td><td>16:48</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:48</td></tr> <tr><td>Control Test</td><td>0.192</td><td>16:49</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:50</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1940</td><td></td></tr> <tr><td>Std Dev</td><td>0.0017</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.8928</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	16:46	Control Test	0.195	16:47	Air Blank	0.000	16:47	Control Test	0.195	16:48	Air Blank	0.000	16:48	Control Test	0.192	16:49	Air Blank	0.000	16:50	Control Test Stats			Average	0.1940		Std Dev	0.0017		Rel Std Dev(%)	0.8928		MARION COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000813 06/30/2025 Software: 8100.27 DGS <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>15:53</td></tr> <tr><td>Control Test</td><td>0.079</td><td>15:54</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:54</td></tr> <tr><td>Control Test</td><td>0.079</td><td>15:54</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:55</td></tr> <tr><td>Control Test</td><td>0.079</td><td>15:55</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:56</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0790</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	15:53	Control Test	0.079	15:54	Air Blank	0.000	15:54	Control Test	0.079	15:54	Air Blank	0.000	15:55	Control Test	0.079	15:55	Air Blank	0.000	15:56	Control Test Stats			Average	0.0790		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	16:36																																																																																																																																																	
Control Test	0.048	16:37																																																																																																																																																	
Air Blank	0.000	16:37																																																																																																																																																	
Control Test	0.047	16:38																																																																																																																																																	
Air Blank	0.000	16:38																																																																																																																																																	
Control Test	0.048	16:39																																																																																																																																																	
Air Blank	0.000	16:40																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0477																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.2112																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	16:31																																																																																																																																																	
Control Test	0.077	16:31																																																																																																																																																	
Air Blank	0.000	16:32																																																																																																																																																	
Control Test	0.077	16:32																																																																																																																																																	
Air Blank	0.000	16:33																																																																																																																																																	
Control Test	0.077	16:34																																																																																																																																																	
Air Blank	0.000	16:34																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0770																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	16:46																																																																																																																																																	
Control Test	0.195	16:47																																																																																																																																																	
Air Blank	0.000	16:47																																																																																																																																																	
Control Test	0.195	16:48																																																																																																																																																	
Air Blank	0.000	16:48																																																																																																																																																	
Control Test	0.192	16:49																																																																																																																																																	
Air Blank	0.000	16:50																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1940																																																																																																																																																		
Std Dev	0.0017																																																																																																																																																		
Rel Std Dev(%)	0.8928																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	15:53																																																																																																																																																	
Control Test	0.079	15:54																																																																																																																																																	
Air Blank	0.000	15:54																																																																																																																																																	
Control Test	0.079	15:54																																																																																																																																																	
Air Blank	0.000	15:55																																																																																																																																																	
Control Test	0.079	15:55																																																																																																																																																	
Air Blank	0.000	15:56																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0790																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		

#2

MARION COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000813
06/30/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	17:01
Control Test	0.197 ✓	17:01
Air Blank	0.000	17:02
Control Test	0.197 ✓	17:03
Air Blank	0.000	17:03
Control Test	0.195 ✓	17:04
Air Blank	0.000	17:04
Control Test Stats		
Average	0.1963	
Std Dev	0.0012	
Rel Std Dev(%)	0.5881	


Operator Signature

80-000813 6/30/25
Sut

Retightened seal and changed
O-ring on simulator
reran and passed
criteria,
Sut

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MARION COUNTY SO
Time of Inspection: 18:57

Date of Inspection: 06/30/2025

Serial Number: 80-000813
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG429602 Exp: 10/22/2026
0.000	0.047	0.079	0.196	0.079
0.000	0.048	0.078	0.195	0.079
0.000	0.047	0.079	0.195	0.079
0.000	0.048	0.079	0.196	0.079
0.000	0.047	0.078	0.196	0.078
0.000	0.048	0.078	0.196	0.079
0.000	0.048	0.078	0.196	0.079
0.000	0.048	0.078	0.196	0.078
0.000	0.048	0.078	0.195	0.079
0.000	0.047	0.079	0.195	0.079


Standard Deviations	0.0005	0.0005	0.0005	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5


Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


LEANDRA HIGGINBOTHAM
Signature and Printed Name

06/30/2025
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000813, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000813</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>MARION COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>06/30/2025</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>18:57</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.


Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

06/30/2025

Date


LEANDRA HIGGINBOTHAM,
Department Inspector

FDLE/ATP Form 69 October 2024

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1