

INSTRUMENT PROCESSING SHEET

Agency Lake CSO

S/N 80-000822

Florida Department of
Law Enforcement

Date In 01/23/2025

DI Completion Date 02/14/2025

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By TDG	Date	Quality Checks	By TDG	Date																																																																																								
☒ Annual ☐ Registration ☒ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____			☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>139</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP101</u> 32 mm <u>0.175*</u> (.139 - .169) 36 mm <u>0.199*</u> (.156 - .190) 53 mm <u>0.265</u> (.228 - .278) 103 mm <u>0.550*</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>26932</u> ☒ Stability Checks																																																																																										
			<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP6286</td> <td>202303K 03/29/2025</td> </tr> <tr> <td>0.080</td> <td>MP6287</td> <td>202303L 03/29/2025</td> </tr> <tr> <td>0.200</td> <td>MP6288</td> <td>202304C 04/05/2025</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG429602 10/22/2026</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	MP6286	202303K 03/29/2025	0.080	MP6287	202303L 03/29/2025	0.200	MP6288	202304C 04/05/2025	0.080 DGS	N/A	AG429602 10/22/2026																																																																											
Simulator	Serial #	Lot #/Exp																																																																																											
0.050	MP6286	202303K 03/29/2025																																																																																											
0.080	MP6287	202303L 03/29/2025																																																																																											
0.200	MP6288	202304C 04/05/2025																																																																																											
0.080 DGS	N/A	AG429602 10/22/2026																																																																																											
			Flow Column # <u>ATP106</u> ☒ 5L/min – 17mm ☒ 15L/min – 53mm ☒ 30L/min – 103mm ☒ R-Value <u>139</u> ☒ Post Calibration Verification (L/s) Flow Column # <u>ATP101</u> 32 mm <u>0.152</u> (.139 - .169) 36 mm <u>0.164</u> (.156 - .190) 53 mm <u>0.234</u> (.228 - .278) 103 mm <u>0.515</u> (.447 - .547)																																																																																										
			Maintenance By _____ Date _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ _____ _____ _____ _____																																																																																										
<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th style="width:40%;">Calibration Adjustment</th> <th style="width:20%;">By TDG</th> <th style="width:40%;">Department Inspection</th> <th style="width:20%;">By TDG</th> </tr> </thead> <tbody> <tr> <td colspan="2"> Barometric Pressure Gauge <u>1022</u> ID # <u>28199</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP5097</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>MP5098</td> <td>23400</td> <td>10/24/2025</td> </tr> <tr> <td>0.100</td> <td>MP5099</td> <td>24110</td> <td>03/05/2026</td> </tr> <tr> <td>0.200</td> <td>MP5100</td> <td>24080</td> <td>02/13/2026</td> </tr> <tr> <td>0.300</td> <td>MP5101</td> <td>23410</td> <td>11/01/2025</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>06723080A5</td> <td>04/05/2025</td> </tr> </tbody> </table> ☒ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP6286</td> <td>202303K</td> <td>03/29/2025</td> </tr> <tr> <td>0.080</td> <td>MP6287</td> <td>202303L</td> <td>03/29/2025</td> </tr> <tr> <td>0.200</td> <td>MP6288</td> <td>202304C</td> <td>04/05/2025</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG429602</td> <td>10/22/2026</td> </tr> </tbody> </table> </td> <td colspan="2"> Barometric Pressure ID# <u>26932</u> Gauge <u>1027</u> Instrument <u>1027</u> Mouth Alcohol Solution Lot # <u>2024-A</u> Acetone Stock Solution Lot # <u>2023-B</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP6284</td> </tr> <tr> <td>Interferent</td> <td>MP6285</td> </tr> <tr> <td>0.050</td> <td>MP6286</td> </tr> <tr> <td>0.080</td> <td>MP6287</td> </tr> <tr> <td>0.200</td> <td>MP6288</td> </tr> </tbody> </table> </td> </tr> <tr> <td colspan="2"> Notes/Suggested Service: <u>*Outside nominal. (TDG 2/5/25)</u> _____ _____ _____ _____ _____ _____ _____ </td> <td colspan="2"> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th colspan="2">Attachments</th> </tr> </thead> <tbody> <tr> <td> ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Calibration Adjustment </td> <td> ☒ Post-Stability Checks ☒ Flow Calibration ☐ Form 40 ☐ Other _____ </td> </tr> </tbody> </table> </td> </tr> <tr> <td colspan="2"></td> <td colspan="2"> ☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use </td> </tr> <tr> <td colspan="2"></td> <td colspan="2"> Destinee Armstrong Digitally signed by Destinee Armstrong Date: 2025.02.21 10:55:46 -05'00' Shayla Platt Digitally signed by Shayla Platt Date: 2025.02.21 15:04:25 -05'00' </td> </tr> <tr> <td colspan="2"></td> <td colspan="2"> Tech Review / Date _____ Admin Review / Date _____ </td> </tr> </tbody> </table>						Calibration Adjustment	By TDG	Department Inspection	By TDG	Barometric Pressure Gauge <u>1022</u> ID # <u>28199</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP5097</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>MP5098</td> <td>23400</td> <td>10/24/2025</td> </tr> <tr> <td>0.100</td> <td>MP5099</td> <td>24110</td> <td>03/05/2026</td> </tr> <tr> <td>0.200</td> <td>MP5100</td> <td>24080</td> <td>02/13/2026</td> </tr> <tr> <td>0.300</td> <td>MP5101</td> <td>23410</td> <td>11/01/2025</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>06723080A5</td> <td>04/05/2025</td> </tr> </tbody> </table> ☒ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP6286</td> <td>202303K</td> <td>03/29/2025</td> </tr> <tr> <td>0.080</td> <td>MP6287</td> <td>202303L</td> <td>03/29/2025</td> </tr> <tr> <td>0.200</td> <td>MP6288</td> <td>202304C</td> <td>04/05/2025</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG429602</td> <td>10/22/2026</td> </tr> </tbody> </table>		Simulator	Serial #	Lot #	Expiration	0.000	MP5097	N/A	N/A	0.040	MP5098	23400	10/24/2025	0.100	MP5099	24110	03/05/2026	0.200	MP5100	24080	02/13/2026	0.300	MP5101	23410	11/01/2025	0.080 DGS	N/A	06723080A5	04/05/2025	Simulator	Serial #	Lot #	Expiration	0.050	MP6286	202303K	03/29/2025	0.080	MP6287	202303L	03/29/2025	0.200	MP6288	202304C	04/05/2025	0.080 DGS	N/A	AG429602	10/22/2026	Barometric Pressure ID# <u>26932</u> Gauge <u>1027</u> Instrument <u>1027</u> Mouth Alcohol Solution Lot # <u>2024-A</u> Acetone Stock Solution Lot # <u>2023-B</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP6284</td> </tr> <tr> <td>Interferent</td> <td>MP6285</td> </tr> <tr> <td>0.050</td> <td>MP6286</td> </tr> <tr> <td>0.080</td> <td>MP6287</td> </tr> <tr> <td>0.200</td> <td>MP6288</td> </tr> </tbody> </table>		Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288	Notes/Suggested Service: <u>*Outside nominal. (TDG 2/5/25)</u> _____ _____ _____ _____ _____ _____ _____		<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th colspan="2">Attachments</th> </tr> </thead> <tbody> <tr> <td> ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Calibration Adjustment </td> <td> ☒ Post-Stability Checks ☒ Flow Calibration ☐ Form 40 ☐ Other _____ </td> </tr> </tbody> </table>		Attachments		☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Calibration Adjustment	☒ Post-Stability Checks ☒ Flow Calibration ☐ Form 40 ☐ Other _____			☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use				Destinee Armstrong Digitally signed by Destinee Armstrong Date: 2025.02.21 10:55:46 -05'00' Shayla Platt Digitally signed by Shayla Platt Date: 2025.02.21 15:04:25 -05'00'				Tech Review / Date _____ Admin Review / Date _____	
Calibration Adjustment	By TDG	Department Inspection	By TDG																																																																																										
Barometric Pressure Gauge <u>1022</u> ID # <u>28199</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP5097</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>MP5098</td> <td>23400</td> <td>10/24/2025</td> </tr> <tr> <td>0.100</td> <td>MP5099</td> <td>24110</td> <td>03/05/2026</td> </tr> <tr> <td>0.200</td> <td>MP5100</td> <td>24080</td> <td>02/13/2026</td> </tr> <tr> <td>0.300</td> <td>MP5101</td> <td>23410</td> <td>11/01/2025</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>06723080A5</td> <td>04/05/2025</td> </tr> </tbody> </table> ☒ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP6286</td> <td>202303K</td> <td>03/29/2025</td> </tr> <tr> <td>0.080</td> <td>MP6287</td> <td>202303L</td> <td>03/29/2025</td> </tr> <tr> <td>0.200</td> <td>MP6288</td> <td>202304C</td> <td>04/05/2025</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG429602</td> <td>10/22/2026</td> </tr> </tbody> </table>		Simulator	Serial #	Lot #	Expiration	0.000	MP5097	N/A	N/A	0.040	MP5098	23400	10/24/2025	0.100	MP5099	24110	03/05/2026	0.200	MP5100	24080	02/13/2026	0.300	MP5101	23410	11/01/2025	0.080 DGS	N/A	06723080A5	04/05/2025	Simulator	Serial #	Lot #	Expiration	0.050	MP6286	202303K	03/29/2025	0.080	MP6287	202303L	03/29/2025	0.200	MP6288	202304C	04/05/2025	0.080 DGS	N/A	AG429602	10/22/2026	Barometric Pressure ID# <u>26932</u> Gauge <u>1027</u> Instrument <u>1027</u> Mouth Alcohol Solution Lot # <u>2024-A</u> Acetone Stock Solution Lot # <u>2023-B</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP6284</td> </tr> <tr> <td>Interferent</td> <td>MP6285</td> </tr> <tr> <td>0.050</td> <td>MP6286</td> </tr> <tr> <td>0.080</td> <td>MP6287</td> </tr> <tr> <td>0.200</td> <td>MP6288</td> </tr> </tbody> </table>		Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288																														
Simulator	Serial #	Lot #	Expiration																																																																																										
0.000	MP5097	N/A	N/A																																																																																										
0.040	MP5098	23400	10/24/2025																																																																																										
0.100	MP5099	24110	03/05/2026																																																																																										
0.200	MP5100	24080	02/13/2026																																																																																										
0.300	MP5101	23410	11/01/2025																																																																																										
0.080 DGS	N/A	06723080A5	04/05/2025																																																																																										
Simulator	Serial #	Lot #	Expiration																																																																																										
0.050	MP6286	202303K	03/29/2025																																																																																										
0.080	MP6287	202303L	03/29/2025																																																																																										
0.200	MP6288	202304C	04/05/2025																																																																																										
0.080 DGS	N/A	AG429602	10/22/2026																																																																																										
Simulator	Serial Number																																																																																												
0.000	MP6284																																																																																												
Interferent	MP6285																																																																																												
0.050	MP6286																																																																																												
0.080	MP6287																																																																																												
0.200	MP6288																																																																																												
Notes/Suggested Service: <u>*Outside nominal. (TDG 2/5/25)</u> _____ _____ _____ _____ _____ _____ _____		<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th colspan="2">Attachments</th> </tr> </thead> <tbody> <tr> <td> ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Calibration Adjustment </td> <td> ☒ Post-Stability Checks ☒ Flow Calibration ☐ Form 40 ☐ Other _____ </td> </tr> </tbody> </table>		Attachments		☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Calibration Adjustment	☒ Post-Stability Checks ☒ Flow Calibration ☐ Form 40 ☐ Other _____																																																																																						
Attachments																																																																																													
☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Calibration Adjustment	☒ Post-Stability Checks ☒ Flow Calibration ☐ Form 40 ☐ Other _____																																																																																												
		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																																																																																											
		Destinee Armstrong Digitally signed by Destinee Armstrong Date: 2025.02.21 10:55:46 -05'00' Shayla Platt Digitally signed by Shayla Platt Date: 2025.02.21 15:04:25 -05'00'																																																																																											
		Tech Review / Date _____ Admin Review / Date _____																																																																																											

Flow Calibration Adjustment(s)

Performed by TDG

LAKE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000822
02/05/2025
Software: 8100.27

Flow Rate Calibration*****

1: Rate (Liters/min) = 5

SQRT(Diff) = 6.926

2: Rate (Liters/min) = 15

SQRT(Diff) = 11.531

3: Rate (Liters/min) = 30

SQRT(Diff) = 20.973

Dependent Data Scale Factor = 100000 L/min

Independent Data Scale Factor = 236

Rounded Slope = 684

Rounded Intercept = -635250

Correlation = 0.99670

Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																				
LAKE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000822 02/05/2025 Software: 8100.27 0.047 to 0.053 ✓	LAKE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000822 02/05/2025 Software: 8100.27 0.077 to 0.083 ✓	LAKE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000822 02/05/2025 Software: 8100.27 0.194 to 0.206 ✓	LAKE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000822 02/05/2025 Software: 8100.27 0.077 to 0.083 ✗ ≤0.003 of Wet ✗																																																																																																																																				
LAKE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000822 02/05/2025 Software: 8100.27 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>10:53</td></tr> <tr><td>Control Test</td><td>0.049</td><td>10:54</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:54</td></tr> <tr><td>Control Test</td><td>0.049</td><td>10:55</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:56</td></tr> <tr><td>Control Test</td><td>0.048</td><td>10:56</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:57</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0487</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1863</td><td></td></tr> </table> Operator's Signature 	Air Blank	0.000	10:53	Control Test	0.049	10:54	Air Blank	0.000	10:54	Control Test	0.049	10:55	Air Blank	0.000	10:56	Control Test	0.048	10:56	Air Blank	0.000	10:57	Control Test Stats			Average	0.0487		Std Dev	0.0006		Rel Std Dev(%)	1.1863		LAKE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000822 02/05/2025 Software: 8100.27 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>11:01</td></tr> <tr><td>Control Test</td><td>0.078</td><td>11:01</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:02</td></tr> <tr><td>Control Test</td><td>0.078</td><td>11:03</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:03</td></tr> <tr><td>Control Test</td><td>0.078</td><td>11:04</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:04</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0780</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </table> Operator's Signature 	Air Blank	0.000	11:01	Control Test	0.078	11:01	Air Blank	0.000	11:02	Control Test	0.078	11:03	Air Blank	0.000	11:03	Control Test	0.078	11:04	Air Blank	0.000	11:04	Control Test Stats			Average	0.0780		Std Dev	0.0000		Rel Std Dev(%)	0.0000		LAKE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000822 02/05/2025 Software: 8100.27 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>11:07</td></tr> <tr><td>Control Test</td><td>0.201</td><td>11:08</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:08</td></tr> <tr><td>Control Test</td><td>0.199</td><td>11:09</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:10</td></tr> <tr><td>Control Test</td><td>0.200</td><td>11:10</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:11</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.2000</td><td></td></tr> <tr><td>Std Dev</td><td>0.0010</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.5000</td><td></td></tr> </table> Operator's Signature 	Air Blank	0.000	11:07	Control Test	0.201	11:08	Air Blank	0.000	11:08	Control Test	0.199	11:09	Air Blank	0.000	11:10	Control Test	0.200	11:10	Air Blank	0.000	11:11	Control Test Stats			Average	0.2000		Std Dev	0.0010		Rel Std Dev(%)	0.5000		LAKE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000822 02/05/2025 Software: 8100.27 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>10:36</td></tr> <tr><td>Control Test</td><td>0.073</td><td>10:37</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:37</td></tr> <tr><td>Control Test</td><td>0.073</td><td>10:37</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:38</td></tr> <tr><td>Control Test</td><td>0.073</td><td>10:38</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:39</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0730</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </table> Operator's Signature 	Air Blank	0.000	10:36	Control Test	0.073	10:37	Air Blank	0.000	10:37	Control Test	0.073	10:37	Air Blank	0.000	10:38	Control Test	0.073	10:38	Air Blank	0.000	10:39	Control Test Stats			Average	0.0730		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Air Blank	0.000	10:53																																																																																																																																					
Control Test	0.049	10:54																																																																																																																																					
Air Blank	0.000	10:54																																																																																																																																					
Control Test	0.049	10:55																																																																																																																																					
Air Blank	0.000	10:56																																																																																																																																					
Control Test	0.048	10:56																																																																																																																																					
Air Blank	0.000	10:57																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0487																																																																																																																																						
Std Dev	0.0006																																																																																																																																						
Rel Std Dev(%)	1.1863																																																																																																																																						
Air Blank	0.000	11:01																																																																																																																																					
Control Test	0.078	11:01																																																																																																																																					
Air Blank	0.000	11:02																																																																																																																																					
Control Test	0.078	11:03																																																																																																																																					
Air Blank	0.000	11:03																																																																																																																																					
Control Test	0.078	11:04																																																																																																																																					
Air Blank	0.000	11:04																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0780																																																																																																																																						
Std Dev	0.0000																																																																																																																																						
Rel Std Dev(%)	0.0000																																																																																																																																						
Air Blank	0.000	11:07																																																																																																																																					
Control Test	0.201	11:08																																																																																																																																					
Air Blank	0.000	11:08																																																																																																																																					
Control Test	0.199	11:09																																																																																																																																					
Air Blank	0.000	11:10																																																																																																																																					
Control Test	0.200	11:10																																																																																																																																					
Air Blank	0.000	11:11																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.2000																																																																																																																																						
Std Dev	0.0010																																																																																																																																						
Rel Std Dev(%)	0.5000																																																																																																																																						
Air Blank	0.000	10:36																																																																																																																																					
Control Test	0.073	10:37																																																																																																																																					
Air Blank	0.000	10:37																																																																																																																																					
Control Test	0.073	10:37																																																																																																																																					
Air Blank	0.000	10:38																																																																																																																																					
Control Test	0.073	10:38																																																																																																																																					
Air Blank	0.000	10:39																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0730																																																																																																																																						
Std Dev	0.0000																																																																																																																																						
Rel Std Dev(%)	0.0000																																																																																																																																						

***** AUTO CAL DATA *****

<<<< CHANNEL 1 >>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.108
Std Dev = 0.01 Rel Std Dev = 9.40
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.859
Std Dev = 0.01 Rel Std Dev = 1.15
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.955
Std Dev = 0.00 Rel Std Dev = 0.21
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.773
Std Dev = 0.03 Rel Std Dev = 0.72
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.518
Std Dev = 0.02 Rel Std Dev = 0.32
Zero Order Coef = -270.36
First Order Coef = 2522.81
Second Order Coef = 20.62
Standard Deviation = 15.386976

<<<< CHANNEL 2 >>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.105
Std Dev = 0.00 Rel Std Dev = 3.43
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.508
Std Dev = 0.01 Rel Std Dev = 0.38
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.536
Std Dev = 0.01 Rel Std Dev = 0.27
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.819
Std Dev = 0.01 Rel Std Dev = 0.12
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 9.946
Std Dev = 0.01 Rel Std Dev = 0.07
Zero Order Coef = -144.50
First Order Coef = 1348.64
Second Order Coef = 10.25
Standard Deviation = 6.664906

<<<< CHANNEL 1 >>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.105
Std Dev = 0.00 Rel Std Dev = 3.43
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.508
Std Dev = 0.01 Rel Std Dev = 0.38
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.536
Std Dev = 0.01 Rel Std Dev = 0.27
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.819
Std Dev = 0.01 Rel Std Dev = 0.12
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 9.946
Std Dev = 0.01 Rel Std Dev = 0.07
Zero Order Coef = -144.50
First Order Coef = 1348.64
Second Order Coef = 10.25
Standard Deviation = 6.664906

<<<< CHANNEL 2 >>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.105
Std Dev = 0.00 Rel Std Dev = 3.43
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.508
Std Dev = 0.01 Rel Std Dev = 0.38
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.536
Std Dev = 0.01 Rel Std Dev = 0.27
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.819
Std Dev = 0.01 Rel Std Dev = 0.12
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 9.946
Std Dev = 0.01 Rel Std Dev = 0.07
Zero Order Coef = -144.50
First Order Coef = 1348.64
Second Order Coef = 10.25
Standard Deviation = 6.664906

Dry Gas H2O Adjust Results *****
Barometric Pressure = 1022
3 um H2O Adjust (mg/l*10,000) = 871
9 um H2O Adjust (mg/l*10,000) = 494
***** AUTO CAL PASS

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 6.8590 (-0.0050)
Sample #2 = 6.8270 (-0.0190)
Sample #3 = 6.8100 (-0.0290)
Sample #4 = 6.8190 (-0.0200)
Avg % Abs = 6.8187 (-0.0227)
STD DEV = 0.0085 (-0.0055)
REL STD DEV = 0.125 (24.298)

<<<< CHANNEL 1 >>>>
Sol Value = 0.300 g/210L ***
Fit Value = 1.4286 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12523, Sum Io = 13539
Sample % Abs (% Abs Ref)
Sample #1 = 5.5380 (-0.0210)
Sample #2 = 5.5160 (-0.0170)
Sample #3 = 5.5020 (-0.0150)
Sample #4 = 5.5370 (-0.0060)
Avg % Abs = 5.5183 (-0.0127)
STD DEV = 0.0176 (-0.0059)
REL STD DEV = 0.319 (46.259)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 9.9680 (-0.0120)
Sample #2 = 9.9410 (-0.0440)
Sample #3 = 9.9540 (-0.0390)
Sample #4 = 9.9440 (-0.0410)
Avg % Abs = 9.9463 (-0.0413)
STD DEV = 0.0068 (-0.0025)
REL STD DEV = 0.068 (6.089)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.5370 (-0.0140)
Sample #2 = 1.5020 (-0.0040)
Sample #3 = 1.5100 (-0.0030)
Sample #4 = 1.5130 (-0.0080)
Avg % Abs = 1.5183 (-0.0050)
STD DEV = 0.0057 (-0.0026)
REL STD DEV = 0.377 (52.915)

<<<< CHANNEL 1 >>>>
Sol Value = 0.100 g/210L ***
Fit Value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12524, Sum Io = 13539
Sample % Abs (% Abs Ref)
Sample #1 = 1.9610 (-0.0190)
Sample #2 = 1.9510 (-0.0030)
Sample #3 = 1.9590 (-0.0060)
Sample #4 = 1.9540 (-0.0020)
Avg % Abs = 1.9547 (-0.0023)
STD DEV = 0.0040 (-0.0040)
REL STD DEV = 0.207 (173.205)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.5540 (-0.0020)
Sample #2 = 3.5260 (-0.0110)
Sample #3 = 3.5450 (-0.0000)
Sample #4 = 3.5380 (-0.0060)
Avg % Abs = 3.5363 (-0.0057)
STD DEV = 0.0096 (-0.0055)
REL STD DEV = 0.272 (97.192)

<<<< CHANNEL 1 >>>>
Sol Value = 0.200 g/210L ***
Fit Value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12522, Sum Io = 13539
Sample % Abs (% Abs Ref)
Sample #1 = 3.7750 (-0.0090)
Sample #2 = 3.7430 (-0.0180)
Sample #3 = 3.7820 (-0.0030)
Sample #4 = 3.7950 (-0.0100)
Avg % Abs = 3.7733 (-0.0037)
STD DEV = 0.0271 (-0.0140)
REL STD DEV = 0.717 (382.143)

<<<< CHANNEL 2 >>>>
Sol Value = 0.040 g/210L ***
Fit Value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12527, Sum Io = 13542
Sample % Abs (% Abs Ref)
Sample #1 = 0.8690 (-0.0130)
Sample #2 = 0.8670 (-0.0210)
Sample #3 = 0.8620 (-0.0080)
Sample #4 = 0.8480 (-0.0140)
Avg % Abs = 0.8590 (-0.0050)
STD DEV = 0.0098 (-0.0177)
REL STD DEV = 1.147 (735.866)

LAKE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
09:22:16
SN 80-000822
02/13/2025

Auto Calibration
Max Power Res Value = 41
Auto Range Res Value = 25

<<<< CHANNEL 1 >>>>
Sol Value = 0.000 g/210L ***
Fit Value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12530, Sum Io = 13544
Sample % Abs (% Abs Ref)
Sample #1 = 0.0980 (-0.0350)
Sample #2 = 0.1190 (-0.0240)
Sample #3 = 0.0990 (-0.0200)
Sample #4 = 0.1060 (-0.0030)
Avg % Abs = 0.1080 (-0.0157)
STD DEV = 0.0101 (-0.0112)
REL STD DEV = 9.397 (71.173)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.0940 (-0.0160)
Sample #2 = 0.1040 (-0.0090)
Sample #3 = 0.1090 (-0.0190)
Sample #4 = 0.1020 (-0.0070)
Avg % Abs = 0.1050 (-0.0117)
STD DEV = 0.0036 (-0.0064)
REL STD DEV = 3.434 (55.107)

<<<< CHANNEL 1 >>>>
Sol Value = 0.040 g/210L ***
Fit Value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12527, Sum Io = 13542
Sample % Abs (% Abs Ref)
Sample #1 = 0.8690 (-0.0130)
Sample #2 = 0.8670 (-0.0210)
Sample #3 = 0.8620 (-0.0080)
Sample #4 = 0.8480 (-0.0140)
Avg % Abs = 0.8590 (-0.0050)
STD DEV = 0.0098 (-0.0177)
REL STD DEV = 1.147 (735.866)

Optical Calibration Adjustment

By: TDG

Post-Cal Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083 ✓ ≤0.003 of Wet ✓
LAKE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000822 02/13/2025 Software: 8100.27 Test g/210L Time Air Blank 0.000 11:18 Control Test 0.052 11:19 Air Blank 0.000 11:19 Control Test 0.051 11:20 Air Blank 0.000 11:20 Control Test 0.051 11:21 Air Blank 0.000 11:22 Control Test Stats Average 0.0513 Std Dev 0.0006 Rel Std Dev(%) 1.1247 Operator's Signature <i>TAL</i>	LAKE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000822 02/13/2025 Software: 8100.27 Test g/210L Time Air Blank 0.000 11:27 Control Test 0.080 11:28 Air Blank 0.000 11:28 Control Test 0.080 11:29 Air Blank 0.000 11:30 Control Test 0.080 11:30 Air Blank 0.000 11:31 Control Test Stats Average 0.0800 Std Dev 0.0000 Rel Std Dev(%) 0.0000 Operator's Signature <i>W</i>	LAKE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000822 02/13/2025 Software: 8100.27 Test g/210L Time Air Blank 0.000 11:35 Control Test 0.198 11:36 Air Blank 0.000 11:37 Control Test 0.197 11:37 Air Blank 0.000 11:38 Control Test 0.197 11:39 Air Blank 0.000 11:39 Control Test Stats Average 0.1973 Std Dev 0.0006 Rel Std Dev(%) 0.2926 Operator's Signature <i>W</i>	LAKE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000822 02/13/2025 Software: 8100.27 Test g/210L Time Air Blank 0.000 11:08 Control Test 0.080 11:08 Air Blank 0.000 11:08 Control Test 0.080 11:09 Air Blank 0.000 11:09 Control Test 0.080 11:09 Air Blank 0.000 11:10 Control Test Stats Average 0.0800 Std Dev 0.0000 Rel Std Dev(%) 0.0000 Operator's Signature <i>W</i>

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKE COUNTY SO
Time of Inspection: 13:00

Date of Inspection: 02/14/2025

Serial Number: 80-000822
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG429602 Exp: 10/22/2026
0.000	0.051	0.080	0.198	0.080
0.000	0.051	0.080	0.198	0.080
0.000	0.051	0.080	0.198	0.080
0.000	0.051	0.080	0.198	0.079
0.000	0.051	0.080	0.198	0.079
0.000	0.051	0.080	0.198	0.079
0.000	0.051	0.081	0.198	0.079
0.000	0.051	0.081	0.198	0.079
0.000	0.051	0.080	0.198	0.079
0.000	0.051	0.080	0.198	0.078


Standard Deviations	0.0000	0.0004	0.0000	0.0006
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



Signature and Printed Name

TAYLOR D GUTSCHOW

02/14/2025
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000822, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000822</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>LAKE COUNTY SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>02/14/2025</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>13:00</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

02/14/2025

Date

Taylor D Gutschow

TAYLOR D GUTSCHOW,

Department Inspector

Service • Integrity • Respect • Quality