



Agency Miami PD

S/N 80-000874

Date In 5/30/25

DI Completion Date 06/25/2025

☐ Ship ☐ P/U ☐ H/D ☒ CMI ☐ EE

Intake By <u>KTS</u> Date <u>6/4/25</u> ☒ Annual ☐ Registration ☒ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____	Quality Checks By <u>KTS</u> Date <u>6/4/25</u> ☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>106</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP106</u> 32 mm <u>0.164</u> (.139 - .169) 36 mm <u>0.187</u> (.156 - .190) 53 mm <u>0.261</u> (.228 - .278) 103 mm <u>0.531</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28199</u> ☒ Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP8231</td> <td>202406K 06/19/2026</td> </tr> <tr> <td>0.080</td> <td>MP8230</td> <td>202406L 06/19/2026</td> </tr> <tr> <td>0.200</td> <td>MP8229</td> <td>202406N 06/20/2026</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG429602 10/22/2026</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	MP8231	202406K 06/19/2026	0.080	MP8230	202406L 06/19/2026	0.200	MP8229	202406N 06/20/2026	0.080 DGS	N/A	AG429602 10/22/2026	Flow Calibration By _____ Date _____ Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) Maintenance By _____ Date _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ _____ _____ _____ _____ _____
Simulator	Serial #	Lot #/Exp															
0.050	MP8231	202406K 06/19/2026															
0.080	MP8230	202406L 06/19/2026															
0.200	MP8229	202406N 06/20/2026															
0.080 DGS	N/A	AG429602 10/22/2026															

Calibration Adjustment				By <u>TDG</u>	
Barometric Pressure Gauge <u>1019</u>		ID # <u>26932</u>			
Simulator	Serial #	Lot #	Expiration		
0.000	MP6284	N/A	N/A		
0.040	MP6285	23400	10/24/2025		
0.100	MP6286	24110	03/05/2026		
0.200	MP6287	24080	02/13/2026		
0.300	MP6288	23410	11/01/2025		
0.080 DGS	N/A	28424080A3	11/05/2026		
☒ Post Calibration Adjustment Stability Checks					
Simulator	Serial #	Lot #	Expiration		
0.050	MP4863	202406K	06/19/2026		
0.080	MP4864	202406L	06/19/2026		
0.200	MP6284	202406N	06/20/2026		
0.080 DGS	N/A	AG429602	10/22/2026		
Notes/Suggested Service: <u>Perfomed root cause analysis.</u> <u>User/equipment error not found. KTS 6/4/25</u>					
<u>Re-used MP6284 for post-cal stability. Rinsed and dried</u> <u>prior to adding new solution. (TDG 6/20/25)</u>					
<u>Discussed low R-value with Agency Inspector. Sending</u> <u>back to CMI. (TDG 6/25/25)</u>					

Department Inspection		By <u>TDG</u>	
Barometric Pressure ID# <u>28199</u>			
Gauge <u>1020</u>	Instrument <u>1020</u>		
Mouth Alcohol Solution Lot # <u>2024-A</u>			
Acetone Stock Solution Lot # <u>2023-B</u>			
Simulator	Serial Number		
0.000	MP5097		
Interferent	MP5098		
0.050	MP5099		
0.080	MP5100		
0.200	MP5101		
Attachments			
☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Calibration Adjustment		☒ Post-Stability Checks ☐ Flow Calibration ☒ Form 40 ☐ Other _____	
☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC			
☐ Return to/Place into Evidentiary Use ☒ Remain Out of Evidentiary Use			
☐ Conduct an Agency Inspection Before Evidentiary Use			
Digitally signed by Phil Nicodemo Date: 2025.07.01 14:20:07 -04'00'		Digitally signed by Shayla Platt Date: 2025.07.01 14:42:08 -04'00'	
Tech Review / Date _____		Admin Review / Date _____	

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MIAMI PD
Time of Inspection: 16:05

Date of Inspection: 06/04/2025

Serial Number: 80-000874
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____

Remarks: Bypassed Agency inspection to operate instrument. Compliance not determined. KTS 6/4/2025

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



KATIE T SPEARIN

Signature and Printed Name

06/04/2025
Date

Stability Checks

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083 ☒ ≤0.003 of Wet																																																																																																																																																
MIAMI PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000874 06/04/2025 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>16:21</td></tr> <tr><td>Control Test</td><td>0.048</td><td>16:22</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:22</td></tr> <tr><td>Control Test</td><td>0.048</td><td>16:23</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:23</td></tr> <tr><td>Control Test</td><td>0.047</td><td>16:24</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:25</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0477</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>1.2112</td><td></td></tr> </tbody> </table> <i>[Signature]</i> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	16:21	Control Test	0.048	16:22	Air Blank	0.000	16:22	Control Test	0.048	16:23	Air Blank	0.000	16:23	Control Test	0.047	16:24	Air Blank	0.000	16:25	Control Test Stats			Average	0.0477		Std Dev	0.0006		Rel. Std Dev(%)	1.2112		MIAMI PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000874 06/04/2025 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>16:26</td></tr> <tr><td>Control Test</td><td>0.078</td><td>16:27</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:27</td></tr> <tr><td>Control Test</td><td>0.078</td><td>16:28</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:29</td></tr> <tr><td>Control Test</td><td>0.078</td><td>16:29</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:30</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0780</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <i>[Signature]</i> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	16:26	Control Test	0.078	16:27	Air Blank	0.000	16:27	Control Test	0.078	16:28	Air Blank	0.000	16:29	Control Test	0.078	16:29	Air Blank	0.000	16:30	Control Test Stats			Average	0.0780		Std Dev	0.0000		Rel. Std Dev(%)	0.0000		MIAMI PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000874 06/04/2025 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>16:31</td></tr> <tr><td>Control Test</td><td>0.200</td><td>16:32</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:32</td></tr> <tr><td>Control Test</td><td>0.200</td><td>16:33</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:33</td></tr> <tr><td>Control Test</td><td>0.199</td><td>16:34</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:35</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1997</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.2892</td><td></td></tr> </tbody> </table> <i>[Signature]</i> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	16:31	Control Test	0.200	16:32	Air Blank	0.000	16:32	Control Test	0.200	16:33	Air Blank	0.000	16:33	Control Test	0.199	16:34	Air Blank	0.000	16:35	Control Test Stats			Average	0.1997		Std Dev	0.0006		Rel. Std Dev(%)	0.2892		MIAMI PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000874 06/04/2025 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>16:16</td></tr> <tr><td>Control Test</td><td>0.076</td><td>16:16</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:17</td></tr> <tr><td>Control Test</td><td>0.075</td><td>16:17</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:18</td></tr> <tr><td>Control Test</td><td>0.076</td><td>16:18</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>16:19</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0757</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.7630</td><td></td></tr> </tbody> </table> <i>[Signature]</i> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	16:16	Control Test	0.076	16:16	Air Blank	0.000	16:17	Control Test	0.075	16:17	Air Blank	0.000	16:18	Control Test	0.076	16:18	Air Blank	0.000	16:19	Control Test Stats			Average	0.0757		Std Dev	0.0006		Rel. Std Dev(%)	0.7630	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	16:21																																																																																																																																																	
Control Test	0.048	16:22																																																																																																																																																	
Air Blank	0.000	16:22																																																																																																																																																	
Control Test	0.048	16:23																																																																																																																																																	
Air Blank	0.000	16:23																																																																																																																																																	
Control Test	0.047	16:24																																																																																																																																																	
Air Blank	0.000	16:25																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0477																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel. Std Dev(%)	1.2112																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	16:26																																																																																																																																																	
Control Test	0.078	16:27																																																																																																																																																	
Air Blank	0.000	16:27																																																																																																																																																	
Control Test	0.078	16:28																																																																																																																																																	
Air Blank	0.000	16:29																																																																																																																																																	
Control Test	0.078	16:29																																																																																																																																																	
Air Blank	0.000	16:30																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0780																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel. Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	16:31																																																																																																																																																	
Control Test	0.200	16:32																																																																																																																																																	
Air Blank	0.000	16:32																																																																																																																																																	
Control Test	0.200	16:33																																																																																																																																																	
Air Blank	0.000	16:33																																																																																																																																																	
Control Test	0.199	16:34																																																																																																																																																	
Air Blank	0.000	16:35																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1997																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel. Std Dev(%)	0.2892																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	16:16																																																																																																																																																	
Control Test	0.076	16:16																																																																																																																																																	
Air Blank	0.000	16:17																																																																																																																																																	
Control Test	0.075	16:17																																																																																																																																																	
Air Blank	0.000	16:18																																																																																																																																																	
Control Test	0.076	16:18																																																																																																																																																	
Air Blank	0.000	16:19																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0757																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel. Std Dev(%)	0.7630																																																																																																																																																		

***** CHANNEL 2 *****
Sample % Abs (% Abs Ref)
Sample #1 = 1.5640 (-0.0120)
Sample #2 = 1.5230 (-0.0240)
Sample #3 = 1.5140 (-0.0430)
Sample #4 = 1.5150 (-0.0470)
Avg % Abs = 1.5140 (-0.0380)
STD DEV = 0.0095 (-0.0123)
REL STD DEV = 0.630 (32.337)

M14M1 PD
Intoxilyzer - Alcohol Analyzer
Model 8000
05/20/2025
SN 80-000674
13:02:28

Auto Calibration
Max Power Res Value = 21
Auto Range Res Value = 9
Sol Value = 0.100 g/210L ***
Fit Value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12886, Sum Io = 13434
***** CHANNEL 1 *****
Sample % Abs (% Abs Ref)
Sample #1 = 1.9920 (-0.0320)
Sample #2 = 1.9560 (-0.0010)
Sample #3 = 1.9160 (-0.0000)
Sample #4 = 1.9150 (-0.0240)
Avg % Abs = 1.9290 (-0.0077)
STD DEV = 0.0234 (-0.0142)
REL STD DEV = 1.212 (184.616)

Sol Value = 0.000 g/210L ***
Fit Value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12880, Sum Io = 13438
***** CHANNEL 1 *****
Sample % Abs (% Abs Ref)
Sample #1 = 0.1370 (-0.0060)
Sample #2 = 0.1340 (-0.0030)
Sample #3 = 0.1390 (-0.0000)
Sample #4 = 0.1340 (-0.0010)
Avg % Abs = 0.1357 (-0.0013)
STD DEV = 0.0029 (-0.0015)
REL STD DEV = 0.128 (114.564)

***** CHANNEL 2 *****
Sample % Abs (% Abs Ref)
Sample #1 = 3.6110 (-0.0010)
Sample #2 = 3.5490 (-0.0610)
Sample #3 = 3.5170 (-0.0750)
Sample #4 = 3.5190 (-0.0860)
Avg % Abs = 3.5280 (-0.0740)
STD DEV = 0.0182 (-0.0125)
REL STD DEV = 0.516 (16.932)

Sol Value = 0.200 g/210L ***
Fit Value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12894, Sum Io = 13433
***** CHANNEL 1 *****
Sample % Abs (% Abs Ref)
Sample #1 = 3.7720 (-0.0170)
Sample #2 = 3.6910 (-0.0780)
Sample #3 = 3.6710 (-0.0820)
Sample #4 = 3.6710 (-0.1050)
Avg % Abs = 3.6777 (-0.0683)
STD DEV = 0.0115 (-0.0145)
REL STD DEV = 0.314 (116.496)

Sol Value = 0.040 g/210L ***
Fit Value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12883, Sum Io = 13437
***** CHANNEL 1 *****
Sample % Abs (% Abs Ref)
Sample #1 = 1.6680 (-0.0200)
Sample #2 = 1.6670 (-0.0210)
Sample #3 = 0.6670 (-0.0140)
Sample #4 = 0.6690 (-0.0000)
Avg % Abs = 0.8677 (-0.0117)
STD DEV = 0.0012 (-0.0107)
REL STD DEV = 0.133 (91.652)

***** CHANNEL 2 *****
Sample % Abs (% Abs Ref)
Sample #1 = 6.9220 (-0.0070)
Sample #2 = 6.7880 (-0.1500)
Sample #3 = 6.7750 (-0.1620)
Sample #4 = 6.7540 (-0.1780)
Avg % Abs = 6.7723 (-0.1633)
STD DEV = 0.0172 (-0.0140)
REL STD DEV = 0.253 (8.601)

Sol Value = 0.300 g/210L ***
Fit Value = 1.4286 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12898, Sum Io = 13430
***** CHANNEL 1 *****
Sample % Abs (% Abs Ref)
Sample #1 = 5.5160 (-0.0200)
Sample #2 = 5.3910 (-0.1070)
Sample #3 = 5.3980 (-0.1000)
Sample #4 = 5.3770 (-0.1270)
Avg % Abs = 5.3867 (-0.1113)
STD DEV = 0.0107 (-0.0140)
REL STD DEV = 0.198 (12.586)

***** CHANNEL 2 *****
Sample % Abs (% Abs Ref)
Sample #1 = 10.1050 (-0.0190)
Sample #2 = 9.8850 (-0.1910)
Sample #3 = 9.8600 (-0.2140)
Sample #4 = 9.8700 (-0.2230)
Avg % Abs = 9.8717 (-0.2793)
STD DEV = 0.0126 (-0.0165)
REL STD DEV = 0.127 (7.683)

***** AUTO CAL DATA *****
***** CHANNEL 1 *****
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.136
Std Dev = 0.00 Rel Std Dev = 2.13
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.868
Std Dev = 0.00 Rel Std Dev = 0.13
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.929
Std Dev = 0.02 Rel Std Dev = 1.21
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.678
Std Dev = 0.01 Rel Std Dev = 0.31
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.389
Std Dev = 0.01 Rel Std Dev = 0.20
Zero Order Coef = -365.76
First Order Coef = 2620.31
Second Order Coef = 18.38
Standard Deviation = 10.358184

***** CHANNEL 2 *****
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.137
Std Dev = 0.02 Rel Std Dev = 13.60
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.514
Std Dev = 0.01 Rel Std Dev = 0.63
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.528
Std Dev = 0.02 Rel Std Dev = 0.52
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.772
Std Dev = 0.02 Rel Std Dev = 0.25
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 9.872
Std Dev = 0.01 Rel Std Dev = 0.13
Zero Order Coef = -165.63
First Order Coef = 1365.38
Second Order Coef = 10.17
Standard Deviation = 2.906255

Solution Stats Quadratic Fit Chan 1
Act Fit Residual
g/210L g/210L g/210L
0.000 0.000 -0.0000
0.040 0.040 -0.0014
0.100 0.100 0.0000
0.200 0.200 0.0001
0.300 0.300 -0.0000

Optical Calibration Adjustment

By: TDG

Solution Stats Quadratic Fit Chan 1
Act Fit Residual
g/210L g/210L g/210L
0.000 0.000 0.0002
0.040 0.040 -0.0014
0.100 0.100 0.0000
0.200 0.200 0.0001
0.300 0.300 -0.0000

Solution Stats Quadratic Fit Chan 2
Act Fit Residual
g/210L g/210L g/210L
0.000 0.000 -0.0000
0.040 0.040 -0.0000
0.100 0.100 0.0001
0.200 0.200 -0.0001
0.300 0.300 0.0000

Sol Value = 0.080 g/210L ***
Fit Value = 0.3810 mg/l %%%
Samples Taken = 4, Discarded = 1
***** CHANNEL 1 *****
Sample #1 = 3006.00
Sample #2 = 2890.00
Sample #3 = 2971.00
Sample #4 = 3029.00
Average Result = 2963.3333
STD DEV = 69.8164
REL STD DEV = 2.356

***** CHANNEL 2 *****
Sample #1 = 3415.00
Sample #2 = 3393.00
Sample #3 = 3415.00
Sample #4 = 3470.60
Average Result = 3426.0000
STD DEV = 39.6611
REL STD DEV = 1.158

Dry Gas H2O Adjust Results *****
Barometric Pressure = 1018
3 um H2O Adjust (mg/l*10.000) = 846
5 um H2O Adjust (mg/l*10.000) = 363
**** AUTO CAL PASS

Post-Cal Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																				
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																				
✓	✓	✓	✓																																																																																				
≤0.003 of Wet	✓	✓	≤0.003 of Wet																																																																																				
✓	✓	✓	✓																																																																																				
MIAMI PD Intoxilyzer - Alcohol Analyzer Model: 8000 06/20/2025 Software: 8100.27 SN 80-000874 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>14:41</td></tr> <tr><td>Control Test</td><td>0.051</td><td>14:42</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:43</td></tr> <tr><td>Control Test</td><td>0.050</td><td>14:43</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:44</td></tr> <tr><td>Control Test</td><td>0.050</td><td>14:45</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:45</td></tr> </table> Control Test Stats Average 0.0503 Std Dev 0.0006 Rel Std Dev(%) 1.1471	Air Blank	0.000	14:41	Control Test	0.051	14:42	Air Blank	0.000	14:43	Control Test	0.050	14:43	Air Blank	0.000	14:44	Control Test	0.050	14:45	Air Blank	0.000	14:45	MIAMI PD Intoxilyzer - Alcohol Analyzer Model: 8000 06/20/2025 Software: 8100.27 SN 80-000874 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>14:57</td></tr> <tr><td>Control Test</td><td>0.082</td><td>14:57</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:58</td></tr> <tr><td>Control Test</td><td>0.082</td><td>14:59</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:59</td></tr> <tr><td>Control Test</td><td>0.080</td><td>15:00</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:00</td></tr> </table> Control Test Stats Average 0.0813 Std Dev 0.0012 Rel Std Dev(%) 1.4197	Air Blank	0.000	14:57	Control Test	0.082	14:57	Air Blank	0.000	14:58	Control Test	0.082	14:59	Air Blank	0.000	14:59	Control Test	0.080	15:00	Air Blank	0.000	15:00	MIAMI PD Intoxilyzer - Alcohol Analyzer Model: 8000 06/20/2025 Software: 8100.27 SN 80-000874 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>15:06</td></tr> <tr><td>Control Test</td><td>0.203</td><td>15:07</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:07</td></tr> <tr><td>Control Test</td><td>0.200</td><td>15:08</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:09</td></tr> <tr><td>Control Test</td><td>0.199</td><td>15:09</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:10</td></tr> </table> Control Test Stats Average 0.2007 Std Dev 0.0021 Rel Std Dev(%) 1.0374	Air Blank	0.000	15:06	Control Test	0.203	15:07	Air Blank	0.000	15:07	Control Test	0.200	15:08	Air Blank	0.000	15:09	Control Test	0.199	15:09	Air Blank	0.000	15:10	MIAMI PD Intoxilyzer - Alcohol Analyzer Model: 8000 06/20/2025 Software: 8100.27 SN 80-000874 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>14:31</td></tr> <tr><td>Control Test</td><td>0.079</td><td>14:32</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:32</td></tr> <tr><td>Control Test</td><td>0.079</td><td>14:33</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:33</td></tr> <tr><td>Control Test</td><td>0.079</td><td>14:33</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:34</td></tr> </table> Control Test Stats Average 0.0793 Std Dev 0.0000 Rel Std Dev(%) 0.0000	Air Blank	0.000	14:31	Control Test	0.079	14:32	Air Blank	0.000	14:32	Control Test	0.079	14:33	Air Blank	0.000	14:33	Control Test	0.079	14:33	Air Blank	0.000	14:34
Air Blank	0.000	14:41																																																																																					
Control Test	0.051	14:42																																																																																					
Air Blank	0.000	14:43																																																																																					
Control Test	0.050	14:43																																																																																					
Air Blank	0.000	14:44																																																																																					
Control Test	0.050	14:45																																																																																					
Air Blank	0.000	14:45																																																																																					
Air Blank	0.000	14:57																																																																																					
Control Test	0.082	14:57																																																																																					
Air Blank	0.000	14:58																																																																																					
Control Test	0.082	14:59																																																																																					
Air Blank	0.000	14:59																																																																																					
Control Test	0.080	15:00																																																																																					
Air Blank	0.000	15:00																																																																																					
Air Blank	0.000	15:06																																																																																					
Control Test	0.203	15:07																																																																																					
Air Blank	0.000	15:07																																																																																					
Control Test	0.200	15:08																																																																																					
Air Blank	0.000	15:09																																																																																					
Control Test	0.199	15:09																																																																																					
Air Blank	0.000	15:10																																																																																					
Air Blank	0.000	14:31																																																																																					
Control Test	0.079	14:32																																																																																					
Air Blank	0.000	14:32																																																																																					
Control Test	0.079	14:33																																																																																					
Air Blank	0.000	14:33																																																																																					
Control Test	0.079	14:33																																																																																					
Air Blank	0.000	14:34																																																																																					
Operator's Signature 	Operator's Signature 	Operator's Signature 	Operator's Signature 																																																																																				

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MIAMI PD

Time of Inspection: 12:43

Date of Inspection: 06/25/2025

Serial Number: 80-000874

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG429602 Exp: 10/22/2026
0.000	0.051	0.082	0.204	0.079
0.000	0.051	0.082	0.205	0.078
0.000	0.051	0.082	0.204	0.079
0.000	0.051	0.082	0.205	0.078
0.000	0.051	0.081	0.205	0.078
0.000	0.051	0.082	0.205	0.079
0.000	0.052	0.082	0.205	0.079
0.000	0.051	0.082	0.205	0.079
0.000	0.051	0.082	0.204	0.079
0.000	0.052	0.082	0.205	0.079


Standard Deviations	0.0004	0.0003	0.0004	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

06/25/2025
Date

Return Material Authorization

Ship to: ☒ CMI, Inc.

☐ Enforcement Electronics

Shipment to repair facility authorized by: Angel Fernandez on 6/25/2025

Items Returned: Instrument ☒ Supplies ☐ Other ☐ Describe: _____

Instrument Model: Intoxilyzer 8000 Serial Number: 80-000874

Bill To Address:

Miami Police Department

Attn: Angel Fernandez

Ship to Address:

Florida Department of Law Enforcement

Fort Myers Regional Operations Center

Attn: Taylor Gutschow

4700 Terminal Drive, Suite 1

Fort Myers, FL 33907

Reason for Return:

Instrument just returned from repair (work order 408888). R-value is near 100.

Please choose one of the following options:

- ☐ 1. I _____, authorize all repairs.
- ☐ 2. I _____, authorize repairs up to \$_____.
- ☒ 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Angel Fernandez

Phone #: 305-710-3239 Email: 28394@miami-police.org

ATP Contact Name: Taylor Gutschow ATP Email: TaylorGutschow@fdle.state.fl.us



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000874, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000874</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>MIAMI PD</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>06/25/2025</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>12:43</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Taylor
Gutschow
Digitally signed by Taylor
Gutschow
Date: 2025.06.27 11:42:41
-04'00'

06/25/2025

Date

TAYLOR D GUTSCHOW,
Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality