



INSTRUMENT PROCESSING SHEET

Agency Calhoun County SOS/N 80-000875Florida Department of
Law EnforcementDate In 01/24/2025 DI Completion Date 02/25/2025☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake By <u>ALL</u> Date <u>01/24/2025</u>		Quality Checks By <u>DA</u> Date <u>02/14/2025</u>		Flow Calibration By <u>DA</u> Date <u>02/25/2025</u>																																																													
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: <u>dropoff no box</u>		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>235</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP102</u> 32 mm <u>0.144</u> (.139 - .169) 36 mm <u>0.156</u> (.156 - .190) 53 mm <u>0.226</u> (.228 - .278) 103 mm <u>0.484</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>30793</u> ☒ Stability Checks		☒ Flow Column # <u>ATP105</u> ☒ 5L/min – 17mm ☒ 15L/min – 53mm ☒ 30L/min – 103mm ☒ R-Value <u>234</u> ☒ Post Calibration Verification (L/s) Flow Column # <u>ATP102</u> 32 mm <u>0.142</u> (.139 - .169) 36 mm <u>0.164</u> (.156 - .190) 53 mm <u>0.230</u> (.228 - .278) 103 mm <u>0.488</u> (.447 - .547)																																																													
		<table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP5088</td><td>202406K 06/19/2026</td></tr><tr><td>0.080</td><td>MP5089</td><td>202406L 06/19/2026</td></tr><tr><td>0.200</td><td>MP5090</td><td>202406N 06/20/2026</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG429602 10/22/2026</td></tr></tbody></table>		Simulator	Serial #	Lot #/Exp	0.050	MP5088	202406K 06/19/2026	0.080	MP5089	202406L 06/19/2026	0.200	MP5090	202406N 06/20/2026	0.080 DGS	N/A	AG429602 10/22/2026	Maintenance By _____ Date _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____																																														
Simulator	Serial #	Lot #/Exp																																																															
0.050	MP5088	202406K 06/19/2026																																																															
0.080	MP5089	202406L 06/19/2026																																																															
0.200	MP5090	202406N 06/20/2026																																																															
0.080 DGS	N/A	AG429602 10/22/2026																																																															
Calibration Adjustment By _____				Department Inspection By <u>DA</u>																																																													
Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table> ☐ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>				Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Barometric Pressure ID# <u>30793</u> Gauge <u>1011</u> Instrument <u>1010</u> Mouth Alcohol Solution Lot # <u>2024-A</u> Acetone Stock Solution Lot # <u>2024-B</u> <table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6290</td></tr><tr><td>Interferent</td><td>MP6292</td></tr><tr><td>0.050</td><td>MP5088</td></tr><tr><td>0.080</td><td>MP5090</td></tr><tr><td>0.200</td><td>MP6289</td></tr></tbody></table>		Simulator	Serial Number	0.000	MP6290	Interferent	MP6292	0.050	MP5088	0.080	MP5090	0.200	MP6289
Simulator	Serial #	Lot #	Expiration																																																														
0.000		N/A	N/A																																																														
0.040																																																																	
0.100																																																																	
0.200																																																																	
0.300																																																																	
0.080 DGS	N/A																																																																
Simulator	Serial #	Lot #	Expiration																																																														
0.050																																																																	
0.080																																																																	
0.200																																																																	
0.080 DGS	N/A																																																																
Simulator	Serial Number																																																																
0.000	MP6290																																																																
Interferent	MP6292																																																																
0.050	MP5088																																																																
0.080	MP5090																																																																
0.200	MP6289																																																																
Notes/Suggested Service: _____ _____ _____ _____ _____ _____				Attachments ☒ Form 41 ☐ Post-Stability Checks ☒ Stability Checks ☒ Flow Calibration ☒ Calibration Certificate ☐ Form 40 ☐ Calibration Adjustment ☐ Other _____																																																													
				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use Digitally signed by Taylor Gutschow Date: 2025.03.04 09:41:03 -05'00' Shayla Platt Date: 2025.03.04 13:02:18 -05'00' Tech Review / Date _____ Admin Review / Date _____																																																													

Stability Checks 80-000875

2/14/25

DA

CALHOUN COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000875
02/14/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:09
Control Test	0.049	12:10
Air Blank	0.000	12:11
Control Test	0.049	12:11
Air Blank	0.000	12:12
Control Test	0.050	12:13
Air Blank	0.000	12:13
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel Std Dev(%)	1.1703	

DA

Operator's Signature

CALHOUN COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000875
02/14/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:59
Control Test	0.081	11:59
Air Blank	0.000	12:00
Control Test	0.081	12:01
Air Blank	0.000	12:01
Control Test	0.080	12:02
Air Blank	0.000	12:02
Control Test Stats		
Average	0.0807	
Std Dev	0.0006	
Rel Std Dev(%)	0.7157	

Wet

DA

Operator's Signature

CALHOUN COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000875
02/14/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:04
Control Test	0.203	12:05
Air Blank	0.000	12:06
Control Test	0.201	12:06
Air Blank	0.000	12:07
Control Test	0.201	12:08
Air Blank	0.000	12:08
Control Test Stats		
Average	0.2017	
Std Dev	0.0012	
Rel Std Dev(%)	0.5726	

DA

Operator's Signature

CALHOUN COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000875
02/14/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:18
Control Test	0.080	12:18
Air Blank	0.000	12:19
Control Test	0.080	12:19
Air Blank	0.000	12:20
Control Test	0.080	12:20
Air Blank	0.000	12:21
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS

DA

Operator's Signature

Stability Checks

80-000875

DA

2/14/25

CALHOUN COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000875
02/14/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:56
Control Test	0.082	11:56
Air Blank	PUR*	11:57
Air Blank	0.000	11:58

*Purge Fail

wet



Operator's Signature

CALHOUN COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000875
02/14/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:15
Control Test	0.000	12:15
Air Blank	0.000	12:16
Control Test	0.000	12:16
Air Blank	0.000	12:16
Control Test	0.000	12:17
Air Blank	0.000	12:17
Control Test Stats		
Average	0.0000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS



Operator's Signature

Purge fail due to flow column
still being attached to breath
hose. I removed and repeated
0.05g/210L stability check with
nominal valves.

DA 2/14/25

DGS was not attached
to calibration inlet.
Ensured connection and
repeated DGS stability
check with nominal
results.

DA 2/14/25

Flow Rate Calibration Adjustment 80-000875 2/25/25

CALHOUN COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000875
02/25/2025
Software: 8100.27

Flow Rate Calibration*****

1: Rate (Liters/min) = 5

 SQRT(Diff)) = 7.547

2: Rate (Liters/min) = 15

 SQRT(Diff)) = 12.246

3: Rate (Liters/min) = 30

 SQRT(Diff)) = 22.113

Dependent Data Scale Factor = 100000 L/min

Independent Data Scale Factor = 256

Rounded Slope = 659

Rounded Intercept = -688639

Correlation = 0.99621

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: CALHOUN COUNTY SO
Time of Inspection: 11:22

Date of Inspection: 02/25/2025

Serial Number: 80-000875
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG429602 Exp: 10/22/2026
0.000	0.049	0.081	0.202	0.079
0.000	0.050	0.081	0.202	0.080
0.000	0.049	0.081	0.203	0.079
0.000	0.050	0.081	0.202	0.079
0.000	0.050	0.081	0.202	0.079
0.000	0.050	0.081	0.202	0.079
0.000	0.050	0.081	0.202	0.079
0.000	0.050	0.081	0.203	0.079
0.000	0.049	0.080	0.202	0.079
0.000	0.050	0.081	0.203	0.079

Standard Deviations	0.0004	0.0003	0.0004	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



DESTINEE N ARMSTRONG

Signature and Printed Name

02/25/2025
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000875, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000875</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>CALHOUN COUNTY SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>02/25/2025</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>11:22</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

02/25/2025

Date

DESTINEE N ARMSTRONG,

Department Inspector

FDLE/ATP Form 69 October 2024

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality