



INSTRUMENT PROCESSING SHEET

Agency Miami Dade SOS/N 80-000882Florida Department of
Law EnforcementDate In 06/09/2025 DI Completion Date 07/11/2025☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake By <u>TDG</u> Date <u>07/10/2025</u>		Quality Checks By <u>TDG</u> Date <u>07/11/2025</u>		Flow Calibration By _____ Date _____																																																															
☒ Annual ☒ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: <u>Agency Inspector Closius reports the instrument displayed "Forms Not Found". Registration needs to be updated.</u>		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>138</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>0.152</u> (.139 - .169) 36 mm <u>0.164</u> (.156 - .190) 53 mm <u>0.242</u> (.228 - .278) 103 mm <u>0.511</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28199</u> ☒ Stability Checks		☐ Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																																															
		<table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6286</td><td>202406K 06/19/2026</td></tr><tr><td>0.080</td><td>MP6287</td><td>202406L 06/19/2026</td></tr><tr><td>0.200</td><td>MP6288</td><td>202406N 06/20/2026</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG429602 10/22/2026</td></tr></tbody></table>		Simulator	Serial #	Lot #/Exp	0.050	MP6286	202406K 06/19/2026	0.080	MP6287	202406L 06/19/2026	0.200	MP6288	202406N 06/20/2026	0.080 DGS	N/A	AG429602 10/22/2026	Maintenance By <u>TDG</u> Date <u>07/10/2025</u> ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☒ Other <u>CMI reloaded forms</u>																																																
Simulator	Serial #	Lot #/Exp																																																																	
0.050	MP6286	202406K 06/19/2026																																																																	
0.080	MP6287	202406L 06/19/2026																																																																	
0.200	MP6288	202406N 06/20/2026																																																																	
0.080 DGS	N/A	AG429602 10/22/2026																																																																	
Calibration Adjustment By _____				Department Inspection By <u>TDG</u>																																																															
Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table> ☐ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>				Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Barometric Pressure ID# <u>28199</u> Gauge <u>1020</u> Instrument <u>1019</u> Mouth Alcohol Solution Lot # <u>2024-A</u> Acetone Stock Solution Lot # <u>2023-B</u> <table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6284</td></tr><tr><td>Interferent</td><td>MP6285</td></tr><tr><td>0.050</td><td>MP6286</td></tr><tr><td>0.080</td><td>MP6287</td></tr><tr><td>0.200</td><td>MP6288</td></tr></tbody></table> Attachments <table border="1" style="width:100%"><tbody><tr><td>☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment</td><td>☐ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☒ Other <u>Form 47</u></td></tr></tbody></table>		Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288	☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment	☐ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☒ Other <u>Form 47</u>
Simulator	Serial #	Lot #	Expiration																																																																
0.000		N/A	N/A																																																																
0.040																																																																			
0.100																																																																			
0.200																																																																			
0.300																																																																			
0.080 DGS	N/A																																																																		
Simulator	Serial #	Lot #	Expiration																																																																
0.050																																																																			
0.080																																																																			
0.200																																																																			
0.080 DGS	N/A																																																																		
Simulator	Serial Number																																																																		
0.000	MP6284																																																																		
Interferent	MP6285																																																																		
0.050	MP6286																																																																		
0.080	MP6287																																																																		
0.200	MP6288																																																																		
☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment	☐ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☒ Other <u>Form 47</u>																																																																		
Notes/Suggested Service: _____ _____ _____ _____ _____ _____				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use Phil Nicodemo Digitally signed by Phil Nicodemo Date: 2025.07.18 12:05:34 -04'00' Shayla Platt Shayla Platt Date: 2025.07.26 11:38:19 -04'00' Tech Review / Date _____ Admin Review / Date _____																																																															



Florida Department of
Law Enforcement

REQUEST FOR REGISTRATION

MAKE AND MODEL OF INSTRUMENT: Intoxilyzer 8000

SERIAL NUMBER: 80-000882

OWNING AGENCY: Miami Dade Sheriff's Office

DATE OF DEPARTMENT INSPECTION: 07/11/2025

AGENCY INSPECTOR: Richard Closius

ADDRESS: 1519 NW 79th Avenue

CITY, STATE, ZIP: Miami, FL 33126

TELEPHONE NUMBER: 305-458-3991

FAX NUMBER: N/A

EMAIL ADDRESS (if available): rlclosius@mdso.com

For Program Office Use Only:

- ☐ Registration Issued
- ☒ Instrument Added to Evidentiary Instrument Database
- ☐ Instrument Added to Monthly Statistics Database
- ☒ Contact Information Added to Instrument Database

Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
MIAMI DDOE SO Intoxilyzer - Alcohol Analyzer Model 8100 07/11/2025 Software: 8100.27 SN 80-000862 Test g/210L Time Air Blank 0.000 12:52 Control Test 0.049 12:52 Air Blank 0.000 12:53 Control Test 0.048 12:54 Air Blank 0.000 12:54 Control Test 0.048 12:55 Air Blank 0.000 12:55 Control Test Stats Average 0.0483 Std Dev 0.0006 Rel Std Dev(%) 1.1945 Operator's Signature 	MIAMI DDOE SO Intoxilyzer - Alcohol Analyzer Model 8100 07/11/2025 Software: 8100.27 SN 80-000862 Test g/210L Time Air Blank 0.000 12:37 Control Test 0.079 12:37 Air Blank 0.000 12:38 Control Test 0.077 12:39 Air Blank 0.000 12:39 Control Test 0.078 12:40 Air Blank 0.000 12:41 Control Test Stats Average 0.0780 Std Dev 0.0010 Rel Std Dev(%) 1.2821 Operator's Signature 	MIAMI DDOE SO Intoxilyzer - Alcohol Analyzer Model 8100 07/11/2025 Software: 8100.27 SN 80-000862 Test g/210L Time Air Blank 0.000 12:45 Control Test 0.200 12:45 Air Blank 0.000 12:46 Control Test 0.198 12:47 Air Blank 0.000 12:47 Control Test 0.198 12:48 Air Blank 0.000 12:49 Control Test Stats Average 0.1987 Std Dev 0.0012 Rel Std Dev(%) 0.5812 Operator's Signature 	MIAMI DDOE SO Intoxilyzer - Alcohol Analyzer Model 8100 07/11/2025 Software: 8100.27 SN 80-000862 Test g/210L Time Air Blank 0.000 12:56 Control Test 0.079 12:57 Air Blank 0.000 12:57 Control Test 0.079 12:58 Air Blank 0.000 12:58 Control Test 0.079 12:58 Air Blank 0.000 12:59 Control Test Stats Average 0.0790 Std Dev 0.0000 Rel Std Dev(%) 0.0000 Operator's Signature 

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MIAMI DADE SO

Time of Inspection: 15:59

Date of Inspection: 07/11/2025

Serial Number: 80-000882

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG429602 Exp: 10/22/2026
0.000	0.048	0.079	0.198	0.080
0.000	0.049	0.079	0.198	0.079
0.000	0.048	0.079	0.198	0.079
0.000	0.049	0.078	0.198	0.080
0.000	0.049	0.079	0.198	0.080
0.000	0.049	0.079	0.198	0.079
0.000	0.048	0.078	0.198	0.079
0.000	0.049	0.079	0.198	0.079
0.000	0.049	0.079	0.199	0.079
0.000	0.049	0.079	0.198	0.079


Standard Deviations	0.0004	0.0004	0.0003	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

07/11/2025
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000882, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000882</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>MIAMI DADE SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>07/11/2025</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>15:59</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Taylor
Gutschow
Digitally signed by Taylor
Gutschow
Date: 2025.07.14 10:18:23
-04'00'

07/11/2025

Date

TAYLOR D GUTSCHOW,
Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality