



INSTRUMENT PROCESSING SHEET

Agency Pinellas CSO

S/N 80-000889

Florida Department of
Law EnforcementDate In 03/24/2025 DI Completion Date 04/23/2025☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By TDG	Date	Quality Checks	By TDG	Date (See Notes)	Flow Calibration	By	Date																																					
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: <u>Needs battery change.</u>		<u>03/27/2025</u>	☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>153</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP106</u> 32 mm <u>0.144</u> (.139 - .169) 36 mm <u>0.160</u> (.156 - .190) 53 mm <u>0.234</u> (.228 - .278) 103 mm <u>0.500</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28199</u> ☒ Stability Checks			Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																							
			<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6286</td><td>202406K 06/19/2026</td></tr><tr><td>0.080</td><td>MP6287</td><td>202406L 06/19/2026</td></tr><tr><td>0.200</td><td>MP6288</td><td>202406N 06/20/2026</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG429602 10/22/2026</td></tr></tbody></table>			Simulator	Serial #	Lot #/Exp	0.050	MP6286	202406K 06/19/2026	0.080	MP6287	202406L 06/19/2026	0.200	MP6288	202406N 06/20/2026	0.080 DGS	N/A	AG429602 10/22/2026	Maintenance By TDG <u> </u> Date (See below) <u> </u> ☒ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ <u>Replaced battery on 4/15. Instrument did not hold date/time. Checked work on 4/21 and secured battery when it was observed to be loose. Instrument now holds date/time. (TDG 4/21/25)</u>																								
Simulator	Serial #	Lot #/Exp																																											
0.050	MP6286	202406K 06/19/2026																																											
0.080	MP6287	202406L 06/19/2026																																											
0.200	MP6288	202406N 06/20/2026																																											
0.080 DGS	N/A	AG429602 10/22/2026																																											
Calibration Adjustment <u> </u> By <u> </u> Barometric Pressure Gauge <u> </u> ID # <u> </u>			Department Inspection <u> </u> By <u> </u> Barometric Pressure ID# <u>28199</u> Gauge <u>1020</u> Instrument <u>1015</u> Mouth Alcohol Solution Lot # <u>2024-A</u> Acetone Stock Solution Lot # <u>2023-B</u>																																										
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>			Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6284</td></tr><tr><td>Interferent</td><td>MP6285</td></tr><tr><td>0.050</td><td>MP6286</td></tr><tr><td>0.080</td><td>MP6287</td></tr><tr><td>0.200</td><td>MP6288</td></tr></tbody></table>			Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288
Simulator	Serial #	Lot #	Expiration																																										
0.000		N/A	N/A																																										
0.040																																													
0.100																																													
0.200																																													
0.300																																													
0.080 DGS	N/A																																												
Simulator	Serial Number																																												
0.000	MP6284																																												
Interferent	MP6285																																												
0.050	MP6286																																												
0.080	MP6287																																												
0.200	MP6288																																												
☐ Post Calibration Adjustment Stability Checks			Attachments ☒ Form 41 ☐ Post-Stability Checks ☒ Stability Checks ☐ Flow Calibration ☒ Calibration Certificate ☐ Form 40 ☐ Calibration Adjustment ☐ Other _____																																										
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A																									
Simulator	Serial #	Lot #	Expiration																																										
0.050																																													
0.080																																													
0.200																																													
0.080 DGS	N/A																																												
Notes/Suggested Service: <u>Checked breath tube screen and replaced o-rings on 3/27. (TDG 3/27/25)</u>			☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																																										
<u>Verified setup during battery change. (TDG 4/15/25)</u>			Phil Nicodemo Digitally signed by Phil Nicodemo Date: 2025.05.02 12:48:43 +04'00'																																										
<u>Finished the Quality Checks on 4/22. (TDG 4/22/25)</u>			Shayla Platt Digitally signed by Shayla Platt Date: 2025.05.03 10:22:27 -04'00'																																										
			Tech Review / Date <u> </u> Admin Review / Date <u> </u>																																										

Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																																
✓	✓	✓	✓																																																																																																																																																
0.003 of Wet			0.003 of Wet																																																																																																																																																
✓			✓																																																																																																																																																
PINELLAS COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000889 04/22/2025 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:36</td></tr> <tr><td>Control Test</td><td>0.049</td><td>14:37</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:37</td></tr> <tr><td>Control Test</td><td>0.048</td><td>14:38</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:39</td></tr> <tr><td>Control Test</td><td>0.048</td><td>14:39</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:40</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0483</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1945</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	14:36	Control Test	0.049	14:37	Air Blank	0.000	14:37	Control Test	0.048	14:38	Air Blank	0.000	14:39	Control Test	0.048	14:39	Air Blank	0.000	14:40	Control Test Stats			Average	0.0483		Std Dev	0.0006		Rel Std Dev(%)	1.1945		PINELLAS COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000889 04/22/2025 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:43</td></tr> <tr><td>Control Test</td><td>0.081</td><td>14:43</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:44</td></tr> <tr><td>Control Test</td><td>0.079</td><td>14:45</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:45</td></tr> <tr><td>Control Test</td><td>0.079</td><td>14:46</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:47</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0797</td><td></td></tr> <tr><td>Std Dev</td><td>0.0012</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.4494</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	14:43	Control Test	0.081	14:43	Air Blank	0.000	14:44	Control Test	0.079	14:45	Air Blank	0.000	14:45	Control Test	0.079	14:46	Air Blank	0.000	14:47	Control Test Stats			Average	0.0797		Std Dev	0.0012		Rel Std Dev(%)	1.4494		PINELLAS COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000889 04/22/2025 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:29</td></tr> <tr><td>Control Test</td><td>0.203</td><td>14:30</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:30</td></tr> <tr><td>Control Test</td><td>0.199</td><td>14:31</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:32</td></tr> <tr><td>Control Test</td><td>0.198</td><td>14:32</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:33</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.2000</td><td></td></tr> <tr><td>Std Dev</td><td>0.0026</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.3229</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	14:29	Control Test	0.203	14:30	Air Blank	0.000	14:30	Control Test	0.199	14:31	Air Blank	0.000	14:32	Control Test	0.198	14:32	Air Blank	0.000	14:33	Control Test Stats			Average	0.2000		Std Dev	0.0026		Rel Std Dev(%)	1.3229		PINELLAS COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000889 04/22/2025 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:50</td></tr> <tr><td>Control Test</td><td>0.078</td><td>14:50</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:50</td></tr> <tr><td>Control Test</td><td>0.078</td><td>14:51</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:51</td></tr> <tr><td>Control Test</td><td>0.080</td><td>14:52</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:52</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0787</td><td></td></tr> <tr><td>Std Dev</td><td>0.0012</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.4578</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	14:50	Control Test	0.078	14:50	Air Blank	0.000	14:50	Control Test	0.078	14:51	Air Blank	0.000	14:51	Control Test	0.080	14:52	Air Blank	0.000	14:52	Control Test Stats			Average	0.0787		Std Dev	0.0012		Rel Std Dev(%)	1.4578	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:36																																																																																																																																																	
Control Test	0.049	14:37																																																																																																																																																	
Air Blank	0.000	14:37																																																																																																																																																	
Control Test	0.048	14:38																																																																																																																																																	
Air Blank	0.000	14:39																																																																																																																																																	
Control Test	0.048	14:39																																																																																																																																																	
Air Blank	0.000	14:40																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0483																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1945																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:43																																																																																																																																																	
Control Test	0.081	14:43																																																																																																																																																	
Air Blank	0.000	14:44																																																																																																																																																	
Control Test	0.079	14:45																																																																																																																																																	
Air Blank	0.000	14:45																																																																																																																																																	
Control Test	0.079	14:46																																																																																																																																																	
Air Blank	0.000	14:47																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0797																																																																																																																																																		
Std Dev	0.0012																																																																																																																																																		
Rel Std Dev(%)	1.4494																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:29																																																																																																																																																	
Control Test	0.203	14:30																																																																																																																																																	
Air Blank	0.000	14:30																																																																																																																																																	
Control Test	0.199	14:31																																																																																																																																																	
Air Blank	0.000	14:32																																																																																																																																																	
Control Test	0.198	14:32																																																																																																																																																	
Air Blank	0.000	14:33																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.2000																																																																																																																																																		
Std Dev	0.0026																																																																																																																																																		
Rel Std Dev(%)	1.3229																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:50																																																																																																																																																	
Control Test	0.078	14:50																																																																																																																																																	
Air Blank	0.000	14:50																																																																																																																																																	
Control Test	0.078	14:51																																																																																																																																																	
Air Blank	0.000	14:51																																																																																																																																																	
Control Test	0.080	14:52																																																																																																																																																	
Air Blank	0.000	14:52																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0787																																																																																																																																																		
Std Dev	0.0012																																																																																																																																																		
Rel Std Dev(%)	1.4578																																																																																																																																																		

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PINELLAS COUNTY SO
Time of Inspection: 13:31

Date of Inspection: 04/23/2025

Serial Number: 80-000889
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG429602 Exp: 10/22/2026
0.000	0.049	0.079	0.200	0.078
0.000	0.049	0.079	0.200	0.077
0.000	0.049	0.078	0.200	0.077
0.000	0.049	0.079	0.200	0.077
0.000	0.049	0.079	0.200	0.077
0.000	0.048	0.078	0.200	0.078
0.000	0.049	0.078	0.200	0.078
0.000	0.049	0.078	0.200	0.077
0.000	0.049	0.079	0.200	0.077
0.000	0.049	0.078	0.200	0.077
Standard Deviations	0.0003	0.0005	0.0000	0.0004

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

04/23/2025
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000889, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000889</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>PINELLAS COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>04/23/2025</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>13:31</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

04/23/2025

Date

Taylor D Gutschow

TAYLOR D GUTSCHOW,
Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality