



INSTRUMENT PROCESSING SHEET

Agency FFWCCS/N 80-000904Florida Department of
Law EnforcementDate In 06/02/2024 DI Completion Date 07/17/2025☐ Ship ☐ P/U ☐ H/D ☒ CMI ☐ EE

Intake	By	TDG	Date	Quality Checks	By	TDG	Date	Flow Calibration	By	TDG	Date																												
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☒ Power cord ☒ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: <u>Oscar Saavedra dropped off in pelican case with DGS and mouthpieces. DGS expired on 6/27/25. Discussed w/ Saavedra and discarded DGS to prevent accidental use. Needs battery change.</u>			<u>07/08/2025</u>	☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>96*</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP106</u> 32 mm <u>0.121**</u> (.139 - .169) 36 mm <u>0.140**</u> (.156 - .190) 53 mm <u>0.226**</u> (.228 - .278) 103 mm <u>0.492</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28199</u> ☒ Stability Checks				Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																															
				<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6286</td><td>202406K 06/19/2026</td></tr><tr><td>0.080</td><td>MP6287</td><td>202406L 06/19/2026</td></tr><tr><td>0.200</td><td>MP6288</td><td>202406N 06/20/2026</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG429602 10/22/2026</td></tr></tbody></table>			Simulator	Serial #	Lot #/Exp	0.050	MP6286	202406K 06/19/2026	0.080	MP6287	202406L 06/19/2026	0.200	MP6288	202406N 06/20/2026	0.080 DGS	N/A	AG429602 10/22/2026	<table border="1"><thead><tr><th colspan="2">Maintenance</th><th>By</th><th>TDG</th><th>Date</th></tr></thead><tbody><tr><td colspan="2">☒ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____</td><td></td><td></td><td><u>07/14/2025</u></td></tr></tbody></table>				Maintenance		By	TDG	Date	☒ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____				<u>07/14/2025</u>				
Simulator	Serial #	Lot #/Exp																																					
0.050	MP6286	202406K 06/19/2026																																					
0.080	MP6287	202406L 06/19/2026																																					
0.200	MP6288	202406N 06/20/2026																																					
0.080 DGS	N/A	AG429602 10/22/2026																																					
Maintenance		By	TDG	Date																																			
☒ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____				<u>07/14/2025</u>																																			
Calibration Adjustment				Department Inspection																																			
Barometric Pressure Gauge _____ ID # _____				Barometric Pressure ID# <u>28199</u>																																			
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>				Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Gauge <u>1019</u> Instrument <u>1016</u> Mouth Alcohol Solution Lot # <u>2024-A</u> Acetone Stock Solution Lot # <u>2023-B</u>							
Simulator	Serial #	Lot #	Expiration																																				
0.000		N/A	N/A																																				
0.040																																							
0.100																																							
0.200																																							
0.300																																							
0.080 DGS	N/A																																						
☐ Post Calibration Adjustment Stability Checks				<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6284</td></tr><tr><td>Interferent</td><td>MP6285</td></tr><tr><td>0.050</td><td>MP6286</td></tr><tr><td>0.080</td><td>MP6287</td></tr><tr><td>0.200</td><td>MP6288</td></tr></tbody></table>								Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288																
Simulator	Serial Number																																						
0.000	MP6284																																						
Interferent	MP6285																																						
0.050	MP6286																																						
0.080	MP6287																																						
0.200	MP6288																																						
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>				Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			<table border="1"><thead><tr><th colspan="2">Attachments</th></tr></thead><tbody><tr><td>☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment</td><td>☐ Post-Stability Checks ☐ Flow Calibration ☒ Form 40 ☒ Other <u>Form 51</u></td></tr></tbody></table>								Attachments		☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment	☐ Post-Stability Checks ☐ Flow Calibration ☒ Form 40 ☒ Other <u>Form 51</u>				
Simulator	Serial #	Lot #	Expiration																																				
0.050																																							
0.080																																							
0.200																																							
0.080 DGS	N/A																																						
Attachments																																							
☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment	☐ Post-Stability Checks ☐ Flow Calibration ☒ Form 40 ☒ Other <u>Form 51</u>																																						
Notes/Suggested Service: <u>Checked breath tube screen and replaced o-rings on 7/8. Verified setup on 7/14. Finished Quality Checks on 7/15. (TDG 7/15/25)</u>				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☐ Return to/Place into Evidentiary Use ☒ Remain Out of Evidentiary Use ☐ Conduct an Agency Inspection Before Evidentiary Use																																			
*R-value is <100. Sending to CMI. (TDG 7/15/25)				Phil Nicodemo Digitally signed by Phil Nicodemo Date: 2025.07.18 12:05:13 -04'00'																																			
**Outside nominal. Performed root cause analysis; no user/equipment error. Did not perform flow cal adjust due to instrument being sent to repair. (TDG 7/15/25)				Shayla Platt Digitally signed by Shayla Platt Date: 2025.07.19 11:49:08 -04'00'																																			
				Tech Review / Date _____ Admin Review / Date _____																																			

Return Material Authorization

Ship to: ☒ CMI, Inc.

☐ Enforcement Electronics

Shipment to repair facility authorized by: Oscar Saavedra on 7/15/2025

Items Returned: Instrument ☒ Supplies ☐ Other ☐ Describe: _____

Instrument Model: Intoxilyzer 8000 Serial Number: 80-000904

Bill To Address:

FFWCC

Attn: Oscar Saavedra

Ship to Address:

Florida Department of Law Enforcement

Fort Myers Regional Operations Center

Attn: Taylor Gutschow

4700 Terminal Drive, Suite 1

Fort Myers, FL 33907

Reason for Return:

R-value is below 100.

Please choose one of the following options:

- ☐ 1. I _____, authorize all repairs.
- ☐ 2. I _____, authorize repairs up to \$_____.
- ☒ 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Oscar Saavedra

Phone #: 786-382-3372 Email: Oscar.Saavedra@MyFWC.com

ATP Contact Name: Taylor Gutschow ATP Email: TaylorGutschow@fdle.state.fl.us

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FFWCC

Time of Inspection: 13:06

Date of Inspection: 07/15/2025

Serial Number: 80-000904

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____

Remarks:

BYPASSED AI TO OPERATE INSTRUMENT. COMPLIANCE UNDETERMINED.

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

07/15/2025
Date

Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																				
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																				
✓	✓	✓	✓																																																																																																																																				
≤0.003 of Wet	✓	✓	≤0.003 of Wet																																																																																																																																				
✓	✓	✓	✓																																																																																																																																				
FFWCC Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000904 07/15/2025 Software: 8100.27 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>14:14</td></tr> <tr><td>Control Test</td><td>0.049</td><td>14:15</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:16</td></tr> <tr><td>Control Test</td><td>0.049</td><td>14:16</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:17</td></tr> <tr><td>Control Test</td><td>0.049</td><td>14:18</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:18</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0490</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </table> Operator's Signature 	Air Blank	0.000	14:14	Control Test	0.049	14:15	Air Blank	0.000	14:16	Control Test	0.049	14:16	Air Blank	0.000	14:17	Control Test	0.049	14:18	Air Blank	0.000	14:18	Control Test Stats			Average	0.0490		Std Dev	0.0000		Rel Std Dev(%)	0.0000		FFWCC Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000904 07/15/2025 Software: 8100.27 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>14:22</td></tr> <tr><td>Control Test</td><td>0.080</td><td>14:22</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:23</td></tr> <tr><td>Control Test</td><td>0.079</td><td>14:24</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:24</td></tr> <tr><td>Control Test</td><td>0.079</td><td>14:25</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:25</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0793</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7277</td><td></td></tr> </table> Operator's Signature 	Air Blank	0.000	14:22	Control Test	0.080	14:22	Air Blank	0.000	14:23	Control Test	0.079	14:24	Air Blank	0.000	14:24	Control Test	0.079	14:25	Air Blank	0.000	14:25	Control Test Stats			Average	0.0793		Std Dev	0.0006		Rel Std Dev(%)	0.7277		FFWCC Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000904 07/15/2025 Software: 8100.27 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>14:06</td></tr> <tr><td>Control Test</td><td>0.198</td><td>14:07</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:08</td></tr> <tr><td>Control Test</td><td>0.197</td><td>14:08</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:09</td></tr> <tr><td>Control Test</td><td>0.197</td><td>14:09</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:10</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1973</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2926</td><td></td></tr> </table> Operator's Signature 	Air Blank	0.000	14:06	Control Test	0.198	14:07	Air Blank	0.000	14:08	Control Test	0.197	14:08	Air Blank	0.000	14:09	Control Test	0.197	14:09	Air Blank	0.000	14:10	Control Test Stats			Average	0.1973		Std Dev	0.0006		Rel Std Dev(%)	0.2926		FFWCC Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000904 07/15/2025 Software: 8100.27 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>14:11</td></tr> <tr><td>Control Test</td><td>0.077</td><td>14:12</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:12</td></tr> <tr><td>Control Test</td><td>0.078</td><td>14:12</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:13</td></tr> <tr><td>Control Test</td><td>0.078</td><td>14:13</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:14</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0777</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7434</td><td></td></tr> </table> Operator's Signature 	Air Blank	0.000	14:11	Control Test	0.077	14:12	Air Blank	0.000	14:12	Control Test	0.078	14:12	Air Blank	0.000	14:13	Control Test	0.078	14:13	Air Blank	0.000	14:14	Control Test Stats			Average	0.0777		Std Dev	0.0006		Rel Std Dev(%)	0.7434	
Air Blank	0.000	14:14																																																																																																																																					
Control Test	0.049	14:15																																																																																																																																					
Air Blank	0.000	14:16																																																																																																																																					
Control Test	0.049	14:16																																																																																																																																					
Air Blank	0.000	14:17																																																																																																																																					
Control Test	0.049	14:18																																																																																																																																					
Air Blank	0.000	14:18																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0490																																																																																																																																						
Std Dev	0.0000																																																																																																																																						
Rel Std Dev(%)	0.0000																																																																																																																																						
Air Blank	0.000	14:22																																																																																																																																					
Control Test	0.080	14:22																																																																																																																																					
Air Blank	0.000	14:23																																																																																																																																					
Control Test	0.079	14:24																																																																																																																																					
Air Blank	0.000	14:24																																																																																																																																					
Control Test	0.079	14:25																																																																																																																																					
Air Blank	0.000	14:25																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0793																																																																																																																																						
Std Dev	0.0006																																																																																																																																						
Rel Std Dev(%)	0.7277																																																																																																																																						
Air Blank	0.000	14:06																																																																																																																																					
Control Test	0.198	14:07																																																																																																																																					
Air Blank	0.000	14:08																																																																																																																																					
Control Test	0.197	14:08																																																																																																																																					
Air Blank	0.000	14:09																																																																																																																																					
Control Test	0.197	14:09																																																																																																																																					
Air Blank	0.000	14:10																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.1973																																																																																																																																						
Std Dev	0.0006																																																																																																																																						
Rel Std Dev(%)	0.2926																																																																																																																																						
Air Blank	0.000	14:11																																																																																																																																					
Control Test	0.077	14:12																																																																																																																																					
Air Blank	0.000	14:12																																																																																																																																					
Control Test	0.078	14:12																																																																																																																																					
Air Blank	0.000	14:13																																																																																																																																					
Control Test	0.078	14:13																																																																																																																																					
Air Blank	0.000	14:14																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0777																																																																																																																																						
Std Dev	0.0006																																																																																																																																						
Rel Std Dev(%)	0.7434																																																																																																																																						

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FFWCC

Time of Inspection: 11:56

Date of Inspection: 07/17/2025

Serial Number: 80-000904

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG429602 Exp: 10/22/2026
0.000	0.050	0.079	0.198	0.079
0.000	0.050	0.079	0.199	0.078
0.000	0.050	0.079	0.198	0.078
0.000	0.050	0.079	0.198	0.077
0.000	0.050	0.079	0.198	0.078
0.000	0.050	0.079	0.198	0.077
0.000	0.051	0.079	0.198	0.078
0.000	0.051	0.079	0.198	0.078
0.000	0.051	0.079	0.198	0.078
0.000	0.051	0.079	0.198	0.078


Standard Deviations	0.0005	0.0000	0.0003	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

07/17/2025
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000904, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000904</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>FFWCC</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>07/17/2025</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>11:56</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Taylor
Gutschow
Digitally signed by Taylor
Gutschow
Date: 2025.07.17 15:12:04
-04'00'

07/17/2025

Date

TAYLOR D GUTSCHOW,
Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality