



INSTRUMENT PROCESSING SHEET

Agency Port St. Joe PDS/N 80-000908Florida Department of
Law EnforcementDate In 4/10/25DI Completion Date 4/21/25☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake By <u>SP</u> Date <u>4/16/25</u>		Quality Checks By <u>SLH/SP</u> Date <u>4/18 & 4/21</u>		Flow Calibration By _____ Date _____																																									
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: Instrument hand delivered on 4/10/25. AI reports DSP failures. Turned on instrument during intake process. No DSP failure obtained. SP		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>247</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP103</u> 32 mm <u>.167</u> (.139 - .169) 36 mm <u>.179</u> (.156 - .190) 53 mm <u>.250</u> (.228 - .278) 103 mm <u>.515</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28427</u> ☒ Stability Checks		☐ Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																									
		<table border="1" style="width:100%"><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr><tr><td>0.050</td><td>MP6291</td><td>202406K 6/19/26</td></tr><tr><td>0.080</td><td>MP6292</td><td>202406L 6/19/26</td></tr><tr><td>0.200</td><td>MP6293</td><td>202406N 6/20/26</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG429602 10/22/26</td></tr></table>		Simulator	Serial #	Lot #/Exp	0.050	MP6291	202406K 6/19/26	0.080	MP6292	202406L 6/19/26	0.200	MP6293	202406N 6/20/26	0.080 DGS	N/A	AG429602 10/22/26	Maintenance By _____ Date _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____																										
Simulator	Serial #	Lot #/Exp																																											
0.050	MP6291	202406K 6/19/26																																											
0.080	MP6292	202406L 6/19/26																																											
0.200	MP6293	202406N 6/20/26																																											
0.080 DGS	N/A	AG429602 10/22/26																																											
Calibration Adjustment By _____		Department Inspection By <u>SP</u>																																											
Barometric Pressure Gauge _____ ID # _____		Barometric Pressure ID# <u>28662</u> Gauge <u>1018</u> Instrument <u>1019</u> Mouth Alcohol Solution Lot # <u>2024-A</u> Acetone Stock Solution Lot # <u>2024-B</u>																																											
<table border="1" style="width:100%"><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			<table border="1" style="width:100%"><tr><th>Simulator</th><th>Serial Number</th></tr><tr><td>0.000</td><td>MP6289</td></tr><tr><td>Interferent</td><td>MP6290</td></tr><tr><td>0.050</td><td>MP6291</td></tr><tr><td>0.080</td><td>MP6292</td></tr><tr><td>0.200</td><td>MP6293</td></tr></table>				Simulator	Serial Number	0.000	MP6289	Interferent	MP6290	0.050	MP6291	0.080	MP6292	0.200	MP6293
Simulator	Serial #	Lot #	Expiration																																										
0.000		N/A	N/A																																										
0.040																																													
0.100																																													
0.200																																													
0.300																																													
0.080 DGS	N/A																																												
Simulator	Serial Number																																												
0.000	MP6289																																												
Interferent	MP6290																																												
0.050	MP6291																																												
0.080	MP6292																																												
0.200	MP6293																																												
☐ Post Calibration Adjustment Stability Checks		Attachments ☒ Form 41 ☐ Post-Stability Checks ☒ Stability Checks ☐ Flow Calibration ☒ Calibration Certificate ☒ Form 40 ☐ Calibration Adjustment ☐ Other _____																																											
<table border="1" style="width:100%"><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></table>		Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																							
Simulator	Serial #	Lot #	Expiration																																										
0.050																																													
0.080																																													
0.200																																													
0.080 DGS	N/A																																												
Notes/Suggested Service: No DSP failures observed during QC or inspection process. SP 4/21/25 Tech Review: All quality checks excluding stability checks performed by SLH. SP 4/23/25		Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2025.04.23 12:45:05 -0400 Phil Nicodemo Digitally signed by Phil Nicodemo Date: 2025.05.01 15:17:24 -0400 Tech Review / Date _____ Admin Review / Date _____																																											

Florida Department of Law Enforcement

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PORT ST JOE PD

Time of Inspection: 11:02

Date of Inspection: 04/21/2025

Serial Number: 80-000908

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG429602 Exp: 10/22/2026
0.000	0.048	0.079	0.199	0.079
0.000	0.048	0.079	0.200	0.079
0.000	0.048	0.080	0.200	0.080
0.000	0.048	0.080	0.200	0.079
0.000	0.048	0.080	0.199	0.079
0.000	0.048	0.080	0.200	0.079
0.000	0.048	0.080	0.200	0.079
0.000	0.048	0.080	0.200	0.079
0.000	0.049	0.079	0.200	0.079
0.000	0.049	0.080	0.200	0.079

Standard Deviations	0.0004	0.0004	0.0004	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt

SHAYLA D PLATT

Signature and Printed Name

04/21/2025
Date

PORT ST JOE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000908
04/21/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:24
Control Test	0.048	07:25
Air Blank	0.000	07:25
Control Test	0.048	07:26
Air Blank	0.000	07:27
Control Test	0.049	07:27
Air Blank	0.000	07:28
Control Test Stats		
Average	0.0483	
Std Dev	0.0006	
Rel Std Dev(%)	1.1945	

SP

Operator's Signature

PORT ST JOE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000908
04/21/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:30
Control Test	0.079	07:30
Air Blank	0.000	07:31
Control Test	0.080	07:31
Air Blank	0.000	07:32
Control Test	0.079	07:33
Air Blank	0.000	07:33
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

SJP

Operator's Signature

PORT ST JOE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000908
04/21/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:36
Control Test	0.200	07:36
Air Blank	0.000	07:37
Control Test	0.199	07:37
Air Blank	0.000	07:38
Control Test	0.200	07:39
Air Blank	0.000	07:39
Control Test Stats		
Average	0.1997	
Std Dev	0.0006	
Rel Std Dev(%)	0.2892	



Operator's Signature

PORT ST JOE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000908
04/21/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:14
Control Test	0.080	07:14
Air Blank	0.000	07:15
Control Test	0.079	07:15
Air Blank	0.000	07:16
Control Test	0.079	07:16
Air Blank	0.000	07:16
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

DGS

SP

Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PORT ST JOE PD
Time of Inspection: 07:08

Date of Inspection: 04/21/2025

Serial Number: 80-000908
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____

Remarks:

BYPASSED AI TO OPERATIVE INSTRUMENT

The above instrument complies (N/A) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt

SHAYLA D PLATT

Signature and Printed Name

04/21/2025
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number **80-000908** , manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	80-000908	UNCERTAINTY* $\pm$
Owning Agency:	PORT ST JOE PD	0.050 g/ 210 L 0.004
Calibration Date:	04/21/2025	0.080 g/ 210 L 0.004
Calibration Time:	11:02	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Digitally signed by

Shayla Platt

Shayla Platt

Date: 2025.04.21

12:28:32 -04'00'

04/21/2025

Date

SHAYLA D PLATT,

Department Inspector

FDLE/ATP Form 69 October 2024

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality