



INSTRUMENT PROCESSING SHEET

Agency Lee CSO

S/N 80-000939

Florida Department of
Law Enforcement

Date In 02/20/2025 DI Completion Date 02/21/2025

☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By TDG	Date 02/20/2025	Quality Checks	By TDG	Date 02/21/2025	Flow Calibration	By	Date																													
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: Dropped off. No box.			☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value 212 ☒ Flow Verification (L/s) Flow Column # ATP106 32 mm 0.152 (.139 - .169) 36 mm 0.167 (.156 - .190) 53 mm 0.246 (.228 - .278) 103 mm 0.531 (.447 - .547) ☒ Barometric Pressure Check Gauge ID # 26932 ☒ Stability Checks			Flow Column # ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value ☐ Post Calibration Verification (L/s) Flow Column # 32 mm (.139 - .169) 36 mm (.156 - .190) 53 mm (.228 - .278) 103 mm (.447 - .547)																															
			<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6286</td><td>202303K 03/29/2025</td></tr><tr><td>0.080</td><td>MP6287</td><td>202303L 03/29/2025</td></tr><tr><td>0.200</td><td>MP6288</td><td>202304C 04/05/2025</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG429602 10/22/2026</td></tr></tbody></table>	Simulator	Serial #	Lot #/Exp	0.050	MP6286	202303K 03/29/2025	0.080	MP6287	202303L 03/29/2025	0.200	MP6288	202304C 04/05/2025	0.080 DGS	N/A	AG429602 10/22/2026																			
Simulator	Serial #	Lot #/Exp																																			
0.050	MP6286	202303K 03/29/2025																																			
0.080	MP6287	202303L 03/29/2025																																			
0.200	MP6288	202304C 04/05/2025																																			
0.080 DGS	N/A	AG429602 10/22/2026																																			
Calibration Adjustment			Department Inspection																																		
By			By TDG																																		
Barometric Pressure Gauge ID #			Barometric Pressure ID# 26932																																		
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>			Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Gauge 1026 Instrument 1021 Mouth Alcohol Solution Lot # 2024-A Acetone Stock Solution Lot # 2023-B						
Simulator	Serial #	Lot #	Expiration																																		
0.000		N/A	N/A																																		
0.040																																					
0.100																																					
0.200																																					
0.300																																					
0.080 DGS	N/A																																				
☐ Post Calibration Adjustment Stability Checks			<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6284</td></tr><tr><td>Interferent</td><td>MP6285</td></tr><tr><td>0.050</td><td>MP6286</td></tr><tr><td>0.080</td><td>MP6287</td></tr><tr><td>0.200</td><td>MP6288</td></tr></tbody></table>					Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288																		
Simulator	Serial Number																																				
0.000	MP6284																																				
Interferent	MP6285																																				
0.050	MP6286																																				
0.080	MP6287																																				
0.200	MP6288																																				
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			<table border="1"><thead><tr><th colspan="2">Attachments</th></tr></thead><tbody><tr><td>☒ Form 41</td><td>☐ Post-Stability Checks</td></tr><tr><td>☒ Stability Checks</td><td>☐ Flow Calibration</td></tr><tr><td>☒ Calibration Certificate</td><td>☐ Form 40</td></tr><tr><td>☐ Calibration Adjustment</td><td>☐ Other</td></tr></tbody></table>					Attachments		☒ Form 41	☐ Post-Stability Checks	☒ Stability Checks	☐ Flow Calibration	☒ Calibration Certificate	☐ Form 40	☐ Calibration Adjustment	☐ Other
Simulator	Serial #	Lot #	Expiration																																		
0.050																																					
0.080																																					
0.200																																					
0.080 DGS	N/A																																				
Attachments																																					
☒ Form 41	☐ Post-Stability Checks																																				
☒ Stability Checks	☐ Flow Calibration																																				
☒ Calibration Certificate	☐ Form 40																																				
☐ Calibration Adjustment	☐ Other																																				
Notes/Suggested Service:			☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																																		
			Digitally signed by Phil Nicodemo Date: 2025.02.21 15:09:46 -05'00'																																		
			Shayla Platt Shayla Platt Date: 2025.02.21 15:43:38 -05'00'																																		
			Tech Review / Date Admin Review / Date																																		

Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
LEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000939 02/21/2025 Software: 8100.27 Test g/210L Time Air Blank 0.000 10:59 Control Test 0.051 10:59 Air Blank 0.000 11:00 Control Test 0.051 11:01 Air Blank 0.000 11:01 Control Test 0.050 11:02 Air Blank 0.000 11:02 Control Test Stats Average 0.0507 Std Dev 0.0006 Rel Std Dev(%) 1.1395	LEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000939 02/21/2025 Software: 8100.27 Test g/210L Time Air Blank 0.000 10:38 Control Test 0.081 10:38 Air Blank 0.000 10:39 Control Test 0.080 10:40 Air Blank 0.000 10:40 Control Test 0.080 10:41 Air Blank 0.000 10:41 Control Test Stats Average 0.0803 Std Dev 0.0006 Rel Std Dev(%) 0.7187	LEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000939 02/21/2025 Software: 8100.27 Test g/210L Time Air Blank 0.000 10:31 Control Test 0.198 10:32 Air Blank 0.000 10:33 Control Test 0.197 10:33 Air Blank 0.000 10:34 Control Test 0.197 10:35 Air Blank 0.000 10:35 Control Test Stats Average 0.1973 Std Dev 0.0006 Rel Std Dev(%) 0.2926	LEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000939 02/21/2025 Software: 8100.27 Test g/210L Time Air Blank 0.000 10:24 Control Test 0.078 10:24 Air Blank 0.000 10:25 Control Test 0.079 10:25 Air Blank 0.000 10:26 Control Test 0.079 10:26 Air Blank 0.000 10:26 Control Test Stats Average 0.0787 Std Dev 0.0006 Rel Std Dev(%) 0.7339
Operator's Signature 	Operator's Signature 	Operator's Signature 	Operator's Signature 

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: LEE COUNTY SO
Time of Inspection: 14:31

Date of Inspection: 02/21/2025

Serial Number: 80-000939
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG429602 Exp: 10/22/2026
0.000	0.049	0.079	0.196	0.080
0.000	0.050	0.078	0.197	0.080
0.000	0.050	0.079	0.197	0.079
0.000	0.050	0.079	0.198	0.079
0.000	0.050	0.079	0.197	0.079
0.000	0.051	0.080	0.197	0.079
0.000	0.050	0.079	0.197	0.078
0.000	0.050	0.079	0.197	0.079
0.000	0.050	0.079	0.197	0.079
0.000	0.050	0.079	0.197	0.078


Standard Deviations	0.0004	0.0004	0.0004	0.0006
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.


I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

02/21/2025
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000939, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	80-000939	UNCERTAINTY* $\pm$
Owning Agency:	LEE COUNTY SO	0.050 g/ 210 L 0.004
Calibration Date:	02/21/2025	0.080 g/ 210 L 0.004
Calibration Time:	14:31	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

02/21/2025

Date

TAYLOR D GUTSCHOW,
Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality