



INSTRUMENT PROCESSING SHEET

Agency COCOA PDS/N 80-001000Florida Department of
Law EnforcementDate In 7/29/2025DI Completion Date 08/06/2025☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By WKP	Date	Quality Checks	By WKP	Date	Flow Calibration	By SLH	Date																																																																	
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☒ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____		<u>7/29/2025</u>	☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>167</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP 105</u> 32 mm <u>0.140</u> (.139 - .169) 36 mm <u>0.152 / 0.152</u> (.156 - .190) 53 mm <u>0.222 / 0.222</u> (.228 - .278) 103 mm <u>0.492</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28427</u> ☒ Stability Checks		<u>7/31/2025</u>	Flow Column # <u>ATP102</u> ☒ 5L/min – 17mm ☒ 15L/min – 53mm ☒ 30L/min – 103mm ☒ R-Value <u>168</u> ☒ Post Calibration Verification (L/s) Flow Column # <u>ATP105</u> 32 mm <u>0.148</u> (.139 - .169) 36 mm <u>0.164</u> (.156 - .190) 53 mm <u>0.238</u> (.228 - .278) 103 mm <u>0.535</u> (.447 - .547)		<u>08/05/2025</u>																																																																	
			<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6291</td><td>202406K 06/19/2026</td></tr><tr><td>0.080</td><td>MP6292</td><td>202406L 06/19/2026</td></tr><tr><td>0.200</td><td>MP6293</td><td>202406N 06/20/2026</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG429602 10/22/2026</td></tr></tbody></table>			Simulator	Serial #	Lot #/Exp	0.050	MP6291	202406K 06/19/2026	0.080	MP6292	202406L 06/19/2026	0.200	MP6293	202406N 06/20/2026	0.080 DGS	N/A	AG429602 10/22/2026	<table border="1"><thead><tr><th colspan="2">Maintenance By _____ Date _____</th></tr><tr><td>☐ Battery Replacement</td><td></td></tr><tr><td>☐ Dry Gas Regulator Replacement</td><td></td></tr><tr><td>☐ Breath Tube Replacement</td><td></td></tr><tr><td>☐ Other _____</td><td></td></tr><tr><td>_____</td><td></td></tr><tr><td>_____</td><td></td></tr><tr><td>_____</td><td></td></tr><tr><td>_____</td><td></td></tr></thead></table>			Maintenance By _____ Date _____		☐ Battery Replacement		☐ Dry Gas Regulator Replacement		☐ Breath Tube Replacement		☐ Other _____		_____		_____		_____		_____																																	
Simulator	Serial #	Lot #/Exp																																																																							
0.050	MP6291	202406K 06/19/2026																																																																							
0.080	MP6292	202406L 06/19/2026																																																																							
0.200	MP6293	202406N 06/20/2026																																																																							
0.080 DGS	N/A	AG429602 10/22/2026																																																																							
Maintenance By _____ Date _____																																																																									
☐ Battery Replacement																																																																									
☐ Dry Gas Regulator Replacement																																																																									
☐ Breath Tube Replacement																																																																									
☐ Other _____																																																																									

Calibration Adjustment By _____			Department Inspection By <u>SLH</u>																																																																						
Barometric Pressure Gauge _____ ID # _____ <table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table> ☐ Post Calibration Adjustment Stability Checks <table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>			Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Barometric Pressure ID# <u>28421</u> Gauge <u>1020</u> Instrument <u>1016</u> Mouth Alcohol Solution Lot # <u>2025-A</u> Acetone Stock Solution Lot # <u>2024-B</u> <table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6289</td></tr><tr><td>Interferent</td><td>MP6290</td></tr><tr><td>0.050</td><td>MP5088</td></tr><tr><td>0.080</td><td>MP5089</td></tr><tr><td>0.200</td><td>MP5090</td></tr></tbody></table> Attachments <table border="1"><tbody><tr><td>☒ Form 41</td><td>☐ Post-Stability Checks</td></tr><tr><td>☒ Stability Checks</td><td>☒ Flow Calibration</td></tr><tr><td>☒ Calibration Certificate</td><td>☐ Form 40</td></tr><tr><td>☐ Calibration Adjustment</td><td>☐ Other _____</td></tr></tbody></table> ☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use Digitally signed by Shayla Platt Date: 2025.08.07 13:08:30 -04'00' Taylor Gutschow Tech Review / Date Digitally signed by Taylor Gutschow Date: 2025.08.07 12:32:52 -04'00' Shayla Platt Admin Review / Date			Simulator	Serial Number	0.000	MP6289	Interferent	MP6290	0.050	MP5088	0.080	MP5089	0.200	MP5090	☒ Form 41	☐ Post-Stability Checks	☒ Stability Checks	☒ Flow Calibration	☒ Calibration Certificate	☐ Form 40	☐ Calibration Adjustment	☐ Other _____
Simulator	Serial #	Lot #	Expiration																																																																						
0.000		N/A	N/A																																																																						
0.040																																																																									
0.100																																																																									
0.200																																																																									
0.300																																																																									
0.080 DGS	N/A																																																																								
Simulator	Serial #	Lot #	Expiration																																																																						
0.050																																																																									
0.080																																																																									
0.200																																																																									
0.080 DGS	N/A																																																																								
Simulator	Serial Number																																																																								
0.000	MP6289																																																																								
Interferent	MP6290																																																																								
0.050	MP5088																																																																								
0.080	MP5089																																																																								
0.200	MP5090																																																																								
☒ Form 41	☐ Post-Stability Checks																																																																								
☒ Stability Checks	☒ Flow Calibration																																																																								
☒ Calibration Certificate	☐ Form 40																																																																								
☐ Calibration Adjustment	☐ Other _____																																																																								
Notes/Suggested Service: <u>Performed root cause analysis on failed flow verifications, reformed flow verifications and results remained the same values as the initial measurements. WKP 7/31/2025</u> _____ _____ _____ _____ _____																																																																									

Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L	
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083	≤0.003 of Wet
Performed Root Case Analysis	Performed Root Case Analysis	Performed Root Case Analysis	Performed Root Case Analysis	
COCOA PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 60-001000 07/31/2025 Software: 8100.27	COCOA PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 60-001000 07/31/2025 Software: 8100.27	COCOA PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 60-001000 07/31/2025 Software: 8100.27	COCOA PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 60-001000 07/31/2025 Software: 8100.27	
Test g/210L Time	Test g/210L Time	Test g/210L Time	Test g/210L Time	
Air Blank 0.000 10:25	Air Blank 0.000 10:31	Air Blank 0.000 10:37	Air Blank 0.000 10:43	
Control Test 0.050 10:25	Control Test 0.080 10:31	Control Test 0.200 10:38	Control Test 0.081 10:44	
Air Blank 0.000 10:26	Air Blank 0.000 10:32	Air Blank 0.000 10:38	Air Blank 0.000 10:44	
Control Test 0.049 10:27	Control Test 0.081 10:33	Control Test 0.199 10:39	Control Test 0.081 10:44	
Air Blank 0.000 10:27	Air Blank 0.000 10:33	Air Blank 0.000 10:40	Air Blank 0.000 10:45	
Control Test 0.050 10:28	Control Test 0.079 10:34	Control Test 0.199 10:40	Control Test 0.081 10:45	
Air Blank 0.000 10:28	Air Blank 0.000 10:34	Air Blank 0.000 10:41	Air Blank 0.000 10:46	
Control Test Stats	Control Test Stats	Control Test Stats	Control Test Stats	
Average 0.0497	Average 0.0800	Average 0.1993	Average 0.0810	
Std Dev 0.0006	Std Dev 0.0010	Std Dev 0.0006	Std Dev 0.0000	
Rel Std Dev(%) 1.1625	Rel Std Dev(%) 1.2500	Rel Std Dev(%) 0.2895	Rel Std Dev(%) 0.0000	
Operator's Signature	Operator's Signature	Operator's Signature	Operator's Signature	

Flow Calibration

80-001000
SWA 8/5/25

Root Cause Analysis Performed Prior



Performed by SLH

INTOXILYZER 8000
Instrument Initialization
10:07 08/05/2025

COCOA PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001000
08/05/2025
Software: 8100.27

Flow Rate Calibration*****

1: Rate (Liters/min) = 5

SQRT(Diff) = 7.141

2: Rate (Liters/min) = 15

SQRT(Diff) = 11.660

3: Rate (Liters/min) = 30

SQRT(Diff) = 20.172

Dependent Data Scale Factor = 100000 L/min

Independent Data Scale Factor = 256

Rounded Slope = 742

Rounded Intercept = -799342

Correlation = 0.99819

Florida Department of Law Enforcement

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: COCOA PD

Time of Inspection: 13:17

Date of Inspection: 08/06/2025

Serial Number: 80-001000

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG429602 Exp: 10/22/2026
0.000	0.049	0.079	0.197	0.079
0.000	0.049	0.079	0.197	0.079
0.000	0.049	0.078	0.198	0.079
0.000	0.049	0.079	0.199	0.079
0.000	0.049	0.079	0.198	0.079
0.000	0.049	0.079	0.197	0.079
0.000	0.049	0.079	0.198	0.078
0.000	0.049	0.079	0.197	0.078
0.000	0.049	0.079	0.198	0.079
0.000	0.049	0.079	0.197	0.078


Standard Deviations	0.0000	0.0003	0.0006	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5



Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


 LEANDRA HIGGINBOTHAM
 Signature and Printed Name

08/06/2025
 Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001000, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001000</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>COCOA PD</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>08/06/2025</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>13:17</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 3%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

08/06/2025

Date

LEANDRA HAGGINBOTHAM,

Department Inspector

FDLE/ATP Form 69 October 2024
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality