



Agency Haines City PD

S/N 80-001007

Date In 03/14/2025 DI Completion Date N/A

☐ Ship ☐ P/U ☐ H/D ☐ CMI ☒ EE

Calibration Adjustment

By_

☐ Post Calibration Adjustment Stability Checks

Notes/Suggested Service: Added static bag. (TDG 3/26/25)

During Set-Up Verification, the time, owning agency, and serial number were incorrect. Corrected the time and owning agency but cannot correct serial number. Will return to EE. (TDG 3/26/25)

Tech Review: Filled in "Return from CMI/EE" box. (TDG 3/28/25)

Department Inspection

By _

Attachments

- ☐ Instrument Complies with Chapter 11D-8, FAC
- ☒ Instrument Does Not Comply with Chapter 11D-8, FAC

- ☐ Return to/Place into Evidentiary Use
- ☒ Remain Out of Evidentiary Use

- ☐ Conduct an Agency Inspection Before Evidentiary Use

Digitally signed by

Phil Nicodemo Digitally signed by Phil Nicodemo
Date: 2025.03.28 15:49:27 -0400

Shayla Platt

Shayla Platt

• Date: 2025.03.31

Tech Review / Date

Admin Review / Date

Return Material Authorization

Ship to:

☐ CMI, Inc.

☒ Enforcement Electronics

Shipment to repair facility authorized by: Sean Bruner on 3/26/2025

Items Returned: Instrument ☒ Supplies ☐ Other ☐ Describe: _____

Instrument Model: Intoxilyzer 8000 Serial Number: 80-001007

Bill To Address:

Haines City Police Department

Attn: Sean Bruner

Ship to Address:

Florida Department of Law Enforcement

Fort Myers Regional Operations Center

Attn: Taylor Gutschow

4700 Terminal Drive, Suite 1

Fort Myers, FL 33907

Reason for Return:

Instrument returned from EE with incorrect agency name and serial number displayed by the
instrument. Corrected the agency name but unable to correct serial number.

Please choose one of the following options:

☐ 1. I _____, authorize all repairs.

☐ 2. I _____, authorize repairs up to \$_____.

☒ 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Sean Bruner

Phone #: 727-641-2144

Email: sean.bruner@hainescity.com

ATP Contact Name: Taylor Gutschow

ATP Email: TaylorGutschow@fdle.state.fl.us