



INSTRUMENT PROCESSING SHEET

Agency Jacksonville SOS/N 80-001018Florida Department of
Law EnforcementDate In 4/14/25DI Completion Date 4/21/2025☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By SP	Date <u>4/16/25</u>	Quality Checks	By SLH/SP	Date <u>4/18 & 4/21</u>	Flow Calibration	By	Date																																																									
☒ Annual ☐ Registration ☒ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☒ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____			☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>112</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP103</u> 32 mm <u>.164</u> (.139 - .169) 36 mm <u>.183</u> (.156 - .190) 53 mm <u>.253</u> (.228 - .278) 103 mm <u>.527</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28427</u> ☒ Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6291</td><td>202406K 6/19/26</td></tr><tr><td>0.080</td><td>MP6292</td><td>202406L 6/19/26</td></tr><tr><td>0.200</td><td>MP6293</td><td>202406N 6/20/26</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG429602 10/22/26</td></tr></tbody></table>			Simulator	Serial #	Lot #/Exp	0.050	MP6291	202406K 6/19/26	0.080	MP6292	202406L 6/19/26	0.200	MP6293	202406N 6/20/26	0.080 DGS	N/A	AG429602 10/22/26	Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																												
			Simulator	Serial #	Lot #/Exp																																																												
			0.050	MP6291	202406K 6/19/26																																																												
			0.080	MP6292	202406L 6/19/26																																																												
0.200	MP6293	202406N 6/20/26																																																															
0.080 DGS	N/A	AG429602 10/22/26																																																															
						Maintenance By _____ Date _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ _____ _____ _____																																																											
Calibration Adjustment By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table> ☐ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>			Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Department Inspection By <u>SP</u> Barometric Pressure ID# <u>28662</u> Gauge <u>1018</u> Instrument <u>1019</u> Mouth Alcohol Solution Lot # <u>2024-A</u> Acetone Stock Solution Lot # <u>2024-B</u> <table border="1" style="width:100%; border-collapse: collapse;"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6289</td></tr><tr><td>Interferent</td><td>MP6290</td></tr><tr><td>0.050</td><td>MP6291</td></tr><tr><td>0.080</td><td>MP6292</td></tr><tr><td>0.200</td><td>MP6293</td></tr></tbody></table> Attachments ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment ☐ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____			Simulator	Serial Number	0.000	MP6289	Interferent	MP6290	0.050	MP6291	0.080	MP6292	0.200	MP6293
Simulator	Serial #	Lot #	Expiration																																																														
0.000		N/A	N/A																																																														
0.040																																																																	
0.100																																																																	
0.200																																																																	
0.300																																																																	
0.080 DGS	N/A																																																																
Simulator	Serial #	Lot #	Expiration																																																														
0.050																																																																	
0.080																																																																	
0.200																																																																	
0.080 DGS	N/A																																																																
Simulator	Serial Number																																																																
0.000	MP6289																																																																
Interferent	MP6290																																																																
0.050	MP6291																																																																
0.080	MP6292																																																																
0.200	MP6293																																																																
Notes/Suggested Service: Tech Review: All quality checks excluding stability checks performed by SLH. SP 4/23/25 _____ _____ _____ _____ _____			☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																																																														
						Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2025.04.23 13:17:38 -04'00' Phil Nicodemo Digitally signed by Phil Nicodemo Date: 2025.05.01 15:16:35 -04'00' Tech Review / Date _____ Admin Review / Date _____																																																											

Florida Department of Law Enforcement

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: JACKSONVILLE SO
Time of Inspection: 11:21

Date of Inspection: 04/21/2025

Serial Number: 80-001018
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) 980421/25 Lot#:AG420602 Exp: 10/22/2026
0.000	0.049	0.081	0.204	0.078
0.000	0.049	0.081	0.205	0.078
0.000	0.049	0.081	0.205	0.078
0.000	0.049	0.081	0.205	0.078
0.000	0.050	0.081	0.205	0.077
0.000	0.049	0.081	0.206	0.078
0.000	0.050	0.081	0.206	0.077
0.000	0.050	0.082	0.205	0.077
0.000	0.050	0.081	0.205	0.077
0.000	0.050	0.081	0.206	0.077

Standard Deviations	0.0005	0.0003	0.0006	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt

SHAYLA D PLATT

Signature and Printed Name

04/21/2025
Date

Stability Checks # 80-001018

JACKSONVILLE SO
Intoxilyzer - Alcotest Analyzer
Model 8000
04/21/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:22
Control Test	0.049	07:23
Air Blank	0.000	07:23
Control Test	0.049	07:24
Air Blank	0.000	07:25
Control Test	0.049	07:25
Air Blank	0.000	07:26
Control Test	0.049	07:26
Average	0.049	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SP

Operator's Signature

JACKSONVILLE SO
Intoxilyzer - Alcotest Analyzer
Model 8000
04/21/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:27
Control Test	0.081	07:28
Air Blank	0.000	07:28
Control Test	0.082	07:29
Air Blank	0.000	07:30
Control Test	0.081	07:30
Air Blank	0.000	07:31
Control Test	0.083	07:31
Average	0.0813	
Std Dev	0.0006	
Rel Std Dev(%)	0.7059	

SP

Operator's Signature

JACKSONVILLE SO
Intoxilyzer - Alcotest Analyzer
Model 8000
04/21/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:32
Control Test	0.205	07:33
Air Blank	0.000	07:34
Control Test	0.204	07:34
Air Blank	0.000	07:35
Control Test	0.205	07:35
Air Blank	0.000	07:36
Control Test	0.204	07:36
Average	0.2047	
Std Dev	0.0006	
Rel Std Dev(%)	0.2821	

SP

Operator's Signature

JACKSONVILLE SO
Intoxilyzer - Alcotest Analyzer
Model 8000
04/21/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:04
Control Test	0.081	07:05
Air Blank	0.000	07:05
Control Test	0.079	07:06
Air Blank	0.000	07:06
Control Test	0.079	07:07
Air Blank	0.000	07:07
Control Test	0.0797	07:07
Average	0.0797	
Std Dev	0.0012	
Rel Std Dev(%)	1.4494	

DGS

SP

Operator's Signature



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001018 , manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001018</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>JACKSONVILLE SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>04/21/2025</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>11:21</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Shayla Platt
Digitally signed by
Shayla Platt
Date: 2025.04.21
12:37:02 -04'00'

04/21/2025

Date

SHAYLA D PLATT,
Department Inspector

FDLE/ATP Form 69 October 2024

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality