

Agency Polk County SO

S/N 80-001058

Date In 4/14/25DI Completion Date 4/21/2025

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Calibration Adjustment

By_

☐ Post Calibration Adjustment Stability ChecksNotes/Suggested Service:

Tech Review: All quality checks excluding stability checks performed by SLH. SP 4/23/25

Department Inspection

By SP

Attachments

Instrument Complies with Chapter 11D-8, FAC

☐ Instrument Does Not Comply with Chapter 11D-8, FAC

 Return to/Place into Evidentiary Use

☐ **Remain Out of Evidentiary Use**

Conduct an Agency Inspection Before Evidentiary Use

Taylor
Gutschow

Digitally signed by Taylor Gutschow
Date: 2025.04.23 14:16:41 -0400

Phil Nicodemo

Digitally signed by Phil
Nicodemo
Date: 2025.05.01 15:16:58 -04'00'

Tech Review / Date

Admin Review / Date

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: POLK CSO
Time of Inspection: 10:08

Date of Inspection: 04/21/2025

Serial Number: 80-001058
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:202406L Exp: 06/19/2026
0.000	0.048	0.079	0.201	0.000 / 0.078
0.000	0.048	0.079	0.201	0.000 / 0.078
0.000	0.047	0.079	0.201	0.000 / 0.079
0.000	0.048	0.079	0.201	0.078 / 0.078
0.000	0.048	0.079	0.201	0.077 / 0.078
0.000	0.047	0.079	0.201	0.077 / 0.078
0.000	0.047	0.079	0.201	0.077 / 0.078
0.000	0.047	0.079	0.201	0.078 / 0.079
0.000	0.048	0.079	0.201	0.078 / 0.078
0.000	0.048	0.079	0.201	0.077 / 0.078
Standard Deviations	0.0005	0.0000	0.0000	0.0374 / 0.0004

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002 Number of Simulators Used: 5

Remarks:

08: Control Outside Tolerance- SIM ATTACHED INSTEAD OF DGS .

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt

SHAYLA D PLATT

Signature and Printed Name

04/21/2025
Date

Stability Checks # 80-001058

POLK CSO
Intoxilyzer - Alcohol Analyzer
Model 8000
04/21/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:29
Control Test	0.049	07:30
Air Blank	0.000	07:31
Control Test	0.049	07:31
Air Blank	0.000	07:32
Control Test	0.049	07:33
Air Blank	0.000	07:33
Control Test Stats		
Average	0.0487	
Std Dev	0.0006	
Rel Std Dev(%)	1.1663	

SP

Operator's Signature

POLK CSO
Intoxilyzer - Alcohol Analyzer
Model 8000
04/21/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:35
Control Test	0.079	07:35
Air Blank	0.000	07:36
Control Test	0.080	07:36
Air Blank	0.000	07:37
Control Test	0.079	07:38
Air Blank	0.000	07:38
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

SP

Operator's Signature

POLK CSO
Intoxilyzer - Alcohol Analyzer
Model 8000
04/21/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:40
Control Test	0.201	07:41
Air Blank	0.000	07:41
Control Test	0.200	07:42
Air Blank	0.000	07:43
Control Test	0.200	07:43
Air Blank	0.000	07:44
Control Test Stats		
Average	0.2003	
Std Dev	0.0006	
Rel Std Dev(%)	0.2882	

SP

Operator's Signature

POLK CSO
Intoxilyzer - Alcohol Analyzer
Model 8000
04/21/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	07:10
Control Test	0.079	07:10
Air Blank	0.000	07:10
Control Test	0.079	07:11
Air Blank	0.000	07:11
Control Test	0.078	07:12
Air Blank	0.000	07:12
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

DGS

SP

Operator's Signature



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number **80-001058**, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	80-001058	Serial Number:	80-001058
Owning Agency:	POLK CSO	0.050 g/ 210 L	0.004
Calibration Date:	04/21/2025	0.080 g/ 210 L	0.004
Calibration Time:	10:08	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Shayla Platt
Digitally signed by Shayla Platt
Date: 2025.04.21 12:41:21 -04'00'

04/21/2025

Date

SHAYLA D PLATT,
Department Inspector

FDLE/ATP Form 69 October 2024

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality