



Agency Hollywood Police Department

S/N 80-001063

Date In 07/17/2025 DI Completion Date 08/04/2025

☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Calibration Adjustment

By_

ID

❑ Post Calibration Adjustment Stability Checks

Notes/Suggested Service:

Department Inspection

By TDG

Acetone Stock Solution Lot # 2023-B

Attachments

- Instrument Complies with Chapter 11D-8, FAC**

- ☐
- Instrument Does Not Comply with Chapter 11D-8, FAC

- **Return to/Place into Evidentiary Use**

- ☐
- Remain Out of Evidentiary Use**

- ### ■ Conduct an Agency Inspection Before Evidentiary Use

Phil Nicodemo	Digitally signed by Phil Nicodemo Date: 2025.08.12 21:46:15 +04'00'	Shayla Platt	Digitally signed by Shayla Platt Date: 2025.08.13 22:31:56 +08'00'
---------------	---	--------------	--

Tech Review / Date

Admin Review / Date

Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																																
☒	☒	☒	☒																																																																																																																																																
☒ ≤0.003 of Wet			☒ ≤0.003 of Wet																																																																																																																																																
HOLLYWOOD PD Intoxilyzer - Alcohol Analyzer Model: 8000 08/04/2025 Software: 8100.27 SN 80-001063 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:27</td></tr> <tr><td>Control Test</td><td>0.049</td><td>11:28</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:28</td></tr> <tr><td>Control Test</td><td>0.049</td><td>11:29</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:30</td></tr> <tr><td>Control Test</td><td>0.049</td><td>11:30</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:31</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0490</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	11:27	Control Test	0.049	11:28	Air Blank	0.000	11:28	Control Test	0.049	11:29	Air Blank	0.000	11:30	Control Test	0.049	11:30	Air Blank	0.000	11:31	Control Test Stats			Average	0.0490		Std Dev	0.0000		Rel Std Dev(%)	0.0000		HOLLYWOOD PD Intoxilyzer - Alcohol Analyzer Model: 8000 08/04/2025 Software: 8100.27 SN 80-001063 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:20</td></tr> <tr><td>Control Test</td><td>0.080</td><td>11:21</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:22</td></tr> <tr><td>Control Test</td><td>0.080</td><td>11:22</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:23</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:23</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:24</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0797</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7247</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	11:20	Control Test	0.080	11:21	Air Blank	0.000	11:22	Control Test	0.080	11:22	Air Blank	0.000	11:23	Control Test	0.079	11:23	Air Blank	0.000	11:24	Control Test Stats			Average	0.0797		Std Dev	0.0006		Rel Std Dev(%)	0.7247		HOLLYWOOD PD Intoxilyzer - Alcohol Analyzer Model: 8000 08/04/2025 Software: 8100.27 SN 80-001063 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:02</td></tr> <tr><td>Control Test</td><td>0.204</td><td>11:02</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:03</td></tr> <tr><td>Control Test</td><td>0.203</td><td>11:04</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:04</td></tr> <tr><td>Control Test</td><td>0.202</td><td>11:05</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:06</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.2030</td><td></td></tr> <tr><td>Std Dev</td><td>0.0010</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.4926</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	11:02	Control Test	0.204	11:02	Air Blank	0.000	11:03	Control Test	0.203	11:04	Air Blank	0.000	11:04	Control Test	0.202	11:05	Air Blank	0.000	11:06	Control Test Stats			Average	0.2030		Std Dev	0.0010		Rel Std Dev(%)	0.4926		HOLLYWOOD PD Intoxilyzer - Alcohol Analyzer Model: 8000 08/04/2025 Software: 8100.27 SN 80-001063 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:32</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:32</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:33</td></tr> <tr><td>Control Test</td><td>0.080</td><td>11:33</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:34</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:34</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:34</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0793</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7277</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	11:32	Control Test	0.079	11:32	Air Blank	0.000	11:33	Control Test	0.080	11:33	Air Blank	0.000	11:34	Control Test	0.079	11:34	Air Blank	0.000	11:34	Control Test Stats			Average	0.0793		Std Dev	0.0006		Rel Std Dev(%)	0.7277	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:27																																																																																																																																																	
Control Test	0.049	11:28																																																																																																																																																	
Air Blank	0.000	11:28																																																																																																																																																	
Control Test	0.049	11:29																																																																																																																																																	
Air Blank	0.000	11:30																																																																																																																																																	
Control Test	0.049	11:30																																																																																																																																																	
Air Blank	0.000	11:31																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0490																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:20																																																																																																																																																	
Control Test	0.080	11:21																																																																																																																																																	
Air Blank	0.000	11:22																																																																																																																																																	
Control Test	0.080	11:22																																																																																																																																																	
Air Blank	0.000	11:23																																																																																																																																																	
Control Test	0.079	11:23																																																																																																																																																	
Air Blank	0.000	11:24																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0797																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7247																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:02																																																																																																																																																	
Control Test	0.204	11:02																																																																																																																																																	
Air Blank	0.000	11:03																																																																																																																																																	
Control Test	0.203	11:04																																																																																																																																																	
Air Blank	0.000	11:04																																																																																																																																																	
Control Test	0.202	11:05																																																																																																																																																	
Air Blank	0.000	11:06																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.2030																																																																																																																																																		
Std Dev	0.0010																																																																																																																																																		
Rel Std Dev(%)	0.4926																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:32																																																																																																																																																	
Control Test	0.079	11:32																																																																																																																																																	
Air Blank	0.000	11:33																																																																																																																																																	
Control Test	0.080	11:33																																																																																																																																																	
Air Blank	0.000	11:34																																																																																																																																																	
Control Test	0.079	11:34																																																																																																																																																	
Air Blank	0.000	11:34																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0793																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7277																																																																																																																																																		
Operator's Signature	Operator's Signature	Operator's Signature	Operator's Signature																																																																																																																																																

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: HOLLYWOOD PD
Time of Inspection: 15:20

Date of Inspection: 08/04/2025

Serial Number: 80-001063
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG429602 Exp: 10/22/2026
0.000	0.049	0.080	0.201	0.078
0.000	0.049	0.078	0.202	0.077
0.000	0.049	0.080	0.202	0.078
0.000	0.049	0.079	0.202	0.077
0.000	0.049	0.079	0.201	0.078
0.000	0.049	0.079	0.202	0.078
0.000	0.049	0.079	0.202	0.078
0.000	0.050	0.080	0.202	0.078
0.000	0.050	0.079	0.202	0.078
0.000	0.050	0.079	0.202	0.078

Standard Deviations	0.0004	0.0006	0.0004	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

08/04/2025
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number **80-001063**, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	80-001063	UNCERTAINTY* $\pm$
Owning Agency:	HOLLYWOOD PD	0.050 g/ 210 L 0.004
Calibration Date:	08/04/2025	0.080 g/ 210 L 0.004
Calibration Time:	15:20	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Taylor
Gutschow
Digitally signed by Taylor
Gutschow
Date: 2025.08.05 14:58:24
-04'00'

08/04/2025

Date

TAYLOR D GUTSCHOW,
Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality



INSTRUMENT PROCESSING SHEET

Agency Hollywood PDS/N 8-001063Florida Department of
Law EnforcementDate In 03/19/2025 DI Completion Date 04/01/2025☐ Ship ☐ P/U ☐ H/D ☒ CMI ☐ EE

Intake	By TDG	Date	Quality Checks	By TDG	Date	Flow Calibration	By	Date																									
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: <u>Dropped off with no box.</u> <u>Took several tries to power up.</u>		<u>03/19/2025</u>	☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>227</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP101</u> 32 mm <u>0.167</u> (.139 - .169) 36 mm <u>0.187</u> (.156 - .190) 53 mm <u>0.253</u> (.228 - .278) 103 mm <u>0.515</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>26932</u> ☒ Stability Checks		<u>03/19/2025</u>	☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																											
			<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6286</td><td>202303K 03/29/2025</td></tr><tr><td>0.080</td><td>MP6287</td><td>202303L 03/29/2025</td></tr><tr><td>0.200</td><td>MP6288</td><td>202304C 04/05/2025</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG429602 10/22/2026</td></tr></tbody></table>			Simulator	Serial #	Lot #/Exp	0.050	MP6286	202303K 03/29/2025	0.080	MP6287	202303L 03/29/2025	0.200	MP6288	202304C 04/05/2025	0.080 DGS	N/A	AG429602 10/22/2026	<table border="1"><thead><tr><th>Maintenance By</th><th>Date</th></tr></thead><tbody><tr><td>☐ Battery Replacement</td><td></td></tr><tr><td>☐ Dry Gas Regulator Replacement</td><td></td></tr><tr><td>☐ Breath Tube Replacement</td><td></td></tr><tr><td>☐ Other _____</td><td></td></tr></tbody></table>			Maintenance By	Date	☐ Battery Replacement		☐ Dry Gas Regulator Replacement		☐ Breath Tube Replacement		☐ Other _____	
Simulator	Serial #	Lot #/Exp																															
0.050	MP6286	202303K 03/29/2025																															
0.080	MP6287	202303L 03/29/2025																															
0.200	MP6288	202304C 04/05/2025																															
0.080 DGS	N/A	AG429602 10/22/2026																															
Maintenance By	Date																																
☐ Battery Replacement																																	
☐ Dry Gas Regulator Replacement																																	
☐ Breath Tube Replacement																																	
☐ Other _____																																	
Calibration Adjustment By TDG			Department Inspection By TDG																														
Barometric Pressure Gauge <u>1017</u> ID # <u>28199</u>			Barometric Pressure ID# <u>26932</u>																														
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td>MP5097</td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td>MP5098</td><td>23400</td><td>10/24/2025</td></tr><tr><td>0.100</td><td>MP5099</td><td>24110</td><td>03/05/2026</td></tr><tr><td>0.200</td><td>MP5100</td><td>24080</td><td>02/13/2026</td></tr><tr><td>0.300</td><td>MP5101</td><td>23410</td><td>11/01/2025</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>06723080A5</td><td>04/05/2025</td></tr></tbody></table>			Simulator	Serial #	Lot #	Expiration	0.000	MP5097	N/A	N/A	0.040	MP5098	23400	10/24/2025	0.100	MP5099	24110	03/05/2026	0.200	MP5100	24080	02/13/2026	0.300	MP5101	23410	11/01/2025	0.080 DGS	N/A	06723080A5	04/05/2025	Gauge <u>1017</u> Instrument <u>1017</u> Mouth Alcohol Solution Lot # <u>2024-A</u> Acetone Stock Solution Lot # <u>2023-B</u>		
Simulator	Serial #	Lot #	Expiration																														
0.000	MP5097	N/A	N/A																														
0.040	MP5098	23400	10/24/2025																														
0.100	MP5099	24110	03/05/2026																														
0.200	MP5100	24080	02/13/2026																														
0.300	MP5101	23410	11/01/2025																														
0.080 DGS	N/A	06723080A5	04/05/2025																														
☒ Post Calibration Adjustment Stability Checks			<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6284</td></tr><tr><td>Interferent</td><td>MP6285</td></tr><tr><td>0.050</td><td>MP6286</td></tr><tr><td>0.080</td><td>MP6287</td></tr><tr><td>0.200</td><td>MP6288</td></tr></tbody></table>			Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288																
Simulator	Serial Number																																
0.000	MP6284																																
Interferent	MP6285																																
0.050	MP6286																																
0.080	MP6287																																
0.200	MP6288																																
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td>MP6286</td><td>202406K</td><td>06/19/2026</td></tr><tr><td>0.080</td><td>MP6287</td><td>202406L</td><td>06/19/2026</td></tr><tr><td>0.200</td><td>MP6288</td><td>202304C</td><td>04/05/2025</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG429602</td><td>10/22/2026</td></tr></tbody></table>			Simulator	Serial #	Lot #	Expiration	0.050	MP6286	202406K	06/19/2026	0.080	MP6287	202406L	06/19/2026	0.200	MP6288	202304C	04/05/2025	0.080 DGS	N/A	AG429602	10/22/2026	Attachments ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Calibration Adjustment ☐ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☒ Other <u>Form 51</u>										
Simulator	Serial #	Lot #	Expiration																														
0.050	MP6286	202406K	06/19/2026																														
0.080	MP6287	202406L	06/19/2026																														
0.200	MP6288	202304C	04/05/2025																														
0.080 DGS	N/A	AG429602	10/22/2026																														
Notes/Suggested Service: <u>Performed root cause analysis on Stability Check. (TDG 3/19/25)</u> <u>Sending to CMI for power-up issue. (TDG 4/1/25)</u>			☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☐ Return to/Place into Evidentiary Use ☒ Remain Out of Evidentiary Use ☐ Conduct an Agency Inspection Before Evidentiary Use																														
			Digitally signed by Phil Nicodemus Date: 2025.04.16 09:12:12 -0400 Phil Nicodemus																														
			Digitally signed by Shayla Platt Date: 2025.04.16 13:53:33 -04'00' Shayla Platt																														
			Tech Review / Date																														
			Admin Review / Date																														

Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																				
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																				
✓ HOLLYWOOD PD Intoxilizer - Alcohol Analyzer Model 8000 03/19/2025 Software: 8100.27 SN 80-001063 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>10:22</td></tr> <tr><td>Control Test</td><td>0.050</td><td>10:23</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:23</td></tr> <tr><td>Control Test</td><td>0.049</td><td>10:24</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:24</td></tr> <tr><td>Control Test</td><td>0.050</td><td>10:25</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:25</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0497</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1625</td><td></td></tr> </table> Operator's Signature <i>ML</i>	Air Blank	0.000	10:22	Control Test	0.050	10:23	Air Blank	0.000	10:23	Control Test	0.049	10:24	Air Blank	0.000	10:24	Control Test	0.050	10:25	Air Blank	0.000	10:25	Control Test Stats			Average	0.0497		Std Dev	0.0006		Rel Std Dev(%)	1.1625		✓ HOLLYWOOD PD Intoxilizer - Alcohol Analyzer Model 8000 03/19/2025 Software: 8100.27 SN 80-001063 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>10:30</td></tr> <tr><td>Control Test</td><td>0.079</td><td>10:31</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:31</td></tr> <tr><td>Control Test</td><td>0.079</td><td>10:32</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:32</td></tr> <tr><td>Control Test</td><td>0.079</td><td>10:33</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:33</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0790</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </table> Operator's Signature <i>ML</i>	Air Blank	0.000	10:30	Control Test	0.079	10:31	Air Blank	0.000	10:31	Control Test	0.079	10:32	Air Blank	0.000	10:32	Control Test	0.079	10:33	Air Blank	0.000	10:33	Control Test Stats			Average	0.0790		Std Dev	0.0000		Rel Std Dev(%)	0.0000		✗ HOLLYWOOD PD Intoxilizer - Alcohol Analyzer Model 8000 03/19/2025 Software: 8100.27 SN 80-001063 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>10:36</td></tr> <tr><td>Control Test</td><td>0.193</td><td>10:37</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:38</td></tr> <tr><td>Control Test</td><td>0.194</td><td>10:38</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:39</td></tr> <tr><td>Control Test</td><td>0.196</td><td>10:39</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:40</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1943</td><td></td></tr> <tr><td>Std Dev</td><td>0.0015</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7860</td><td></td></tr> </table> Operator's Signature <i>ML</i>	Air Blank	0.000	10:36	Control Test	0.193	10:37	Air Blank	0.000	10:38	Control Test	0.194	10:38	Air Blank	0.000	10:39	Control Test	0.196	10:39	Air Blank	0.000	10:40	Control Test Stats			Average	0.1943		Std Dev	0.0015		Rel Std Dev(%)	0.7860		0.077 to 0.083 ✓ HOLLYWOOD PD Intoxilizer - Alcohol Analyzer Model 8000 03/19/2025 Software: 8100.27 SN 80-001063 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>10:18</td></tr> <tr><td>Control Test</td><td>0.080</td><td>10:19</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:19</td></tr> <tr><td>Control Test</td><td>0.080</td><td>10:20</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:20</td></tr> <tr><td>Control Test</td><td>0.081</td><td>10:20</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:21</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0803</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7187</td><td></td></tr> </table> Operator's Signature <i>ML</i>	Air Blank	0.000	10:18	Control Test	0.080	10:19	Air Blank	0.000	10:19	Control Test	0.080	10:20	Air Blank	0.000	10:20	Control Test	0.081	10:20	Air Blank	0.000	10:21	Control Test Stats			Average	0.0803		Std Dev	0.0006		Rel Std Dev(%)	0.7187	
Air Blank	0.000	10:22																																																																																																																																					
Control Test	0.050	10:23																																																																																																																																					
Air Blank	0.000	10:23																																																																																																																																					
Control Test	0.049	10:24																																																																																																																																					
Air Blank	0.000	10:24																																																																																																																																					
Control Test	0.050	10:25																																																																																																																																					
Air Blank	0.000	10:25																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0497																																																																																																																																						
Std Dev	0.0006																																																																																																																																						
Rel Std Dev(%)	1.1625																																																																																																																																						
Air Blank	0.000	10:30																																																																																																																																					
Control Test	0.079	10:31																																																																																																																																					
Air Blank	0.000	10:31																																																																																																																																					
Control Test	0.079	10:32																																																																																																																																					
Air Blank	0.000	10:32																																																																																																																																					
Control Test	0.079	10:33																																																																																																																																					
Air Blank	0.000	10:33																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0790																																																																																																																																						
Std Dev	0.0000																																																																																																																																						
Rel Std Dev(%)	0.0000																																																																																																																																						
Air Blank	0.000	10:36																																																																																																																																					
Control Test	0.193	10:37																																																																																																																																					
Air Blank	0.000	10:38																																																																																																																																					
Control Test	0.194	10:38																																																																																																																																					
Air Blank	0.000	10:39																																																																																																																																					
Control Test	0.196	10:39																																																																																																																																					
Air Blank	0.000	10:40																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.1943																																																																																																																																						
Std Dev	0.0015																																																																																																																																						
Rel Std Dev(%)	0.7860																																																																																																																																						
Air Blank	0.000	10:18																																																																																																																																					
Control Test	0.080	10:19																																																																																																																																					
Air Blank	0.000	10:19																																																																																																																																					
Control Test	0.080	10:20																																																																																																																																					
Air Blank	0.000	10:20																																																																																																																																					
Control Test	0.081	10:20																																																																																																																																					
Air Blank	0.000	10:21																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0803																																																																																																																																						
Std Dev	0.0006																																																																																																																																						
Rel Std Dev(%)	0.7187																																																																																																																																						

Stability Checks

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083																																	
During root cause analysis, the lid felt slightly loose. Secured the seal and repeated. The third result of the repeated stability was still outside the range. Will perform an optical cal adjust. JML 3/19/25	#2 HOLLYWOOD PD Intoxilyzer - Alcohol Analyzer Model 8000 03/19/2025 Software: 8100.27 SN 80-001063 Test g/210L Time <table border="1"> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>10:42</td></tr> <tr><td>Control Test</td><td>0.197</td><td>10:42</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:43</td></tr> <tr><td>Control Test</td><td>0.196</td><td>10:44</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:44</td></tr> <tr><td>Control Test</td><td>0.193</td><td>10:45</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:45</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1953</td><td></td></tr> <tr><td>Std Dev</td><td>0.0021</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.0657</td><td></td></tr> </tbody> </table> Operator's Signature JML	Air Blank	0.000	10:42	Control Test	0.197	10:42	Air Blank	0.000	10:43	Control Test	0.196	10:44	Air Blank	0.000	10:44	Control Test	0.193	10:45	Air Blank	0.000	10:45	Control Test Stats			Average	0.1953		Std Dev	0.0021		Rel Std Dev(%)	1.0657		X	≤0.003 of Wet
Air Blank	0.000	10:42																																		
Control Test	0.197	10:42																																		
Air Blank	0.000	10:43																																		
Control Test	0.196	10:44																																		
Air Blank	0.000	10:44																																		
Control Test	0.193	10:45																																		
Air Blank	0.000	10:45																																		
Control Test Stats																																				
Average	0.1953																																			
Std Dev	0.0021																																			
Rel Std Dev(%)	1.0657																																			

***** AUTO CAL DATA *****

<<<< CHANNEL 1 >>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.142
Std Dev = 0.01 Rel Std Dev = 5.88
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.872
Std Dev = 0.01 Rel Std Dev = 1.24
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.940
Std Dev = 0.01 Rel Std Dev = 0.46
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.704
Std Dev = 0.02 Rel Std Dev = 0.50
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.366
Std Dev = 0.02 Rel Std Dev = 0.44
Zero Order Coef = -360.43
First Order Coef = 2578.19
Second Order Coef = 25.94
Standard Deviation = 16.077234

<<<< CHANNEL 2 >>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.116
Std Dev = 0.01 Rel Std Dev = 9.19
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.542
Std Dev = 0.01 Rel Std Dev = 0.58
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.586
Std Dev = 0.00 Rel Std Dev = 0.14
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.880
Std Dev = 0.00 Rel Std Dev = 0.05
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 9.990
Std Dev = 0.02 Rel Std Dev = 0.24
Zero Order Coef = -157.71
First Order Coef = 1325.58
Second Order Coef = 12.01
Standard Deviation = 8.896792

<<<< CHANNEL 1 >>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.142
Std Dev = 0.01 Rel Std Dev = 5.88
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.872
Std Dev = 0.01 Rel Std Dev = 1.24
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.940
Std Dev = 0.01 Rel Std Dev = 0.46
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.704
Std Dev = 0.02 Rel Std Dev = 0.50
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.366
Std Dev = 0.02 Rel Std Dev = 0.44
Zero Order Coef = -360.43
First Order Coef = 2578.19
Second Order Coef = 25.94
Standard Deviation = 16.077234

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 6.8880 (-0.0120)
Sample #2 = 6.8820 (-0.0100)
Sample #3 = 6.8760 (-0.0100)
Sample #4 = 6.8810 (-0.0200)
Avg % Abs = 6.8757 (-0.0103)
STD DEV = 0.0032 (-0.0095)
REL STD DEV = 0.047 (91.978)

<<<< CHANNEL 1 >>>>
Sol Value = 0.300 g/210L ***
Fit Value = 1.4286 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12616, Sum Io = 13668
Sum Io = 12616, Sum Io = 13668
Sample % Abs (% Abs Ref)
Sample #1 = 5.4120 (-0.0150)
Sample #2 = 5.3740 (-0.0000)
Sample #3 = 5.4130 (-0.0180)
Sample #4 = 5.3710 (-0.0180)
Avg % Abs = 5.3860 (-0.0087)
STD DEV = 0.0234 (-0.0090)
REL STD DEV = 0.435 (104.160)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 10.0300 (-0.0150)
Sample #2 = 9.9740 (-0.0070)
Sample #3 = 10.0180 (-0.0090)
Sample #4 = 9.9780 (-0.0050)
Avg % Abs = 9.9900 (-0.0070)
STD DEV = 0.0243 (-0.0020)
REL STD DEV = 0.244 (28.571)

<<<< CHANNEL 1 >>>>
Sol Value = 0.200 g/210L ***
Fit Value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12639, Sum Io = 13679
Sum Io = 12639, Sum Io = 13679
Sample % Abs (% Abs Ref)
Sample #1 = 3.7300 (-0.0120)
Sample #2 = 3.7170 (-0.0190)
Sample #3 = 3.6830 (-0.0400)
Sample #4 = 3.7120 (-0.0460)
Avg % Abs = 3.7040 (-0.0350)
STD DEV = 0.0184 (-0.0142)
REL STD DEV = 0.496 (40.507)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.5460 (-0.0110)
Sample #2 = 1.5330 (-0.0070)
Sample #3 = 1.5510 (-0.0040)
Sample #4 = 1.5430 (-0.0010)
Avg % Abs = 1.5423 (-0.0013)
STD DEV = 0.0090 (-0.0055)
REL STD DEV = 0.585 (413.068)

<<<< CHANNEL 1 >>>>
Sol Value = 0.100 g/210L ***
Fit Value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12645, Sum Io = 13666
Sum Io = 12645, Sum Io = 13666
Sample % Abs (% Abs Ref)
Sample #1 = 1.9840 (-0.0420)
Sample #2 = 1.9450 (-0.0160)
Sample #3 = 1.9460 (-0.0200)
Sample #4 = 1.9300 (-0.0170)
Avg % Abs = 1.9403 (-0.0070)
STD DEV = 0.0090 (-0.0200)
REL STD DEV = 0.462 (285.357)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.5780 (-0.0110)
Sample #2 = 3.5890 (-0.0020)
Sample #3 = 3.5880 (-0.0140)
Sample #4 = 3.5800 (-0.0140)
Avg % Abs = 3.5857 (-0.0087)
STD DEV = 0.0049 (-0.0092)
REL STD DEV = 0.138 (106.588)

<<<< CHANNEL 1 >>>>
Sol Value = 0.200 g/210L ***
Fit Value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12639, Sum Io = 13679
Sum Io = 12639, Sum Io = 13679
Sample % Abs (% Abs Ref)
Sample #1 = 3.7300 (-0.0120)
Sample #2 = 3.7170 (-0.0190)
Sample #3 = 3.6830 (-0.0400)
Sample #4 = 3.7120 (-0.0460)
Avg % Abs = 3.7040 (-0.0350)
STD DEV = 0.0184 (-0.0142)
REL STD DEV = 0.496 (40.507)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.5460 (-0.0110)
Sample #2 = 1.5330 (-0.0070)
Sample #3 = 1.5510 (-0.0040)
Sample #4 = 1.5430 (-0.0010)
Avg % Abs = 1.5423 (-0.0013)
STD DEV = 0.0090 (-0.0055)
REL STD DEV = 0.585 (413.068)

<<<< CHANNEL 1 >>>>
Sol Value = 0.100 g/210L ***
Fit Value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12645, Sum Io = 13666
Sum Io = 12645, Sum Io = 13666
Sample % Abs (% Abs Ref)
Sample #1 = 1.9840 (-0.0420)
Sample #2 = 1.9450 (-0.0160)
Sample #3 = 1.9460 (-0.0200)
Sample #4 = 1.9300 (-0.0170)
Avg % Abs = 1.9403 (-0.0070)
STD DEV = 0.0090 (-0.0200)
REL STD DEV = 0.462 (285.357)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.5780 (-0.0110)
Sample #2 = 3.5890 (-0.0020)
Sample #3 = 3.5880 (-0.0140)
Sample #4 = 3.5800 (-0.0140)
Avg % Abs = 3.5857 (-0.0087)
STD DEV = 0.0049 (-0.0092)
REL STD DEV = 0.138 (106.588)

<<<< CHANNEL 1 >>>>
Sol Value = 0.200 g/210L ***
Fit Value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12639, Sum Io = 13679
Sum Io = 12639, Sum Io = 13679
Sample % Abs (% Abs Ref)
Sample #1 = 3.7300 (-0.0120)
Sample #2 = 3.7170 (-0.0190)
Sample #3 = 3.6830 (-0.0400)
Sample #4 = 3.7120 (-0.0460)
Avg % Abs = 3.7040 (-0.0350)
STD DEV = 0.0184 (-0.0142)
REL STD DEV = 0.496 (40.507)

Optical Calibration
Adjustment
By: TDG

Post-Cal Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																				
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																				
✓ HOLLYWOOD PD Intoxilyzer - Alcohol Analyzer Model 8000 04/01/2025 Software: 8100.27 SN 80-001063 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>12:49</td></tr> <tr><td>Control Test</td><td>0.049</td><td>12:49</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:50</td></tr> <tr><td>Control Test</td><td>0.049</td><td>12:51</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:51</td></tr> <tr><td>Control Test</td><td>0.049</td><td>12:52</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:52</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0490</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </table>	Air Blank	0.000	12:49	Control Test	0.049	12:49	Air Blank	0.000	12:50	Control Test	0.049	12:51	Air Blank	0.000	12:51	Control Test	0.049	12:52	Air Blank	0.000	12:52	Control Test Stats			Average	0.0490		Std Dev	0.0000		Rel Std Dev(%)	0.0000		✓ HOLLYWOOD PD Intoxilyzer - Alcohol Analyzer Model 8000 04/01/2025 Software: 8100.27 SN 80-001063 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>12:59</td></tr> <tr><td>Control Test</td><td>0.079</td><td>13:00</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:00</td></tr> <tr><td>Control Test</td><td>0.079</td><td>13:01</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:01</td></tr> <tr><td>Control Test</td><td>0.079</td><td>13:02</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:03</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0790</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </table>	Air Blank	0.000	12:59	Control Test	0.079	13:00	Air Blank	0.000	13:00	Control Test	0.079	13:01	Air Blank	0.000	13:01	Control Test	0.079	13:02	Air Blank	0.000	13:03	Control Test Stats			Average	0.0790		Std Dev	0.0000		Rel Std Dev(%)	0.0000		✓ HOLLYWOOD PD Intoxilyzer - Alcohol Analyzer Model 8000 04/01/2025 Software: 8100.27 SN 80-001063 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>13:09</td></tr> <tr><td>Control Test</td><td>0.198</td><td>13:09</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:10</td></tr> <tr><td>Control Test</td><td>0.199</td><td>13:11</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:11</td></tr> <tr><td>Control Test</td><td>0.198</td><td>13:12</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:12</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1983</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2911</td><td></td></tr> </table>	Air Blank	0.000	13:09	Control Test	0.198	13:09	Air Blank	0.000	13:10	Control Test	0.199	13:11	Air Blank	0.000	13:11	Control Test	0.198	13:12	Air Blank	0.000	13:12	Control Test Stats			Average	0.1983		Std Dev	0.0006		Rel Std Dev(%)	0.2911		HOLLYWOOD PD Intoxilyzer - Alcohol Analyzer Model 8000 04/01/2025 Software: 8100.27 SN 80-001063 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>12:42</td></tr> <tr><td>Control Test</td><td>0.080</td><td>12:42</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:43</td></tr> <tr><td>Control Test</td><td>0.080</td><td>12:43</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:44</td></tr> <tr><td>Control Test</td><td>0.080</td><td>12:44</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:45</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0800</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </table>	Air Blank	0.000	12:42	Control Test	0.080	12:42	Air Blank	0.000	12:43	Control Test	0.080	12:43	Air Blank	0.000	12:44	Control Test	0.080	12:44	Air Blank	0.000	12:45	Control Test Stats			Average	0.0800		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Air Blank	0.000	12:49																																																																																																																																					
Control Test	0.049	12:49																																																																																																																																					
Air Blank	0.000	12:50																																																																																																																																					
Control Test	0.049	12:51																																																																																																																																					
Air Blank	0.000	12:51																																																																																																																																					
Control Test	0.049	12:52																																																																																																																																					
Air Blank	0.000	12:52																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0490																																																																																																																																						
Std Dev	0.0000																																																																																																																																						
Rel Std Dev(%)	0.0000																																																																																																																																						
Air Blank	0.000	12:59																																																																																																																																					
Control Test	0.079	13:00																																																																																																																																					
Air Blank	0.000	13:00																																																																																																																																					
Control Test	0.079	13:01																																																																																																																																					
Air Blank	0.000	13:01																																																																																																																																					
Control Test	0.079	13:02																																																																																																																																					
Air Blank	0.000	13:03																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0790																																																																																																																																						
Std Dev	0.0000																																																																																																																																						
Rel Std Dev(%)	0.0000																																																																																																																																						
Air Blank	0.000	13:09																																																																																																																																					
Control Test	0.198	13:09																																																																																																																																					
Air Blank	0.000	13:10																																																																																																																																					
Control Test	0.199	13:11																																																																																																																																					
Air Blank	0.000	13:11																																																																																																																																					
Control Test	0.198	13:12																																																																																																																																					
Air Blank	0.000	13:12																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.1983																																																																																																																																						
Std Dev	0.0006																																																																																																																																						
Rel Std Dev(%)	0.2911																																																																																																																																						
Air Blank	0.000	12:42																																																																																																																																					
Control Test	0.080	12:42																																																																																																																																					
Air Blank	0.000	12:43																																																																																																																																					
Control Test	0.080	12:43																																																																																																																																					
Air Blank	0.000	12:44																																																																																																																																					
Control Test	0.080	12:44																																																																																																																																					
Air Blank	0.000	12:45																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0800																																																																																																																																						
Std Dev	0.0000																																																																																																																																						
Rel Std Dev(%)	0.0000																																																																																																																																						
Operator's Signature <i>[Signature]</i>	Operator's Signature <i>[Signature]</i>	Operator's Signature <i>[Signature]</i>	Operator's Signature <i>[Signature]</i>																																																																																																																																				

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: HOLLYWOOD PD
Time of Inspection: 15:49

Date of Inspection: 04/01/2025

Serial Number: 80-001063
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG429602 Exp: 10/22/2026
0.000	0.049	0.078	0.197	0.079
0.000	0.049	0.078	0.196	0.080
0.000	0.048	0.079	0.196	0.081
0.000	0.049	0.078	0.197	0.080
0.000	0.049	0.078	0.196	0.080
0.000	0.049	0.078	0.197	0.081
0.000	0.049	0.078	0.197	0.081
0.000	0.049	0.078	0.196	0.081
0.000	0.049	0.078	0.197	0.081
0.000	0.049	0.079	0.197	0.081


Standard Deviations	0.0003	0.0004	0.0005	0.0007
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



Signature and Printed Name

TAYLOR D GUTSCHOW

04/01/2025
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001063, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001063</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>HOLLYWOOD PD</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>04/01/2025</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>15:49</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

04/01/2025

Date
TAYLOR D GUTSCHOW,
Department Inspector

FDLE/ATP Form 69 December 2021
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Return Material Authorization

Ship to:



CMI, Inc.



Enforcement Electronics

Shipment to repair facility authorized by: Rafael Deriet on 04/01/2025

Items Returned: Instrument ☒ Supplies ☐ Other ☐ Describe: _____

Instrument Model: Intoxilyzer 8000 Serial Number: 80-001063

Bill To Address:

Hollywood Police Department

Attn: Rafael Deriet

Ship to Address:

Florida Department of Law Enforcement

Fort Myers Regional Operations Center

Attn: Taylor Gutschow

4700 Terminal Drive, Suite 1

Fort Myers, FL 33907

Reason for Return:

Instrument requires multiple attempts to power on.

Please choose one of the following options:

☐ 1. I _____, authorize all repairs.

☐ 2. I _____, authorize repairs up to \$_____.

☒ 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Rafael Deriet

Phone #: 1-305-4090626

Email: RDeriet@hollywoodfl.org

ATP Contact Name: Taylor Gutschow

ATP Email: TaylorGutschow@fdle.state.fl.us