



Agency City of Miramar Police

S/N 80-001067

Date In 06/12/2025 DI Completion Date 07/11/2025

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake By TDG Date 07/10/2025

- ☒ Annual
- ☐ Registration
- ☐ Return from CMI / EE

- Case
- Keyboard
- Feet
- Ports
- Handle
- Dry Gas Shelf
- Breath Tube
- Screws Tight

- Power cord
- Static Bag
- Printer Cable
- 12V DC Cable

Notes: _____

Quality Checks By TDG Date 07/11/2025

☒ Breath Tube Screen
☒ Replace External O-Rings
☒ Instrument Set Up Verified
☒ R-Value 216
☒ Flow Verification (L/s)
 Flow Column # ATP104
 32 mm 0.152 (.139 - .169)
 36 mm 0.167 (.156 - .190)
 53 mm 0.242 (.228 - .278)
 103 mm 0.496 (.447 - .547)

■ Barometric Pressure Check
Gauge ID # 28199

Simulator	Serial #	Lot #/Exp
0.050	MP6286	202406K 06/19/2026
0.080	MP6287	202406L 06/19/2026
0.200	MP6288	202406N 06/20/2026
0.080 DGS	N/A	AG429602 10/22/2026

Flow Calibration By_____ Date_____

Maintenance By _____ Date _____

☐ Battery Replacement

☐ Dry Gas Regulator Replacement

☐ Breath Tube Replacement

☐ Other _____

Barometric Pressure Gauge _____ ID # _____

☐ Post Calibration Adjustment Stability Checks

Simulator	Serial #	Lot #	Expiration
0.050			
0.080			
0.200			
0.080 DGS	N/A		

Notes/Suggested Service: _____

Barometric Pressure ID# 28199

Gauge 1020 Instrument 1019
Mouth Alcohol Solution Lot # 2024-A
Acetone Stock Solution Lot # 2023-B

Simulator	Serial Number
0.000	MP6284
Interferent	MP6285
0.050	MP6286
0.080	MP6287
0.200	MP6288

☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment	☐ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____
---	---

- ☒ Instrument Complies with Chapter 11D-8, FAC
- ☐ Instrument Does Not Comply with Chapter 11D-8, FAC

- ☒ Return to/Place into Evidentiary Use
- ☐ Remain Out of Evidentiary Use

- ☒ Conduct an Agency Inspection Before Evidentiary Use

Phil Nicodemo Digitally signed by Phil Nicodemo
Date: 2025.07.15 12:12:10 -0400

Tech Review / Date

Shayla Platt

Admin Review / Date

Stability Checks

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083 ✓ ≤0.003 of Wet ✓																																																																																																																																																
CITY OF MIRAMAR POL Intoxilyzer - Alcohol Analyzer Model 8000 07/11/2025 Software: 8100.27 SN 80-001067 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:46</td></tr> <tr><td>Control Test</td><td>0.049</td><td>12:47</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:47</td></tr> <tr><td>Control Test</td><td>0.049</td><td>12:48</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:48</td></tr> <tr><td>Control Test</td><td>0.049</td><td>12:49</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:50</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0490</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	12:46	Control Test	0.049	12:47	Air Blank	0.000	12:47	Control Test	0.049	12:48	Air Blank	0.000	12:48	Control Test	0.049	12:49	Air Blank	0.000	12:50	Control Test Stats			Average	0.0490		Std Dev	0.0000		Rel Std Dev(%)	0.0000		CITY OF MIRAMAR POL Intoxilyzer - Alcohol Analyzer Model 8000 07/11/2025 Software: 8100.27 SN 80-001067 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:53</td></tr> <tr><td>Control Test</td><td>0.080</td><td>12:53</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:54</td></tr> <tr><td>Control Test</td><td>0.079</td><td>12:55</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:55</td></tr> <tr><td>Control Test</td><td>0.080</td><td>12:56</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:56</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0797</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7247</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	12:53	Control Test	0.080	12:53	Air Blank	0.000	12:54	Control Test	0.079	12:55	Air Blank	0.000	12:55	Control Test	0.080	12:56	Air Blank	0.000	12:56	Control Test Stats			Average	0.0797		Std Dev	0.0006		Rel Std Dev(%)	0.7247		CITY OF MIRAMAR POL Intoxilyzer - Alcohol Analyzer Model 8000 07/11/2025 Software: 8100.27 SN 80-001067 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:38</td></tr> <tr><td>Control Test</td><td>0.200</td><td>12:39</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:39</td></tr> <tr><td>Control Test</td><td>0.198</td><td>12:40</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:40</td></tr> <tr><td>Control Test</td><td>0.198</td><td>12:41</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:42</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1987</td><td></td></tr> <tr><td>Std Dev</td><td>0.0012</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.5812</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	12:38	Control Test	0.200	12:39	Air Blank	0.000	12:39	Control Test	0.198	12:40	Air Blank	0.000	12:40	Control Test	0.198	12:41	Air Blank	0.000	12:42	Control Test Stats			Average	0.1987		Std Dev	0.0012		Rel Std Dev(%)	0.5812		CITY OF MIRAMAR POL Intoxilyzer - Alcohol Analyzer Model 8000 07/11/2025 Software: 8100.27 SN 80-001067 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:58</td></tr> <tr><td>Control Test</td><td>0.077</td><td>12:58</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:58</td></tr> <tr><td>Control Test</td><td>0.077</td><td>12:59</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:59</td></tr> <tr><td>Control Test</td><td>0.078</td><td>13:00</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:00</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0773</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7456</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	12:58	Control Test	0.077	12:58	Air Blank	0.000	12:58	Control Test	0.077	12:59	Air Blank	0.000	12:59	Control Test	0.078	13:00	Air Blank	0.000	13:00	Control Test Stats			Average	0.0773		Std Dev	0.0006		Rel Std Dev(%)	0.7456	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:46																																																																																																																																																	
Control Test	0.049	12:47																																																																																																																																																	
Air Blank	0.000	12:47																																																																																																																																																	
Control Test	0.049	12:48																																																																																																																																																	
Air Blank	0.000	12:48																																																																																																																																																	
Control Test	0.049	12:49																																																																																																																																																	
Air Blank	0.000	12:50																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0490																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:53																																																																																																																																																	
Control Test	0.080	12:53																																																																																																																																																	
Air Blank	0.000	12:54																																																																																																																																																	
Control Test	0.079	12:55																																																																																																																																																	
Air Blank	0.000	12:55																																																																																																																																																	
Control Test	0.080	12:56																																																																																																																																																	
Air Blank	0.000	12:56																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0797																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7247																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:38																																																																																																																																																	
Control Test	0.200	12:39																																																																																																																																																	
Air Blank	0.000	12:39																																																																																																																																																	
Control Test	0.198	12:40																																																																																																																																																	
Air Blank	0.000	12:40																																																																																																																																																	
Control Test	0.198	12:41																																																																																																																																																	
Air Blank	0.000	12:42																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1987																																																																																																																																																		
Std Dev	0.0012																																																																																																																																																		
Rel Std Dev(%)	0.5812																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:58																																																																																																																																																	
Control Test	0.077	12:58																																																																																																																																																	
Air Blank	0.000	12:58																																																																																																																																																	
Control Test	0.077	12:59																																																																																																																																																	
Air Blank	0.000	12:59																																																																																																																																																	
Control Test	0.078	13:00																																																																																																																																																	
Air Blank	0.000	13:00																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0773																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7456																																																																																																																																																		

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: CITY OF MIRAMAR POL
Time of Inspection: 15:50

Date of Inspection: 07/11/2025

Serial Number: 80-001067
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG429602 Exp: 10/22/2026
0.000	0.049	0.078	0.199	0.077
0.000	0.049	0.079	0.198	0.078
0.000	0.049	0.079	0.199	0.077
0.000	0.049	0.078	0.198	0.077
0.000	0.049	0.079	0.198	0.077
0.000	0.049	0.078	0.198	0.077
0.000	0.049	0.079	0.198	0.077
0.000	0.049	0.079	0.198	0.077
0.000	0.049	0.078	0.197	0.077
0.000	0.049	0.078	0.197	0.077

Standard Deviations	0.0000	0.0005	0.0006	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

07/11/2025
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001067, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001067</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>CITY OF MIRAMAR POL</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>07/11/2025</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>15:50</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. The supplier of dry gas standard controls is Taylor Gutschow. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Taylor Gutschow
Digitally signed by
Taylor Gutschow
Date: 2025.07.14
10:38:40 -04'00'

07/11/2025 Date
TAYLOR D GUTSCHOW,
Department Inspector

FDLE/ATP Form 69 December 2021
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality