



INSTRUMENT PROCESSING SHEET

Agency Clearwater PDS/N 80-001068Florida Department of
Law EnforcementDate In 01/23/2025 DI Completion Date 02/14/2025☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By	TDG	Date	02/04/2025	Quality Checks	By	TDG	Date	02/05/2025	Flow Calibration	By	Date																								
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☒ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: <u>AI reports the instrument will not pass diagnostics after a subject spit into the breath hose.</u> <u>Powered on. Instrument passed diagnostics.</u>					☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>150</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP101</u> 32 mm <u>0.152</u> (.139 - .169) 36 mm <u>0.171</u> (.156 - .190) 53 mm <u>0.242</u> (.228 - .278) 103 mm <u>0.515</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>26932</u> ☒ Stability Checks					Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																										
					<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6286</td><td>202303K 03/29/2025</td></tr><tr><td>0.080</td><td>MP6287</td><td>202303L 03/29/2025</td></tr><tr><td>0.200</td><td>MP6288</td><td>202304C 04/05/2025</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG429602 10/22/2026</td></tr></tbody></table>			Simulator	Serial #	Lot #/Exp	0.050	MP6286	202303K 03/29/2025	0.080	MP6287	202303L 03/29/2025	0.200	MP6288	202304C 04/05/2025	0.080 DGS	N/A	AG429602 10/22/2026	<table border="1"><thead><tr><th colspan="2">Maintenance By _____ Date _____</th></tr><tr><td>☐ Battery Replacement</td><td></td></tr><tr><td>☐ Dry Gas Regulator Replacement</td><td></td></tr><tr><td>☐ Breath Tube Replacement</td><td></td></tr><tr><td>☐ Other _____</td><td></td></tr></thead></table>				Maintenance By _____ Date _____		☐ Battery Replacement		☐ Dry Gas Regulator Replacement		☐ Breath Tube Replacement		☐ Other _____	
Simulator	Serial #	Lot #/Exp																																		
0.050	MP6286	202303K 03/29/2025																																		
0.080	MP6287	202303L 03/29/2025																																		
0.200	MP6288	202304C 04/05/2025																																		
0.080 DGS	N/A	AG429602 10/22/2026																																		
Maintenance By _____ Date _____																																				
☐ Battery Replacement																																				
☐ Dry Gas Regulator Replacement																																				
☐ Breath Tube Replacement																																				
☐ Other _____																																				

Calibration Adjustment By TDG					**Department Inspection** By TDG																																																																				
Barometric Pressure Gauge 1022 ID # 28199					Barometric Pressure ID# 26932 Gauge 1027 Instrument 1027 Mouth Alcohol Solution Lot # 2024-A Acetone Stock Solution Lot # 2023-B																																																																				
	Simulator	Serial #	Lot #	Expiration		-----------	----------	------------	------------		0.000	MP5097	N/A	N/A		0.040	MP5098	23400	10/24/2025		0.100	MP5099	24110	03/05/2026		0.200	MP5100	24080	02/13/2026		0.300	MP5101	23410	11/01/2025		0.080 DGS	N/A	06723080A5	04/05/2025							Simulator	Serial Number		-------------	---------------		0.000	MP6284		Interferent	MP6285		0.050	MP6286		0.080	MP6287		0.200	MP6288								
☒ Post Calibration Adjustment Stability Checks					**Attachments**																																																																				
	Simulator	Serial #	Lot #	Expiration		-----------	----------	----------	------------		0.050	MP6286	202303K	03/29/2025		0.080	MP6287	202303L	03/29/2025		0.200	MP6288	202304C	04/05/2025		0.080 DGS	N/A	AG429602	10/22/2026										---	---		☒ Form 41	☒ Post-Stability Checks		☒ Stability Checks	☐ Flow Calibration		☒ Calibration Certificate	☒ Form 40		☒ Calibration Adjustment	☐ Other _____																					
Notes/Suggested Service: Added static bag. (TDG) Instrument passed warm-up diagnostics on 2/4, 2/5, 2/13, and 2/14. (TDG)					☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																																																																				
					Digitally signed by Phil Nicodemo Date: 2025.02.21 15:11:28 -05'00'																																																																				
					Shayla Platt Date: 2025.02.21 15:44:59 -05'00'																																																																				
					Tech Review / Date _____ Admin Review / Date _____																																																																				

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: CLEARWATER PD
Time of Inspection: 14:43

Date of Inspection: 02/04/2025

Serial Number: 80-001068
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____

Remarks:

AI NOT CONDUCTED, BYPASSED TO BRING OUT OF DISABLED MODE.

Not determined ^{ML} 2/4/25

The above instrument complies (☒) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

02/04/2025
Date

Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																				
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																				
✓ CLEARWATER PD Intoxilyzer - Alcohol Analyzer Model 8000 02/05/2025 Software: 8100.27 SN 80-001068 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>10:37</td></tr> <tr><td>Control Test</td><td>0.049</td><td>10:37</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:38</td></tr> <tr><td>Control Test</td><td>0.050</td><td>10:39</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:39</td></tr> <tr><td>Control Test</td><td>0.050</td><td>10:40</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:40</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0497</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1625</td><td></td></tr> </table>	Air Blank	0.000	10:37	Control Test	0.049	10:37	Air Blank	0.000	10:38	Control Test	0.050	10:39	Air Blank	0.000	10:39	Control Test	0.050	10:40	Air Blank	0.000	10:40	Control Test Stats			Average	0.0497		Std Dev	0.0006		Rel Std Dev(%)	1.1625		✓ CLEARWATER PD Intoxilyzer - Alcohol Analyzer Model 8000 02/05/2025 Software: 8100.27 SN 80-001068 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>10:43</td></tr> <tr><td>Control Test</td><td>0.078</td><td>10:44</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:44</td></tr> <tr><td>Control Test</td><td>0.079</td><td>10:45</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:45</td></tr> <tr><td>Control Test</td><td>0.080</td><td>10:46</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:47</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0790</td><td></td></tr> <tr><td>Std Dev</td><td>0.0010</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.2658</td><td></td></tr> </table>	Air Blank	0.000	10:43	Control Test	0.078	10:44	Air Blank	0.000	10:44	Control Test	0.079	10:45	Air Blank	0.000	10:45	Control Test	0.080	10:46	Air Blank	0.000	10:47	Control Test Stats			Average	0.0790		Std Dev	0.0010		Rel Std Dev(%)	1.2658		✗ CLEARWATER PD Intoxilyzer - Alcohol Analyzer Model 8000 02/05/2025 Software: 8100.27 SN 80-001068 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>10:28</td></tr> <tr><td>Control Test</td><td>0.191</td><td>10:28</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:29</td></tr> <tr><td>Control Test</td><td>0.193</td><td>10:30</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:30</td></tr> <tr><td>Control Test</td><td>0.194</td><td>10:31</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:31</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1927</td><td></td></tr> <tr><td>Std Dev</td><td>0.0015</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7928</td><td></td></tr> </table>	Air Blank	0.000	10:28	Control Test	0.191	10:28	Air Blank	0.000	10:29	Control Test	0.193	10:30	Air Blank	0.000	10:30	Control Test	0.194	10:31	Air Blank	0.000	10:31	Control Test Stats			Average	0.1927		Std Dev	0.0015		Rel Std Dev(%)	0.7928		✓ CLEARWATER PD Intoxilyzer - Alcohol Analyzer Model 8000 02/05/2025 Software: 8100.27 SN 80-001068 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>10:01</td></tr> <tr><td>Control Test</td><td>0.077</td><td>10:02</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:02</td></tr> <tr><td>Control Test</td><td>0.077</td><td>10:03</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:03</td></tr> <tr><td>Control Test</td><td>0.078</td><td>10:03</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:04</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0773</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7466</td><td></td></tr> </table>	Air Blank	0.000	10:01	Control Test	0.077	10:02	Air Blank	0.000	10:02	Control Test	0.077	10:03	Air Blank	0.000	10:03	Control Test	0.078	10:03	Air Blank	0.000	10:04	Control Test Stats			Average	0.0773		Std Dev	0.0006		Rel Std Dev(%)	0.7466	
Air Blank	0.000	10:37																																																																																																																																					
Control Test	0.049	10:37																																																																																																																																					
Air Blank	0.000	10:38																																																																																																																																					
Control Test	0.050	10:39																																																																																																																																					
Air Blank	0.000	10:39																																																																																																																																					
Control Test	0.050	10:40																																																																																																																																					
Air Blank	0.000	10:40																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0497																																																																																																																																						
Std Dev	0.0006																																																																																																																																						
Rel Std Dev(%)	1.1625																																																																																																																																						
Air Blank	0.000	10:43																																																																																																																																					
Control Test	0.078	10:44																																																																																																																																					
Air Blank	0.000	10:44																																																																																																																																					
Control Test	0.079	10:45																																																																																																																																					
Air Blank	0.000	10:45																																																																																																																																					
Control Test	0.080	10:46																																																																																																																																					
Air Blank	0.000	10:47																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0790																																																																																																																																						
Std Dev	0.0010																																																																																																																																						
Rel Std Dev(%)	1.2658																																																																																																																																						
Air Blank	0.000	10:28																																																																																																																																					
Control Test	0.191	10:28																																																																																																																																					
Air Blank	0.000	10:29																																																																																																																																					
Control Test	0.193	10:30																																																																																																																																					
Air Blank	0.000	10:30																																																																																																																																					
Control Test	0.194	10:31																																																																																																																																					
Air Blank	0.000	10:31																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.1927																																																																																																																																						
Std Dev	0.0015																																																																																																																																						
Rel Std Dev(%)	0.7928																																																																																																																																						
Air Blank	0.000	10:01																																																																																																																																					
Control Test	0.077	10:02																																																																																																																																					
Air Blank	0.000	10:02																																																																																																																																					
Control Test	0.077	10:03																																																																																																																																					
Air Blank	0.000	10:03																																																																																																																																					
Control Test	0.078	10:03																																																																																																																																					
Air Blank	0.000	10:04																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0773																																																																																																																																						
Std Dev	0.0006																																																																																																																																						
Rel Std Dev(%)	0.7466																																																																																																																																						
Operator's Signature <i>ML</i>	Operator's Signature <i>ML</i>	Operator's Signature <i>ML</i>	Operator's Signature <i>ML</i>																																																																																																																																				

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.5680 (-0.0020)
Sample #2 = 1.5730 (-0.0140)
Sample #3 = 1.5850 (-0.0100)
Sample #4 = 1.5680 (-0.0070)
Avg % Abs = 1.5733 (-0.0103)
STD DEV = 0.0087 (0.0035)
REL STD DEV = 0.555 (33.966)

CLEARWATER PD
IncoXalyzer - Alcohol Analyzer
Model 8000
02/13/2025
SN 80-001068
09:30:24

Sol Value = 0.100 g/210L ***
Fit value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12553, Sum Io = 14234
<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.9060 (-0.0030)
Sample #2 = 1.9050 (0.0220)
Sample #3 = 1.9230 (0.0040)
Sample #4 = 1.8930 (0.0360)
Avg % Abs = 1.9070 (0.0207)
STD DEV = 0.0151 (0.0160)
REL STD DEV = 0.792 (77.621)

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.6830 (-0.0070)
Sample #2 = 3.6640 (0.0210)
Sample #3 = 3.6940 (0.0050)
Sample #4 = 3.6790 (0.0160)
Avg % Abs = 3.6790 (0.0140)
STD DEV = 0.0150 (0.0082)
REL STD DEV = 0.408 (58.467)

Sol Value = 0.200 g/210L ***
Fit value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12547, Sum Io = 14231
<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.7100 (-0.0310)
Sample #2 = 3.6950 (-0.0130)
Sample #3 = 3.7050 (-0.0070)
Sample #4 = 3.6960 (0.0180)
Avg % Abs = 3.6987 (-0.0007)
STD DEV = 0.0055 (0.0154)
REL STD DEV = 0.149 (2466.271)

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.1330 (0.0000)
Sample #2 = 0.0870 (0.0390)
Sample #3 = 0.0670 (0.0420)
Sample #4 = 0.0890 (0.0440)
Avg % Abs = 0.0810 (0.0417)
STD DEV = 0.0122 (0.0025)
REL STD DEV = 15.019 (6.040)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12554, Sum Io = 14234
<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.8620 (-0.0260)
Sample #2 = 0.8180 (-0.0120)
Sample #3 = 0.8560 (-0.0120)
Sample #4 = 0.8230 (0.0020)
Avg % Abs = 0.8323 (-0.0073)
STD DEV = 0.0206 (0.0081)
REL STD DEV = 2.481 (110.221)

***** AUTO CAL DATA *****
<<<<< CHANNEL 1 >>>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.063
Std Dev = 0.00 Rel Std Dev = 3.17
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.832
Std Dev = 0.02 Rel Std Dev = 2.48
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.917
Std Dev = 0.02 Rel Std Dev = 0.79
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.699
Std Dev = 0.01 Rel Std Dev = 0.15
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.441
Std Dev = 0.02 Rel Std Dev = 0.33
First Order Coef = -183.95
Second Order Coef = 2543.57
Standard Deviation = 26.38326

<<<<< CHANNEL 2 >>>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.081
Std Dev = 0.01 Rel Std Dev = 15.02
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.575
Std Dev = 0.01 Rel Std Dev = 0.55
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.679
Std Dev = 0.01 Rel Std Dev = 0.41
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 7.073
Std Dev = 0.02 Rel Std Dev = 0.21
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 10.325
Std Dev = 0.01 Rel Std Dev = 0.08
Zero Order Coef = -122.06
First Order Coef = 1287.97
Second Order Coef = 10.46
Standard Deviation = 17.955306

Solution Stats Quadratic Fit Chan 1
Act Fit Residual
g/210L g/210L g/210L
0.000 -0.000 0.0005
0.040 0.041 -0.0009
0.100 0.100 0.0004
0.200 0.200 0.0001
0.300 0.300 -0.0001

Solution Stats Quadratic Fit Chan 2
Act Fit Residual
g/210L g/210L g/210L
0.000 -0.000 0.0004
0.040 0.041 -0.0006
0.100 0.100 0.0001
0.200 0.200 0.0003
0.300 0.300 -0.0001

Sol Value = 0.060 g/210L ***
Fit value = 0.3810 mg/l %%%
Samples Taken = 4, Discarded = 1
***** CHANNEL 1 *****
Sample #1 = 3254.00
Sample #2 = 3241.00
Sample #3 = 3215.00
Sample #4 = 3181.00
Average Result = 3212.3333
STD DEV = 30.0888
REL STD DEV = 0.937

***** CHANNEL 2 *****
Sample #1 = 3460.00
Sample #2 = 3468.00
Sample #3 = 3465.00
Sample #4 = 3457.00
Average Result = 3463.3333
STD DEV = 5.6862
REL STD DEV = 0.164

Dry Gas H2O Adjust Results *****
Barometric Pressure = 1022
3 um H2O Adjust (mg/l*10,000) = 597
9 um H2O Adjust (mg/l*10,000) = 346
***** AUTO CAL PASS *****

Optical Calibration
Adjustment
By: TDG

Post-Cal Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
✓ CLEARWATER PD Intoxilizer - Alcohol Analyzer Model 8000 SN 80-001068 02/13/2025 Software: 8100.27 Test g/210L Time Air Blank 0.000 11:12 Control Test 0.050 11:12 Air Blank 0.000 11:13 Control Test 0.049 11:14 Air Blank 0.000 11:14 Control Test 0.050 11:15 Air Blank 0.000 11:15 Control Test Stats Average 0.0497 Std Dev 0.0006 Rel Std Dev(%) 1.1625	✓ CLEARWATER PD Intoxilizer - Alcohol Analyzer Model 8000 SN 80-001068 02/13/2025 Software: 8100.27 Test g/210L Time Air Blank 0.000 11:18 Control Test 0.079 11:19 Air Blank 0.000 11:20 Control Test 0.079 11:20 Air Blank 0.000 11:21 Control Test 0.079 11:21 Air Blank 0.000 11:22 Control Test Stats Average 0.0790 Std Dev 0.0000 Rel Std Dev(%) 0.0000	✓ CLEARWATER PD Intoxilizer - Alcohol Analyzer Model 8000 SN 80-001068 02/13/2025 Software: 8100.27 Test g/210L Time Air Blank 0.000 11:28 Control Test 0.197 11:29 Air Blank 0.000 11:29 Control Test 0.197 11:30 Air Blank 0.000 11:30 Control Test 0.197 11:31 Air Blank 0.000 11:32 Control Test Stats Average 0.1970 Std Dev 0.0000 Rel Std Dev(%) 0.0000	06/25 CLEARWATER PD Intoxilizer - Alcohol Analyzer Model 8000 SN 80-001068 02/13/2025 Software: 8100.27 Test g/210L Time Air Blank 0.000 11:02 Control Test 0.080 11:03 Air Blank 0.000 11:03 Control Test 0.080 11:03 Air Blank 0.000 11:04 Control Test 0.080 11:04 Air Blank 0.000 11:05 Control Test Stats Average 0.0800 Std Dev 0.0000 Rel Std Dev(%) 0.0000
Operator's Signature <i>W</i>	Operator's Signature <i>W</i>	Operator's Signature <i>W</i>	Operator's Signature <i>W</i>

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: CLEARWATER PD
Time of Inspection: 12:43

Date of Inspection: 02/14/2025

Serial Number: 80-001068
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG429602 Exp: 10/22/2026
0.000	0.050	0.078	0.196	0.081
0.000	0.049	0.079	0.196	0.080
0.000	0.050	0.078	0.195	0.080
0.000	0.050	0.078	0.196	0.080
0.000	0.049	0.078	0.197	0.080
0.000	0.050	0.078	0.196	0.080
0.000	0.049	0.079	0.197	0.081
0.000	0.050	0.079	0.196	0.080
0.000	0.050	0.079	0.196	0.080
0.000	0.050	0.079	0.197	0.080

Standard Deviations	0.0004	0.0005	0.0006	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.


I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

02/14/2025
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001068, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001068</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>CLEARWATER PD</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>02/14/2025</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>12:43</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

02/14/2025

Date

TAYLOR D GUTSCHOW,
Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality