



Agency Florida Highway Patrol Troop C

S/N 80-001109

Date In 06/25/2025 DI Completion Date 07/17/2025

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Calibration Adjustment

By_

ID

☐ Post Calibration Adjustment Stability ChecksNotes/Suggested Service:

Department Inspection

By TDG

Acetone Stock Solution Lot # 2023-B

Attachments

- ☐
- Instrument Does Not Comply with Chapter 11D-8, FAC

- ☐
- Remain Out of Evidentiary Use**

- ### ■ Conduct an Agency Inspection Before Evidentiary Use

Phil Nicodemo

Shayla Platt
Date: 2025.07.19

Tech Review / Date

Admin Review / Date

Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																				
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																				
FL HIGHWAY PATROL Intoxilyzer - Alconol Analyzer Model 8000 07/15/2025 Software: 8100.27 SN 80-001109 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>14:04</td></tr> <tr><td>Control Test</td><td>0.048</td><td>14:05</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:06</td></tr> <tr><td>Control Test</td><td>0.048</td><td>14:06</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:07</td></tr> <tr><td>Control Test</td><td>0.048</td><td>14:08</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:08</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0480</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </table>	Air Blank	0.000	14:04	Control Test	0.048	14:05	Air Blank	0.000	14:06	Control Test	0.048	14:06	Air Blank	0.000	14:07	Control Test	0.048	14:08	Air Blank	0.000	14:08	Control Test Stats			Average	0.0480		Std Dev	0.0000		Rel Std Dev(%)	0.0000		FL HIGHWAY PATROL Intoxilyzer - Alconol Analyzer Model 8000 07/15/2025 Software: 8100.27 SN 80-001109 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>14:17</td></tr> <tr><td>Control Test</td><td>0.080</td><td>14:18</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:18</td></tr> <tr><td>Control Test</td><td>0.079</td><td>14:19</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:19</td></tr> <tr><td>Control Test</td><td>0.079</td><td>14:20</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:21</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0793</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7277</td><td></td></tr> </table>	Air Blank	0.000	14:17	Control Test	0.080	14:18	Air Blank	0.000	14:18	Control Test	0.079	14:19	Air Blank	0.000	14:19	Control Test	0.079	14:20	Air Blank	0.000	14:21	Control Test Stats			Average	0.0793		Std Dev	0.0006		Rel Std Dev(%)	0.7277		FL HIGHWAY PATROL Intoxilyzer - Alconol Analyzer Model 8000 07/15/2025 Software: 8100.27 SN 80-001109 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>14:24</td></tr> <tr><td>Control Test</td><td>0.201</td><td>14:25</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:25</td></tr> <tr><td>Control Test</td><td>0.200</td><td>14:26</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:27</td></tr> <tr><td>Control Test</td><td>0.199</td><td>14:27</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:28</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.2000</td><td></td></tr> <tr><td>Std Dev</td><td>0.0010</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.5000</td><td></td></tr> </table>	Air Blank	0.000	14:24	Control Test	0.201	14:25	Air Blank	0.000	14:25	Control Test	0.200	14:26	Air Blank	0.000	14:27	Control Test	0.199	14:27	Air Blank	0.000	14:28	Control Test Stats			Average	0.2000		Std Dev	0.0010		Rel Std Dev(%)	0.5000		FL HIGHWAY PATROL Intoxilyzer - Alconol Analyzer Model 8000 07/15/2025 Software: 8100.27 SN 80-001109 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>14:12</td></tr> <tr><td>Control Test</td><td>0.079</td><td>14:13</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:13</td></tr> <tr><td>Control Test</td><td>0.079</td><td>14:13</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:14</td></tr> <tr><td>Control Test</td><td>0.080</td><td>14:14</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:15</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0793</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7277</td><td></td></tr> </table>	Air Blank	0.000	14:12	Control Test	0.079	14:13	Air Blank	0.000	14:13	Control Test	0.079	14:13	Air Blank	0.000	14:14	Control Test	0.080	14:14	Air Blank	0.000	14:15	Control Test Stats			Average	0.0793		Std Dev	0.0006		Rel Std Dev(%)	0.7277	
Air Blank	0.000	14:04																																																																																																																																					
Control Test	0.048	14:05																																																																																																																																					
Air Blank	0.000	14:06																																																																																																																																					
Control Test	0.048	14:06																																																																																																																																					
Air Blank	0.000	14:07																																																																																																																																					
Control Test	0.048	14:08																																																																																																																																					
Air Blank	0.000	14:08																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0480																																																																																																																																						
Std Dev	0.0000																																																																																																																																						
Rel Std Dev(%)	0.0000																																																																																																																																						
Air Blank	0.000	14:17																																																																																																																																					
Control Test	0.080	14:18																																																																																																																																					
Air Blank	0.000	14:18																																																																																																																																					
Control Test	0.079	14:19																																																																																																																																					
Air Blank	0.000	14:19																																																																																																																																					
Control Test	0.079	14:20																																																																																																																																					
Air Blank	0.000	14:21																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0793																																																																																																																																						
Std Dev	0.0006																																																																																																																																						
Rel Std Dev(%)	0.7277																																																																																																																																						
Air Blank	0.000	14:24																																																																																																																																					
Control Test	0.201	14:25																																																																																																																																					
Air Blank	0.000	14:25																																																																																																																																					
Control Test	0.200	14:26																																																																																																																																					
Air Blank	0.000	14:27																																																																																																																																					
Control Test	0.199	14:27																																																																																																																																					
Air Blank	0.000	14:28																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.2000																																																																																																																																						
Std Dev	0.0010																																																																																																																																						
Rel Std Dev(%)	0.5000																																																																																																																																						
Air Blank	0.000	14:12																																																																																																																																					
Control Test	0.079	14:13																																																																																																																																					
Air Blank	0.000	14:13																																																																																																																																					
Control Test	0.079	14:13																																																																																																																																					
Air Blank	0.000	14:14																																																																																																																																					
Control Test	0.080	14:14																																																																																																																																					
Air Blank	0.000	14:15																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0793																																																																																																																																						
Std Dev	0.0006																																																																																																																																						
Rel Std Dev(%)	0.7277																																																																																																																																						
Operator's Signature <i>ML</i>	Operator's Signature <i>ML</i>	Operator's Signature <i>ML</i>	Operator's Signature <i>ML</i>																																																																																																																																				

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 12:55

Date of Inspection: 07/17/2025

Serial Number: 80-001109
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG429602 Exp: 10/22/2026
0.000	0.048	0.078	0.200	0.079
0.000	0.048	0.078	0.201	0.078
0.000	0.048	0.078	0.201	0.079
0.000	0.048	0.078	0.201	0.079
0.000	0.048	0.079	0.200	0.078
0.000	0.048	0.078	0.201	0.078
0.000	0.048	0.079	0.201	0.079
0.000	0.048	0.078	0.201	0.078
0.000	0.048	0.078	0.201	0.079
0.000	0.048	0.079	0.201	0.078

Standard Deviations	0.0000	0.0004	0.0004	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

07/17/2025
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001109 , manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001109</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>FL HIGHWAY PATROL</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>07/17/2025</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>12:55</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Taylor
Gutschow
Digitally signed by Taylor Gutschow
Date: 2025.07.17 15:20:23 -04'00'

07/17/2025

Date

TAYLOR D GUTSCHOW,
Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality