



INSTRUMENT PROCESSING SHEET

Agency Florida Highway PatrolS/N 80-001115Florida Department of
Law EnforcementDate In 6/10/2025DI Completion Date N/A☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By	PN	Date	Quality Checks	By	PN	Date	Flow Calibration	By	Date																											
<input type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: Instrument dropped off by Agency Inspector. Reported 0.200 g/210L readings were low. PN 6/10/2025			6/10/2025	<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>155</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP 103</u> 32 mm <u>.148</u> (.139 - .169) 36 mm <u>.171</u> (.156 - .190) 53 mm <u>.242</u> (.228 - .278) 103 mm <u>.500</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28662</u> <input checked="" type="checkbox"/> Stability Checks			7/11/2025	Flow Column # _____ <input type="checkbox"/> 5L/min – 17mm <input type="checkbox"/> 15L/min – 53mm <input type="checkbox"/> 30L/min – 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																													
				<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP5088</td><td>202406K 06/19/2026</td></tr><tr><td>0.080</td><td>MP5089</td><td>202406L 06/19/2026</td></tr><tr><td>0.200</td><td>MP5090</td><td>202406N 06/20/2026</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG429602 10/22/2026</td></tr></tbody></table>			Simulator	Serial #	Lot #/Exp	0.050	MP5088	202406K 06/19/2026	0.080	MP5089	202406L 06/19/2026	0.200	MP5090	202406N 06/20/2026	0.080 DGS	N/A	AG429602 10/22/2026				<table border="1"><thead><tr><th colspan="2">Maintenance By _____ Date _____</th></tr><tr><td><input type="checkbox"/> Battery Replacement</td><td></td></tr><tr><td><input type="checkbox"/> Dry Gas Regulator Replacement</td><td></td></tr><tr><td><input type="checkbox"/> Breath Tube Replacement</td><td></td></tr><tr><td><input type="checkbox"/> Other _____</td><td></td></tr></thead></table>			Maintenance By _____ Date _____		<input type="checkbox"/> Battery Replacement		<input type="checkbox"/> Dry Gas Regulator Replacement		<input type="checkbox"/> Breath Tube Replacement		<input type="checkbox"/> Other _____	
Simulator	Serial #	Lot #/Exp																																			
0.050	MP5088	202406K 06/19/2026																																			
0.080	MP5089	202406L 06/19/2026																																			
0.200	MP5090	202406N 06/20/2026																																			
0.080 DGS	N/A	AG429602 10/22/2026																																			
Maintenance By _____ Date _____																																					
<input type="checkbox"/> Battery Replacement																																					
<input type="checkbox"/> Dry Gas Regulator Replacement																																					
<input type="checkbox"/> Breath Tube Replacement																																					
<input type="checkbox"/> Other _____																																					
Calibration Adjustment By _____				Department Inspection By _____																																	
Barometric Pressure Gauge _____ ID # _____				Barometric Pressure ID# _____																																	
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>				Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Gauge _____ Instrument _____					
Simulator	Serial #	Lot #	Expiration																																		
0.000		N/A	N/A																																		
0.040																																					
0.100																																					
0.200																																					
0.300																																					
0.080 DGS	N/A																																				
				Mouth Alcohol Solution Lot # _____																																	
				Acetone Stock Solution Lot # _____																																	
				<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td></td></tr><tr><td>Interferent</td><td></td></tr><tr><td>0.050</td><td></td></tr><tr><td>0.080</td><td></td></tr><tr><td>0.200</td><td></td></tr></tbody></table>						Simulator	Serial Number	0.000		Interferent		0.050		0.080		0.200																	
Simulator	Serial Number																																				
0.000																																					
Interferent																																					
0.050																																					
0.080																																					
0.200																																					
<input type="checkbox"/> Post Calibration Adjustment Stability Checks				Attachments																																	
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>				Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			<table border="1"><tbody><tr><td><input type="checkbox"/> Form 41</td><td><input type="checkbox"/> Post-Stability Checks</td></tr><tr><td><input checked="" type="checkbox"/> Stability Checks</td><td><input type="checkbox"/> Flow Calibration</td></tr><tr><td><input type="checkbox"/> Calibration Certificate</td><td><input type="checkbox"/> Form 40</td></tr><tr><td><input type="checkbox"/> Calibration Adjustment</td><td><input type="checkbox"/> Other _____</td></tr></tbody></table>						<input type="checkbox"/> Form 41	<input type="checkbox"/> Post-Stability Checks	<input checked="" type="checkbox"/> Stability Checks	<input type="checkbox"/> Flow Calibration	<input type="checkbox"/> Calibration Certificate	<input type="checkbox"/> Form 40	<input type="checkbox"/> Calibration Adjustment	<input type="checkbox"/> Other _____
Simulator	Serial #	Lot #	Expiration																																		
0.050																																					
0.080																																					
0.200																																					
0.080 DGS	N/A																																				
<input type="checkbox"/> Form 41	<input type="checkbox"/> Post-Stability Checks																																				
<input checked="" type="checkbox"/> Stability Checks	<input type="checkbox"/> Flow Calibration																																				
<input type="checkbox"/> Calibration Certificate	<input type="checkbox"/> Form 40																																				
<input type="checkbox"/> Calibration Adjustment	<input type="checkbox"/> Other _____																																				
Notes/Suggested Service: _____ Instrument previously had 2025 inspection. Evaluated instrument by performing quality checks. Unable to detect items that would require additional service or repair. Returning to agency without performing additional calibration. Agency Inspector was informed to monitor instrument. PN 7/11/2025				<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use																																	
				Digitally signed by Shayla Platt Date: 2025.07.14 08:52:15 -04'00'																																	
				Digitally signed by Taylor Gutschow Date: 2025.07.14 12:12:16 -04'00'																																	
				Tech Review / Date _____ Admin Review / Date _____																																	

0.05 g/210L stability check measurements have been reviewed on hard copy printout and are all 0.049.
PN 7/11/2025

Test	Grade	Time
Pre-Work	0.000	10:55
Control Test	0.140	10:57
Pre-Work	0.000	10:57
Control Test	0.040	10:59
Pre-Work	0.000	10:59
Control Test	0.040	11:00
Pre-Work	0.000	11:00
Control Test	0.040	11:01
Pre-Work	0.000	11:01
Control Test	0.040	11:02
Pre-Work	0.000	11:02
Control Test	0.040	11:03
Pre-Work	0.000	11:03
Control Test	0.040	11:04
Pre-Work	0.000	11:04
Control Test	0.040	11:05
Pre-Work	0.000	11:05
Control Test	0.040	11:06
Pre-Work	0.000	11:06
Control Test	0.040	11:07
Pre-Work	0.000	11:07
Control Test	0.040	11:08
Pre-Work	0.000	11:08
Control Test	0.040	11:09
Pre-Work	0.000	11:09
Control Test	0.040	11:10
Pre-Work	0.000	11:10
Control Test	0.040	11:11
Pre-Work	0.000	11:11
Control Test	0.040	11:12
Pre-Work	0.000	11:12
Control Test	0.040	11:13
Pre-Work	0.000	11:13
Control Test	0.040	11:14
Pre-Work	0.000	11:14
Control Test	0.040	11:15
Pre-Work	0.000	11:15
Control Test	0.040	11:16
Pre-Work	0.000	11:16
Control Test	0.040	11:17
Pre-Work	0.000	11:17
Control Test	0.040	11:18
Pre-Work	0.000	11:18
Control Test	0.040	11:19
Pre-Work	0.000	11:19
Control Test	0.040	11:20
Pre-Work	0.000	11:20
Control Test	0.040	11:21
Pre-Work	0.000	11:21
Control Test	0.040	11:22
Pre-Work	0.000	11:22
Control Test	0.040	11:23
Pre-Work	0.000	11:23
Control Test	0.040	11:24
Pre-Work	0.000	11:24
Control Test	0.040	11:25
Pre-Work	0.000	11:25
Control Test	0.040	11:26
Pre-Work	0.000	11:26
Control Test	0.040	11:27
Pre-Work	0.000	11:27
Control Test	0.040	11:28
Pre-Work	0.000	11:28
Control Test	0.040	11:29
Pre-Work	0.000	11:29
Control Test	0.040	11:30
Pre-Work	0.000	11:30
Control Test	0.040	11:31
Pre-Work	0.000	11:31
Control Test	0.040	11:32
Pre-Work	0.000	11:32
Control Test	0.040	11:33
Pre-Work	0.000	11:33
Control Test	0.040	11:34
Pre-Work	0.000	11:34
Control Test	0.040	11:35
Pre-Work	0.000	11:35
Control Test	0.040	11:36
Pre-Work	0.000	11:36
Control Test	0.040	11:37
Pre-Work	0.000	11:37
Control Test	0.040	11:38
Pre-Work	0.000	11:38
Control Test	0.040	11:39
Pre-Work	0.000	11:39
Control Test	0.040	11:40
Pre-Work	0.000	11:40
Control Test	0.040	11:41
Pre-Work	0.000	11:41
Control Test	0.040	11:42
Pre-Work	0.000	11:42
Control Test	0.040	11:43
Pre-Work	0.000	11:43
Control Test	0.040	11:44
Pre-Work	0.000	11:44
Control Test	0.040	11:45
Pre-Work	0.000	11:45
Control Test	0.040	11:46
Pre-Work	0.000	11:46
Control Test	0.040	11:47
Pre-Work	0.000	11:47
Control Test	0.040	11:48
Pre-Work	0.000	11:48
Control Test	0.040	11:49
Pre-Work	0.000	11:49
Control Test	0.040	11:50
Pre-Work	0.000	11:50
Control Test	0.040	11:51

Test	Std Dev	Time
Air Blank	0.074	10:12
Control Test	0.074	10:13
Air Blank	0.074	10:13
Control Test	0.074	10:14
Air Blank	0.074	10:14
Control Test	0.080	10:15
Air Blank	0.080	10:16
Control Test		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

Test	g/210L	Time
Air Blank	0.000	10:19
Control Test	0.199	10:20
Air Blank	0.000	10:20
Control Test	0.199	10:21
Air Blank	0.000	10:22
Control Test	0.199	10:22
Air Blank	0.000	10:23
Control Test Stats		
Average	0.1990	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Test	g/210L	Time
Air Blank	0.000	10:24
Control Test	0.079	10:25
Air Blank	0.000	10:26
Control Test	0.079	10:26
Air Blank	0.000	10:26
Control Test	0.079	10:26
Air Blank	0.000	10:27
Control Test	0.079	10:27
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

DGS