



Agency Polk County Sheriff's Office

S/N 80-001127

Date In 04/01/2025 DI Completion Date 05/01/2025

☐Ship ☐P/U ☐H/D ☒CMI ☐EE

Calibration Adjustment

By_

ID

☐ Post Calibration Adjustment Stability Checks

Notes/Suggested Service: Green Start Test button must be pushed several times before it functions. Observed this while powering up and exiting Standby Mode. Also has quiet tones. (TDG 5/1/25)

Sending to CMI to evaluate button/tones. (TDG 5/5/25)

Department Inspection

By TDG

Acetone Stock Solution Lot # 2023-B

Attachments

- | | |
|---|---|
| ☒ Form 41 | ☐ Post-Stability Checks |
| ☒ Stability Checks | ☐ Flow Calibration |
| ☒ Calibration Certificate | ☒ Form 40 |
| ☐ Calibration Adjustment | ☒ Other Form 51 |

- ☒ Instrument Complies with Chapter 11D-8, FAC
- ☐ Instrument Does Not Comply with Chapter 11D-8, FAC
- ☐ Return to/Place into Evidentiary Use
- ☒ Remain Out of Evidentiary Use
- ☐ Conduct an Agency Inspection Before Evidentiary Use

Phil Nicodemo
Digitally signed by Phil Nicodemo
Date: 2025.05.15 15:03:29
-04'00'

Shayla Platt
Date: 2025.05.15 20:38:28
-04'00'

Tech Review / Date

Admin Review / Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT – INTOXILYZER 8000

Agency: POLK COUNTY SO

Time of Inspection: 10:04

Date of Inspection: 05/01/2025

Serial Number: 80-001127

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08g/210L Test (g/210L) Lot#:_____ Exp:_____	0.20g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:_____ Exp:_____

Number of Simulators Used: _____

Remarks:

AI NOT CONDUCTED. BYPASSED TO BRING OUT OF DISABLED MODE.

Not determined ML 5/1/25

The above instrument complies (☒) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Taylor D Gutschow

TAYLOR D GUTSCHOW

Signature and Printed Name

05/01/2025

Date

Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
✓	✓	✓	✓
POLK COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 05/01/2025 Software: 8100.27 SN 80-001127 Test g/210L Time Air Blank 0.000 10:56 Control Test 0.050 10:57 Air Blank 0.000 10:58 Control Test 0.049 10:58 Air Blank 0.000 10:59 Control Test 0.049 11:00 Air Blank 0.000 11:00 Control Test Stats Average 0.0493 Std Dev 0.0006 Rel Std Dev(%) 1.1703	POLK COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 05/01/2025 Software: 8100.27 SN 80-001127 Test g/210L Time Air Blank 0.000 11:11 Control Test 0.081 11:12 Air Blank 0.000 11:12 Control Test 0.080 11:13 Air Blank 0.000 11:13 Control Test 0.080 11:14 Air Blank 0.000 11:15 Control Test Stats Average 0.0803 Std Dev 0.0006 Rel Std Dev(%) 0.7187	POLK COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 05/01/2025 Software: 8100.27 SN 80-001127 Test g/210L Time Air Blank 0.000 11:17 Control Test 0.202 11:18 Air Blank 0.000 11:19 Control Test 0.200 11:19 Air Blank 0.000 11:20 Control Test 0.200 11:21 Air Blank 0.000 11:21 Control Test Stats Average 0.2007 Std Dev 0.0012 Rel Std Dev(%) 0.5754	POLK COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 05/01/2025 Software: 8100.27 SN 80-001127 Test g/210L Time Air Blank 0.000 11:27 Control Test 0.078 11:27 Air Blank 0.000 11:27 Control Test 0.079 11:28 Air Blank 0.000 11:28 Control Test 0.080 11:29 Air Blank 0.000 11:29 Control Test Stats Average 0.0790 Std Dev 0.0010 Rel Std Dev(%) 1.2658
Operator's Signature 	Operator's Signature 	Operator's Signature 	Operator's Signature 

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: POLK COUNTY SO
Time of Inspection: 14:40

Date of Inspection: 05/01/2025

Serial Number: 80-001127
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG429602 Exp: 10/22/2026
0.000	0.049	0.080	0.200	0.078
0.000	0.050	0.081	0.200	0.079
0.000	0.049	0.080	0.200	0.079
0.000	0.049	0.080	0.200	0.078
0.000	0.050	0.081	0.201	0.079
0.000	0.049	0.081	0.200	0.078
0.000	0.049	0.080	0.200	0.079
0.000	0.050	0.081	0.201	0.079
0.000	0.049	0.080	0.200	0.079
0.000	0.050	0.081	0.200	0.079

Standard Deviations	0.0005	0.0005	0.0004	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

05/01/2025
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001127 , manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001127</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>POLK COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>05/01/2025</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>14:40</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Taylor
Gutschow
Digitally signed by Taylor
Gutschow
Date: 2025.05.01 15:35:04
+0400

05/01/2025

Date

TAYLOR D GUTSCHOW,
Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Return Material Authorization

Ship to: ☒ CMI, Inc.

☐ Enforcement Electronics

Shipment to repair facility authorized by: Kevin Cole on 05/02/2025

Items Returned: Instrument ☒ Supplies ☐ Other ☐ Describe: _____

Instrument Model: Intoxilyzer 8000 Serial Number: 80-001127

Bill To Address:

Polk County Sheriff's Office

Attn: Kevin Cole

Ship to Address:

Florida Department of Law Enforcement

Fort Myers Regional Operations Center

Attn: Taylor Gutschow

4700 Terminal Drive, Suite 1

Fort Myers, FL 33907

Reason for Return:

Green Start Test button must be pressed multiple times to function. Instrument signal tones are barely audible.

Please choose one of the following options:

- ☐ 1. I _____, authorize all repairs.
- ☐ 2. I _____, authorize repairs up to \$_____.
- ☒ 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Kevin Cole

Phone #: (863) 247-1891

Email: KCole@polksheriff.org

ATP Contact Name: Taylor Gutschow ATP Email: TaylorGutschow@fdle.state.fl.us