



INSTRUMENT PROCESSING SHEET

Agency Leon County SOS/N 80-001153Florida Department of
Law EnforcementDate In 2/11/25DI Completion Date 3/4/2025☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake By <u>ALL</u> Date <u>2/11/25</u>		Quality Checks By <u>PN</u> Date <u>3/4/2025</u>		Flow Calibration By <u>PN</u> Date <u>3/4/2025</u>																													
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: <u>dropoff, no box, uploaded upon receipt by request of agency</u>		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>219</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP 102</u> 32 mm <u>.144</u> (.139 - .169) 36 mm <u>.160</u> (.156 - .190) 53 mm <u>.218*</u> (.228 - .278) 103 mm <u>.433</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>30793</u> ☒ Stability Checks		☒ Flow Column # <u>ATP 103</u> ☒ 5L/min – 17mm ☒ 15L/min – 53mm ☒ 30L/min – 103mm ☒ R-Value <u>220</u> ☒ Post Calibration Verification (L/s) Flow Column # <u>ATP 102</u> 32 mm <u>.140</u> (.139 - .169) 36 mm <u>.160</u> (.156 - .190) 53 mm <u>.230</u> (.228 - .278) 103 mm <u>.480</u> (.447 - .547)																													
		<table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6291</td><td>202406K 06/19/2026</td></tr><tr><td>0.080</td><td>MP6292</td><td>202406L 06/19/2026</td></tr><tr><td>0.200</td><td>MP6293</td><td>202406N 06/20/2026</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG429602 10/22/2026</td></tr></tbody></table>		Simulator	Serial #	Lot #/Exp	0.050	MP6291	202406K 06/19/2026	0.080	MP6292	202406L 06/19/2026	0.200	MP6293	202406N 06/20/2026	0.080 DGS	N/A	AG429602 10/22/2026	Maintenance By _____ Date _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____														
Simulator	Serial #	Lot #/Exp																															
0.050	MP6291	202406K 06/19/2026																															
0.080	MP6292	202406L 06/19/2026																															
0.200	MP6293	202406N 06/20/2026																															
0.080 DGS	N/A	AG429602 10/22/2026																															
Calibration Adjustment By _____		Department Inspection By <u>PN</u>																															
Barometric Pressure Gauge _____ ID # _____		Barometric Pressure ID# <u>30793</u>																															
<table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Gauge <u>1013</u> Instrument <u>1008</u> Mouth Alcohol Solution Lot # <u>2024-A</u> Acetone Stock Solution Lot # <u>2024-B</u>			
Simulator	Serial #	Lot #	Expiration																														
0.000		N/A	N/A																														
0.040																																	
0.100																																	
0.200																																	
0.300																																	
0.080 DGS	N/A																																
☐ Post Calibration Adjustment Stability Checks		<table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6294</td></tr><tr><td>Interferent</td><td>MP6290</td></tr><tr><td>0.050</td><td>MP6291</td></tr><tr><td>0.080</td><td>MP6292</td></tr><tr><td>0.200</td><td>MP6293</td></tr></tbody></table>				Simulator	Serial Number	0.000	MP6294	Interferent	MP6290	0.050	MP6291	0.080	MP6292	0.200	MP6293																
Simulator	Serial Number																																
0.000	MP6294																																
Interferent	MP6290																																
0.050	MP6291																																
0.080	MP6292																																
0.200	MP6293																																
<table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Attachments ☒ Form 41 ☐ Post-Stability Checks ☒ Stability Checks ☒ Flow Calibration ☒ Calibration Certificate ☐ Form 40 ☐ Calibration Adjustment ☐ Other _____											
Simulator	Serial #	Lot #	Expiration																														
0.050																																	
0.080																																	
0.200																																	
0.080 DGS	N/A																																
Notes/Suggested Service: _____ <u>Instrument produced flow measurements outside of acceptable range. Performed root cause, and conducted flow calibration adjustment. PN 3/4/2025</u>		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																															
		Digitally signed by _____ Destinee Armstrong Digitally signed by Destinee Armstrong Date: 2025.03.05 14:33:32 -05'00' Shayla Platt Shayla Platt Tech Review / Date _____ Admin Review / Date <u>2025.03.06 11:00:15 -05'00'</u>																															

FLOW CALIBRATION
ADJUSTMENT
PN

LEON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001153
03/04/2025
Software: 8100.27

Flow Rate Calibration*****

1: Rate (Liters/min) = 5

SQRT(Diff)) = 6.855

2: Rate (Liters/min) = 15

SQRT(Diff)) = 11.180

3: Rate (Liters/min) = 30

SQRT(Diff)) = 19.285

Dependent Data Scale Factor = 100000 L/min

Independent Data Scale Factor = 256

Rounded Slope = 778

Rounded Intercept = -809778

Correlation = 0.99827

STABILITY CHECK

LEON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001153
03/04/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:46
Control Test	0.049	10:46
Air Blank	0.000	10:47
Control Test	0.049	10:47
Air Blank	0.000	10:48
Control Test	0.049	10:49
Air Blank	0.000	10:49
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

LEON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001153
03/04/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:50
Control Test	0.079	10:51
Air Blank	0.000	10:52
Control Test	0.079	10:52
Air Blank	0.000	10:53
Control Test	0.079	10:54
Air Blank	0.000	10:54
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

LEON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001153
03/04/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:55
Control Test	0.200	10:56
Air Blank	0.000	10:56
Control Test	0.201	10:57
Air Blank	0.000	10:58
Control Test	0.201	10:58
Air Blank	0.000	10:59
Control Test Stats		
Average	0.2007	
Std Dev	0.0006	
Rel Std Dev(%)	0.2877	

Operator's Signature

LEON COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001153
03/04/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:01
Control Test	0.081	11:02
Air Blank	0.000	11:02
Control Test	0.081	11:02
Air Blank	0.000	11:03
Control Test	0.081	11:03
Air Blank	0.000	11:04
Control Test Stats		
Average	0.0810	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

D65

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: LEON COUNTY SO

Time of Inspection: 13:53

Date of Inspection: 03/04/2025

Serial Number: 80-001153

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG429602 Exp: 10/22/2026
0.000	0.049	0.079	0.201	0.081
0.000	0.049	0.079	0.201	0.080
0.000	0.049	0.079	0.201	0.081
0.000	0.049	0.080	0.200	0.080
0.000	0.049	0.079	0.200	0.080
0.000	0.049	0.079	0.201	0.080
0.000	0.049	0.079	0.201	0.080
0.000	0.050	0.079	0.201	0.080
0.000	0.049	0.079	0.201	0.080
0.000	0.049	0.079	0.201	0.080

Standard Deviations	0.0003	0.0003	0.0004	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.


I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



PHIL NICODEMUS

Signature and Printed Name

03/04/2025
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001153, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001153</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>LEON COUNTY SO</u>	0.050 g/210 L	0.004
Calibration Date:	<u>03/04/2025</u>	0.080 g/210 L	0.004
Calibration Time:	<u>13:53</u>	0.200 g/210 L	0.007
		0.080 g/210 L Dry Gas Control	0.005

All results are reported in g/210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

03/04/2025

Date

PHIL NICODEMO,

Department Inspector

FDLE/ATP Form 69 October 2024

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality