

Flow Calibration Adjustment(s)

Root Cause Analysis Performed Prior? ☒

Performed by TDG

Several lines on the printout
did not print legibly. Flow
cal adjust was passing.

TDG
3/28/25

MARTIN COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001189
03/27/2025
Software: 8100.27

Flow Rate Calibration*****

1: Rate (Liters/min) = 5

SQRT(Diff) = 6.082

2: Rate (Liters/min) = 15

SQRT(Diff) = 11.355

3: Rate (Liters/min) = 30

SQRT(Diff) = 20.000

4: Rate (Liters/min) = 45

SQRT(Diff) = 22.500

Rounded Slope = 651

Rounded Intercept = -466615

Correlation = 0.99869

Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
MARTIN COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001189 03/27/2025 Software: 8100.27 Test g/210L Time Air Blank 0.000 14:58 Control Test 0.050 14:58 Air Blank 0.000 14:59 Control Test 0.050 15:00 Air Blank 0.000 15:00 Control Test 0.049 15:01 Air Blank 0.000 15:02 Control Test Stats Average 0.0497 Std Dev 0.0006 Rel Std Dev(%) 1.1625 Operator's Signature 	MARTIN COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001189 03/27/2025 Software: 8100.27 Test g/210L Time Air Blank 0.000 15:04 Control Test 0.080 15:05 Air Blank 0.000 15:05 Control Test 0.080 15:06 Air Blank 0.000 15:06 Control Test 0.079 15:07 Air Blank 0.000 15:08 Control Test Stats Average 0.0797 Std Dev 0.0006 Rel Std Dev(%) 0.7247 Operator's Signature 	MARTIN COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001189 03/27/2025 Software: 8100.27 Test g/210L Time Air Blank 0.000 15:13 Control Test 0.197 15:14 Air Blank 0.000 15:14 Control Test 0.197 15:15 Air Blank 0.000 15:15 Control Test 0.196 15:16 Air Blank 0.000 15:17 Control Test Stats Average 0.1967 Std Dev 0.0006 Rel Std Dev(%) 0.2936 Operator's Signature 	MARTIN COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001189 03/27/2025 Software: 8100.27 Test g/210L Time Air Blank 0.000 14:52 Control Test 0.080 14:52 Air Blank 0.000 14:52 Control Test 0.080 14:53 Air Blank 0.000 14:53 Control Test 0.080 14:54 Air Blank 0.000 14:54 Control Test Stats Average 0.0800 Std Dev 0.0000 Rel Std Dev(%) 0.0000 Operator's Signature 

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MARTIN COUNTY SO
Time of Inspection: 12:32

Date of Inspection: 03/28/2025

Serial Number: 80-001189
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG429602 Exp: 10/22/2026
0.000	0.049	0.079	0.198	0.079
0.000	0.050	0.080	0.198	0.079
0.000	0.050	0.079	0.198	0.080
0.000	0.049	0.080	0.198	0.079
0.000	0.050	0.080	0.199	0.079
0.000	0.050	0.080	0.198	0.080
0.000	0.050	0.080	0.199	0.079
0.000	0.050	0.079	0.198	0.080
0.000	0.050	0.079	0.199	0.080
0.000	0.050	0.080	0.199	0.079


Standard Deviations	0.0004	0.0005	0.0005	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.


I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

03/28/2025
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001189, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001189</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>MARTIN COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>03/28/2025</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>12:32</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Date

03/28/2025

Taylor D Gutschow

TAYLOR D GUTSCHOW,

Department Inspector

Service • Integrity • Respect • Quality