



INSTRUMENT PROCESSING SHEET

Agency Panama City Beach PDS/N 80-001192Florida Department of
Law EnforcementDate In 01/13/2025 DI Completion Date 03/31/2025☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake By <u>ALL</u> Date <u>01/13/2025</u>		Quality Checks By <u>DA</u> Date <u>02/25/2025</u>		Flow Calibration By <u>DA</u> Date <u>02/25/2025</u>																																																																			
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>191</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP102</u> 32 mm <u>0.144</u> (.139 - .169) 36 mm <u>0.160</u> (.156 - .190) 53 mm <u>0.226</u> (.228 - .278) 103 mm <u>0.488</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>30793</u> ☒ Stability Checks <table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP5088</td><td>202406K 06/19/2026</td></tr><tr><td>0.080</td><td>MP5090</td><td>202406L 06/19/2026</td></tr><tr><td>0.200</td><td>MP6289</td><td>202406N 06/20/2026</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG429602 10/22/2026</td></tr></tbody></table>		Simulator	Serial #	Lot #/Exp	0.050	MP5088	202406K 06/19/2026	0.080	MP5090	202406L 06/19/2026	0.200	MP6289	202406N 06/20/2026	0.080 DGS	N/A	AG429602 10/22/2026	Flow Column # <u>ATP105</u> ☒ 5L/min – 17mm ☒ 15L/min – 53mm ☒ 30L/min – 103mm ☒ R-Value <u>195</u> ☒ Post Calibration Verification (L/s) Flow Column # <u>ATP102</u> 32 mm <u>0.144</u> (.139 - .169) 36 mm <u>0.160</u> (.156 - .190) 53 mm <u>0.230</u> (.228 - .278) 103 mm <u>0.484</u> (.447 - .547) Maintenance By _____ Date _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ _____ _____ _____ _____ _____																																																				
Simulator	Serial #	Lot #/Exp																																																																					
0.050	MP5088	202406K 06/19/2026																																																																					
0.080	MP5090	202406L 06/19/2026																																																																					
0.200	MP6289	202406N 06/20/2026																																																																					
0.080 DGS	N/A	AG429602 10/22/2026																																																																					
Calibration Adjustment By <u>DA</u>		Department Inspection By <u>DA</u>																																																																					
Barometric Pressure Gauge <u>1021</u> ID # <u>28427</u> <table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td>MP6294</td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td>MP6295</td><td>24330</td><td>09/10/2026</td></tr><tr><td>0.100</td><td>MP6296</td><td>24390</td><td>10/29/2026</td></tr><tr><td>0.200</td><td>MP6297</td><td>24080</td><td>02/13/2026</td></tr><tr><td>0.300</td><td>MP6298</td><td>24430</td><td>12/10/2026</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>06723080A5</td><td>04/05/2025</td></tr></tbody></table> ☒ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td>MP5088</td><td>202406K</td><td>06/19/2026</td></tr><tr><td>0.080</td><td>MP5089</td><td>202406L</td><td>06/19/2026</td></tr><tr><td>0.200</td><td>MP5090</td><td>202406N</td><td>06/20/2026</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG429602</td><td>10/22/2026</td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000	MP6294	N/A	N/A	0.040	MP6295	24330	09/10/2026	0.100	MP6296	24390	10/29/2026	0.200	MP6297	24080	02/13/2026	0.300	MP6298	24430	12/10/2026	0.080 DGS	N/A	06723080A5	04/05/2025	Simulator	Serial #	Lot #	Expiration	0.050	MP5088	202406K	06/19/2026	0.080	MP5089	202406L	06/19/2026	0.200	MP5090	202406N	06/20/2026	0.080 DGS	N/A	AG429602	10/22/2026	Barometric Pressure ID# <u>30793</u> Gauge <u>1010</u> Instrument <u>1008</u> Mouth Alcohol Solution Lot # <u>2024-A</u> Acetone Stock Solution Lot # <u>2024-B</u> <table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP5086</td></tr><tr><td>Interferent</td><td>MP5087</td></tr><tr><td>0.050</td><td>MP5088</td></tr><tr><td>0.080</td><td>MP5089</td></tr><tr><td>0.200</td><td>MP5090</td></tr></tbody></table> Attachments <table border="1" style="width:100%"><tbody><tr><td>☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Calibration Adjustment</td><td>☒ Post-Stability Checks ☒ Flow Calibration ☒ Form 40 ☐ Other _____</td></tr></tbody></table> ☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use <table border="1" style="width:100%"><tbody><tr><td>Taylor Gutschow Tech Review / Date</td><td>Digitally signed by Taylor Gutschow Date: 2025.04.02 14:40:54 -04'00'</td><td>Shayla Platt Admin Review / Date</td><td>Digitally signed by Shayla Platt Date: 2025.04.03 09:12:20 -04'00'</td></tr></tbody></table>				Simulator	Serial Number	0.000	MP5086	Interferent	MP5087	0.050	MP5088	0.080	MP5089	0.200	MP5090	☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Calibration Adjustment	☒ Post-Stability Checks ☒ Flow Calibration ☒ Form 40 ☐ Other _____	Taylor Gutschow Tech Review / Date	Digitally signed by Taylor Gutschow Date: 2025.04.02 14:40:54 -04'00'	Shayla Platt Admin Review / Date	Digitally signed by Shayla Platt Date: 2025.04.03 09:12:20 -04'00'
Simulator	Serial #	Lot #	Expiration																																																																				
0.000	MP6294	N/A	N/A																																																																				
0.040	MP6295	24330	09/10/2026																																																																				
0.100	MP6296	24390	10/29/2026																																																																				
0.200	MP6297	24080	02/13/2026																																																																				
0.300	MP6298	24430	12/10/2026																																																																				
0.080 DGS	N/A	06723080A5	04/05/2025																																																																				
Simulator	Serial #	Lot #	Expiration																																																																				
0.050	MP5088	202406K	06/19/2026																																																																				
0.080	MP5089	202406L	06/19/2026																																																																				
0.200	MP5090	202406N	06/20/2026																																																																				
0.080 DGS	N/A	AG429602	10/22/2026																																																																				
Simulator	Serial Number																																																																						
0.000	MP5086																																																																						
Interferent	MP5087																																																																						
0.050	MP5088																																																																						
0.080	MP5089																																																																						
0.200	MP5090																																																																						
☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Calibration Adjustment	☒ Post-Stability Checks ☒ Flow Calibration ☒ Form 40 ☐ Other _____																																																																						
Taylor Gutschow Tech Review / Date	Digitally signed by Taylor Gutschow Date: 2025.04.02 14:40:54 -04'00'	Shayla Platt Admin Review / Date	Digitally signed by Shayla Platt Date: 2025.04.03 09:12:20 -04'00'																																																																				
Notes/Suggested Service: <u>I performed root-cause analysis for out-of-range flow verification and stability checks on 02/25/2025. No user/equipment error was identified.</u> <u>DA 03/27/2025</u> _____ _____ _____ _____ _____																																																																							

Stability Checks

80-001192

2/25/25

DA

PANAMA CITY BEACH PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001192
02/25/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:35
Control Test	0.049	12:36
Air Blank	0.000	12:36
Control Test	0.049	12:37
Air Blank	0.000	12:37
Control Test	0.049	12:38
Air Blank	0.000	12:39
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	



Operator's Signature

PANAMA CITY BEACH PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001192
02/25/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:41
Control Test	0.079	12:41
Air Blank	0.000	12:42
Control Test	0.079	12:43
Air Blank	0.000	12:43
Control Test	0.079	12:44
Air Blank	0.000	12:45
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Wet

DA

Operator's Signature

PANAMA CITY BEACH PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001192
02/25/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:48
Control Test	0.200	12:49
Air Blank	0.000	12:49
Control Test	0.199	12:50
Air Blank	0.000	12:51
Control Test	0.199	12:51
Air Blank	0.000	12:52
Control Test Stats		
Average	0.1993	
Std Dev	0.0006	
Rel Std Dev(%)	0.2896	



Operator's Signature

PANAMA CITY BEACH PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001192
02/25/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:54
Control Test	0.076	12:54
Air Blank	0.000	12:55
Control Test	0.076	12:55
Air Blank	0.000	12:56
Control Test	0.076	12:56
Air Blank	0.000	12:56
Control Test Stats		
Average	0.0760	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS



Operator's Signature

Flow Rate Calibration Adjustment DA 2/25/25

PANAMA CITY BEACH PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001192
02/25/2025
Software: 8100.27

Flow Rate Calibration*****

1: Rate (Liters/min) = 5

SQRT(Diff)) = 7.070

2: Rate (Liters/min) = 15

SQRT(Diff)) = 11.957

3: Rate (Liters/min) = 30

SQRT(Diff)) = 21.398

Dependent Data Scale Factor = 100000 L/min

Independent Data Scale Factor = 256

Rounded Slope = 673

Rounded Intercept = -656111

Correlation = 0.99779

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PANAMA CITY BEACH PD

Serial Number: 80-001192

Time of Inspection: 09:50

Date of Inspection: 03/27/2025

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot #: Exp:

Number of Simulators Used: _____

Remarks:

BYPASS AI TO OPERATE INSTRUMENT COMPLIANCE NOT DETERMINED

N/A DA 3/27/25

The above instrument complies (☒) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

LA

DESTINEE N ARMSTRONG

Signature and Printed Name

03/27/2025

Date

Optical Bench Calibration Adjustment DA 3/27/25

80-001192 Panama City Beach PD

PANAMA CITY BEACH PD
 Instrument = Alcohol Analyzer
 Model 8000 SN 80-001192
 03/27/2025 19:53:28

Auto Calibration
 Max Power Res Value = 37
 Auto Range Res Value = 21
 Sol Value = 0.000 g/210L ***
 Fit value = 0.0000 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum lo = 12553, Sum lo = 13353

***** CHANNEL 1 *****

Sample	% Abs	(% Abs Ref)
Sample #1 = 0.1440	(-0.0150)	
Sample #2 = 0.1430	(0.0100)	
Sample #3 = 0.1420	(0.0190)	
Sample #4 = 0.1630	(0.0120)	
Avg % Abs = 0.1493	(0.0137)	
STD DEV = 0.0118	(0.0047)	
REL STD DEV = 7.933	(34.579)	

***** CHANNEL 2 *****

Sample	% Abs	(% Abs Ref)
Sample #1 = 0.1670	(-0.0100)	
Sample #2 = 0.1610	(0.0000)	
Sample #3 = 0.1740	(-0.0050)	
Sample #4 = 0.1810	(-0.0070)	
Avg % Abs = 0.1720	(-0.0040)	
STD DEV = 0.0101	(0.0036)	
REL STD DEV = 5.901	(90.139)	

Sol Value = 0.040 g/210L ***
 Fit value = 0.1905 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum lo = 12548, Sum lo = 13353

***** CHANNEL 1 *****

Sample	% Abs	(% Abs Ref)
Sample #1 = 0.8510	(-0.0290)	
Sample #2 = 0.8370	(-0.0090)	
Sample #3 = 0.8420	(-0.0120)	
Sample #4 = 0.8450	(-0.0010)	
Avg % Abs = 0.8413	(-0.0073)	
STD DEV = 0.0040	(0.0057)	
REL STD DEV = 0.480	(77.540)	

***** CHANNEL 2 *****

Sample	% Abs	(% Abs Ref)
Sample #1 = 1.6620	(-0.0050)	
Sample #2 = 1.6550	(0.0000)	
Sample #3 = 1.6470	(0.0000)	
Sample #4 = 1.6460	(0.0050)	
Avg % Abs = 1.6493	(0.0030)	
STD DEV = 0.0049	(0.0052)	
REL STD DEV = 0.299	(173.205)	

Sol Value = 0.100 g/210L ***
 Fit value = 0.4762 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum lo = 12548, Sum lo = 13354

***** CHANNEL 1 *****

Sample	% Abs	(% Abs Ref)
Sample #1 = 1.8770	(-0.0050)	
Sample #2 = 1.8640	(0.0000)	
Sample #3 = 1.8610	(0.0030)	
Sample #4 = 1.8550	(0.0140)	
Avg % Abs = 1.8600	(0.0057)	
STD DEV = 0.0046	(0.0074)	
REL STD DEV = 0.246	(130.078)	

***** CHANNEL 2 *****

Sample	% Abs	(% Abs Ref)
Sample #1 = 3.8040	(0.0040)	
Sample #2 = 3.7870	(0.0320)	
Sample #3 = 3.7560	(0.0480)	
Sample #4 = 3.7670	(0.0370)	
Avg % Abs = 3.7700	(0.0390)	
STD DEV = 0.0157	(0.0082)	
REL STD DEV = 0.417	(20.988)	

Sol Value = 0.200 g/210L ***
 Fit value = 0.9524 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum lo = 12561, Sum lo = 13355

***** CHANNEL 1 *****

Sample	% Abs	(% Abs Ref)
Sample #1 = 3.5770	(-0.0190)	
Sample #2 = 3.5550	(0.0110)	
Sample #3 = 3.5270	(0.0310)	
Sample #4 = 3.5360	(0.0300)	
Avg % Abs = 3.5393	(0.0240)	
STD DEV = 0.0143	(0.0113)	
REL STD DEV = 0.404	(46.956)	

***** CHANNEL 2 *****

Sample	% Abs	(% Abs Ref)
Sample #1 = 7.2130	(-0.0130)	
Sample #2 = 7.1620	(0.0350)	
Sample #3 = 7.1520	(0.0530)	
Sample #4 = 7.1610	(0.0540)	
Avg % Abs = 7.1583	(0.0473)	
STD DEV = 0.0055	(0.0107)	
REL STD DEV = 0.077	(22.590)	

Sol Value = 0.300 g/210L ***
 Fit value = 1.4286 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum lo = 12563, Sum lo = 13356

***** CHANNEL 1 *****

Sample	% Abs	(% Abs Ref)
Sample #1 = 5.2070	(-0.0180)	
Sample #2 = 5.1690	(0.0330)	
Sample #3 = 5.1480	(0.0490)	
Sample #4 = 5.1540	(0.0430)	
Avg % Abs = 5.1570	(0.0417)	
STD DEV = 0.0108	(0.0081)	
REL STD DEV = 0.210	(19.399)	

***** CHANNEL 2 *****

Sample	% Abs	(% Abs Ref)
Sample #1 = 10.4460	(-0.0070)	
Sample #2 = 10.3580	(0.0860)	
Sample #3 = 10.3580	(0.0960)	
Sample #4 = 10.3450	(0.0990)	
Avg % Abs = 10.3537	(0.0937)	
STD DEV = 0.0075	(0.0068)	
REL STD DEV = 0.072	(7.267)	

***** AUTO CAL DATA *****
 ***** CHANNEL 1 *****
 Sol Val = 0.0000 mg/l or 0.000 g/210L
 % Abs = 0.149
 Std Dev = 0.01 Rel Std Dev = 7.93
 Sol Val = 0.1905 mg/l or 0.040 g/210L
 % Abs = 0.841
 Std Dev = 0.00 Rel Std Dev = 0.48
 Sol Val = 0.4762 mg/l or 0.100 g/210L
 % Abs = 1.860
 Std Dev = 0.00 Rel Std Dev = 0.25
 Sol Val = 0.9524 mg/l or 0.200 g/210L
 % Abs = 3.539
 Std Dev = 0.01 Rel Std Dev = 0.40
 Sol Val = 1.4286 mg/l or 0.300 g/210L
 % Abs = 5.157
 Std Dev = 0.01 Rel Std Dev = 0.21
 Zero Order Coef = -403.39
 First Order Coef = 2724.83
 Second Order Coef = 23.76
 Standard Deviation = 10.937756

***** CHANNEL 2 *****
 Sol Val = 0.0000 mg/l or 0.000 g/210L
 % Abs = 0.172
 Std Dev = 0.01 Rel Std Dev = 5.90
 Sol Val = 0.1905 mg/l or 0.040 g/210L
 % Abs = 1.649
 Std Dev = 0.00 Rel Std Dev = 0.30
 Sol Val = 0.4762 mg/l or 0.100 g/210L
 % Abs = 3.770
 Std Dev = 0.02 Rel Std Dev = 0.42
 Sol Val = 0.9524 mg/l or 0.200 g/210L
 % Abs = 7.158
 Std Dev = 0.01 Rel Std Dev = 0.08
 Sol Val = 1.4286 mg/l or 0.300 g/210L
 % Abs = 10.354
 Std Dev = 0.01 Rel Std Dev = 0.07
 Zero Order Coef = -222.39
 First Order Coef = 1273.96
 Second Order Coef = 12.28
 Standard Deviation = 5.463410

Solution Stats Quadratic Fit Chan 1		
Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.000	-0.0001
0.040	0.040	-0.0000
0.100	0.100	0.0003
0.200	0.200	-0.0003
0.300	0.300	0.0001

Solution Stats Quadratic Fit Chan 2		
Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	-0.000	0.0001
0.040	0.040	-0.0002
0.100	0.100	0.0001
0.200	0.200	-0.0001
0.300	0.300	0.0000

Sol Value = 0.080 g/210L ***
 Fit value = 0.3810 mg/l %%%
 Samples Taken = 4, Discarded = 1
 ***** CHANNEL 1 *****
 Sample #1 = 3053.00
 Sample #2 = 3050.00
 Sample #3 = 3006.00
 Sample #4 = 3071.00
 Average Result = 3042.3333
 STD DEV = 33.1713
 REL STD DEV = 1.090

***** CHANNEL 2 *****
 Sample #1 = 3396.00
 Sample #2 = 3431.00
 Sample #3 = 3420.00
 Sample #4 = 3462.00
 Average Result = 3437.6667
 STD DEV = 21.7792
 REL STD DEV = 0.634

 Dry Gas H2O Adjust Results *****
 Barometric Pressure = 1021
 3 um H2O Adjust (mg/l*10,000) = 767
 9 um H2O Adjust (mg/l*10,000) = 372
 ***** AUTO CAL PASS *****

Post-Stability Checks 80-001192 DA 3/27/25

PANAMA CITY BEACH PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001192
03/27/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:21
Control Test	0.048	11:22
Air Blank	0.000	11:22
Control Test	0.049	11:23
Air Blank	0.000	11:24
Control Test	0.050	11:24
Air Blank	0.000	11:25
Control Test Stats		
Average	0.0490	
Std Dev	0.0010	
Rel Std Dev(%)	2.0408	



Operator's Signature

PANAMA CITY BEACH PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001192
03/27/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:08
Control Test	0.081	11:09
Air Blank	0.000	11:09
Control Test	0.080	11:10
Air Blank	0.000	11:10
Control Test	0.080	11:11
Air Blank	0.000	11:12
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

Wet



Operator's Signature

PANAMA CITY BEACH PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001192
03/27/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:14
Control Test	0.201	11:14
Air Blank	0.000	11:15
Control Test	0.201	11:16
Air Blank	0.000	11:16
Control Test	0.201	11:17
Air Blank	0.000	11:17
Control Test Stats		
Average	0.2010	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	



Operator's Signature


PANAMA CITY BEACH PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001192
03/27/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:28
Control Test	0.078	11:28
Air Blank	0.000	11:29
Control Test	0.079	11:29
Air Blank	0.000	11:30
Control Test	0.078	11:30
Air Blank	0.000	11:30
Control Test Stats		
Average	0.0783	
Std Dev	0.0006	
Rel Std Dev(%)	0.7370	

DGS

DA 3/27/25





Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PANAMA CITY BEACH PD
Time of Inspection: 10:15

Date of Inspection: 03/31/2025

Serial Number: 80-001192
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG429602 Exp: 10/22/2026
0.000	0.048	0.078	0.199	0.077
0.000	0.048	0.078	0.199	0.077
0.000	0.048	0.079	0.199	0.077
0.000	0.048	0.078	0.199	0.077
0.000	0.048	0.078	0.199	0.076
0.000	0.048	0.078	0.199	0.076
0.000	0.048	0.078	0.198	0.076
0.000	0.048	0.078	0.198	0.076
0.000	0.048	0.078	0.199	0.076
0.000	0.048	0.078	0.199	0.076

Standard Deviations	0.0000	0.0003	0.0004	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.


I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



DESTINEE N ARMSTRONG

Signature and Printed Name

03/31/2025
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001192, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001192</u>	UNCERTAINTY * $\pm$	
Owning Agency:	<u>PANAMA CITY BEACH PD</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>03/31/2025</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>10:15</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

03/31/2025

Date

DESTINEE N ARMSTRONG,
Department Inspector

FDLE/ATP Form 69 October 2024
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality