



INSTRUMENT PROCESSING SHEET

Agency Hendry County Sheriff's OfficeS/N 80-001207Florida Department of
Law EnforcementDate In 06/02/2025 DI Completion Date 06/13/2025☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake By <u>TDG</u> Date <u>06/06/2025</u>		Quality Checks By <u>TDG</u> Date <u>06/13/2025</u>		Flow Calibration By _____ Date _____																																									
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☐ Handle ☒ Keyboard ☐ Dry Gas Shelf ☒ Feet ☐ Breath Tube ☒ Ports ☐ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: <u>Dropped off by Dep. Link</u>		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>132</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP106</u> 32 mm <u>0.140</u> (.139 - .169) 36 mm <u>0.160</u> (.156 - .190) 53 mm <u>0.230</u> (.228 - .278) 103 mm <u>0.500</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28199</u> ☒ Stability Checks		☐ Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																									
		<table border="1" style="width:100%"><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr><tr><td>0.050</td><td>MP6286</td><td>202406K 06/19/2026</td></tr><tr><td>0.080</td><td>MP6287</td><td>202406L 06/19/2026</td></tr><tr><td>0.200</td><td>MP6288</td><td>202406N 06/20/2026</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG429602 10/22/2026</td></tr></table>		Simulator	Serial #	Lot #/Exp	0.050	MP6286	202406K 06/19/2026	0.080	MP6287	202406L 06/19/2026	0.200	MP6288	202406N 06/20/2026	0.080 DGS	N/A	AG429602 10/22/2026	Maintenance By _____ Date _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____																										
Simulator	Serial #	Lot #/Exp																																											
0.050	MP6286	202406K 06/19/2026																																											
0.080	MP6287	202406L 06/19/2026																																											
0.200	MP6288	202406N 06/20/2026																																											
0.080 DGS	N/A	AG429602 10/22/2026																																											
Calibration Adjustment By _____		Department Inspection By <u>TDG</u>																																											
Barometric Pressure Gauge _____ ID # _____		Barometric Pressure ID# <u>28199</u> Gauge <u>1018</u> Instrument <u>1018</u> Mouth Alcohol Solution Lot # <u>2024-A</u> Acetone Stock Solution Lot # <u>2023-B</u>																																											
<table border="1" style="width:100%"><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			<table border="1" style="width:100%"><tr><th>Simulator</th><th>Serial Number</th></tr><tr><td>0.000</td><td>MP6284</td></tr><tr><td>Interferent</td><td>MP6285</td></tr><tr><td>0.050</td><td>MP6286</td></tr><tr><td>0.080</td><td>MP6287</td></tr><tr><td>0.200</td><td>MP6288</td></tr></table>				Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288
Simulator	Serial #	Lot #	Expiration																																										
0.000		N/A	N/A																																										
0.040																																													
0.100																																													
0.200																																													
0.300																																													
0.080 DGS	N/A																																												
Simulator	Serial Number																																												
0.000	MP6284																																												
Interferent	MP6285																																												
0.050	MP6286																																												
0.080	MP6287																																												
0.200	MP6288																																												
☐ Post Calibration Adjustment Stability Checks		Attachments ☒ Form 41 ☐ Post-Stability Checks ☒ Stability Checks ☐ Flow Calibration ☒ Calibration Certificate ☐ Form 40 ☐ Calibration Adjustment ☐ Other _____																																											
<table border="1" style="width:100%"><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></table>		Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																							
Simulator	Serial #	Lot #	Expiration																																										
0.050																																													
0.080																																													
0.200																																													
0.080 DGS	N/A																																												
Notes/Suggested Service: _____		Digitally signed by Phil Nicodemo Date: 2025.06.17 15:55:19 -0400 Shayla Platt Shayla Platt Date: 2025.06.17 19:32:59 -0400 Tech Review / Date Admin Review / Date																																											

Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																																
HENDRY COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 06/13/2025 Software: 8100.27 SN 80-001207 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:34</td></tr> <tr><td>Control Test</td><td>0.049</td><td>13:35</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:36</td></tr> <tr><td>Control Test</td><td>0.048</td><td>13:36</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:37</td></tr> <tr><td>Control Test</td><td>0.048</td><td>13:38</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:38</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0483</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1945</td><td></td></tr> </tbody> </table> Operator's Signature _____	Test	g/210L	Time	Air Blank	0.000	13:34	Control Test	0.049	13:35	Air Blank	0.000	13:36	Control Test	0.048	13:36	Air Blank	0.000	13:37	Control Test	0.048	13:38	Air Blank	0.000	13:38	Control Test Stats			Average	0.0483		Std Dev	0.0006		Rel Std Dev(%)	1.1945		HENDRY COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 06/13/2025 Software: 8100.27 SN 80-001207 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:41</td></tr> <tr><td>Control Test</td><td>0.079</td><td>13:41</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:42</td></tr> <tr><td>Control Test</td><td>0.079</td><td>13:43</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:43</td></tr> <tr><td>Control Test</td><td>0.079</td><td>13:44</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:45</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0790</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature _____	Test	g/210L	Time	Air Blank	0.000	13:41	Control Test	0.079	13:41	Air Blank	0.000	13:42	Control Test	0.079	13:43	Air Blank	0.000	13:43	Control Test	0.079	13:44	Air Blank	0.000	13:45	Control Test Stats			Average	0.0790		Std Dev	0.0000		Rel Std Dev(%)	0.0000		HENDRY COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 06/13/2025 Software: 8100.27 SN 80-001207 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:15</td></tr> <tr><td>Control Test</td><td>0.201</td><td>14:15</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:16</td></tr> <tr><td>Control Test</td><td>0.200</td><td>14:17</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:17</td></tr> <tr><td>Control Test</td><td>0.199</td><td>14:18</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:18</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.2000</td><td></td></tr> <tr><td>Std Dev</td><td>0.0010</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.5000</td><td></td></tr> </tbody> </table> Operator's Signature _____	Test	g/210L	Time	Air Blank	0.000	14:15	Control Test	0.201	14:15	Air Blank	0.000	14:16	Control Test	0.200	14:17	Air Blank	0.000	14:17	Control Test	0.199	14:18	Air Blank	0.000	14:18	Control Test Stats			Average	0.2000		Std Dev	0.0010		Rel Std Dev(%)	0.5000		HENDRY COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 06/13/2025 Software: 8100.27 SN 80-001207 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:19</td></tr> <tr><td>Control Test</td><td>0.078</td><td>13:20</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:20</td></tr> <tr><td>Control Test</td><td>0.078</td><td>13:21</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:21</td></tr> <tr><td>Control Test</td><td>0.079</td><td>13:21</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:22</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0783</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7370</td><td></td></tr> </tbody> </table> Operator's Signature _____	Test	g/210L	Time	Air Blank	0.000	13:19	Control Test	0.078	13:20	Air Blank	0.000	13:20	Control Test	0.078	13:21	Air Blank	0.000	13:21	Control Test	0.079	13:21	Air Blank	0.000	13:22	Control Test Stats			Average	0.0783		Std Dev	0.0006		Rel Std Dev(%)	0.7370	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:34																																																																																																																																																	
Control Test	0.049	13:35																																																																																																																																																	
Air Blank	0.000	13:36																																																																																																																																																	
Control Test	0.048	13:36																																																																																																																																																	
Air Blank	0.000	13:37																																																																																																																																																	
Control Test	0.048	13:38																																																																																																																																																	
Air Blank	0.000	13:38																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0483																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1945																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:41																																																																																																																																																	
Control Test	0.079	13:41																																																																																																																																																	
Air Blank	0.000	13:42																																																																																																																																																	
Control Test	0.079	13:43																																																																																																																																																	
Air Blank	0.000	13:43																																																																																																																																																	
Control Test	0.079	13:44																																																																																																																																																	
Air Blank	0.000	13:45																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0790																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:15																																																																																																																																																	
Control Test	0.201	14:15																																																																																																																																																	
Air Blank	0.000	14:16																																																																																																																																																	
Control Test	0.200	14:17																																																																																																																																																	
Air Blank	0.000	14:17																																																																																																																																																	
Control Test	0.199	14:18																																																																																																																																																	
Air Blank	0.000	14:18																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.2000																																																																																																																																																		
Std Dev	0.0010																																																																																																																																																		
Rel Std Dev(%)	0.5000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:19																																																																																																																																																	
Control Test	0.078	13:20																																																																																																																																																	
Air Blank	0.000	13:20																																																																																																																																																	
Control Test	0.078	13:21																																																																																																																																																	
Air Blank	0.000	13:21																																																																																																																																																	
Control Test	0.079	13:21																																																																																																																																																	
Air Blank	0.000	13:22																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0783																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7370																																																																																																																																																		

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: HENDRY COUNTY SO
Time of Inspection: 16:23

Date of Inspection: 06/13/2025

Serial Number: 80-001207
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG429602 Exp: 10/22/2026
0.000	0.049	0.079	0.200	0.077
0.000	0.049	0.079	0.199	0.076
0.000	0.049	0.078	0.199	0.077
0.000	0.048	0.078	0.199	0.076
0.000	0.049	0.079	0.199	0.076
0.000	0.048	0.078	0.199	0.076
0.000	0.049	0.078	0.199	0.076
0.000	0.048	0.078	0.199	0.076
0.000	0.049	0.078	0.198	0.075
0.000	0.048	0.078	0.198	0.075

Standard Deviations	0.0005	0.0004	0.0005	0.0006
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



 Signature and Printed Name

TAYLOR D GUTSCHOW

06/13/2025
 Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001207, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001207</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>HENDRY COUNTY SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>06/13/2025</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>16:23</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Taylor
Gutschow
Digitally signed by Taylor
Gutschow
Date: 2025.06.14 14:48:09
-04'00'

06/13/2025

Date

TAYLOR D GUTSCHOW,

Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality