



Agency West Palm Beach Police Department

S/N 80-001235

Date In 04/14/2025 DI Completion Date 05/07/2025

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Calibration Adjustment

By_

ID

☐ Post Calibration Adjustment Stability ChecksNotes/Suggested Service:

Department Inspection

By TDG

Acetone Stock Solution Lot # 2023-B

Attachments

- Instrument Complies with Chapter 11D-8, FAC**

- ☐
- Instrument Does Not Comply with Chapter 11D-8, FAC

- #### Return to/Place into Evidentiary Use

- ☐
- Remain Out of Evidentiary Use**

- ### ■ Conduct an Agency Inspection Before Evidentiary Use

Digitally signed by

Phil Nicodemo

Shayla Platt

Shayla Platt
Date: 2025.05.15
20:41:32 -04'00'

Tech Review / Date

Admin Review / Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: WEST PALM BEACH PD

Time of Inspection: 13:22

Date of Inspection: 05/06/2025

Serial Number: 80-001235

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08g/210L Test (g/210L) Lot#:_____ Exp:_____	0.20g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:_____ Exp:_____

Number of Simulators Used: _____

Remarks:

AI NOT CONDUCTED. BYPASSED TO BRING OUT OF DISABLED MODE.

Not determined *7/6/25*

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Taylor D Gutschow

TAYLOR D GUTSCHOW

Signature and Printed Name

05/06/2025
Date

Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																												
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																												
✓	✓	✓	✓																																																																																																												
WEST PALM BEACH PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001235 05/06/2025 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:12</td></tr> <tr><td>Control Test</td><td>0.050</td><td>14:13</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:14</td></tr> <tr><td>Control Test</td><td>0.049</td><td>14:14</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:15</td></tr> <tr><td>Control Test</td><td>0.049</td><td>14:16</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:16</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0493</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1703</td><td></td></tr> </tbody> </table> Operator's Signature _____	Test	g/210L	Time	Air Blank	0.000	14:12	Control Test	0.050	14:13	Air Blank	0.000	14:14	Control Test	0.049	14:14	Air Blank	0.000	14:15	Control Test	0.049	14:16	Air Blank	0.000	14:16	Control Test Stats			Average	0.0493		Std Dev	0.0006		Rel Std Dev(%)	1.1703		WEST PALM BEACH PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001235 05/06/2025 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:22</td></tr> <tr><td>Control Test</td><td>0.079</td><td>14:22</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:23</td></tr> <tr><td>Control Test</td><td>0.079</td><td>14:24</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:24</td></tr> <tr><td>Control Test</td><td>0.079</td><td>14:25</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:25</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0790</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature _____	Test	g/210L	Time	Air Blank	0.000	14:22	Control Test	0.079	14:22	Air Blank	0.000	14:23	Control Test	0.079	14:24	Air Blank	0.000	14:24	Control Test	0.079	14:25	Air Blank	0.000	14:25	Control Test Stats			Average	0.0790		Std Dev	0.0000		Rel Std Dev(%)	0.0000		WEST PALM BEACH PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001235 05/06/2025 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:30</td></tr> <tr><td>Control Test</td><td>0.200</td><td>14:31</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:32</td></tr> <tr><td>Control Test</td><td>0.198</td><td>14:32</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:33</td></tr> <tr><td>Control Test</td><td>0.198</td><td>14:34</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:34</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1987</td><td></td></tr> <tr><td>Std Dev</td><td>0.0012</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.5812</td><td></td></tr> </tbody> </table> Operator's Signature _____	Test	g/210L	Time	Air Blank	0.000	14:30	Control Test	0.200	14:31	Air Blank	0.000	14:32	Control Test	0.198	14:32	Air Blank	0.000	14:33	Control Test	0.198	14:34	Air Blank	0.000	14:34	Control Test Stats			Average	0.1987		Std Dev	0.0012		Rel Std Dev(%)	0.5812		Printed to external printer. Will attach. ML SLC/25
Test	g/210L	Time																																																																																																													
Air Blank	0.000	14:12																																																																																																													
Control Test	0.050	14:13																																																																																																													
Air Blank	0.000	14:14																																																																																																													
Control Test	0.049	14:14																																																																																																													
Air Blank	0.000	14:15																																																																																																													
Control Test	0.049	14:16																																																																																																													
Air Blank	0.000	14:16																																																																																																													
Control Test Stats																																																																																																															
Average	0.0493																																																																																																														
Std Dev	0.0006																																																																																																														
Rel Std Dev(%)	1.1703																																																																																																														
Test	g/210L	Time																																																																																																													
Air Blank	0.000	14:22																																																																																																													
Control Test	0.079	14:22																																																																																																													
Air Blank	0.000	14:23																																																																																																													
Control Test	0.079	14:24																																																																																																													
Air Blank	0.000	14:24																																																																																																													
Control Test	0.079	14:25																																																																																																													
Air Blank	0.000	14:25																																																																																																													
Control Test Stats																																																																																																															
Average	0.0790																																																																																																														
Std Dev	0.0000																																																																																																														
Rel Std Dev(%)	0.0000																																																																																																														
Test	g/210L	Time																																																																																																													
Air Blank	0.000	14:30																																																																																																													
Control Test	0.200	14:31																																																																																																													
Air Blank	0.000	14:32																																																																																																													
Control Test	0.198	14:32																																																																																																													
Air Blank	0.000	14:33																																																																																																													
Control Test	0.198	14:34																																																																																																													
Air Blank	0.000	14:34																																																																																																													
Control Test Stats																																																																																																															
Average	0.1987																																																																																																														
Std Dev	0.0012																																																																																																														
Rel Std Dev(%)	0.5812																																																																																																														

WEST PALM BEACH PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001235
05/06/2025
Software: 8100.27

0.08
DLS

Test	g/210L	Time
Air Blank	0.000	13:36
Control Test	0.078	13:36
Air Blank	0.000	13:37
Control Test	0.078	13:37
Air Blank	0.000	13:37
Control Test	0.078	13:38
Air Blank	0.000	13:38
Control Test Stats		
Average	0.0780	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

ML

Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: WEST PALM BEACH PD
Time of Inspection: 13:40

Date of Inspection: 05/07/2025

Serial Number: 80-001235
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG429602 Exp: 10/22/2026
0.000	0.050	0.079	0.198	0.077
0.000	0.049	0.079	0.199	0.077
0.000	0.049	0.079	0.198	0.077
0.000	0.049	0.079	0.199	0.077
0.000	0.049	0.080	0.199	0.077
0.000	0.049	0.079	0.198	0.077
0.000	0.049	0.079	0.199	0.077
0.000	0.049	0.079	0.199	0.077
0.000	0.049	0.079	0.199	0.077
0.000	0.050	0.079	0.198	0.077
0.000	0.049	0.079	0.198	0.077

Standard Deviations	0.0004	0.0003	0.0005	0.0000
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

05/07/2025
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001235, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001235</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>WEST PALM BEACH PD</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>05/07/2025</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>13:40</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Taylor
Gutschow
Digitally signed by Taylor
Gutschow
Date: 2025.05.08 10:51:40
+04'00'

05/07/2025

Date

TAYLOR D GUTSCHOW,

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Department Inspector

Service • Integrity • Respect • Quality