



## INSTRUMENT PROCESSING SHEET

Agency Jacksonville Sheriff's Office (JSO)S/N 80-001279Florida Department of  
Law EnforcementDate In 6/2/2025DI Completion Date 06/17/2025☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By SLH	Date	Quality Checks	By SLH	Date	Flow Calibration	By	Date															
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE  Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight  Other Equipment/ Accessories: <input checked="" type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable  Notes: <u>Memo for inspection from</u> <u>JSO included. SLH 6/19/25</u>		<u>6/2/2025</u>	<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>134</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP105</u> 32 mm <u>0.148</u> (.139 - .169) 36 mm <u>0.160</u> (.156 - .190) 53 mm <u>0.230</u> (.228 - .278) 103 mm <u>0.503</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28421</u> <input checked="" type="checkbox"/> Stability Checks		<u>6/2/2025</u>	Flow Column # _____ <input type="checkbox"/> 5L/min – 17mm <input type="checkbox"/> 15L/min – 53mm <input type="checkbox"/> 30L/min – 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																	
			<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP5088</td><td>202406K 06/19/2026</td></tr><tr><td>0.080</td><td>MP5089</td><td>202406L 06/19/2026</td></tr><tr><td>0.200</td><td>MP5090</td><td>202406N 06/20/2026</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG429602 10/22/2026</td></tr></tbody></table>	Simulator	Serial #	Lot #/Exp	0.050	MP5088	202406K 06/19/2026	0.080	MP5089	202406L 06/19/2026	0.200	MP5090	202406N 06/20/2026	0.080 DGS	N/A	AG429602 10/22/2026					
Simulator	Serial #	Lot #/Exp																					
0.050	MP5088	202406K 06/19/2026																					
0.080	MP5089	202406L 06/19/2026																					
0.200	MP5090	202406N 06/20/2026																					
0.080 DGS	N/A	AG429602 10/22/2026																					

Calibration Adjustment	By SLH	Department Inspection	By SLH																																																														
Barometric Pressure Gauge <u>1012/1013</u> ID # <u>28662/28427</u> <table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td>MP6294/5086</td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td>MP6295</td><td>24330</td><td>09/10/2026</td></tr><tr><td>0.100</td><td>MP6296</td><td>24390</td><td>10/29/2026</td></tr><tr><td>0.200</td><td>MP6297</td><td>24080</td><td>02/13/2026</td></tr><tr><td>0.300</td><td>MP6298</td><td>24430</td><td>12/10/2026</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>28424080A3</td><td>11/05/2026</td></tr></tbody></table> <input checked="" type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td>MP5088</td><td>202406K</td><td>06/19/2026</td></tr><tr><td>0.080</td><td>MP5089</td><td>202406L</td><td>06/19/2026</td></tr><tr><td>0.200</td><td>MP5090</td><td>202406N</td><td>06/20/2026</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG429602</td><td>10/22/2026</td></tr></tbody></table> Notes/Suggested Service: <u>Instrument stalled on 'Autorang-</u> <u>ing' portion of both days of adjustments (adj) and stalled</u> <u>on the selection of 'Inspection' during the Department</u> <u>Inspection. Had faults detected at 0.080g/210L DGS for</u> <u>1st adj and 0.04 and 0.100g/210L during 2nd adj so re-</u> <u>peated when prompted. Good connections of the</u> <u>simulator to the instrument was difficult to achieve.SLH</u> <u>6/17/2025</u>	Simulator	Serial #	Lot #	Expiration	0.000	MP6294/5086	N/A	N/A	0.040	MP6295	24330	09/10/2026	0.100	MP6296	24390	10/29/2026	0.200	MP6297	24080	02/13/2026	0.300	MP6298	24430	12/10/2026	0.080 DGS	N/A	28424080A3	11/05/2026	Simulator	Serial #	Lot #	Expiration	0.050	MP5088	202406K	06/19/2026	0.080	MP5089	202406L	06/19/2026	0.200	MP5090	202406N	06/20/2026	0.080 DGS	N/A	AG429602	10/22/2026		Barometric Pressure ID# <u>28421</u> Gauge <u>1013</u> Instrument <u>1012</u> Mouth Alcohol Solution Lot # <u>2025A</u> Acetone Stock Solution Lot # <u>2024B</u> <table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6294</td></tr><tr><td>Interferent</td><td>MP5087</td></tr><tr><td>0.050</td><td>MP5088</td></tr><tr><td>0.080</td><td>MP5089</td></tr><tr><td>0.200</td><td>MP5090</td></tr></tbody></table> <b>Attachments</b> <table border="1"><tbody><tr><td><input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Calibration Certificate <input checked="" type="checkbox"/> Calibration Adjustment x2</td><td><input checked="" type="checkbox"/> Post-Stability Checks x2 <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Form 40 <input checked="" type="checkbox"/> Other <u>Memo</u> SLH 6/19/25</td></tr></tbody></table> <input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use  Taylor Gutschow <small>Digitally signed by Taylor Gutschow Date: 2025.06.24 10:12:49 -0400</small> Phil Nicodemo <small>Digitally signed by Phil Nicodemo Date: 2025.06.26 10:29:15 -0400</small> Tech Review / Date _____ Admin Review / Date _____	Simulator	Serial Number	0.000	MP6294	Interferent	MP5087	0.050	MP5088	0.080	MP5089	0.200	MP5090	<input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Calibration Certificate <input checked="" type="checkbox"/> Calibration Adjustment x2	<input checked="" type="checkbox"/> Post-Stability Checks x2 <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Form 40 <input checked="" type="checkbox"/> Other <u>Memo</u> SLH 6/19/25	
Simulator	Serial #	Lot #	Expiration																																																														
0.000	MP6294/5086	N/A	N/A																																																														
0.040	MP6295	24330	09/10/2026																																																														
0.100	MP6296	24390	10/29/2026																																																														
0.200	MP6297	24080	02/13/2026																																																														
0.300	MP6298	24430	12/10/2026																																																														
0.080 DGS	N/A	28424080A3	11/05/2026																																																														
Simulator	Serial #	Lot #	Expiration																																																														
0.050	MP5088	202406K	06/19/2026																																																														
0.080	MP5089	202406L	06/19/2026																																																														
0.200	MP5090	202406N	06/20/2026																																																														
0.080 DGS	N/A	AG429602	10/22/2026																																																														
Simulator	Serial Number																																																																
0.000	MP6294																																																																
Interferent	MP5087																																																																
0.050	MP5088																																																																
0.080	MP5089																																																																
0.200	MP5090																																																																
<input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Calibration Certificate <input checked="" type="checkbox"/> Calibration Adjustment x2	<input checked="" type="checkbox"/> Post-Stability Checks x2 <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Form 40 <input checked="" type="checkbox"/> Other <u>Memo</u> SLH 6/19/25																																																																

Tech Review: Added note in Intake area and added checkmark and 'Memo' in the Attachments area SLH 6/19/25



## Jacksonville Sheriff's Office

T.K. WATERS, SHERIFF

Serve • Protect • Community

ATTENTION FDLE : 80-001279

Instrument is here for annual Departmental Inspection.

ATTN : Sgt M. Potter #7599

Office of the Sheriff

500 East Adams St.

Jacksonville FL. 32202





Alcohol Testing Department

Rec'd in lab 6/2/2025

SHA

# Stability Checks

SWS 6/2/25  
80-001279

0.050 g/210L 0.047 to 0.053 g/210L	0.080 g/210L 0.077 to 0.083 g/210L	0.200 g/210L 0.194 to 0.206 g/210L	DGS 0.080 g/210L 0.077 to 0.083 g/210L 50.003 g/210L of Wet																																																																																																																																																
<p>Performed Root Case Analysis <input checked="" type="checkbox"/></p> <p>Ensured connection w/ return port and retrain see following page. Also verified seal of simulator.</p>	<p>Performed Root Case Analysis <input checked="" type="checkbox"/></p>	<p>Performed Root Case Analysis <input checked="" type="checkbox"/></p> <p>no user or equipment error - will perform calibration adjustment.</p>	<p>Performed Root Case Analysis <input checked="" type="checkbox"/></p>																																																																																																																																																
<p>JACKSONVILLE SO Intoxilyzer - Alcohol Analyzer Model 8000 06/02/2025 Software: 8100.27</p> <p>#1</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>15:00</td></tr> <tr><td>Control Test</td><td>0.035</td><td>15:00</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:01</td></tr> <tr><td>Control Test</td><td>0.035</td><td>15:02</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:02</td></tr> <tr><td>Control Test</td><td>0.034</td><td>15:03</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:03</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0347</td><td></td></tr> <tr><td>Std Dev</td><td>0.0106</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.6654</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	15:00	Control Test	0.035	15:00	Air Blank	0.000	15:01	Control Test	0.035	15:02	Air Blank	0.000	15:02	Control Test	0.034	15:03	Air Blank	0.000	15:03	Control Test Stats			Average	0.0347		Std Dev	0.0106		Rel Std Dev(%)	1.6654		<p>JACKSONVILLE SO Intoxilyzer - Alcohol Analyzer Model 8000 06/02/2025 Software: 8100.27</p> <p>JACKSONVILLE SO Intoxilyzer - Alcohol Analyzer Model 8000 06/02/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>15:06</td></tr> <tr><td>Control Test</td><td>0.079</td><td>15:07</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:07</td></tr> <tr><td>Control Test</td><td>0.079</td><td>15:08</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:09</td></tr> <tr><td>Control Test</td><td>0.078</td><td>15:09</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:10</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0787</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7339</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	15:06	Control Test	0.079	15:07	Air Blank	0.000	15:07	Control Test	0.079	15:08	Air Blank	0.000	15:09	Control Test	0.078	15:09	Air Blank	0.000	15:10	Control Test Stats			Average	0.0787		Std Dev	0.0006		Rel Std Dev(%)	0.7339		<p>JACKSONVILLE SO Intoxilyzer - Alcohol Analyzer Model 8000 06/02/2025 Software: 8100.27</p> <p>JACKSONVILLE SO Intoxilyzer - Alcohol Analyzer Model 8000 06/02/2025 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>15:16</td></tr> <tr><td>Control Test</td><td>0.135</td><td>15:17</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:18</td></tr> <tr><td>Control Test</td><td>0.133</td><td>15:18</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:19</td></tr> <tr><td>Control Test</td><td>0.134</td><td>15:19</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:20</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1340</td><td></td></tr> <tr><td>Std Dev</td><td>0.0010</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7463</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	15:16	Control Test	0.135	15:17	Air Blank	0.000	15:18	Control Test	0.133	15:18	Air Blank	0.000	15:19	Control Test	0.134	15:19	Air Blank	0.000	15:20	Control Test Stats			Average	0.1340		Std Dev	0.0010		Rel Std Dev(%)	0.7463		<p>JACKSONVILLE SO Intoxilyzer - Alcohol Analyzer Model 8000 06/02/2025 Software: 8100.27</p> <p>DGS</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>15:22</td></tr> <tr><td>Control Test</td><td>0.079</td><td>15:23</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:23</td></tr> <tr><td>Control Test</td><td>0.079</td><td>15:23</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:24</td></tr> <tr><td>Control Test</td><td>0.079</td><td>15:24</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:25</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0790</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	15:22	Control Test	0.079	15:23	Air Blank	0.000	15:23	Control Test	0.079	15:23	Air Blank	0.000	15:24	Control Test	0.079	15:24	Air Blank	0.000	15:25	Control Test Stats			Average	0.0790		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	15:00																																																																																																																																																	
Control Test	0.035	15:00																																																																																																																																																	
Air Blank	0.000	15:01																																																																																																																																																	
Control Test	0.035	15:02																																																																																																																																																	
Air Blank	0.000	15:02																																																																																																																																																	
Control Test	0.034	15:03																																																																																																																																																	
Air Blank	0.000	15:03																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0347																																																																																																																																																		
Std Dev	0.0106																																																																																																																																																		
Rel Std Dev(%)	1.6654																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	15:06																																																																																																																																																	
Control Test	0.079	15:07																																																																																																																																																	
Air Blank	0.000	15:07																																																																																																																																																	
Control Test	0.079	15:08																																																																																																																																																	
Air Blank	0.000	15:09																																																																																																																																																	
Control Test	0.078	15:09																																																																																																																																																	
Air Blank	0.000	15:10																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0787																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7339																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	15:16																																																																																																																																																	
Control Test	0.135	15:17																																																																																																																																																	
Air Blank	0.000	15:18																																																																																																																																																	
Control Test	0.133	15:18																																																																																																																																																	
Air Blank	0.000	15:19																																																																																																																																																	
Control Test	0.134	15:19																																																																																																																																																	
Air Blank	0.000	15:20																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1340																																																																																																																																																		
Std Dev	0.0010																																																																																																																																																		
Rel Std Dev(%)	0.7463																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	15:22																																																																																																																																																	
Control Test	0.079	15:23																																																																																																																																																	
Air Blank	0.000	15:23																																																																																																																																																	
Control Test	0.079	15:23																																																																																																																																																	
Air Blank	0.000	15:24																																																																																																																																																	
Control Test	0.079	15:24																																																																																																																																																	
Air Blank	0.000	15:25																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0790																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
<p>Operator's Signature</p> 	<p>Operator's Signature</p> 	<p>Operator's Signature</p> 	<p>Operator's Signature</p> 																																																																																																																																																

S/N: 80-001279

6/2/2025 SUT

JACKSONVILLE SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001279  
06/02/2025  
Software: 8100.27

#2 for 0.050 g/210L will perform calibration adjustment.

Test	g/210L	Time
Air Blank	0.000	15:12
Control Test	0.039	15:12
Air Blank	0.000	15:13
Control Test	0.041	15:13
Air Blank	0.000	15:14
Control Test	0.041	15:15
Air Blank	0.000	15:15
Control Test Stats		
Average	0.0403	
Std Dev	0.0012	
Rel Std Dev(%)	2.9629	

  
Operator's Signature

80-001279  
6/10/25 SLS

Auto Calibration  
Max Power Res Value = 42  
Auto Range Res Value = 4

Sol Value = 0.000 g/210L \*\*\*  
Fit value = 0.0000 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12850, Sum Io = 13598

Channel 1 Data:  
Sample % Abs (% Abs Ref)  
Sample #1 = 0.1580 (-0.0290)  
Sample #2 = 0.1320 (0.0460)  
Sample #3 = 0.1400 (0.0620)  
Sample #4 = 0.1260 (0.1150)  
Avg % Abs = 0.1327 (0.0743)  
STD DEV = 0.0070 (0.0361)  
REL STD DEV = 5.294 (48.586)

Channel 2 Data:  
Sample % Abs (% Abs Ref)  
Sample #1 = 0.1590 (-0.0010)  
Sample #2 = 0.1540 (0.0370)  
Sample #3 = 0.1340 (0.0460)  
Sample #4 = 0.1400 (0.0480)  
Avg % Abs = 0.1427 (0.0437)  
STD DEV = 0.0103 (0.0059)  
REL STD DEV = 7.194 (13.419)

Sol Value = 0.040 g/210L \*\*\*  
Fit value = 0.1905 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12825, Sum Io = 13586

Channel 1 Data:  
Sample % Abs (% Abs Ref)  
Sample #1 = 0.8670 (-0.0330)  
Sample #2 = 0.8650 (0.0000)  
Sample #3 = 0.8620 (0.0310)  
Sample #4 = 0.8760 (0.0440)  
Avg % Abs = 0.8677 (0.0250)  
STD DEV = 0.0074 (0.0226)  
REL STD DEV = 0.850 (90.421)

Channel 2 Data:  
Sample % Abs (% Abs Ref)  
Sample #1 = 1.5580 (-0.0160)  
Sample #2 = 1.5710 (-0.0070)  
Sample #3 = 1.5560 (-0.0030)  
Sample #4 = 1.5350 (0.0250)  
Avg % Abs = 1.5540 (0.0050)  
STD DEV = 0.0181 (0.0174)  
REL STD DEV = 1.164 (348.712)

Sol Value = 0.100 g/210L \*\*\*  
Fit value = 0.4762 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12809, Sum Io = 13581

Channel 1 Data:  
Sample % Abs (% Abs Ref)  
Sample #1 = 1.9180 (-0.0120)  
Sample #2 = 1.9220 (0.0290)  
Sample #3 = 1.9180 (0.0540)  
Sample #4 = 1.9090 (0.0890)  
Avg % Abs = 1.9163 (0.0573)  
STD DEV = 0.0067 (0.0301)  
REL STD DEV = 0.347 (52.567)

Channel 2 Data:  
Sample % Abs (% Abs Ref)  
Sample #1 = 3.6010 (-0.0160)  
Sample #2 = 3.5760 (0.0240)  
Sample #3 = 3.5580 (0.0460)  
Sample #4 = 3.5830 (0.0470)  
Avg % Abs = 3.5723 (0.0390)  
STD DEV = 0.0129 (0.0130)  
REL STD DEV = 0.361 (33.333)

Sol Value = 0.200 g/210L \*\*\*  
Fit value = 0.9524 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12792, Sum Io = 13574

Channel 1 Data:  
Sample % Abs (% Abs Ref)  
Sample #1 = 3.6860 (-0.0180)  
Sample #2 = 3.6850 (0.0060)  
Sample #3 = 3.6750 (0.0300)  
Sample #4 = 3.6510 (0.0700)  
Avg % Abs = 3.6703 (0.0353)  
STD DEV = 0.0175 (0.0323)  
REL STD DEV = 0.476 (91.505)

Channel 2 Data:  
Sample % Abs (% Abs Ref)  
Sample #1 = 6.9230 (-0.0130)  
Sample #2 = 6.8690 (0.0330)  
Sample #3 = 6.8460 (0.0650)  
Sample #4 = 6.8570 (0.0680)  
Avg % Abs = 6.8573 (0.0557)  
STD DEV = 0.0115 (0.0197)  
REL STD DEV = 0.168 (35.309)

Sol Value = 0.300 g/210L \*\*\*  
Fit value = 1.4286 mg/l %%%  
Samples Taken = 4, Discarded = 1  
Sum Io = 12781, Sum Io = 13570

Channel 1 Data:  
Sample % Abs (% Abs Ref)  
Sample #1 = 5.4020 (-0.0100)  
Sample #2 = 5.3460 (0.0490)  
Sample #3 = 5.3370 (0.0850)  
Sample #4 = 5.3420 (0.0920)  
Avg % Abs = 5.3417 (0.0753)  
STD DEV = 0.0045 (0.0231)  
REL STD DEV = 0.084 (30.627)

Channel 2 Data:  
Sample % Abs (% Abs Ref)  
Sample #1 = 10.0510 (-0.0070)  
Sample #2 = 9.9530 (0.0980)  
Sample #3 = 9.9460 (0.1070)  
Sample #4 = 9.9290 (0.1290)  
Avg % Abs = 9.9427 (0.1113)  
STD DEV = 0.0123 (0.0159)  
REL STD DEV = 0.124 (14.324)

Auto Calibration Data:  
Channel 1 Data:  
Sol Val = 0.0000 mg/l or 0.000 g/210L  
% Abs = 0.133  
Std Dev = 0.01 Rel Std Dev = 5.29  
Sol Val = 0.1905 mg/l or 0.040 g/210L  
% Abs = 0.868  
Std Dev = 0.01 Rel Std Dev = 0.85  
Sol Val = 0.4762 mg/l or 0.100 g/210L  
% Abs = 1.916  
Std Dev = 0.01 Rel Std Dev = 0.35  
Sol Val = 0.9524 mg/l or 0.200 g/210L  
% Abs = 3.670  
Std Dev = 0.02 Rel Std Dev = 0.48  
Sol Val = 1.4286 mg/l or 0.300 g/210L  
% Abs = 5.342  
Std Dev = 0.00 Rel Std Dev = 0.08  
Zero Order Coef = -350.08  
First Order Coef = 2601.73  
Second Order Coef = 25.64  
Standard Deviation = 22.359337

Channel 2 Data:  
Sol Val = 0.0000 mg/l or 0.000 g/210L  
% Abs = 0.143  
Std Dev = 0.01 Rel Std Dev = 7.19  
Sol Val = 0.1905 mg/l or 0.040 g/210L  
% Abs = 1.554  
Std Dev = 0.02 Rel Std Dev = 1.16  
Sol Val = 0.4762 mg/l or 0.100 g/210L  
% Abs = 3.572  
Std Dev = 0.01 Rel Std Dev = 0.36  
Sol Val = 0.9524 mg/l or 0.200 g/210L  
% Abs = 6.857  
Std Dev = 0.01 Rel Std Dev = 0.17  
Sol Val = 1.4286 mg/l or 0.300 g/210L  
% Abs = 9.943  
Std Dev = 0.01 Rel Std Dev = 0.12  
Zero Order Coef = -192.10  
First Order Coef = 1338.55  
Second Order Coef = 11.77  
Standard Deviation = 15.197507

Solution Stats Quadratic Fit Chan 1

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	-0.000	0.0001
0.040	0.040	-0.0005
0.100	0.099	0.0007
0.200	0.200	-0.0004
0.300	0.300	0.0001

80-001279

6/10/25 SW

Solution Stats Quadratic Fit Chan 2		
Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	-0.000	0.0000
0.040	0.040	-0.0002
0.100	0.100	0.0005
0.200	0.200	-0.0003
0.300	0.300	0.0001

Sol Value = 0.080 g/210L \*\*\*  
Fit value = 0.3810 mg/l %%%  
Samples Taken = 4, Discarded = 1  
\*\*\*\*\* CHANNEL 1  
Sample #1 = 3104.00  
Sample #2 = 3088.00  
Sample #3 = 3116.00  
Sample #4 = 3000.00  
Average Result = 3068.0000  
STD DEV = 60.5310  
REL STD DEV = 1.973

\*\*\*\*\*  
\*\*\*\*\* CHANNEL 2  
Sample #1 = 3424.00  
Sample #2 = 3709.00  
Sample #3 = 3433.00  
Sample #4 = 3410.00  
Average Result = 3517.3333  
STD DEV = 166.3861  
REL STD DEV = 4.730

\*\*\*\*\*  
Sol Value = 0.080 g/210L \*\*\*  
Fit value = 0.3810 mg/l %%%  
Samples Taken = 4, Discarded = 1  
\*\*\*\*\* CHANNEL 1  
Sample #1 = 3187.00  
Sample #2 = 3096.00  
Sample #3 = 3098.00  
Sample #4 = 3125.00  
Average Result = 3106.3333  
STD DEV = 16.1967  
REL STD DEV = 0.521





\*\*\*\*\*  
\*\*\*\*\* CHANNEL 2  
Sample #1 = 3400.00  
Sample #2 = 3416.00  
Sample #3 = 3443.00  
Sample #4 = 3461.00  
Average Result = 3440.0000  
STD DEV = 22.6495  
REL STD DEV = 0.658

\*\*\*\*\*  
Dry Gas H2O Adjust Results \*\*\*\*\*  
Barometric Pressure = 1012  
3 um H2O Adjust (mg/l\*10,000) = 703  
9 um H2O Adjust (mg/l\*10,000) = 369  
\*\*\*\* AUTO CAL PASS



80-001279  
6/10/2025 SUT

# Post-Calibration Adjustment Stability Checks

0.050 g/210L	0.080 g/210L	0.200 g/210L	DGS 0.080 g/210L																																																																																																																														
0.047 to 0.053 g/210L NO -	0.077 to 0.083 g/210L	0.194 to 0.206 g/210L	0.077 to 0.083 g/210L ≤ 0.003 g/210L of Wet																																																																																																																														
Performed Root Case Analysis	Performed Root Case Analysis	Performed Root Case Analysis	Performed Root Case Analysis																																																																																																																														
<p>JACKSONVILLE SO Intoxilyzer - Alcohol Analyzer Model 8000 06/10/2025 Software: 8100.27</p> <p>#1</p> <p>SN 80-001279</p> <p>Test g/210L Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>17:28</td></tr> <tr><td>Control Test</td><td>0.035</td><td>17:28</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>17:29</td></tr> <tr><td>Control Test</td><td>0.037</td><td>17:29</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>17:30</td></tr> <tr><td>Control Test</td><td>0.037</td><td>17:31</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>17:31</td></tr> <tr><td>Control Test</td><td>0.037</td><td>17:31</td></tr> <tr><td>Average</td><td>0.0363</td><td></td></tr> <tr><td>Std Dev</td><td>0.0012</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>3.1781</td><td></td></tr> </table>	Air Blank	0.000	17:28	Control Test	0.035	17:28	Air Blank	0.000	17:29	Control Test	0.037	17:29	Air Blank	0.000	17:30	Control Test	0.037	17:31	Air Blank	0.000	17:31	Control Test	0.037	17:31	Average	0.0363		Std Dev	0.0012		Rel Std Dev(%)	3.1781		<p>JACKSONVILLE SO Intoxilyzer - Alcohol Analyzer Model 8000 06/10/2025 Software: 8100.27</p> <p>SN 81-001279</p> <p>Test g/210L Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>17:42</td></tr> <tr><td>Control Test</td><td>0.080</td><td>17:43</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>17:43</td></tr> <tr><td>Control Test</td><td>0.080</td><td>17:44</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>17:45</td></tr> <tr><td>Control Test</td><td>0.080</td><td>17:45</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>17:46</td></tr> <tr><td>Control Test</td><td>0.080</td><td>17:46</td></tr> <tr><td>Average</td><td>0.0800</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </table>	Air Blank	0.000	17:42	Control Test	0.080	17:43	Air Blank	0.000	17:43	Control Test	0.080	17:44	Air Blank	0.000	17:45	Control Test	0.080	17:45	Air Blank	0.000	17:46	Control Test	0.080	17:46	Average	0.0800		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>JACKSONVILLE SO Intoxilyzer - Alcohol Analyzer Model 8000 06/10/2025 Software: 8100.27</p> <p>SN 80-001279</p> <p>Test g/210L Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>17:37</td></tr> <tr><td>Control Test</td><td>0.197</td><td>17:38</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>17:39</td></tr> <tr><td>Control Test</td><td>0.199</td><td>17:39</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>17:40</td></tr> <tr><td>Control Test</td><td>0.199</td><td>17:41</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>17:41</td></tr> <tr><td>Control Test</td><td>0.000</td><td>17:41</td></tr> <tr><td>Average</td><td>0.1980</td><td></td></tr> <tr><td>Std Dev</td><td>0.0010</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.5051</td><td></td></tr> </table>	Air Blank	0.000	17:37	Control Test	0.197	17:38	Air Blank	0.000	17:39	Control Test	0.199	17:39	Air Blank	0.000	17:40	Control Test	0.199	17:41	Air Blank	0.000	17:41	Control Test	0.000	17:41	Average	0.1980		Std Dev	0.0010		Rel Std Dev(%)	0.5051		<p>JACKSONVILLE SO Intoxilyzer - Alcohol Analyzer Model 8000 06/10/2025 Software: 8100.27</p> <p>SN 80-001279</p> <p>Test g/210L Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>17:48</td></tr> <tr><td>Control Test</td><td>0.079</td><td>17:48</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>17:49</td></tr> <tr><td>Control Test</td><td>0.078</td><td>17:49</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>17:50</td></tr> <tr><td>Control Test</td><td>0.0783</td><td>17:50</td></tr> <tr><td>Average</td><td>0.0783</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7370</td><td></td></tr> </table>	Air Blank	0.000	17:48	Control Test	0.079	17:48	Air Blank	0.000	17:49	Control Test	0.078	17:49	Air Blank	0.000	17:50	Control Test	0.0783	17:50	Average	0.0783		Std Dev	0.0006		Rel Std Dev(%)	0.7370	
Air Blank	0.000	17:28																																																																																																																															
Control Test	0.035	17:28																																																																																																																															
Air Blank	0.000	17:29																																																																																																																															
Control Test	0.037	17:29																																																																																																																															
Air Blank	0.000	17:30																																																																																																																															
Control Test	0.037	17:31																																																																																																																															
Air Blank	0.000	17:31																																																																																																																															
Control Test	0.037	17:31																																																																																																																															
Average	0.0363																																																																																																																																
Std Dev	0.0012																																																																																																																																
Rel Std Dev(%)	3.1781																																																																																																																																
Air Blank	0.000	17:42																																																																																																																															
Control Test	0.080	17:43																																																																																																																															
Air Blank	0.000	17:43																																																																																																																															
Control Test	0.080	17:44																																																																																																																															
Air Blank	0.000	17:45																																																																																																																															
Control Test	0.080	17:45																																																																																																																															
Air Blank	0.000	17:46																																																																																																																															
Control Test	0.080	17:46																																																																																																																															
Average	0.0800																																																																																																																																
Std Dev	0.0000																																																																																																																																
Rel Std Dev(%)	0.0000																																																																																																																																
Air Blank	0.000	17:37																																																																																																																															
Control Test	0.197	17:38																																																																																																																															
Air Blank	0.000	17:39																																																																																																																															
Control Test	0.199	17:39																																																																																																																															
Air Blank	0.000	17:40																																																																																																																															
Control Test	0.199	17:41																																																																																																																															
Air Blank	0.000	17:41																																																																																																																															
Control Test	0.000	17:41																																																																																																																															
Average	0.1980																																																																																																																																
Std Dev	0.0010																																																																																																																																
Rel Std Dev(%)	0.5051																																																																																																																																
Air Blank	0.000	17:48																																																																																																																															
Control Test	0.079	17:48																																																																																																																															
Air Blank	0.000	17:49																																																																																																																															
Control Test	0.078	17:49																																																																																																																															
Air Blank	0.000	17:50																																																																																																																															
Control Test	0.0783	17:50																																																																																																																															
Average	0.0783																																																																																																																																
Std Dev	0.0006																																																																																																																																
Rel Std Dev(%)	0.7370																																																																																																																																
<p>Operator Signature</p> 	<p>Operator Signature</p> 	<p>Operator Signature</p> 	<p>Operator Signature</p> 																																																																																																																														

80-001279

6/10/2025 SUT

JACKSONVILLE SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001279  
06/10/2025  
Software: 8100.27

#2

Test	g/210L	Time
Air Blank	0.000	17:33
Control Test	0.041	17:33
Air Blank	0.000	17:34
Control Test	0.042	17:35
Air Blank	0.000	17:35
Control Test	0.043	17:36
Air Blank	0.000	17:36
Control Test Stats		
Average	0.0420	
Std Dev	0.0010	
Rel Std Dev(%)	2.3810	

SUT

checked seal &  
tightened.

- noted - did not  
meet acceptance criteria SUT

  
Operator's Signature



JACKSONVILLE SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001279  
06/16/2025 17:11:15

*SW*

Auto Calibration *Calibration adjustment #2*

pg 1 of 2

<<<<< 3um >>>>> <<<<< 9um >>>>>

-----  
Solution = 0.000 g/210L or 0.0000 mg/l, Samples = 4, Discarded = 1  
Sample % Abs (% Abs Ref) % Abs (% Abs Ref)  
Sample #1 0.1290 (-0.0050) 0.1260 (-0.0070)  
Sample #2 0.1120 (0.0530) 0.1770 (0.3210)  
Sample #3 0.1250 (0.0920) 0.1540 (0.3360)  
Sample #4 0.1330 (0.1300) 0.1960 (0.3310)  
Avg % Abs 0.1233 (0.0917) 0.1757 (0.3293)  
STD DEV 0.0106 (0.0385) 0.0210 (0.0076)  
REL STD DEV 8.594 (42.001) 11.973 (2.319)  
-----

-----  
Solution = 0.040 g/210L or 0.1905 mg/l, Samples = 4, Discarded = 1  
Sample % Abs (% Abs Ref) % Abs (% Abs Ref)  
Sample #1 0.8720 (-0.0230) 1.5700 (-0.0050)  
Sample #2 0.8490 (0.0080) 1.6520 (0.0050)  
Sample #3 0.8380 (0.0430) 1.6330 (0.0240)  
Sample #4 0.8550 (0.0570) 1.5800 (0.0120)  
Avg % Abs 0.8473 (0.0360) 1.6217 (0.0137)  
STD DEV 0.0086 (0.0252) 0.0373 (0.0096)  
REL STD DEV 1.018 (70.108) 2.301 (70.310)  
-----

-----  
Solution = 0.100 g/210L or 0.4762 mg/l, Samples = 4, Discarded = 1  
Sample % Abs (% Abs Ref) % Abs (% Abs Ref)  
Sample #1 1.9530 (-0.0010) 3.7160 (-0.0090)  
Sample #2 1.9330 (0.0300) 3.7170 (0.0200)  
Sample #3 1.9240 (0.0590) 3.6940 (0.0430)  
Sample #4 1.9350 (0.0720) 3.6460 (0.0910)  
Avg % Abs 1.9307 (0.0537) 3.6857 (0.0513)  
STD DEV 0.0059 (0.0215) 0.0362 (0.0362)  
REL STD DEV 0.303 (40.066) 0.983 (70.570)  
-----

-----  
Solution = 0.200 g/210L or 0.9524 mg/l, Samples = 4, Discarded = 1  
Sample % Abs (% Abs Ref) % Abs (% Abs Ref)  
Sample #1 3.7040 (-0.0190) 6.9580 (0.0010)  
Sample #2 3.6600 (0.0200) 6.8610 (-0.0430)  
Sample #3 3.6540 (0.0440) 6.8670 (-0.0320)  
Sample #4 3.6460 (0.0730) 6.8560 (-0.0280)  
Avg % Abs 3.6533 (0.0457) 6.8613 (-0.0343)  
STD DEV 0.0070 (0.0265) 0.0055 (0.0078)  
REL STD DEV 0.192 (58.115) 0.080 (22.624)  
-----

-----  
Solution = 0.300 g/210L or 1.4286 mg/l, Samples = 4, Discarded = 1  
Sample % Abs (% Abs Ref) % Abs (% Abs Ref)  
Sample #1 5.4150 (-0.0340) 10.3060 (-0.0220)  
Sample #2 5.3390 (0.0600) 9.8960 (0.2320)  
Sample #3 5.3430 (0.0640) 9.9540 (0.2080)  
Sample #4 5.3370 (0.0700) 9.8960 (0.3010)  
Avg % Abs 5.3397 (0.0647) 9.9153 (0.2470)  
STD DEV 0.0031 (0.0050) 0.0335 (0.0483)  
REL STD DEV 0.057 (7.783) 0.338 (19.547)  
-----

JACKSONVILLE SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001279  
06/16/2025 17:11:15

SW

Auto Calibration

pg 2 of 2

<<<< 3um >>>>  
-----  
Zero Order Coef -319.52  
First Order Coef 2589.83  
Second Order Coef 27.40  
-----

Act (g/210L)	Fit (g/210L)	Residual (g/210L)
0.000	0.000	-0.0000
0.040	0.040	0.0002
0.100	0.100	-0.0004
0.200	0.200	0.0003
0.300	0.300	-0.0001

<<<< 9um >>>>  
-----  
-247.52  
1308.50  
16.00  
-----

Act (g/210L)	Fit (g/210L)	Residual (g/210L)
0.000	-0.000	0.0004
0.040	0.040	-0.0002
0.100	0.101	-0.0006
0.200	0.199	0.0008
0.300	0.300	-0.0003

<<<< 3um >>>> <<<< 9um >>>>  
-----  
Solution = 0.080 g/210L or 0.3810 mg/l, Samples = 4, Discarded = 1  
Sample  
Sample #1 3155.00 3274.00  
Sample #2 3056.00 3307.00  
Sample #3 3088.00 3314.00  
Sample #4 3109.00 3343.00  
Avg 3084.3333 3321.3333  
STD DEV 26.6896 19.0875  
REL STD DEV 0.865 0.575  
H2O adjust (mg/l\*10k) 725 488

Barometric Pressure = 1012


\*\*\*\*\*CALIBRATION SUCCESSFUL\*\*\*\*\*



JACKSONVILLE SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000  
06/16/2025  
Software: 8100.27

stabilizes  
post calibration^  
6/16/2025  
Sut

Test	g/210L	Time
Air Blank	0.000	18:17
Control Test	0.049	18:17
Air Blank	0.000	18:18
Control Test	0.043 —	18:19
Air Blank	0.000	18:19
Control Test	0.046 —	18:20
Air Blank	0.000	18:20
Control Test Stats		
Average	0.0460	
Std Dev	0.0030	
Rel Std Dev(%)	6.5217	


  
Operator's Signature

Root cause analysis - checked seal on Simulator and was good. The connection with the calibration inlet didn't seal was not good. Sut

JACKSONVILLE SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001279  
06/16/2025  
Software: 8100.27

#2

Test	g/210L	Time
Air Blank	0.000	18:28
Control Test	0.051	18:29
Air Blank	0.000	18:30
Control Test	0.050	18:30
Air Blank	0.000	18:31
Control Test	0.047	18:31
Air Blank	0.000	18:32
Control Test Stats		
Average	0.0493	
Std Dev	0.0021	
Rel Std Dev(%)	4.2196	



Operator's Signature

JACKSONVILLE SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001279  
06/16/2025  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	18:33
Control Test	0.082	18:34
Air Blank	0.000	18:34
Control Test	0.083	18:35
Air Blank	0.000	18:36
Control Test	0.081	18:36
Air Blank	0.000	18:37
Control Test Stats		
Average	0.0820 ✓	
Std Dev	0.0010	
Rel Std Dev(%)	1.2195	

  
Operator's Signature

JACKSONVILLE SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001279  
06/16/2025  
Software: 8100.27

DGS

Test	g/210L	Time
Air Blank	0.000	18:23
Control Test	0.079	18:23
Air Blank	0.000	18:24
Control Test	0.079	18:24
Air Blank	0.000	18:25
Control Test	0.080	18:25
Air Blank	0.000	18:25
Control Test Stats		
Average	0.0793 ✓	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

  
Operator's Signature



JACKSONVILLE SO  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-001279  
06/16/2025  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	18:40
Control Test	0.202	18:41
Air Blank	0.000	18:41
Control Test	0.202	18:42
Air Blank	0.000	18:42
Control Test	0.202	18:43
Air Blank	0.000	18:44
Control Test Stats		
Average	0.2020	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

  
Operator's Signature

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: JACKSONVILLE SO  
Time of Inspection: 16:45

Date of Inspection: 06/17/2025

Serial Number: 80-001279  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG429602 Exp: 10/22/2026
0.000	0.046	0.077	0.198	0.081
0.000	0.047	0.077	0.199	0.081
0.000	0.047	0.077	0.199	0.081
0.000	0.047	0.077	0.199	0.081
0.000	0.047	0.077	0.199	0.081
0.000	0.047	0.077	0.199	0.081
0.000	0.047	0.077	0.199	0.081
0.000	0.047	0.077	0.199	0.081
0.000	0.047	0.077	0.199	0.081
0.000	0.047	0.078	0.199	0.081


Standard Deviations	0.0003	0.0003	0.0003	0.0000
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002 Number of Simulators Used: 5

Remarks:

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

  
LEANDRA HIGGINBOTHAM  
Signature and Printed Name

06/17/2025  
Date



# Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
2331 Phillips Road  
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001279, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001279</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>JACKSONVILLE SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>06/17/2025</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>16:45</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within  $\pm 0.005$  or 5%, whichever is greater, of the target alcohol concentration.

\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

## TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full,

without written approval of the Florida Department of

Law Enforcement Alcohol Testing Program.

06/17/2025

Date

LEANDRA HIGGINBOTHAM,

Department Inspector

FDLE/ATP Form 69 October 2024

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality