



INSTRUMENT PROCESSING SHEET

Agency Union County SOS/N 80-001288Florida Department of
Law EnforcementDate In 12/04/2024 DI Completion Date 02/06/2025☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By <u>ALL</u>	Date <u>12/04/2024</u>	Quality Checks	By <u>ALL/DA</u>	Date <u>12/04/2024/</u> <u>02/05/2025</u>	Flow Calibration	By _____	Date _____																									
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: <u>dropoff, no box</u>			☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>301/199</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP-103 /ATP-102</u> 32 mm <u>.000/0.148</u> (.139 - .169) 36 mm <u>.000/0.168</u> (.156 - .190) 53 mm <u>.082/0.234</u> (.228 - .278) 103 mm <u>.429/0.514</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28421/30793</u> ☒ Stability Checks			Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																											
			<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6291</td><td>202406K 06/19/2026</td></tr><tr><td>0.080</td><td>MP6292</td><td>202406L 06/19/2026</td></tr><tr><td>0.200</td><td>MP6293</td><td>202406N 06/20/2026</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG310901/see notes 04/19/2025/see notes</td></tr></tbody></table>			Simulator	Serial #	Lot #/Exp	0.050	MP6291	202406K 06/19/2026	0.080	MP6292	202406L 06/19/2026	0.200	MP6293	202406N 06/20/2026	0.080 DGS	N/A	AG310901/see notes 04/19/2025/see notes	<table border="1"><thead><tr><th colspan="2">Maintenance By _____ Date _____</th></tr><tr><td>☐ Battery Replacement</td><td></td></tr><tr><td>☐ Dry Gas Regulator Replacement</td><td></td></tr><tr><td>☐ Breath Tube Replacement</td><td></td></tr><tr><td>☐ Other _____</td><td></td></tr></thead></table>			Maintenance By _____ Date _____		☐ Battery Replacement		☐ Dry Gas Regulator Replacement		☐ Breath Tube Replacement		☐ Other _____	
Simulator	Serial #	Lot #/Exp																															
0.050	MP6291	202406K 06/19/2026																															
0.080	MP6292	202406L 06/19/2026																															
0.200	MP6293	202406N 06/20/2026																															
0.080 DGS	N/A	AG310901/see notes 04/19/2025/see notes																															
Maintenance By _____ Date _____																																	
☐ Battery Replacement																																	
☐ Dry Gas Regulator Replacement																																	
☐ Breath Tube Replacement																																	
☐ Other _____																																	

Calibration Adjustment	By _____	Department Inspection	By <u>DA</u>																																																																
Barometric Pressure Gauge _____ ID # _____ <table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table> ☐ Post Calibration Adjustment Stability Checks <table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Barometric Pressure ID# <u>30793</u> Gauge <u>1019/1019</u> Instrument <u>1018/1018</u> Mouth Alcohol Solution Lot # <u>2024-A</u> Acetone Stock Solution Lot # <u>2024-B</u> <table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP5086</td></tr><tr><td>Interferent</td><td>MP5087</td></tr><tr><td>0.050</td><td>MP5088</td></tr><tr><td>0.080</td><td>MP5089</td></tr><tr><td>0.200</td><td>MP5090</td></tr></tbody></table> <table border="1"><thead><tr><th colspan="2">Attachments</th></tr><tr><td>☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment</td><td>☐ Post-Stability Checks ☐ Flow Calibration ☒ Form 40 ☒ Other <u>Stability checks</u></td></tr></thead></table> ☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use Shayla Platt Digitally signed by Shayla Platt Date: 2025.03.05 17:54:06 -05'00' Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2025.03.07 13:08:00 -05'00' Tech Review / Date _____ Admin Review / Date _____		Simulator	Serial Number	0.000	MP5086	Interferent	MP5087	0.050	MP5088	0.080	MP5089	0.200	MP5090	Attachments		☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment	☐ Post-Stability Checks ☐ Flow Calibration ☒ Form 40 ☒ Other <u>Stability checks</u>
Simulator	Serial #	Lot #	Expiration																																																																
0.000		N/A	N/A																																																																
0.040																																																																			
0.100																																																																			
0.200																																																																			
0.300																																																																			
0.080 DGS	N/A																																																																		
Simulator	Serial #	Lot #	Expiration																																																																
0.050																																																																			
0.080																																																																			
0.200																																																																			
0.080 DGS	N/A																																																																		
Simulator	Serial Number																																																																		
0.000	MP5086																																																																		
Interferent	MP5087																																																																		
0.050	MP5088																																																																		
0.080	MP5089																																																																		
0.200	MP5090																																																																		
Attachments																																																																			
☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment	☐ Post-Stability Checks ☐ Flow Calibration ☒ Form 40 ☒ Other <u>Stability checks</u>																																																																		
Notes/Suggested Service: First .050 stability was outside of acceptable range, reran stability with similar results. 12/04/2024 ALL First .080 DGS produced all zeros, ensured connections were secure and reran stability with same results. 12/04/2024 Added lot numbers and expirations dates for .200 and DGS 01/08/2025 ALL Repeated quality checks performed on 12/04/2024 to ensure instrument																																																																			

is operating nominally. All results within range. All lot numbers for alcohol reference solutions were the same as noted by ALL. Lot number for the DGS used was: AG429602 with an expiration date of 10/22/2026. DA 02/05/2025. Power cord came loose during department inspection and instrument turned off. Ensured proper connection and repeated inspection. DA 02/06/2025.

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: UNION COUNTY SO

Time of Inspection: 11:48

Date of Inspection: 12/04/2024

Serial Number: 80-001288

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08g/210L Test (g/210L) Lot#:_____ Exp:_____	0.20g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:_____ Exp:_____

Number of Simulators Used: 0

Remarks:

BYPASS AI FOR INSPECTION COMPLIANCE NOT DETERMINED

12/04/2024 *AJ*

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Alfonso Lowry

ALFONSO L LOWRY

Signature and Printed Name

12/04/2024
Date

Stability checks

UNION COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001268
12/04/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:59
Control Test	0.043	15:00
Air Blank	0.000	15:00
Control Test	0.044	15:01
Air Blank	0.000	15:02
Control Test	0.043	15:02
Air Blank	0.000	15:03
Control Test Stats		
Average	0.0433	
Std Dev	0.0006	
Rel Std Dev(%)	1.3323	

Operator's Signature

UNION COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001268
12/04/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:04
Control Test	0.043	15:04
Air Blank	0.000	15:05
Control Test	0.044	15:05
Air Blank	0.000	15:06
Control Test	0.044	15:07
Air Blank	0.000	15:07
Control Test Stats		
Average	0.0437	
Std Dev	0.0006	
Rel Std Dev(%)	1.3222	

Operator's Signature

UNION COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001268
12/04/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:08
Control Test	0.080	15:09
Air Blank	0.000	15:09
Control Test	0.080	15:10
Air Blank	0.000	15:10
Control Test	0.080	15:11
Air Blank	0.000	15:12
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

UNION COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001268
12/04/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:13
Control Test	0.198	15:14
Air Blank	0.000	15:14
Control Test	0.198	15:15
Air Blank	0.000	15:15
Control Test	0.198	15:16
Air Blank	0.000	15:17
Control Test Stats		
Average	0.1980	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

UNION COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001268
12/04/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:18
Control Test	0.000	15:19
Air Blank	0.000	15:19
Control Test	0.000	15:20
Air Blank	0.000	15:20
Control Test	0.000	15:21
Air Blank	0.000	15:21
Control Test Stats		
Average	0.0000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

UNION COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001268
12/04/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:22
Control Test	0.000	15:22
Air Blank	0.000	15:22
Control Test	0.000	15:23
Air Blank	0.000	15:23
Control Test	0.000	15:24
Air Blank	0.000	15:24
Control Test Stats		
Average	0.0000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

Performed discretionary stability checks on instrument to assess calibration status 2/25/2025 PN

UNION COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001288
12/20/2024
Software: 8100.27

Test	g/21.0L	Time
Air Blank	0.000	08:39
Control Test	0.079	08:40
Air Blank	0.000	08:40
Control Test	0.080	08:41
Air Blank	0.000	08:41
Control Test	0.079	08:41
Air Blank	0.000	08:42
Control Test	0.079	08:42
Air Blank	0.000	08:43
Control Test	0.080	08:43
Air Blank	0.000	08:43
Control Test	0.079	08:44
Air Blank	0.000	08:44
Control Test	0.080	08:45
Air Blank	0.000	08:45
Control Test	0.079	08:45
Air Blank	0.000	08:46
Control Test	0.080	08:46
Air Blank	0.000	08:47
Control Test	0.080	08:47
Air Blank	0.000	08:48
Control Test Stats		
Average	0.0795	
Std Dev	0.0005	
Rel Std Dev(%)	0.6629	

D65



Operator's Signature

ADDITIONAL . D65
STABILITY CHECKS

12/20/2024



Performed discretionary stability checks on instrument to assess calibration status 2/25/2025 PN

UNION COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001288
12/20/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:20
Control Test	0.048	10:20
Air Blank	0.000	10:21
Control Test	0.048	10:21
Air Blank	0.000	10:22
Control Test	0.048	10:23
Air Blank	0.000	10:23
Control Test	0.048	10:24
Air Blank	0.000	10:24
Control Test	0.048	10:25
Air Blank	0.000	10:26
Control Test	0.048	10:26
Air Blank	0.000	10:27
Control Test	0.048	10:27
Air Blank	0.000	10:28
Control Test	0.049	10:29
Air Blank	0.000	10:29
Control Test	0.049	10:30
Air Blank	0.000	10:30
Control Test	0.048	10:31
Air Blank	0.000	10:31
Control Test Stats		
Average	0.0482	
Std Dev	0.0014	
Rel Std Dev(%)	0.8748	

ADDITIONAL 0.05 g/210L
STABILITY CHECKS

12/20/2024

LOT 202406K
EXP 6/19/2026

MP 5088

Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: UNION COUNTY SO

Serial Number: 80-001288

Time of Inspection: 13:37

Date of Inspection: 02/05/2025

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot #: Exp:

Number of Simulators Used: _____

Remarks:

BYPASS AI TO OPERATE INSTRUMENT COMPLIANCE NOT DETERMINED

The above instrument complies (N/A) does not comply (DA 2/5/25) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

DA

DESTINEE N ARMSTRONG

Signature and Printed Name


02/05/2025

Date

Stability Checks 80-001288 DA 2/5/25

UNION COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001288
02/05/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:50
Control Test	0.049	13:51
Air Blank	0.000	13:51
Control Test	0.048	13:52
Air Blank	0.000	13:52
Control Test	0.049	13:53
Air Blank	0.000	13:53
Control Test Stats		
Average	0.0487	
Std Dev	0.0006	
Rel Std Dev(%)	1.1863	



Operator's Signature

Repeated stability checks performed on 12/4/24
to ensure instrument is functioning nominally.

DA 2/5/25

UNION COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001288
02/15/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:45
Control Test	0.077	13:46
Air Blank	0.000	13:46
Control Test	0.078	13:47
Air Blank	0.000	13:48
Control Test	0.078	13:48
Air Blank	0.000	13:49
Control Test Stats		
Average	0.0777	
Std Dev	0.0006	
Rel Std Dev(%)	0.7434	

wet

DA

Operator's Signature

External printer came unplugged during
stability test. DA 2/5/25

UNION COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001288
02/05/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:39
Control Test	0.195	13:40
Air Blank	0.000	13:40
Control Test	0.195	13:41
Air Blank	0.000	13:41
Control Test	0.195	13:42
Air Blank	0.000	13:43
Control Test Stats		
Average	0.1950	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	



Operator's Signature

UNION COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001288
02/05/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:56
Control Test	0.078	13:57
Air Blank	0.000	13:57
Control Test	0.079	13:57
Air Blank	0.000	13:58
Control Test	0.079	13:58
Air Blank	0.000	13:59
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

DGS

DA

Operator's Signature

Lot # of DGS used: AG429602

Expiration date: 10/22/2024

DA 2/5/25

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: UNION COUNTY SO
Time of Inspection: 11:17

Date of Inspection: 02/06/2025

Serial Number: 80-001288
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG429602 Exp: 10/22/2026
0.000	0.047	0.078	0.197	0.078
0.000	0.047	0.078	0.197	0.078
0.000	0.047	0.078	0.197	0.078
0.000	0.047	0.078	0.197	0.079
0.000	0.047	0.078	0.197	0.078
0.000	0.047	0.078	0.197	0.078
0.000	0.047	0.079	0.197	0.078
0.000	0.047	0.078	0.197	0.078
0.000	0.047	0.078	0.197	0.078
0.000	0.047	0.078	0.197	0.078

Standard Deviations	0.0000	0.0003	0.0000	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0001 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.


I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



DESTINEE N ARMSTRONG

Signature and Printed Name

02/06/2025
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001288, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001288</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>UNION COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>02/06/2025</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>11:17</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

02/06/2025

Date

DESTINEE N ARMSTRONG,
Department Inspector

FDLE/ATP Form 69 October 2024

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1