



INSTRUMENT PROCESSING SHEET

Agency LEVY COUNTY SOS/N 80-001292Florida Department of
Law EnforcementDate In 6/30/25DI Completion Date 07/09/2025☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By KTS	Date <u>6/30/25</u>	Quality Checks	By KTS	Date <u>7/2/25</u>	Flow Calibration	By	Date																									
☒ Annual ☐ Registration ☒ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☒ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: <u>Intoxilyzer box came inside an additional box for extra support.</u>			☒ Breath Tube Screen <u>7/1/25</u> ☒ Replace External O-Rings <u>KTS 7/16/25</u> ☒ Instrument Set Up Verified ☒ R-Value <u>197</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP102</u> 32 mm <u>0.156</u> (.139 - .169) 36 mm <u>0.179</u> (.156 - .190) 53 mm <u>0.242</u> (.228 - .278) 103 mm <u>0.496</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28421</u> ☒ Stability Checks			Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																											
			<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP5088</td><td>202406K 6/19/26</td></tr><tr><td>0.080</td><td>MP5089</td><td>202406L 6/19/26</td></tr><tr><td>0.200</td><td>MP5090</td><td>202406N 6/20/26</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG429602 10/22/26</td></tr></tbody></table>			Simulator	Serial #	Lot #/Exp	0.050	MP5088	202406K 6/19/26	0.080	MP5089	202406L 6/19/26	0.200	MP5090	202406N 6/20/26	0.080 DGS	N/A	AG429602 10/22/26	<table border="1"><thead><tr><th colspan="2">Maintenance By _____ Date _____</th></tr><tr><td>☐ Battery Replacement</td><td></td></tr><tr><td>☐ Dry Gas Regulator Replacement</td><td></td></tr><tr><td>☐ Breath Tube Replacement</td><td></td></tr><tr><td>☐ Other _____</td><td></td></tr></thead></table>			Maintenance By _____ Date _____		☐ Battery Replacement		☐ Dry Gas Regulator Replacement		☐ Breath Tube Replacement		☐ Other _____	
Simulator	Serial #	Lot #/Exp																															
0.050	MP5088	202406K 6/19/26																															
0.080	MP5089	202406L 6/19/26																															
0.200	MP5090	202406N 6/20/26																															
0.080 DGS	N/A	AG429602 10/22/26																															
Maintenance By _____ Date _____																																	
☐ Battery Replacement																																	
☐ Dry Gas Regulator Replacement																																	
☐ Breath Tube Replacement																																	
☐ Other _____																																	
Calibration Adjustment By <u>SLH</u>			Department Inspection By <u>SLH</u>																														
Barometric Pressure Gauge <u>1013</u> ID # <u>28427</u>			Barometric Pressure ID# <u>28662</u>																														
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td>MP6294</td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td>MP6295</td><td>24330</td><td>09/10/2026</td></tr><tr><td>0.100</td><td>MP6296</td><td>24390</td><td>10/29/2026</td></tr><tr><td>0.200</td><td>MP6297</td><td>24080</td><td>02/13/2026</td></tr><tr><td>0.300</td><td>MP6298</td><td>24430</td><td>12/10/2026</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>28424080A3</td><td>11/05/2026</td></tr></tbody></table>			Simulator	Serial #	Lot #	Expiration	0.000	MP6294	N/A	N/A	0.040	MP6295	24330	09/10/2026	0.100	MP6296	24390	10/29/2026	0.200	MP6297	24080	02/13/2026	0.300	MP6298	24430	12/10/2026	0.080 DGS	N/A	28424080A3	11/05/2026	Gauge <u>1016</u> Instrument <u>1014</u> Mouth Alcohol Solution Lot # <u>2025-A</u> Acetone Stock Solution Lot # <u>2024-B</u>		
Simulator	Serial #	Lot #	Expiration																														
0.000	MP6294	N/A	N/A																														
0.040	MP6295	24330	09/10/2026																														
0.100	MP6296	24390	10/29/2026																														
0.200	MP6297	24080	02/13/2026																														
0.300	MP6298	24430	12/10/2026																														
0.080 DGS	N/A	28424080A3	11/05/2026																														
☒ Post Calibration Adjustment Stability Checks			<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6294</td></tr><tr><td>Interferent</td><td>MP5087</td></tr><tr><td>0.050</td><td>MP5088</td></tr><tr><td>0.080</td><td>MP5089</td></tr><tr><td>0.200</td><td>MP5090</td></tr></tbody></table>			Simulator	Serial Number	0.000	MP6294	Interferent	MP5087	0.050	MP5088	0.080	MP5089	0.200	MP5090																
Simulator	Serial Number																																
0.000	MP6294																																
Interferent	MP5087																																
0.050	MP5088																																
0.080	MP5089																																
0.200	MP5090																																
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td>MP5088</td><td>202406K</td><td>06/19/2026</td></tr><tr><td>0.080</td><td>MP5089</td><td>202406L</td><td>06/19/2026</td></tr><tr><td>0.200</td><td>MP5090</td><td>202406N</td><td>06/20/2026</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG429602</td><td>10/22/2026</td></tr></tbody></table>			Simulator	Serial #	Lot #	Expiration	0.050	MP5088	202406K	06/19/2026	0.080	MP5089	202406L	06/19/2026	0.200	MP5090	202406N	06/20/2026	0.080 DGS	N/A	AG429602	10/22/2026	<table border="1"><thead><tr><th colspan="2">Attachments</th></tr><tr><td>☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Calibration Adjustment</td><td>☒ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____</td></tr></thead></table>			Attachments		☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Calibration Adjustment	☒ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____				
Simulator	Serial #	Lot #	Expiration																														
0.050	MP5088	202406K	06/19/2026																														
0.080	MP5089	202406L	06/19/2026																														
0.200	MP5090	202406N	06/20/2026																														
0.080 DGS	N/A	AG429602	10/22/2026																														
Attachments																																	
☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Calibration Adjustment	☒ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____																																
Notes/Suggested Service: <u>Root cause analysis performed. Equipment or user error not detected. KTS 7/2/25 7/1/25 KTS 7/16/25</u>			☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																														
Tech Review: <u>Corrected Quality Check/Note dates. KTS 7/16/25</u>			Shayla Platt Digitally signed by Shayla Platt Date: 2025.07.16 14:32:50 -04'00'																														
Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2025.07.16 14:04:54 -04'00'			Tech Review / Date Admin Review / Date																														

Stability Checks

0.050 g/210L	0.080 g/210L	0.200 g/210L	DGS 0.080 g/210L																																																																																																																																																
0.047 to 0.053 g/210L ☒	0.077 to 0.083 g/210L ☒	0.194 to 0.206 g/210L ☒	0.077 to 0.083 g/210L ☒ ≤0.003 g/210L of Wet ☒																																																																																																																																																
Performed Root Case Analysis ☒	Performed Root Case Analysis ☒	Performed Root Case Analysis ☒	Performed Root Case Analysis ☒																																																																																																																																																
LEUW COUNTY SO Intoxilyzer - Alcohol Analyzer Model: 8000 SN 60-001292 07/01/2025 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>08:55</td></tr> <tr><td>Control Test</td><td>0.049</td><td>08:56</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:56</td></tr> <tr><td>Control Test</td><td>0.049</td><td>08:57</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:57</td></tr> <tr><td>Control Test</td><td>0.050</td><td>08:58</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:58</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0493</td><td></td></tr> <tr><td>Std Dev</td><td>0.0016</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1703</td><td></td></tr> </tbody> </table> Kattis Operator's Signature	Test	g/210L	Time	Air Blank	0.000	08:55	Control Test	0.049	08:56	Air Blank	0.000	08:56	Control Test	0.049	08:57	Air Blank	0.000	08:57	Control Test	0.050	08:58	Air Blank	0.000	08:58	Control Test Stats			Average	0.0493		Std Dev	0.0016		Rel Std Dev(%)	1.1703		LEUW COUNTY SO Intoxilyzer - Alcohol Analyzer Model: 8000 SN 60-001292 07/01/2025 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>09:07</td></tr> <tr><td>Control Test</td><td>0.079</td><td>09:08</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:08</td></tr> <tr><td>Control Test</td><td>0.079</td><td>09:09</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:09</td></tr> <tr><td>Control Test</td><td>0.078</td><td>09:10</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:11</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0787</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7339</td><td></td></tr> </tbody> </table> Kattis Operator's Signature	Test	g/210L	Time	Air Blank	0.000	09:07	Control Test	0.079	09:08	Air Blank	0.000	09:08	Control Test	0.079	09:09	Air Blank	0.000	09:09	Control Test	0.078	09:10	Air Blank	0.000	09:11	Control Test Stats			Average	0.0787		Std Dev	0.0006		Rel Std Dev(%)	0.7339		LEUW COUNTY SO Intoxilyzer - Alcohol Analyzer Model: 8000 SN 60-001292 07/01/2025 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>09:02</td></tr> <tr><td>Control Test</td><td>0.198</td><td>09:03</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:03</td></tr> <tr><td>Control Test</td><td>0.198</td><td>09:04</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:04</td></tr> <tr><td>Control Test</td><td>0.198</td><td>09:05</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:06</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1980</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Kattis Operator's Signature	Test	g/210L	Time	Air Blank	0.000	09:02	Control Test	0.198	09:03	Air Blank	0.000	09:03	Control Test	0.198	09:04	Air Blank	0.000	09:04	Control Test	0.198	09:05	Air Blank	0.000	09:06	Control Test Stats			Average	0.1980		Std Dev	0.0000		Rel Std Dev(%)	0.0000		LEUW COUNTY SO Intoxilyzer - Alcohol Analyzer Model: 8000 SN 60-001292 07/01/2025 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>08:47</td></tr> <tr><td>Control Test</td><td>0.076</td><td>08:48</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:48</td></tr> <tr><td>Control Test</td><td>0.076</td><td>08:49</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:49</td></tr> <tr><td>Control Test</td><td>0.076</td><td>08:49</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>08:50</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0760</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Kattis Operator's Signature	Test	g/210L	Time	Air Blank	0.000	08:47	Control Test	0.076	08:48	Air Blank	0.000	08:48	Control Test	0.076	08:49	Air Blank	0.000	08:49	Control Test	0.076	08:49	Air Blank	0.000	08:50	Control Test Stats			Average	0.0760		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	08:55																																																																																																																																																	
Control Test	0.049	08:56																																																																																																																																																	
Air Blank	0.000	08:56																																																																																																																																																	
Control Test	0.049	08:57																																																																																																																																																	
Air Blank	0.000	08:57																																																																																																																																																	
Control Test	0.050	08:58																																																																																																																																																	
Air Blank	0.000	08:58																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0493																																																																																																																																																		
Std Dev	0.0016																																																																																																																																																		
Rel Std Dev(%)	1.1703																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	09:07																																																																																																																																																	
Control Test	0.079	09:08																																																																																																																																																	
Air Blank	0.000	09:08																																																																																																																																																	
Control Test	0.079	09:09																																																																																																																																																	
Air Blank	0.000	09:09																																																																																																																																																	
Control Test	0.078	09:10																																																																																																																																																	
Air Blank	0.000	09:11																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0787																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7339																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	09:02																																																																																																																																																	
Control Test	0.198	09:03																																																																																																																																																	
Air Blank	0.000	09:03																																																																																																																																																	
Control Test	0.198	09:04																																																																																																																																																	
Air Blank	0.000	09:04																																																																																																																																																	
Control Test	0.198	09:05																																																																																																																																																	
Air Blank	0.000	09:06																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1980																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	08:47																																																																																																																																																	
Control Test	0.076	08:48																																																																																																																																																	
Air Blank	0.000	08:48																																																																																																																																																	
Control Test	0.076	08:49																																																																																																																																																	
Air Blank	0.000	08:49																																																																																																																																																	
Control Test	0.076	08:49																																																																																																																																																	
Air Blank	0.000	08:50																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0760																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		

LEVY COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001292
07/07/2025 14:00:44

Sub

Auto Calibration

pg 1 of 2

<<<<< 3um >>>>> <<<<< 9um >>>>>

Solution = 0.000 g/210L or 0.0000 mg/l, Samples = 4, Discarded = 1
Sample % Abs (% Abs Ref) % Abs (% Abs Ref)
Sample #1 0.1740 (-0.0080) 0.2280 (0.0000)
Sample #2 0.1390 (0.0670) 0.2210 (0.0220)
Sample #3 0.1110 (0.1110) 0.2010 (0.0470)
Sample #4 0.0910 (0.1390) 0.2140 (0.0530)
Avg % Abs 0.1137 (0.1057) 0.2120 (0.0407)
STD DEV 0.0241 (0.0363) 0.0101 (0.0164)
REL STD DEV 21.212 (34.349) 4.787 (40.431)

Solution = 0.040 g/210L or 0.1905 mg/l, Samples = 4, Discarded = 1
Sample % Abs (% Abs Ref) % Abs (% Abs Ref)
Sample #1 0.8740 (-0.0170) 1.5290 (-0.0090)
Sample #2 0.8020 (0.0260) 1.4960 (0.0180)
Sample #3 0.8410 (0.0110) 1.5190 (0.0090)
Sample #4 0.8090 (0.0430) 1.5050 (0.0150)
Avg % Abs 0.8173 (0.0267) 1.5067 (0.0140)
STD DEV 0.0208 (0.0160) 0.0116 (0.0046)
REL STD DEV 2.544 (60.039) 0.769 (32.733)

Solution = 0.100 g/210L or 0.4762 mg/l, Samples = 4, Discarded = 1
Sample % Abs (% Abs Ref) % Abs (% Abs Ref)
Sample #1 1.8610 (-0.0040) 3.3970 (-0.0080)
Sample #2 1.8740 (-0.0090) 3.3920 (0.0030)
Sample #3 1.8580 (0.0150) 3.3940 (0.0170)
Sample #4 1.8970 (0.0080) 3.3990 (0.0060)
Avg % Abs 1.8763 (0.0047) 3.3950 (0.0087)
STD DEV 0.0196 (0.0123) 0.0036 (0.0074)
REL STD DEV 1.045 (264.479) 0.106 (85.051)

Solution = 0.200 g/210L or 0.9524 mg/l, Samples = 4, Discarded = 1
Sample % Abs (% Abs Ref) % Abs (% Abs Ref)
Sample #1 3.6260 (-0.0140) 6.4920 (-0.0150)
Sample #2 3.6240 (-0.0100) 6.4800 (-0.0020)
Sample #3 3.5730 (0.0310) 6.4560 (0.0140)
Sample #4 3.6060 (0.0220) 6.4490 (0.0130)
Avg % Abs 3.6010 (0.0143) 6.4617 (0.0083)
STD DEV 0.0259 (0.0215) 0.0163 (0.0090)
REL STD DEV 0.718 (150.338) 0.252 (107.555)

Solution = 0.300 g/210L or 1.4286 mg/l, Samples = 4, Discarded = 1
Sample % Abs (% Abs Ref) % Abs (% Abs Ref)
Sample #1 5.3180 (-0.0150) 9.4130 (0.0000)
Sample #2 5.2470 (0.0400) 9.3560 (0.0420)
Sample #3 5.2620 (0.0350) 9.3710 (0.0320)
Sample #4 5.2530 (0.0500) 9.3650 (0.0400)
Avg % Abs 5.2540 (0.0417) 9.3640 (0.0380)
STD DEV 0.0075 (0.0076) 0.0075 (0.0053)
REL STD DEV 0.144 (18.330) 0.081 (13.925)

LEVY COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001292
07/07/2025 14:00:44

SW

Auto Calibration

pg 2 of 2

<<<<< 3um >>>>>

Zero Order Coef -284.82
First Order Coef 2636.31
Second Order Coef 25.79

<<<<< 9um >>>>>

-302.44
1447.27
11.74

Act Fit Residual
(g/210L) (g/210L) (g/210L)
0.000 0.000 -0.0003
0.040 0.040 0.0004
0.100 0.100 0.0002
0.200 0.200 -0.0004
0.300 0.300 0.0002

Act Fit Residual
(g/210L) (g/210L) (g/210L)
0.000 0.000 -0.0001
0.040 0.040 -0.0000
0.100 0.100 0.0003
0.200 0.200 -0.0003
0.300 0.300 0.0001

<<<<< 3um >>>>> <<<<< 9um >>>>>

Solution = 0.080 g/210L or 0.3810 mg/l, Samples = 4, Discarded = 1
Sample
Sample #1 2962.00 3242.00
Sample #2 2859.00 3212.00
Sample #3 2861.00 3221.00
Sample #4 2824.00 3232.00
Avg 2848.0000 3221.6667
STD DEV 20.8087 10.0167
REL STD DEV 0.731 0.311
H2O adjust (mg/l*10k) 961 588

Barometric Pressure = 1012

*****CALIBRATION SUCCESSFUL*****

0.050 g/210L	0.080 g/210L	0.200 g/210L	DGS 0.080 g/210L																																																																																																																																																
0.047 to 0.053 g/210L	0.077 to 0.083 g/210L	0.194 to 0.206 g/210L	0.077 to 0.083 g/210L 50.003 g/210L of Wet																																																																																																																																																
Performed Root Case Analysis	Performed Root Case Analysis	Performed Root Case Analysis	Performed Root Case Analysis																																																																																																																																																
LEUY COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 07/07/2025 Software: 8100.27	LEUY COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 07/07/2025 Software: 8100.27	LEUY COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 07/07/2025 Software: 8100.27	LEUY COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 07/07/2025 Software: 8100.27																																																																																																																																																
<table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>16:51</td></tr><tr><td>Control Test</td><td>0.050</td><td>16:51</td></tr><tr><td>Air Blank</td><td>0.000</td><td>16:52</td></tr><tr><td>Control Test</td><td>0.049</td><td>16:52</td></tr><tr><td>Air Blank</td><td>0.000</td><td>16:53</td></tr><tr><td>Control Test</td><td>0.049</td><td>16:54</td></tr><tr><td>Air Blank</td><td>0.000</td><td>16:54</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0493</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>1.1713</td><td></td></tr></table>	Test	g/210L	Time	Air Blank	0.000	16:51	Control Test	0.050	16:51	Air Blank	0.000	16:52	Control Test	0.049	16:52	Air Blank	0.000	16:53	Control Test	0.049	16:54	Air Blank	0.000	16:54	Control Test Stats			Average	0.0493		Std Dev	0.0006		Rel Std Dev(%)	1.1713		<table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>16:39</td></tr><tr><td>Control Test</td><td>0.079</td><td>16:39</td></tr><tr><td>Air Blank</td><td>0.000</td><td>16:40</td></tr><tr><td>Control Test</td><td>0.080</td><td>16:41</td></tr><tr><td>Air Blank</td><td>0.000</td><td>16:41</td></tr><tr><td>Control Test</td><td>0.080</td><td>16:42</td></tr><tr><td>Air Blank</td><td>0.000</td><td>16:42</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0797</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7247</td><td></td></tr></table>	Test	g/210L	Time	Air Blank	0.000	16:39	Control Test	0.079	16:39	Air Blank	0.000	16:40	Control Test	0.080	16:41	Air Blank	0.000	16:41	Control Test	0.080	16:42	Air Blank	0.000	16:42	Control Test Stats			Average	0.0797		Std Dev	0.0006		Rel Std Dev(%)	0.7247		<table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>16:45</td></tr><tr><td>Control Test</td><td>0.202</td><td>16:46</td></tr><tr><td>Air Blank</td><td>0.000</td><td>16:46</td></tr><tr><td>Control Test</td><td>0.200</td><td>16:47</td></tr><tr><td>Air Blank</td><td>0.000</td><td>16:48</td></tr><tr><td>Control Test</td><td>0.199</td><td>16:48</td></tr><tr><td>Air Blank</td><td>0.000</td><td>16:49</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.2003</td><td></td></tr><tr><td>Std Dev</td><td>0.0015</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7625</td><td></td></tr></table>	Test	g/210L	Time	Air Blank	0.000	16:45	Control Test	0.202	16:46	Air Blank	0.000	16:46	Control Test	0.200	16:47	Air Blank	0.000	16:48	Control Test	0.199	16:48	Air Blank	0.000	16:49	Control Test Stats			Average	0.2003		Std Dev	0.0015		Rel Std Dev(%)	0.7625		<table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>15:47</td></tr><tr><td>Control Test</td><td>0.078</td><td>15:48</td></tr><tr><td>Air Blank</td><td>0.000</td><td>15:48</td></tr><tr><td>Control Test</td><td>0.079</td><td>15:48</td></tr><tr><td>Air Blank</td><td>0.000</td><td>15:49</td></tr><tr><td>Control Test</td><td>0.079</td><td>15:49</td></tr><tr><td>Air Blank</td><td>0.000</td><td>15:50</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0787</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7339</td><td></td></tr></table>	Test	g/210L	Time	Air Blank	0.000	15:47	Control Test	0.078	15:48	Air Blank	0.000	15:48	Control Test	0.079	15:48	Air Blank	0.000	15:49	Control Test	0.079	15:49	Air Blank	0.000	15:50	Control Test Stats			Average	0.0787		Std Dev	0.0006		Rel Std Dev(%)	0.7339	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	16:51																																																																																																																																																	
Control Test	0.050	16:51																																																																																																																																																	
Air Blank	0.000	16:52																																																																																																																																																	
Control Test	0.049	16:52																																																																																																																																																	
Air Blank	0.000	16:53																																																																																																																																																	
Control Test	0.049	16:54																																																																																																																																																	
Air Blank	0.000	16:54																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0493																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1713																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	16:39																																																																																																																																																	
Control Test	0.079	16:39																																																																																																																																																	
Air Blank	0.000	16:40																																																																																																																																																	
Control Test	0.080	16:41																																																																																																																																																	
Air Blank	0.000	16:41																																																																																																																																																	
Control Test	0.080	16:42																																																																																																																																																	
Air Blank	0.000	16:42																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0797																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7247																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	16:45																																																																																																																																																	
Control Test	0.202	16:46																																																																																																																																																	
Air Blank	0.000	16:46																																																																																																																																																	
Control Test	0.200	16:47																																																																																																																																																	
Air Blank	0.000	16:48																																																																																																																																																	
Control Test	0.199	16:48																																																																																																																																																	
Air Blank	0.000	16:49																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.2003																																																																																																																																																		
Std Dev	0.0015																																																																																																																																																		
Rel Std Dev(%)	0.7625																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	15:47																																																																																																																																																	
Control Test	0.078	15:48																																																																																																																																																	
Air Blank	0.000	15:48																																																																																																																																																	
Control Test	0.079	15:48																																																																																																																																																	
Air Blank	0.000	15:49																																																																																																																																																	
Control Test	0.079	15:49																																																																																																																																																	
Air Blank	0.000	15:50																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0787																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7339																																																																																																																																																		
Operator's Signature	Operator's Signature	Operator's Signature	Operator's Signature																																																																																																																																																

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: LEVY COUNTY SO
Time of Inspection: 14:42

Date of Inspection: 07/09/2025

Serial Number: 80-001292
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG429602 Exp: 10/22/2026
0.000	0.048	0.079	0.199	0.078
0.000	0.048	0.079	0.199	0.078
0.000	0.048	0.079	0.199	0.078
0.000	0.048	0.079	0.199	0.078
0.000	0.048	0.079	0.199	0.077
0.000	0.048	0.079	0.199	0.077
0.000	0.048	0.079	0.199	0.078
0.000	0.048	0.079	0.199	0.078
0.000	0.047	0.079	0.199	0.078
0.000	0.047	0.079	0.200	0.078

Standard Deviations	0.0004	0.0000	0.0003	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002 Number of Simulators Used: 5


Remarks:

~~CLEARED ROOM WITH FAN AND REPLACED WATER IN SIMULATOR.~~

Sut no remark for this
instrument Department
Inspection.
Sut

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.


I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



LEANDRA HIGGINBOTHAM

Signature and Printed Name

07/09/2025
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001292, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001292</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>LEVY COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>07/09/2025</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>14:42</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form 69 October 2024
Issuing Authority: Alcohol Testing Program

07/09/2025
Date

Michael
LEANDRA HIGGINBOTHAM,
Department Inspector

Service • Integrity • Respect • Quality