

Agency Santa Rosa County SO

S/N 80-001302

Date In 12/30/2024 DI Completion Date 01/08/2025

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Calibration Adjustment

By_

❑ Post Calibration Adjustment Stability Checks

Notes/Suggested Service:

Department Inspection

By DA

Attachments

- ☒ Instrument Complies with Chapter 11D-8, FAC
- ☐ Instrument Does Not Comply with Chapter 11D-8, FAC
- ☒ Return to/Place into Evidentiary Use
- ☐ Remain Out of Evidentiary Use
- ☒ Conduct an Agency Inspection Before Evidentiary Use

Phil Nicodemo Digitally signed by Phil Nicodemo
Date: 2025.01.14 13:52:37 -05'00'

Digitally signed by Phil
Nicodemo
Date: 2025.01.14 13:52:37 -05'00'

Taylor Gutschow

Digitally signed by Taylor
Gutschow
Date: 2025.01.16 15:11:39 -05'00'


Tech Review / Date

Admin Review / Date

Stability Checks 80-001302 DA 1/3/25

SANTA ROSA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001302
01/03/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:25
Control Test	0.049	11:25
Air Blank	0.000	11:26
Control Test	0.049	11:27
Air Blank	0.000	11:27
Control Test	0.048	11:28
Air Blank	0.000	11:29
Control Test Stats		
Average	0.0487	
Std Dev	0.0006	
Rel Std Dev(%)	1.1863	



Operator's Signature

SANTA ROSA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001302
01/03/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:30
Control Test	0.079	11:30
Air Blank	0.000	11:31
Control Test	0.079	11:31
Air Blank	0.000	11:32
Control Test	0.078	11:33
Air Blank	0.000	11:33
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	


Wet



Operator's Signature

SANTA ROSA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001302
01/03/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:36
Control Test	0.203	11:37
Air Blank	0.000	11:38
Control Test	0.202	11:38
Air Blank	0.000	11:39
Control Test	0.201	11:40
Air Blank	0.000	11:40
Control Test Stats		
Average	0.2020	
Std Dev	0.0010	
Rel Std Dev(%)	0.4950	



Operator's Signature

SANTA ROSA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001302
01/03/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:46
Control Test	0.078	11:46
Air Blank	0.000	11:46
Control Test	0.078	11:47
Air Blank	0.000	11:47
Control Test	0.079	11:48
Air Blank	0.000	11:48
Control Test Stats		
Average	0.0783	
Std Dev	0.0006	
Rel Std Dev(%)	0.7370	

DGS

Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: SANTA ROSA COUNTY SO

Serial Number: 80-001302

Time of Inspection: 14:36

Date of Inspection: 01/08/2025

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#: 202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#: 202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: AG429602 Exp: 10/22/2026
0.000	0.049	0.079	0.202	0.078
0.000	0.049	0.080	0.202	0.078
0.000	0.049	0.079	0.202	0.078
0.000	0.050	0.080	0.202	0.077
0.000	0.050	0.079	0.202	0.078
0.000	0.049	0.080	0.202	0.078
0.000	0.050	0.080	0.202	0.078
0.000	0.050	0.080	0.202	0.078
0.000	0.050	0.080	0.202	0.078
0.000	0.050	0.080	0.203	0.077


Standard Deviations	0.0005	0.0004	0.0003	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



DESTINEE N ARMSTRONG

Signature and Printed Name

01/08/2025

Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001302, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001302</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>SANTA ROSA COUNTY SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>01/08/2025</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>14:36</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Destinee
Armstrong
Digitally signed by Destinee
Armstrong
Date: 2025.01.14 09:40:05
+05'00'

01/08/2025

Date

DESTINEE N ARMSTRONG,
Department Inspector

FDLE/ATP Form 69 October 2024

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality