



INSTRUMENT PROCESSING SHEET

Agency Okeechobee CSOS/N 80-001321Florida Department of
Law EnforcementDate In 01/06/2025 DI Completion Date 02/14/2025☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By TDG	Date	Quality Checks	By TDG	Date	Flow Calibration	By	Date																												
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☒ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: <u>Instrument failed the Dec. 2024 Agency Inspection during the 0.08 ARS Test.</u>		<u>01/09/2025</u>	☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>211</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP101</u> 32 mm <u>0.156</u> (.139 - .169) 36 mm <u>0.171</u> (.156 - .190) 53 mm <u>0.242</u> (.228 - .278) 103 mm <u>0.503</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>26932</u> ☒ Stability Checks		<u>01/13/2025</u>	Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																														
			<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6286</td><td>202303K 03/29/2025</td></tr><tr><td>0.080</td><td>MP6287</td><td>202303L 03/29/2025</td></tr><tr><td>0.200</td><td>MP6288</td><td>202304C 04/05/2025</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG429602 10/22/2026</td></tr></tbody></table>	Simulator	Serial #	Lot #/Exp	0.050	MP6286	202303K 03/29/2025	0.080	MP6287	202303L 03/29/2025	0.200	MP6288	202304C 04/05/2025	0.080 DGS	N/A	AG429602 10/22/2026																		
Simulator	Serial #	Lot #/Exp																																		
0.050	MP6286	202303K 03/29/2025																																		
0.080	MP6287	202303L 03/29/2025																																		
0.200	MP6288	202304C 04/05/2025																																		
0.080 DGS	N/A	AG429602 10/22/2026																																		
Calibration Adjustment			By TDG		Department Inspection																															
Barometric Pressure Gauge <u>1022</u> ID # <u>28199</u>					Barometric Pressure ID# <u>26932</u>																															
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td>MP5097</td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td>MP5098</td><td>23400</td><td>10/24/2025</td></tr><tr><td>0.100</td><td>MP5099</td><td>24110</td><td>03/05/2026</td></tr><tr><td>0.200</td><td>MP5100</td><td>24080</td><td>02/13/2026</td></tr><tr><td>0.300</td><td>MP5101</td><td>23410</td><td>11/01/2025</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>06723080A5</td><td>04/05/2025</td></tr></tbody></table>			Simulator	Serial #	Lot #	Expiration	0.000	MP5097	N/A	N/A	0.040	MP5098	23400	10/24/2025	0.100	MP5099	24110	03/05/2026	0.200	MP5100	24080	02/13/2026	0.300	MP5101	23410	11/01/2025	0.080 DGS	N/A	06723080A5	04/05/2025			Gauge <u>1027</u> Instrument <u>1027</u> Mouth Alcohol Solution Lot # <u>2024-A</u> Acetone Stock Solution Lot # <u>2023-B</u>			
Simulator	Serial #	Lot #	Expiration																																	
0.000	MP5097	N/A	N/A																																	
0.040	MP5098	23400	10/24/2025																																	
0.100	MP5099	24110	03/05/2026																																	
0.200	MP5100	24080	02/13/2026																																	
0.300	MP5101	23410	11/01/2025																																	
0.080 DGS	N/A	06723080A5	04/05/2025																																	
☒ Post Calibration Adjustment Stability Checks					<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6284</td></tr><tr><td>Interferent</td><td>MP6285</td></tr><tr><td>0.050</td><td>MP6286</td></tr><tr><td>0.080</td><td>MP6287</td></tr><tr><td>0.200</td><td>MP6288</td></tr></tbody></table>				Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288			Attachments													
Simulator	Serial Number																																			
0.000	MP6284																																			
Interferent	MP6285																																			
0.050	MP6286																																			
0.080	MP6287																																			
0.200	MP6288																																			
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td>MP6286</td><td>202303K</td><td>03/29/2025</td></tr><tr><td>0.080</td><td>MP6287</td><td>202303L</td><td>03/29/2025</td></tr><tr><td>0.200</td><td>MP6288</td><td>202304C</td><td>04/05/2025</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG429602</td><td>10/22/2026</td></tr></tbody></table>			Simulator	Serial #	Lot #	Expiration	0.050	MP6286	202303K	03/29/2025	0.080	MP6287	202303L	03/29/2025	0.200	MP6288	202304C	04/05/2025	0.080 DGS	N/A	AG429602	10/22/2026			☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Calibration Adjustment ☐ Post-Stability Checks ☐ Flow Calibration ☒ Form 40 ☐ Other _____											
Simulator	Serial #	Lot #	Expiration																																	
0.050	MP6286	202303K	03/29/2025																																	
0.080	MP6287	202303L	03/29/2025																																	
0.200	MP6288	202304C	04/05/2025																																	
0.080 DGS	N/A	AG429602	10/22/2026																																	
Notes/Suggested Service: _____ _____ _____ _____ _____ _____ _____					☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																															
					Digitally signed by Shayla Platt Date: 2025.02.21 15:08:22 -05'00'																															
					Tech Review / Date _____ Admin: Review / Date _____																															

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OKEECHOBEE COUNTY SO

Time of Inspection: 09:40

Date of Inspection: 01/13/2025

Serial Number: 80-001321

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____

Remarks:

AI NOT CONDUCTED. BYPASSED TO BRING OUT OF DISABLED MODE.

Not determined

76
01/13/25

The above instrument complies (☒) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Taylor D Gutschow

TAYLOR D GUTSCHOW

Signature and Printed Name

01/13/2025
Date

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083 ☒ ≤0.003 of Wet ☒																																																																																																															
OKEECHOBEE COUNTY SO Intoxilyzer - Alcotest Analyzer Model 8000 01/13/2025 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>10:18</td></tr> <tr><td>Control Test</td><td>0.048</td><td>10:19</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:19</td></tr> <tr><td>Control Test</td><td>0.048</td><td>10:20</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:20</td></tr> <tr><td>Control Test</td><td>0.048</td><td>10:21</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:21</td></tr> <tr><td>Control Test</td><td>0.048</td><td>10:21</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0480</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	10:18	Control Test	0.048	10:19	Air Blank	0.000	10:19	Control Test	0.048	10:20	Air Blank	0.000	10:20	Control Test	0.048	10:21	Air Blank	0.000	10:21	Control Test	0.048	10:21	Control Test Stats			Average	0.0480		Std Dev	0.0000		Rel Std Dev(%)	0.0000		OKEECHOBEE COUNTY SO Intoxilyzer - Alcotest Analyzer Model 8000 01/13/2025 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>10:27</td></tr> <tr><td>Control Test</td><td>0.077</td><td>10:27</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:28</td></tr> <tr><td>Control Test</td><td>0.077</td><td>10:29</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:29</td></tr> <tr><td>Control Test</td><td>0.077</td><td>10:30</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:30</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0770</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	10:27	Control Test	0.077	10:27	Air Blank	0.000	10:28	Control Test	0.077	10:29	Air Blank	0.000	10:29	Control Test	0.077	10:30	Air Blank	0.000	10:30	Control Test Stats			Average	0.0770		Std Dev	0.0000		Rel Std Dev(%)	0.0000		OKEECHOBEE COUNTY SO Intoxilyzer - Alcotest Analyzer Model 8000 01/13/2025 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>10:34</td></tr> <tr><td>Control Test</td><td>0.197</td><td>10:34</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:35</td></tr> <tr><td>Control Test</td><td>0.195</td><td>10:36</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:36</td></tr> <tr><td>Control Test</td><td>0.196</td><td>10:37</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:37</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1960</td><td></td></tr> <tr><td>Std Dev</td><td>0.0010</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.5102</td><td></td></tr> </tbody> </table> Operator's Signature	Test	g/210L	Time	Air Blank	0.000	10:34	Control Test	0.197	10:34	Air Blank	0.000	10:35	Control Test	0.195	10:36	Air Blank	0.000	10:36	Control Test	0.196	10:37	Air Blank	0.000	10:37	Control Test Stats			Average	0.1960		Std Dev	0.0010		Rel Std Dev(%)	0.5102		DGS 0.08g/210L ☒ 0.077 to 0.083 ☒ ≤0.003 of Wet ☒ Printed to external printer. Results in range. Will attach. ML 01/13/25
Test	g/210L	Time																																																																																																																
Air Blank	0.000	10:18																																																																																																																
Control Test	0.048	10:19																																																																																																																
Air Blank	0.000	10:19																																																																																																																
Control Test	0.048	10:20																																																																																																																
Air Blank	0.000	10:20																																																																																																																
Control Test	0.048	10:21																																																																																																																
Air Blank	0.000	10:21																																																																																																																
Control Test	0.048	10:21																																																																																																																
Control Test Stats																																																																																																																		
Average	0.0480																																																																																																																	
Std Dev	0.0000																																																																																																																	
Rel Std Dev(%)	0.0000																																																																																																																	
Test	g/210L	Time																																																																																																																
Air Blank	0.000	10:27																																																																																																																
Control Test	0.077	10:27																																																																																																																
Air Blank	0.000	10:28																																																																																																																
Control Test	0.077	10:29																																																																																																																
Air Blank	0.000	10:29																																																																																																																
Control Test	0.077	10:30																																																																																																																
Air Blank	0.000	10:30																																																																																																																
Control Test Stats																																																																																																																		
Average	0.0770																																																																																																																	
Std Dev	0.0000																																																																																																																	
Rel Std Dev(%)	0.0000																																																																																																																	
Test	g/210L	Time																																																																																																																
Air Blank	0.000	10:34																																																																																																																
Control Test	0.197	10:34																																																																																																																
Air Blank	0.000	10:35																																																																																																																
Control Test	0.195	10:36																																																																																																																
Air Blank	0.000	10:36																																																																																																																
Control Test	0.196	10:37																																																																																																																
Air Blank	0.000	10:37																																																																																																																
Control Test Stats																																																																																																																		
Average	0.1960																																																																																																																	
Std Dev	0.0010																																																																																																																	
Rel Std Dev(%)	0.5102																																																																																																																	

OKEECHOBEE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001321
01/13/2025
Software: 8100.27

065

Test	g/210L	Time
Air Blank	0.000	10:02
Control Test	0.081	10:02
Air Blank	0.000	10:03
Control Test	0.081	10:03
Air Blank	0.000	10:04
Control Test	0.080	10:04
Air Blank	0.000	10:05
Control Test Stats		
Average	0.0807	
Std Dev	0.0006	
Rel Std Dev(%)	0.7157	

mc

Operator's Signature

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.5470 (-0.0010)
Sample #2 = 1.5430 (-0.0020)
Sample #3 = 1.5540 (-0.0050)
Sample #4 = 1.5590 (-0.0020)
Avg % Abs = 1.5520 (-0.0030)
STD DEV = 0.0082 (0.0017)
REL STD DEV = 0.527 (57.735)

OKEECHOBEE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001321
02/13/2025 09:22:55

Sol Value = 0.100 g/210L ***
Fit Value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12731, 9um Io = 13225

<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.8580 (-0.0200)
Sample #2 = 1.8750 (-0.0290)
Sample #3 = 1.8610 (-0.0080)
Sample #4 = 1.8880 (-0.0120)
Avg % Abs = 1.8750 (-0.0030)
STD DEV = 0.0135 (0.0225)
REL STD DEV = 0.721 (753.510)

Auto Calibration
Max Power Res Value = 39
Auto Range Res Value = 35

Sol Value = 0.000 g/210L ***
Fit Value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12734, 9um Io = 13228

<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.0480 (-0.0120)
Sample #2 = 0.0530 (-0.0300)
Sample #3 = 0.0810 (-0.0420)
Sample #4 = 0.0810 (-0.0220)
Avg % Abs = 0.0717 (-0.0313)
STD DEV = 0.0162 (0.0101)
REL STD DEV = 22.557 (32.127)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.5620 (-0.0010)
Sample #2 = 3.5820 (-0.0060)
Sample #3 = 3.5790 (-0.0140)
Sample #4 = 3.6050 (-0.0080)
Avg % Abs = 3.5687 (0.0053)
STD DEV = 0.0142 (0.0103)
REL STD DEV = 0.396 (192.435)

Sol Value = 0.200 g/210L ***
Fit Value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12731, 9um Io = 13224

<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 3.6470 (-0.0010)
Sample #2 = 3.6480 (-0.0040)
Sample #3 = 3.6260 (-0.0020)
Sample #4 = 3.6400 (-0.0200)
Avg % Abs = 3.6380 (0.0060)
STD DEV = 0.0111 (0.0125)
REL STD DEV = 0.306 (208.167)

Sol Value = 0.040 g/210L ***
Fit Value = 0.1915 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12733, 9um Io = 13228

<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.7650 (-0.0080)
Sample #2 = 0.7630 (-0.0170)
Sample #3 = 0.8170 (-0.0240)
Sample #4 = 0.7990 (-0.0130)
Avg % Abs = 0.7927 (-0.0180)
STD DEV = 0.0274 (0.0056)
REL STD DEV = 3.456 (30.932)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 6.9120 (-0.0110)
Sample #2 = 6.8810 (-0.0070)
Sample #3 = 6.8350 (-0.1080)
Sample #4 = 6.8640 (-0.1240)
Avg % Abs = 6.8600 (0.0130)
STD DEV = 0.0233 (0.0095)
REL STD DEV = 0.339 (73.380)

Sol Value = 0.300 g/210L ***
Fit Value = 1.4286 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12729, 9um Io = 13223

<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 5.3420 (-0.0140)
Sample #2 = 5.3430 (0.0060)
Sample #3 = 5.3820 (-0.0170)
Sample #4 = 5.3850 (-0.0100)
Avg % Abs = 5.3700 (-0.0070)
STD DEV = 0.0234 (0.0118)
REL STD DEV = 0.436 (168.426)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 10.0040 (-0.0140)
Sample #2 = 10.0070 (0.0150)
Sample #3 = 10.0260 (0.0090)
Sample #4 = 10.0430 (0.0070)
Avg % Abs = 10.0253 (0.0103)
STD DEV = 0.0180 (0.0042)
REL STD DEV = 0.180 (40.290)

***** AUTO CAL DATA *****
<<<< CHANNEL 1 >>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.072
Std Dev = 0.02 Rel Std Dev = 22.56
Sol Val = 0.1915 mg/l or 0.040 g/210L
% Abs = 0.793
Std Dev = 0.03 Rel Std Dev = 3.46
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.875
Std Dev = 0.01 Rel Std Dev = 0.72
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.638
Std Dev = 0.01 Rel Std Dev = 0.31
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.370
Std Dev = 0.02 Rel Std Dev = 0.44
Zero Order Coef = -183.51
First Order Coef = 2611.89
Second Order Coef = 15.38
Standard Deviation = 5.488423

<<<< CHANNEL 2 >>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.131
Std Dev = 0.01 Rel Std Dev = 10.34
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.552
Std Dev = 0.01 Rel Std Dev = 0.53
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.589
Std Dev = 0.01 Rel Std Dev = 0.40
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.860
Std Dev = 0.02 Rel Std Dev = 0.34
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 10.025
Std Dev = 0.02 Rel Std Dev = 0.18
Zero Order Coef = -191.06
First Order Coef = 1346.81
Second Order Coef = 9.76
Standard Deviation = 14.708405

Sol Value = 0.080 g/210L ***
Fit Value = 0.3810 mg/l %%%
Samples Taken = 4, Discarded = 1
***** CHANNEL 1
Sample #1 = 3274.00
Sample #2 = 3355.00
Sample #3 = 3297.00
Sample #4 = 3339.00
Average Result = 3330.3333
STD DEV = 29.9555
REL STD DEV = 0.899

***** CHANNEL 2
Sample #1 = 3273.00
Sample #2 = 3293.00
Sample #3 = 3304.00
Sample #4 = 3287.00
Average Result = 3294.6667
STD DEV = 8.6217
REL STD DEV = 0.262

Dry Gas H2O Adjust Results *****
Barometric Pressure = 1022
3 um H2O Adjust (mg/l*10,000) = 479
9 um H2O Adjust (mg/l*10,000) = 515
***** AUTO CAL PASS

Sol Value = 0.080 g/210L ***
Fit Value = 0.3810 mg/l %%%
Samples Taken = 4, Discarded = 1
***** CHANNEL 1
Sample #1 = 3274.00
Sample #2 = 3355.00
Sample #3 = 3297.00
Sample #4 = 3339.00
Average Result = 3330.3333
STD DEV = 29.9555
REL STD DEV = 0.899

***** CHANNEL 2
Sample #1 = 3273.00
Sample #2 = 3293.00
Sample #3 = 3304.00
Sample #4 = 3287.00
Average Result = 3294.6667
STD DEV = 8.6217
REL STD DEV = 0.262

Dry Gas H2O Adjust Results *****
Barometric Pressure = 1022
3 um H2O Adjust (mg/l*10,000) = 479
9 um H2O Adjust (mg/l*10,000) = 515
***** AUTO CAL PASS

Optical Calibration
Adjustment

By: TDG

Post-Cal Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
✓	✓	✓	✓
≤0.003 of Wet	✓	✓	✓
OKEECHOBEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 02/13/2025 Software: 8100.27 SN 80-001321 Test g/210L Time Air Blank 0.000 11:43 Control Test 0.051 11:44 Air Blank 0.000 11:45 Control Test 0.051 11:45 Air Blank 0.000 11:46 Control Test 0.050 11:47 Air Blank 0.000 11:47 Control Test Stats Average 0.0507 Std Dev 0.0006 Rel Std Dev(%) 1.1395 Operator's Signature 	OKEECHOBEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 02/13/2025 Software: 8100.27 SN 80-001321 Test g/210L Time Air Blank 0.000 11:57 Control Test 0.079 11:57 Air Blank 0.000 11:58 Control Test 0.079 11:59 Air Blank 0.000 11:59 Control Test 0.079 12:00 Air Blank 0.000 12:00 Control Test Stats Average 0.0790 Std Dev 0.0000 Rel Std Dev(%) 0.0000 Operator's Signature 	OKEECHOBEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 02/13/2025 Software: 8100.27 SN 80-001321 Test g/210L Time Air Blank 0.000 12:03 Control Test 0.196 12:04 Air Blank 0.000 12:04 Control Test 0.196 12:05 Air Blank 0.000 12:06 Control Test 0.195 12:06 Air Blank 0.000 12:07 Control Test Stats Average 0.1957 Std Dev 0.0006 Rel Std Dev(%) 0.2951 Operator's Signature 	ALS OKEECHOBEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 02/13/2025 Software: 8100.27 SN 80-001321 Test g/210L Time Air Blank 0.000 11:38 Control Test 0.080 11:38 Air Blank 0.000 11:39 Control Test 0.081 11:39 Air Blank 0.000 11:40 Control Test 0.080 11:40 Air Blank 0.000 11:40 Control Test Stats Average 0.0803 Std Dev 0.0006 Rel Std Dev(%) 0.7187 Operator's Signature 

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: OKEECHOBEE COUNTY SO
Time of Inspection: 13:40

Date of Inspection: 02/14/2025

Serial Number: 80-001321
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG429602 Exp: 10/22/2026
0.000	0.049	0.080	0.197	0.079
0.000	0.050	0.080	0.197	0.079
0.000	0.050	0.080	0.198	0.079
0.000	0.050	0.080	0.197	0.078
0.000	0.049	0.080	0.198	0.079
0.000	0.050	0.080	0.198	0.079
0.000	0.050	0.080	0.198	0.079
0.000	0.050	0.080	0.199	0.078
0.000	0.050	0.080	0.198	0.078
0.000	0.050	0.080	0.198	0.079

Standard Deviations	0.0004	0.0000	0.0006	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.


I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

02/14/2025
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001321, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001321</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>OKEECHOBEE COUNTY SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>02/14/2025</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>13:40</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

02/14/2025

Date

Taylor D Gutschow

TAYLOR D GUTSCHOW,
Department Inspector

Service • Integrity • Respect • Quality