



INSTRUMENT PROCESSING SHEET

Agency Port St. Lucie PDS/N 80-001323Florida Department of
Law EnforcementDate In 6/3/25DI Completion Date 6/25/25☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By KTS	Date <u>6/4/25</u>	Quality Checks	By KTS	Date <u>6/4/25</u>	Flow Calibration	By	Date																																							
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: <u>Placed plastic caps in box.</u>			☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>216</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP106</u> 32 mm <u>0.156</u> (.139 - .169) 36 mm <u>0.167</u> (.156 - .190) 53 mm <u>0.242</u> (.228 - .278) 103 mm <u>0.488</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28199</u> ☒ Stability Checks			Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																									
			<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP8231</td><td>202406K 06/19/2026</td></tr><tr><td>0.080</td><td>MP8230</td><td>202406L 06/19/2026</td></tr><tr><td>0.200</td><td>MP8229</td><td>202406N 06/20/2026</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG429602 10/22/2026</td></tr></tbody></table>	Simulator	Serial #	Lot #/Exp	0.050	MP8231	202406K 06/19/2026	0.080	MP8230	202406L 06/19/2026	0.200	MP8229	202406N 06/20/2026	0.080 DGS	N/A	AG429602 10/22/2026																													
Simulator	Serial #	Lot #/Exp																																													
0.050	MP8231	202406K 06/19/2026																																													
0.080	MP8230	202406L 06/19/2026																																													
0.200	MP8229	202406N 06/20/2026																																													
0.080 DGS	N/A	AG429602 10/22/2026																																													
Calibration Adjustment By <u>TDG</u>			Department Inspection By <u>TDG</u>																																												
Barometric Pressure Gauge <u>1019</u> ID # <u>26932</u>			Barometric Pressure ID# <u>28199</u> Gauge <u>1020</u> Instrument <u>1020</u> Mouth Alcohol Solution Lot # <u>2024-A</u> Acetone Stock Solution Lot # <u>2023-B</u>																																												
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td>MP6284</td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td>MP6285</td><td>23400</td><td>10/24/2025</td></tr><tr><td>0.100</td><td>MP6286</td><td>24110</td><td>03/05/2026</td></tr><tr><td>0.200</td><td>MP6287</td><td>24080</td><td>02/13/2026</td></tr><tr><td>0.300</td><td>MP6288</td><td>23410</td><td>11/01/2025</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>28424080A3</td><td>11/05/2026</td></tr></tbody></table>			Simulator	Serial #	Lot #	Expiration	0.000	MP6284	N/A	N/A	0.040	MP6285	23400	10/24/2025	0.100	MP6286	24110	03/05/2026	0.200	MP6287	24080	02/13/2026	0.300	MP6288	23410	11/01/2025	0.080 DGS	N/A	28424080A3	11/05/2026	<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP5097</td></tr><tr><td>Interferent</td><td>MP5098</td></tr><tr><td>0.050</td><td>MP5099</td></tr><tr><td>0.080</td><td>MP5100</td></tr><tr><td>0.200</td><td>MP5101</td></tr></tbody></table>					Simulator	Serial Number	0.000	MP5097	Interferent	MP5098	0.050	MP5099	0.080	MP5100	0.200	MP5101
Simulator	Serial #	Lot #	Expiration																																												
0.000	MP6284	N/A	N/A																																												
0.040	MP6285	23400	10/24/2025																																												
0.100	MP6286	24110	03/05/2026																																												
0.200	MP6287	24080	02/13/2026																																												
0.300	MP6288	23410	11/01/2025																																												
0.080 DGS	N/A	28424080A3	11/05/2026																																												
Simulator	Serial Number																																														
0.000	MP5097																																														
Interferent	MP5098																																														
0.050	MP5099																																														
0.080	MP5100																																														
0.200	MP5101																																														
☒ Post Calibration Adjustment Stability Checks			Attachments																																												
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td>MP4863</td><td>202406K</td><td>06/19/2026</td></tr><tr><td>0.080</td><td>MP4864</td><td>202406L</td><td>06/19/2026</td></tr><tr><td>0.200</td><td>MP6284</td><td>202406N</td><td>06/20/2026</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG429602</td><td>10/22/2026</td></tr></tbody></table>			Simulator	Serial #	Lot #	Expiration	0.050	MP4863	202406K	06/19/2026	0.080	MP4864	202406L	06/19/2026	0.200	MP6284	202406N	06/20/2026	0.080 DGS	N/A	AG429602	10/22/2026	<table border="1"><tbody><tr><td>☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Calibration Adjustment</td><td>☒ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____</td></tr></tbody></table>					☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Calibration Adjustment	☒ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____																		
Simulator	Serial #	Lot #	Expiration																																												
0.050	MP4863	202406K	06/19/2026																																												
0.080	MP4864	202406L	06/19/2026																																												
0.200	MP6284	202406N	06/20/2026																																												
0.080 DGS	N/A	AG429602	10/22/2026																																												
☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Calibration Adjustment	☒ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____																																														
Notes/Suggested Service: <u>Performed root cause analysis on failed DGS stability check. Results were not due to user error or external equipment. KTS/TDG 6/4/25</u> <u>Re-used MP6284 for post-cal stability. Rinsed and dried prior to adding new solution. (TDG 6/20/25)</u>			☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																																												
Phil Nicodemo Digitally signed by Phil Nicodemo Date: 2025.07.01 14:18:46 -04'00'			LeAndra Higginbotham Digitally signed by LeAndra Higginbotham Date: 2025.07.08 08:11:54 -04'00'																																												
Tech Review / Date			Admin Review / Date																																												

Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																				
PORT ST LUCIE PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001323 06/04/2025 Software: 8100.27 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>14:17</td></tr> <tr><td>Control Test</td><td>0.048</td><td>14:17</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:18</td></tr> <tr><td>Control Test</td><td>0.048</td><td>14:19</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:19</td></tr> <tr><td>Control Test</td><td>0.048</td><td>14:20</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:20</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0480</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </table> Operator's Signature _____	Air Blank	0.000	14:17	Control Test	0.048	14:17	Air Blank	0.000	14:18	Control Test	0.048	14:19	Air Blank	0.000	14:19	Control Test	0.048	14:20	Air Blank	0.000	14:20	Control Test Stats			Average	0.0480		Std Dev	0.0000		Rel Std Dev(%)	0.0000		PORT ST LUCIE PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001323 06/04/2025 Software: 8100.27 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>14:22</td></tr> <tr><td>Control Test</td><td>0.079</td><td>14:22</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:23</td></tr> <tr><td>Control Test</td><td>0.079</td><td>14:24</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:24</td></tr> <tr><td>Control Test</td><td>0.079</td><td>14:25</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:25</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0790</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </table> Operator's Signature _____	Air Blank	0.000	14:22	Control Test	0.079	14:22	Air Blank	0.000	14:23	Control Test	0.079	14:24	Air Blank	0.000	14:24	Control Test	0.079	14:25	Air Blank	0.000	14:25	Control Test Stats			Average	0.0790		Std Dev	0.0000		Rel Std Dev(%)	0.0000		PORT ST LUCIE PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001323 06/04/2025 Software: 8100.27 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>14:27</td></tr> <tr><td>Control Test</td><td>0.201</td><td>14:27</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:28</td></tr> <tr><td>Control Test</td><td>0.200</td><td>14:28</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:29</td></tr> <tr><td>Control Test</td><td>0.201</td><td>14:30</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:30</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.2007</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2877</td><td></td></tr> </table> Operator's Signature _____	Air Blank	0.000	14:27	Control Test	0.201	14:27	Air Blank	0.000	14:28	Control Test	0.200	14:28	Air Blank	0.000	14:29	Control Test	0.201	14:30	Air Blank	0.000	14:30	Control Test Stats			Average	0.2007		Std Dev	0.0006		Rel Std Dev(%)	0.2877		PORT ST LUCIE PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001323 06/04/2025 Software: 8100.27 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>13:50</td></tr> <tr><td>Control Test</td><td>0.076</td><td>13:50</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:51</td></tr> <tr><td>Control Test</td><td>0.076</td><td>13:51</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:51</td></tr> <tr><td>Control Test</td><td>0.076</td><td>13:52</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:52</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0760</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </table> Operator's Signature _____	Air Blank	0.000	13:50	Control Test	0.076	13:50	Air Blank	0.000	13:51	Control Test	0.076	13:51	Air Blank	0.000	13:51	Control Test	0.076	13:52	Air Blank	0.000	13:52	Control Test Stats			Average	0.0760		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Air Blank	0.000	14:17																																																																																																																																					
Control Test	0.048	14:17																																																																																																																																					
Air Blank	0.000	14:18																																																																																																																																					
Control Test	0.048	14:19																																																																																																																																					
Air Blank	0.000	14:19																																																																																																																																					
Control Test	0.048	14:20																																																																																																																																					
Air Blank	0.000	14:20																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0480																																																																																																																																						
Std Dev	0.0000																																																																																																																																						
Rel Std Dev(%)	0.0000																																																																																																																																						
Air Blank	0.000	14:22																																																																																																																																					
Control Test	0.079	14:22																																																																																																																																					
Air Blank	0.000	14:23																																																																																																																																					
Control Test	0.079	14:24																																																																																																																																					
Air Blank	0.000	14:24																																																																																																																																					
Control Test	0.079	14:25																																																																																																																																					
Air Blank	0.000	14:25																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0790																																																																																																																																						
Std Dev	0.0000																																																																																																																																						
Rel Std Dev(%)	0.0000																																																																																																																																						
Air Blank	0.000	14:27																																																																																																																																					
Control Test	0.201	14:27																																																																																																																																					
Air Blank	0.000	14:28																																																																																																																																					
Control Test	0.200	14:28																																																																																																																																					
Air Blank	0.000	14:29																																																																																																																																					
Control Test	0.201	14:30																																																																																																																																					
Air Blank	0.000	14:30																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.2007																																																																																																																																						
Std Dev	0.0006																																																																																																																																						
Rel Std Dev(%)	0.2877																																																																																																																																						
Air Blank	0.000	13:50																																																																																																																																					
Control Test	0.076	13:50																																																																																																																																					
Air Blank	0.000	13:51																																																																																																																																					
Control Test	0.076	13:51																																																																																																																																					
Air Blank	0.000	13:51																																																																																																																																					
Control Test	0.076	13:52																																																																																																																																					
Air Blank	0.000	13:52																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0760																																																																																																																																						
Std Dev	0.0000																																																																																																																																						
Rel Std Dev(%)	0.0000																																																																																																																																						
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083 ☒ ≤0.003 of Wet ☒																																																																																																																																				

<<<<< CHANNEL 2 >>>>>					
Sample	% Abs	(% Abs Ref)			
Sample #1 =	1.5920	(-0.0020)	Sample	% Abs	(% Abs Ref)
Sample #2 =	1.5720	(0.0160)	Sample #1 =	6.9070	(0.0000)
Sample #3 =	1.5970	(0.0000)	Sample #2 =	6.9420	(0.0190)
Sample #4 =	1.5970	(0.0130)	Sample #3 =	6.9520	(0.0000)
			Sample #4 =	6.9410	(0.0180)
			Avg % Abs =	6.9450	(0.0123)
			STD DEV =	0.0061	(0.0107)
			REL STD DEV =	0.0085	(0.0192)

601 value = 0.100 g/210L ***
Fit value = 0.4762 mg/l %%%
Sample Taken = 4 Dec 2004

```

<<<< CHANNEL 1 >>>>
Sample      % Abs      (% Abs Ref)
Sample #1 = 1.9520   (-0.0250)
Sample #2 = 1.9500   (-0.0120)
Sample #3 = 1.9350   (0.0080)
Sample #4 = 1.9510   (0.0080)
Avg % Abs = 1.9487 (-0.0040)
STD DEV = 0.0127 (0.0106)
REL STD DEV = 0.650 (264.575)
-----
<<<< CHANNEL 2 >>>>

```

```

Sample #2 = 3.6300 (-0.0010)
Sample #3 = 3.6300 (0.0030)
Sample #4 = 3.6380 (0.0030)
Avg & Sds = 3.6397 (0.0017)
STD DEV = 0.0106 (0.0023)
REL STD DEV = 0.29; (138.564)
-----
Sol. Value = 0.290 g/210L ***
Sol. Value = 0.290 g/210L ***

```

```

3um Io = 12817, 9um Io = 13593
<<<< CHANNEL 1 >>>>
Sample      % Abs      (% Abs Ref)
Sample #1 = 3.5740  (-0.0050)
Sample #2 = 3.6870  (0.0190)
Sample #3 = 3.7010  (0.0050)
Sample #4 = 3.6920  (0.0230)
Avg % Abs = 3.6333  (0.0157)
STD DEV = 0.0071  (0.0095)
REL STD DEV = 0.192  (0.330)
-----

```

Solution Stats Quadratic Fit: Chan 2			
Act	Fit	Residual	
9/210L	9/210L	9/210L	
0.000	0.000	-0.0001	
0.040	0.040	0.0001	
0.100	0.100	0.0002	
0.200	0.200	-0.0003	
0.300	0.300	0.0001	

```

Sol. Value = 0.080 g/210L ***
Fit. Value = 0.3810 mg/L ****
Samples Taken = 4, Discarded = 1
***** CHANNEL 1 *****
Sample #1 = 2879.00
Sample #2 = 2985.00
Sample #3 = 2965.00
Sample #4 = 2908.00
Average Result = 2952.6667
STD DEV = 39.9541
REL STD DEV = 1.353
***** CHANNEL 2 *****

```

```

Sample #2 = 3387.000
Sample #3 = 3403.00
Sample #4 = 3398.00
Average Result = 3392.6667
STD DEV = 8.9629
REL STD DEV = 0.264

*****
Dry Gas H2O Adjust Results *****
Barometric Pressure = 1018
3 um H2O Adjust (mg/l*10,000) = 857

```

***** AUTO CAL PASS

By: TDG

By: TDG

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083																																																																																																																																																
PORT ST LUCIE PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001323 06/20/2025 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>15:04</td></tr> <tr><td>Control Test</td><td>0.049</td><td>15:04</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:05</td></tr> <tr><td>Control Test</td><td>0.049</td><td>15:06</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:06</td></tr> <tr><td>Control Test</td><td>0.049</td><td>15:07</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:07</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0490</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature _____	Test	g/210L	Time	Air Blank	0.000	15:04	Control Test	0.049	15:04	Air Blank	0.000	15:05	Control Test	0.049	15:06	Air Blank	0.000	15:06	Control Test	0.049	15:07	Air Blank	0.000	15:07	Control Test Stats			Average	0.0490		Std Dev	0.0000		Rel Std Dev(%)	0.0000		PORT ST LUCIE PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001323 06/20/2025 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>15:11</td></tr> <tr><td>Control Test</td><td>0.080</td><td>15:12</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:12</td></tr> <tr><td>Control Test</td><td>0.080</td><td>15:13</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:14</td></tr> <tr><td>Control Test</td><td>0.080</td><td>15:14</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:15</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0800</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature _____	Test	g/210L	Time	Air Blank	0.000	15:11	Control Test	0.080	15:12	Air Blank	0.000	15:12	Control Test	0.080	15:13	Air Blank	0.000	15:14	Control Test	0.080	15:14	Air Blank	0.000	15:15	Control Test Stats			Average	0.0800		Std Dev	0.0000		Rel Std Dev(%)	0.0000		PORT ST LUCIE PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001323 06/20/2025 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>15:20</td></tr> <tr><td>Control Test</td><td>0.198</td><td>15:21</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:22</td></tr> <tr><td>Control Test</td><td>0.198</td><td>15:22</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:23</td></tr> <tr><td>Control Test</td><td>0.198</td><td>15:23</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:24</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1980</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature _____	Test	g/210L	Time	Air Blank	0.000	15:20	Control Test	0.198	15:21	Air Blank	0.000	15:22	Control Test	0.198	15:22	Air Blank	0.000	15:23	Control Test	0.198	15:23	Air Blank	0.000	15:24	Control Test Stats			Average	0.1980		Std Dev	0.0000		Rel Std Dev(%)	0.0000		PORT ST LUCIE PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001323 06/20/2025 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:42</td></tr> <tr><td>Control Test</td><td>0.078</td><td>14:42</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:42</td></tr> <tr><td>Control Test</td><td>0.078</td><td>14:43</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:43</td></tr> <tr><td>Control Test</td><td>0.078</td><td>14:44</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:44</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0780</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature _____	Test	g/210L	Time	Air Blank	0.000	14:42	Control Test	0.078	14:42	Air Blank	0.000	14:42	Control Test	0.078	14:43	Air Blank	0.000	14:43	Control Test	0.078	14:44	Air Blank	0.000	14:44	Control Test Stats			Average	0.0780		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	15:04																																																																																																																																																	
Control Test	0.049	15:04																																																																																																																																																	
Air Blank	0.000	15:05																																																																																																																																																	
Control Test	0.049	15:06																																																																																																																																																	
Air Blank	0.000	15:06																																																																																																																																																	
Control Test	0.049	15:07																																																																																																																																																	
Air Blank	0.000	15:07																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0490																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	15:11																																																																																																																																																	
Control Test	0.080	15:12																																																																																																																																																	
Air Blank	0.000	15:12																																																																																																																																																	
Control Test	0.080	15:13																																																																																																																																																	
Air Blank	0.000	15:14																																																																																																																																																	
Control Test	0.080	15:14																																																																																																																																																	
Air Blank	0.000	15:15																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0800																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	15:20																																																																																																																																																	
Control Test	0.198	15:21																																																																																																																																																	
Air Blank	0.000	15:22																																																																																																																																																	
Control Test	0.198	15:22																																																																																																																																																	
Air Blank	0.000	15:23																																																																																																																																																	
Control Test	0.198	15:23																																																																																																																																																	
Air Blank	0.000	15:24																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1980																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:42																																																																																																																																																	
Control Test	0.078	14:42																																																																																																																																																	
Air Blank	0.000	14:42																																																																																																																																																	
Control Test	0.078	14:43																																																																																																																																																	
Air Blank	0.000	14:43																																																																																																																																																	
Control Test	0.078	14:44																																																																																																																																																	
Air Blank	0.000	14:44																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0780																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PORT ST LUCIE PD
Time of Inspection: 13:02

Date of Inspection: 06/25/2025

Serial Number: 80-001323
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG429602 Exp: 10/22/2026
0.000	0.048	0.078	0.199	0.077
0.000	0.048	0.079	0.199	0.078
0.000	0.047	0.079	0.200	0.078
0.000	0.048	0.079	0.199	0.077
0.000	0.048	0.078	0.199	0.077
0.000	0.049	0.079	0.199	0.077
0.000	0.049	0.079	0.199	0.077
0.000	0.048	0.079	0.199	0.077
0.000	0.049	0.079	0.199	0.077
0.000	0.049	0.079	0.199	0.077


Standard Deviations	0.0006	0.0004	0.0003	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.


I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

06/25/2025
Date



Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

Calibration Certificate

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001323, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001323</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>PORT ST LUCIE PD</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>06/25/2025</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>13:02</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. The supplier of dry gas standard controls is Taylor Gutschow. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Taylor Gutschow
Digitally signed by Taylor Gutschow
Date: 2025.06.27 11:58:35 -04'00'

06/25/2025

Date

TAYLOR D GUTSCHOW,
Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality