



INSTRUMENT PROCESSING SHEET

Agency INDIAN RIVER CO SOS/N 80-001328Florida Department of
Law EnforcementDate In 05/30/2025 DI Completion Date 06/25/2025☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By WKP	Date	Quality Checks	By WKP	Date	Flow Calibration	By	Date																																					
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☒ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: <u>Added plastic caps.</u>		<u>06/04/2025</u>	☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>129</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>0.148</u> (.139 - .169) 36 mm <u>0.164</u> (.156 - .190) 53 mm <u>0.234</u> (.228 - .278) 103 mm <u>0.492</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28199</u> ☒ Stability Checks		<u>06/04/2025</u>	☐ Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																							
			<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6286</td><td>202406K 06/19/2026</td></tr><tr><td>0.080</td><td>MP6287</td><td>202406L 06/19/2026</td></tr><tr><td>0.200</td><td>MP6288</td><td>202406N 06/20/2026</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG429602 10/22/2026</td></tr></tbody></table>			Simulator	Serial #	Lot #/Exp	0.050	MP6286	202406K 06/19/2026	0.080	MP6287	202406L 06/19/2026	0.200	MP6288	202406N 06/20/2026	0.080 DGS	N/A	AG429602 10/22/2026	<table border="1"><thead><tr><th colspan="2">Maintenance By TDG</th><th>Date</th></tr></thead><tbody><tr><td colspan="2">☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☒ Other <u>Replaced internal printer paper after post-cal stabilities.</u></td><td><u>06/20/2025</u></td></tr></tbody></table>			Maintenance By TDG		Date	☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☒ Other <u>Replaced internal printer paper after post-cal stabilities.</u>		<u>06/20/2025</u>																
Simulator	Serial #	Lot #/Exp																																											
0.050	MP6286	202406K 06/19/2026																																											
0.080	MP6287	202406L 06/19/2026																																											
0.200	MP6288	202406N 06/20/2026																																											
0.080 DGS	N/A	AG429602 10/22/2026																																											
Maintenance By TDG		Date																																											
☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☒ Other <u>Replaced internal printer paper after post-cal stabilities.</u>		<u>06/20/2025</u>																																											
Calibration Adjustment By <u>TDG</u>			Department Inspection By <u>TDG</u>																																										
Barometric Pressure Gauge <u>1019</u> ID # <u>26932</u>			Barometric Pressure ID# <u>28199</u> Gauge <u>1020</u> Instrument <u>1019</u> Mouth Alcohol Solution Lot # <u>2024-A</u> Acetone Stock Solution Lot # <u>2023-B</u>																																										
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td>MP6284</td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td>MP6285</td><td>23400</td><td>10/24/2025</td></tr><tr><td>0.100</td><td>MP6286</td><td>24110</td><td>03/05/2026</td></tr><tr><td>0.200</td><td>MP6287</td><td>24080</td><td>02/13/2026</td></tr><tr><td>0.300</td><td>MP6288</td><td>23410</td><td>11/01/2025</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>28424080A3</td><td>11/05/2026</td></tr></tbody></table>			Simulator	Serial #	Lot #	Expiration	0.000	MP6284	N/A	N/A	0.040	MP6285	23400	10/24/2025	0.100	MP6286	24110	03/05/2026	0.200	MP6287	24080	02/13/2026	0.300	MP6288	23410	11/01/2025	0.080 DGS	N/A	28424080A3	11/05/2026	<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP5097</td></tr><tr><td>Interferent</td><td>MP5098</td></tr><tr><td>0.050</td><td>MP5099</td></tr><tr><td>0.080</td><td>MP5100</td></tr><tr><td>0.200</td><td>MP5101</td></tr></tbody></table>			Simulator	Serial Number	0.000	MP5097	Interferent	MP5098	0.050	MP5099	0.080	MP5100	0.200	MP5101
Simulator	Serial #	Lot #	Expiration																																										
0.000	MP6284	N/A	N/A																																										
0.040	MP6285	23400	10/24/2025																																										
0.100	MP6286	24110	03/05/2026																																										
0.200	MP6287	24080	02/13/2026																																										
0.300	MP6288	23410	11/01/2025																																										
0.080 DGS	N/A	28424080A3	11/05/2026																																										
Simulator	Serial Number																																												
0.000	MP5097																																												
Interferent	MP5098																																												
0.050	MP5099																																												
0.080	MP5100																																												
0.200	MP5101																																												
☒ Post Calibration Adjustment Stability Checks			Attachments																																										
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td>MP4863</td><td>202406K</td><td>06/19/2026</td></tr><tr><td>0.080</td><td>MP4864</td><td>202406L</td><td>06/19/2026</td></tr><tr><td>0.200</td><td>MP6284</td><td>202406N</td><td>06/20/2026</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG429602</td><td>10/22/2026</td></tr></tbody></table>			Simulator	Serial #	Lot #	Expiration	0.050	MP4863	202406K	06/19/2026	0.080	MP4864	202406L	06/19/2026	0.200	MP6284	202406N	06/20/2026	0.080 DGS	N/A	AG429602	10/22/2026	<table border="1"><tbody><tr><td>☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Calibration Adjustment</td><td>☒ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____</td></tr></tbody></table>			☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Calibration Adjustment	☒ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____																		
Simulator	Serial #	Lot #	Expiration																																										
0.050	MP4863	202406K	06/19/2026																																										
0.080	MP4864	202406L	06/19/2026																																										
0.200	MP6284	202406N	06/20/2026																																										
0.080 DGS	N/A	AG429602	10/22/2026																																										
☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Calibration Adjustment	☒ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____																																												
Notes/Suggested Service: <u>Gauge 28199 reported baro pressure of 1019 during Quality Checks. Instrument reported baro pressure of 1003. WKP/TDG 6/4/25</u>			☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC																																										
<u>Performed root cause analysis on failed stability check. Secured DGS on regulator and detached/reattached rear DGS connector. Repeated stability check results were within range. WKP/TDG 06/04/2025</u>			☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use																																										
<u>Re-used MP6284 for post-cal stability. Rinsed/dried prior to adding new solution. (TDG 6/20/25)</u>			☒ Conduct an Agency Inspection Before Evidentiary Use																																										
			Digitally signed by																																										
			Phil Nicodemo Date: 2025.07.01 13:28:29 -04'00'																																										
			Shayla Platt Date: 2025.07.07 16:37:04 -04'00'																																										
			Tech Review / Date																																										
			Admin Review / Date																																										

80-001328

Stability Checks

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083 ☒ ≤0.003 of Wet ☒
INDIAN RIVER CSO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001328 06/04/2025 Software: 8100.27 Test g/210L Time Air Blank 0.000 11:56 Control Test 0.049 11:56 Air Blank 0.000 11:57 Control Test 0.049 11:57 Air Blank 0.000 11:58 Control Test 0.049 11:58 Air Blank 0.000 11:59 Control Test 0.049 11:59 Control Test Stats Average 0.0490 Std Dev 0.0000 Rel Std Dev(%) 0.0000 <i>Wagner</i> Operator's Signature	INDIAN RIVER CSO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001328 06/04/2025 Software: 8100.27 Test g/210L Time Air Blank 0.000 11:38 Control Test 0.079 11:38 Air Blank 0.000 11:39 Control Test 0.079 11:39 Air Blank 0.000 11:40 Control Test 0.079 11:40 Air Blank 0.000 11:41 Control Test 0.079 11:41 Air Blank 0.000 11:42 Control Test 0.079 11:42 Control Test Stats Average 0.0790 Std Dev 0.0000 Rel Std Dev(%) 0.0000 <i>Wagner</i> Operator's Signature	INDIAN RIVER CSO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001328 06/04/2025 Software: 8100.27 Test g/210L Time Air Blank 0.000 11:47 Control Test 0.198 11:47 Air Blank 0.000 11:48 Control Test 0.198 11:48 Air Blank 0.000 11:49 Control Test 0.198 11:49 Air Blank 0.000 11:50 Control Test 0.197 11:50 Air Blank 0.000 11:51 Control Test 0.197 11:51 Average 0.1977 Std Dev 0.0006 Rel Std Dev(%) 0.2921 Control Test Stats Average 0.1977 Std Dev 0.0006 Rel Std Dev(%) 0.2921 <i>Wagner</i> Operator's Signature	INDIAN RIVER CSO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001328 06/04/2025 Software: 8100.27 Test g/210L Time Air Blank 0.000 12:06 Control Test 0.000 12:06 Air Blank 0.000 12:06 Control Test 0.000 12:07 Air Blank 0.000 12:07 Control Test 0.000 12:07 Air Blank 0.000 12:08 Control Test 0.000 12:08 Control Test Stats Average 0.0000 Std Dev 0.0000 Rel Std Dev(%) 0.0000 <i>Wagner</i> Operator's Signature

80-001328

Stability Checks

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083 ☒ ≤0.003 of Wet ☒																																				
#2 INDIAN RIVER CSO IntoxiMizer - Alcohol Analyzer Model 8000 06/04/2025 Software: 8100.27 SN 80-001328 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>12:19</td></tr><tr><td>Control Test</td><td>0.079</td><td>12:19</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:19</td></tr><tr><td>Control Test</td><td>0.079</td><td>12:20</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:20</td></tr><tr><td>Control Test</td><td>0.079</td><td>12:21</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:21</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0790</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></tbody></table> Operator's Signature <i>marp</i>				Test	g/210L	Time	Air Blank	0.000	12:19	Control Test	0.079	12:19	Air Blank	0.000	12:19	Control Test	0.079	12:20	Air Blank	0.000	12:20	Control Test	0.079	12:21	Air Blank	0.000	12:21	Control Test Stats			Average	0.0790		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																					
Air Blank	0.000	12:19																																					
Control Test	0.079	12:19																																					
Air Blank	0.000	12:19																																					
Control Test	0.079	12:20																																					
Air Blank	0.000	12:20																																					
Control Test	0.079	12:21																																					
Air Blank	0.000	12:21																																					
Control Test Stats																																							
Average	0.0790																																						
Std Dev	0.0000																																						
Rel Std Dev(%)	0.0000																																						

By: TDG

Post-Cal Stability Checks

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083 ✓ ≤0.003 of Wet ✓																																																																																																																																																
INDIAN RIVER CSO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001328 06/20/2025 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>15:14</td></tr> <tr><td>Control Test</td><td>0.050</td><td>15:15</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:15</td></tr> <tr><td>Control Test</td><td>0.050</td><td>15:16</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:17</td></tr> <tr><td>Control Test</td><td>0.048</td><td>15:17</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:18</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0493</td><td></td></tr> <tr><td>Std Dev</td><td>0.0012</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>2.3406</td><td></td></tr> </tbody> </table> Operator's Signature <i>ML</i>	Test	g/210L	Time	Air Blank	0.000	15:14	Control Test	0.050	15:15	Air Blank	0.000	15:15	Control Test	0.050	15:16	Air Blank	0.000	15:17	Control Test	0.048	15:17	Air Blank	0.000	15:18	Control Test Stats			Average	0.0493		Std Dev	0.0012		Rel Std Dev(%)	2.3406		INDIAN RIVER CSO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001328 06/20/2025 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>15:04</td></tr> <tr><td>Control Test</td><td>0.080</td><td>15:04</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:05</td></tr> <tr><td>Control Test</td><td>0.080</td><td>15:06</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:06</td></tr> <tr><td>Control Test</td><td>0.080</td><td>15:07</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:07</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0800</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature <i>ML</i>	Test	g/210L	Time	Air Blank	0.000	15:04	Control Test	0.080	15:04	Air Blank	0.000	15:05	Control Test	0.080	15:06	Air Blank	0.000	15:06	Control Test	0.080	15:07	Air Blank	0.000	15:07	Control Test Stats			Average	0.0800		Std Dev	0.0000		Rel Std Dev(%)	0.0000		INDIAN RIVER CSO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001328 06/20/2025 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>15:26</td></tr> <tr><td>Control Test</td><td>0.199</td><td>15:27</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:28</td></tr> <tr><td>Control Test</td><td>0.198</td><td>15:28</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:29</td></tr> <tr><td>Control Test</td><td>0.197</td><td>15:29</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:30</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1980</td><td></td></tr> <tr><td>Std Dev</td><td>0.0010</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.5051</td><td></td></tr> </tbody> </table> Operator's Signature <i>ML</i>	Test	g/210L	Time	Air Blank	0.000	15:26	Control Test	0.199	15:27	Air Blank	0.000	15:28	Control Test	0.198	15:28	Air Blank	0.000	15:29	Control Test	0.197	15:29	Air Blank	0.000	15:30	Control Test Stats			Average	0.1980		Std Dev	0.0010		Rel Std Dev(%)	0.5051		INDIAN RIVER CSO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001328 06/20/2025 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>15:00</td></tr> <tr><td>Control Test</td><td>0.079</td><td>15:01</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:01</td></tr> <tr><td>Control Test</td><td>0.079</td><td>15:01</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:02</td></tr> <tr><td>Control Test</td><td>0.078</td><td>15:02</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:03</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0787</td><td></td></tr> <tr><td>Std Dev</td><td>0.0016</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7339</td><td></td></tr> </tbody> </table> Operator's Signature <i>ML</i>	Test	g/210L	Time	Air Blank	0.000	15:00	Control Test	0.079	15:01	Air Blank	0.000	15:01	Control Test	0.079	15:01	Air Blank	0.000	15:02	Control Test	0.078	15:02	Air Blank	0.000	15:03	Control Test Stats			Average	0.0787		Std Dev	0.0016		Rel Std Dev(%)	0.7339	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	15:14																																																																																																																																																	
Control Test	0.050	15:15																																																																																																																																																	
Air Blank	0.000	15:15																																																																																																																																																	
Control Test	0.050	15:16																																																																																																																																																	
Air Blank	0.000	15:17																																																																																																																																																	
Control Test	0.048	15:17																																																																																																																																																	
Air Blank	0.000	15:18																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0493																																																																																																																																																		
Std Dev	0.0012																																																																																																																																																		
Rel Std Dev(%)	2.3406																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	15:04																																																																																																																																																	
Control Test	0.080	15:04																																																																																																																																																	
Air Blank	0.000	15:05																																																																																																																																																	
Control Test	0.080	15:06																																																																																																																																																	
Air Blank	0.000	15:06																																																																																																																																																	
Control Test	0.080	15:07																																																																																																																																																	
Air Blank	0.000	15:07																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0800																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	15:26																																																																																																																																																	
Control Test	0.199	15:27																																																																																																																																																	
Air Blank	0.000	15:28																																																																																																																																																	
Control Test	0.198	15:28																																																																																																																																																	
Air Blank	0.000	15:29																																																																																																																																																	
Control Test	0.197	15:29																																																																																																																																																	
Air Blank	0.000	15:30																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1980																																																																																																																																																		
Std Dev	0.0010																																																																																																																																																		
Rel Std Dev(%)	0.5051																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	15:00																																																																																																																																																	
Control Test	0.079	15:01																																																																																																																																																	
Air Blank	0.000	15:01																																																																																																																																																	
Control Test	0.079	15:01																																																																																																																																																	
Air Blank	0.000	15:02																																																																																																																																																	
Control Test	0.078	15:02																																																																																																																																																	
Air Blank	0.000	15:03																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0787																																																																																																																																																		
Std Dev	0.0016																																																																																																																																																		
Rel Std Dev(%)	0.7339																																																																																																																																																		

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: INDIAN RIVER CSO
Time of Inspection: 13:21

Date of Inspection: 06/25/2025

Serial Number: 80-001328
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG429602 Exp: 10/22/2026
0.000	0.047	0.079	0.199	0.079
0.000	0.048	0.079	0.199	0.079
0.000	0.048	0.078	0.199	0.078
0.000	0.048	0.079	0.199	0.078
0.000	0.048	0.079	0.199	0.078
0.000	0.048	0.079	0.199	0.078
0.000	0.048	0.079	0.199	0.079
0.000	0.048	0.079	0.199	0.079
0.000	0.048	0.079	0.199	0.078
0.000	0.048	0.079	0.199	0.078


Standard Deviations	0.0003	0.0003	0.0000	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

06/25/2025
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001328, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001328</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>INDIAN RIVER CSO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>06/25/2025</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>13:21</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Taylor
Gutschow
Digitally signed by Taylor
Gutschow
Date: 2025.06.27 12:09:56
-04'00'

06/25/2025

Date

TAYLOR D GUTSCHOW,
Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality