



INSTRUMENT PROCESSING SHEET

Agency Orange County SOS/N 80-001416Florida Department of
Law EnforcementDate In 01/24/2025 DI Completion Date 02/05/2025☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake By <u>ALL</u> Date <u>01/24/2025</u> ☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____	Quality Checks By <u>DA</u> Date <u>02/04/2025</u> ☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>225</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP102</u> 32 mm <u>0.140</u> (.139 - .169) 36 mm <u>0.156</u> (.156 - .190) 53 mm <u>0.226</u> (.228 - .278) 103 mm <u>0.476</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>30793</u> ☒ Stability Checks <table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP5088</td><td>202406K 06/19/2026</td></tr><tr><td>0.080</td><td>MP5089</td><td>202406L 06/19/2026</td></tr><tr><td>0.200</td><td>MP5090</td><td>202406N 06/20/2026</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG429602 10/22/2026</td></tr></tbody></table>	Simulator	Serial #	Lot #/Exp	0.050	MP5088	202406K 06/19/2026	0.080	MP5089	202406L 06/19/2026	0.200	MP5090	202406N 06/20/2026	0.080 DGS	N/A	AG429602 10/22/2026	Flow Calibration By <u>DA</u> Date <u>02/04/2025</u> Flow Column # <u>ATP105</u> ☒ 5L/min – 17mm ☒ 15L/min – 53mm ☒ 30L/min – 103mm ☒ R-Value <u>231</u> ☒ Post Calibration Verification (L/s) Flow Column # <u>ATP102</u> 32 mm <u>0.144</u> (.139 - .169) 36 mm <u>0.160</u> (.156 - .190) 53 mm <u>0.230</u> (.228 - .278) 103 mm <u>0.492</u> (.447 - .547) Maintenance By _____ Date _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ _____ _____ _____ _____ _____																																	
Simulator	Serial #	Lot #/Exp																																																
0.050	MP5088	202406K 06/19/2026																																																
0.080	MP5089	202406L 06/19/2026																																																
0.200	MP5090	202406N 06/20/2026																																																
0.080 DGS	N/A	AG429602 10/22/2026																																																
Calibration Adjustment By <u>DA</u> Barometric Pressure Gauge <u>1018</u> ID # <u>28427</u> <table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td>MP6289</td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td>MP6290</td><td>24231</td><td>06/25/2026</td></tr><tr><td>0.100</td><td>MP6291</td><td>24110</td><td>03/05/2026</td></tr><tr><td>0.200</td><td>MP6292</td><td>24080</td><td>02/13/2026</td></tr><tr><td>0.300</td><td>MP6293</td><td>24430</td><td>12/10/2026</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>06723080A5</td><td>04/05/2025</td></tr></tbody></table> ☒ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td>MP5088</td><td>202406K</td><td>06/19/2026</td></tr><tr><td>0.080</td><td>MP5089</td><td>202406L</td><td>06/19/2026</td></tr><tr><td>0.200</td><td>MP5090</td><td>202406N</td><td>06/20/2026</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG429602</td><td>10/22/2026</td></tr></tbody></table> Notes/Suggested Service: _____ _____ _____ _____ _____ _____ _____ _____ _____			Simulator	Serial #	Lot #	Expiration	0.000	MP6289	N/A	N/A	0.040	MP6290	24231	06/25/2026	0.100	MP6291	24110	03/05/2026	0.200	MP6292	24080	02/13/2026	0.300	MP6293	24430	12/10/2026	0.080 DGS	N/A	06723080A5	04/05/2025	Simulator	Serial #	Lot #	Expiration	0.050	MP5088	202406K	06/19/2026	0.080	MP5089	202406L	06/19/2026	0.200	MP5090	202406N	06/20/2026	0.080 DGS	N/A	AG429602	10/22/2026
Simulator	Serial #	Lot #	Expiration																																															
0.000	MP6289	N/A	N/A																																															
0.040	MP6290	24231	06/25/2026																																															
0.100	MP6291	24110	03/05/2026																																															
0.200	MP6292	24080	02/13/2026																																															
0.300	MP6293	24430	12/10/2026																																															
0.080 DGS	N/A	06723080A5	04/05/2025																																															
Simulator	Serial #	Lot #	Expiration																																															
0.050	MP5088	202406K	06/19/2026																																															
0.080	MP5089	202406L	06/19/2026																																															
0.200	MP5090	202406N	06/20/2026																																															
0.080 DGS	N/A	AG429602	10/22/2026																																															
Department Inspection By <u>DA</u> Barometric Pressure ID# <u>30793</u> Gauge <u>1019</u> Instrument <u>1018</u> Mouth Alcohol Solution Lot # <u>2024-A</u> Acetone Stock Solution Lot # <u>2024-B</u> <table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP5086</td></tr><tr><td>Interferent</td><td>MP5087</td></tr><tr><td>0.050</td><td>MP5088</td></tr><tr><td>0.080</td><td>MP5089</td></tr><tr><td>0.200</td><td>MP5090</td></tr></tbody></table> Attachments <table border="1" style="width:100%"><tbody><tr><td>☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Calibration Adjustment</td><td>☒ Post-Stability Checks ☒ Flow Calibration ☐ Form 40 ☐ Other _____</td></tr></tbody></table> ☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2025.02.06 10:20:11 -05'00' Shayla Platt Digitally signed by Shayla Platt Date: 2025.02.10 09:26:35 -05'00' Tech Review / Date Admin Review / Date			Simulator	Serial Number	0.000	MP5086	Interferent	MP5087	0.050	MP5088	0.080	MP5089	0.200	MP5090	☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Calibration Adjustment	☒ Post-Stability Checks ☒ Flow Calibration ☐ Form 40 ☐ Other _____																																		
Simulator	Serial Number																																																	
0.000	MP5086																																																	
Interferent	MP5087																																																	
0.050	MP5088																																																	
0.080	MP5089																																																	
0.200	MP5090																																																	
☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Calibration Adjustment	☒ Post-Stability Checks ☒ Flow Calibration ☐ Form 40 ☐ Other _____																																																	

Stability Checks 80-001416 DA 2/4/25

ORANGE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001416
02/04/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:40
Control Test	0.048	11:41
Air Blank	0.000	11:41
Control Test	0.048	11:42
Air Blank	0.000	11:42
Control Test	0.048	11:43
Air Blank	0.000	11:44
Control Test Stats		
Average	0.0480	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	



Operator's Signature

ORANGE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001416
02/04/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:15
Control Test	0.079	11:15
Air Blank	0.000	11:16
Control Test	0.079	11:17
Air Blank	0.000	11:17
Control Test	0.078	11:18
Air Blank	0.000	11:19
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

Wet



Operator's Signature

ORANGE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001416
02/04/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:20
Control Test	0.198	11:20
Air Blank	0.000	11:21
Control Test	0.196	11:22
Air Blank	0.000	11:22
Control Test	0.195	11:23
Air Blank	0.000	11:23
Control Test Stats		
Average	0.1963	
Std Dev	0.0015	
Rel Std Dev(%)	0.7780	



Operator's Signature

ORANGE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001416
02/04/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:25
Control Test	0.076	11:26
Air Blank	0.000	11:26
Control Test	0.076	11:27
Air Blank	0.000	11:27
Control Test	0.077	11:27
Air Blank	0.000	11:28
Control Test Stats		
Average	0.0763	
Std Dev	0.0016	
Rel Std Dev(%)	0.7564	

DGS



Operator's Signature

Flow Rate Calibration Adjustment 80-001416 DA 2/4/25

ORANGE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001416
02/04/2025
Software: 8100.27

Flow Rate Calibration*****

1: Rate (Liters/min) = 5

SQRT(Diff) = 7.000

2: Rate (Liters/min) = 15

SQRT(Diff) = 12.000

3: Rate (Liters/min) = 30

SQRT(Diff) = 21.070

Dependent Data Scale Factor = 100000 L/min

Independent Data Scale Factor = 256

Rounded Slope = 688

Rounded Intercept = -686673

Correlation = 0.99872

Optical Bench Calibration Adjustment

2/5/25 DA

ORANGE COUNTY SO

Intoxilyzer - Alcohol Analyzer

Model 8000

SN 80-001416

02/05/2025

08:52:36

Auto Calibration

pg 1 of 2

<<<<< 3um >>>>>			<<<<< 9um >>>>>		
Solution = 0.000 g/210L or 0.0000 mg/l, Samples = 4, Discarded = 1					
Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)	
Sample #1	0.1430	(-0.0130)	0.1320	(-0.0030)	
Sample #2	0.5670	(0.0260)	0.1870	(0.0090)	
Sample #3	0.1470	(0.4810)	0.1510	(0.0600)	
Sample #4	0.1580	(0.4900)	0.1350	(0.0670)	
Avg % Abs	0.2907	(0.3323)	0.1577	(0.0453)	
STD DEV	0.2394	(0.2653)	0.0266	(0.0317)	
REL STD DEV	82.354	(79.839)	16.892	(69.837)	
Solution = 0.040 g/210L or 0.1905 mg/l, Samples = 4, Discarded = 1					
Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)	
Sample #1	0.5090	(-0.0280)	1.6840	(-0.0190)	
Sample #2	0.8260	(-0.4080)	1.4830	(-0.0260)	
Sample #3	0.8120	(-0.3890)	1.4690	(-0.0040)	
Sample #4	0.8190	(-0.3750)	1.4750	(0.0030)	
Avg % Abs	0.8190	(-0.3907)	1.4757	(-0.0090)	
STD DEV	0.0070	(0.0166)	0.0070	(0.0151)	
REL STD DEV	0.855	(4.240)	0.476	(168.142)	
Solution = 0.100 g/210L or 0.4762 mg/l, Samples = 4, Discarded = 1					
Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)	
Sample #1	1.8390	(-0.0140)	3.4610	(-0.0150)	
Sample #2	1.8630	(0.0010)	3.4690	(-0.0060)	
Sample #3	1.8690	(0.0030)	3.4790	(-0.0180)	
Sample #4	1.8580	(0.0300)	3.4690	(-0.0060)	
Avg % Abs	1.8633	(0.0113)	3.4723	(-0.0100)	
STD DEV	0.0055	(0.0162)	0.0058	(0.0069)	
REL STD DEV	0.296	(142.912)	0.166	(69.282)	
Solution = 0.200 g/210L or 0.9524 mg/l, Samples = 4, Discarded = 1					
Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)	
Sample #1	3.5940	(-0.0200)	6.6890	(-0.0090)	
Sample #2	3.5420	(0.0320)	6.6150	(0.0950)	
Sample #3	3.5140	(0.0540)	6.5840	(0.1040)	
Sample #4	3.4900	(0.0780)	6.5670	(0.1130)	
Avg % Abs	3.5153	(0.0547)	6.5887	(0.1040)	
STD DEV	0.0260	(0.0230)	0.0243	(0.0090)	
REL STD DEV	0.740	(42.086)	0.369	(8.654)	
Solution = 0.300 g/210L or 1.4286 mg/l, Samples = 4, Discarded = 1					
Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)	
Sample #1	5.1780	(-0.0190)	9.6950	(-0.0090)	
Sample #2	5.1260	(0.0620)	9.5590	(0.1500)	
Sample #3	5.1220	(0.0850)	9.5320	(0.1840)	
Sample #4	5.1030	(0.0930)	9.5200	(0.1810)	
Avg % Abs	5.1170	(0.0800)	9.5370	(0.1717)	
STD DEV	0.0123	(0.0161)	0.0200	(0.0188)	
REL STD DEV	0.240	(20.117)	0.209	(10.965)	

XORANGE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001416
02/05/2025 08:52:36

Auto Calibration

pg 2 of 2

<<<< 3um >>>>

Zero Order Coef -712.31
First Order Coef 2954.69
Second Order Coef -5.90

Act (g/210L)	Fit (g/210L)	Residual (g/210L)
0.000	0.003	-0.0031
0.040	0.036	0.0042
0.100	0.100	-0.0002
0.200	0.202	-0.0016
0.300	0.299	0.0007

<<<< 9um >>>>

-199.86
1383.11
14.19

Act (g/210L)	Fit (g/210L)	Residual (g/210L)
0.000	0.000	-0.0004
0.040	0.039	0.0007
0.100	0.100	-0.0003
0.200	0.200	-0.0001
0.300	0.300	0.0001

<<<< 3um >>>> <<<< 9um >>>>

Solution = 0.080 g/210L or 0.3810 mg/l, Samples = 4, Discarded = 1
Sample
Sample #1 2963.00 3343.00
Sample #2 2977.00 3390.00
Sample #3 2977.00 3398.00
Sample #4 3018.00 3408.00
Avg 2990.6667 3398.6667
STD DEV 23.6714 9.0185
REL STD DEV 0.792 0.265
H2O adjust (mg/l*10k) 819 411

Barometric Pressure = 1018

*****CALIBRATION SUCCESSFUL*****

Post-Stability Checks

80- 001414

DA

2/5/25

ORANGE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001416
02/05/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:45
Control Test	0.049	09:46
Air Blank	0.000	09:46
Control Test	0.048	09:47
Air Blank	0.000	09:48
Control Test	0.048	09:48
Air Blank	0.000	09:49
Control Test Stats		
Average	0.0483	
Std Dev	0.0006	
Rel Std Dev(%)	1.1945	

DA

Operator's Signature

ORANGE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001416
02/05/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:51
Control Test	0.079	09:52
Air Blank	0.000	09:53
Control Test	0.079	09:53
Air Blank	0.000	09:54
Control Test	0.079	09:55
Air Blank	0.000	09:55
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Wet

DA

Operator's Signature

ORANGE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001416
02/05/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:58
Control Test	0.202	09:58
Air Blank	0.000	09:59
Control Test	0.200	10:00
Air Blank	0.000	10:00
Control Test	0.200	10:01
Air Blank	0.000	10:02
Control Test Stats		
Average	0.2007	
Std Dev	0.0012	
Rel Std Dev(%)	0.5754	

DA

Operator's Signature

ORANGE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001416
02/05/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:39
Control Test	0.078	10:39
Air Blank	0.000	10:40
Control Test	0.078	10:40
Air Blank	0.000	10:41
Control Test	0.078	10:41
Air Blank	0.000	10:41
Control Test Stats		
Average	0.0780	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS

DA

Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: ORANGE COUNTY SO
Time of Inspection: 13:12

Date of Inspection: 02/05/2025

Serial Number: 80-001416
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG429602 Exp: 10/22/2026
0.000	0.049	0.080	0.202	0.078
0.000	0.048	0.079	0.202	0.077
0.000	0.048	0.079	0.202	0.077
0.000	0.049	0.079	0.201	0.078
0.000	0.049	0.080	0.202	0.078
0.000	0.049	0.080	0.202	0.077
0.000	0.049	0.079	0.201	0.077
0.000	0.049	0.080	0.201	0.078
0.000	0.049	0.080	0.202	0.078
0.000	0.049	0.080	0.202	0.077


Standard Deviations	0.0004	0.0005	0.0004	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.


I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



DESTINEE N ARMSTRONG

Signature and Printed Name

02/05/2025
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001416, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001416</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>ORANGE COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>02/05/2025</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>13:12</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

2A

02/05/2025

Date

DESTINEE N ARMSTRONG,
Department Inspector

FDLE/ATP Form 69 October 2024
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality