



Agency Osceola County SO

S/N 80-001715

Date In 01/24/2025 DI Completion Date 02/04/2025

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Calibration Adjustment

By

☐ Post Calibration Adjustment Stability ChecksNotes/Suggested Service:

Department Inspection

By DA

Attachments

- ☒ Instrument Complies with Chapter 11D-8, FAC
- ☐ Instrument Does Not Comply with Chapter 11D-8, FAC
- ☒ Return to/Place into Evidentiary Use
- ☐ Remain Out of Evidentiary Use
- ☒ Conduct an Agency Inspection Before Evidentiary Use

Taylor
Gutschow

Digitally signed by Taylor
Gutschow
Date: 2025.02.10 10:36:38
+05'00'

Tech Review / Date

Shayla Platt

Admin Review / Date

Stability Checks

80-001715 DA

1/29/25

OSCEOLA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001715
01/29/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:44
Control Test	0.050	14:45
Air Blank	0.000	14:45
Control Test	0.049	14:46
Air Blank	0.000	14:46
Control Test	0.050	14:47
Air Blank	0.000	14:48
Control Test Stats		
Average	0.0497	
Std Dev	0.0006	
Rel Std Dev(%)	1.1625	

DA

Operator's Signature

OSCEOLA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001715
01/29/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:51
Control Test	0.082	14:52
Air Blank	0.000	14:52
Control Test	0.081	14:53
Air Blank	0.000	14:53
Control Test	0.082	14:54
Air Blank	0.000	14:55
Control Test Stats		
Average	0.0817	
Std Dev	0.0006	
Rel Std Dev(%)	0.7070	

Wet

DA

Operator's Signature

OSCEOLA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001715
01/29/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:12
Control Test	0.197	15:13
Air Blank	0.000	15:13
Control Test	0.203	15:14
Air Blank	0.000	15:15
Control Test	0.203	15:15
Air Blank	0.000	15:16
Control Test Stats		
Average	0.2010	
Std Dev	0.0035	
Rel Std Dev(%)	1.7234	

DA

Operator's Signature

OSCEOLA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001715
01/29/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:37
Control Test	0.079	14:38
Air Blank	0.000	14:38
Control Test	0.079	14:38
Air Blank	0.000	14:39
Control Test	0.080	14:39
Air Blank	0.000	14:40
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

DGS

DA

Operator's Signature

Florida Department of Law Enforcement

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: OSCEOLA COUNTY SO

Serial Number: 80-001715

Time of Inspection: 09:52

Date of Inspection: 02/04/2025

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#: 202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#: 202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: AG429602 Exp: 10/22/2026
0.000	0.050	0.081	0.197	0.080
0.000	0.050	0.079	0.200	0.079
0.000	0.049	0.079	0.196	0.079
0.000	0.050	0.082	0.197	0.079
0.000	0.051	0.081	0.197	0.079
0.000	0.050	0.079	0.197	0.079
0.000	0.051	0.080	0.199	0.080
0.000	0.050	0.082	0.198	0.079
0.000	0.051	0.079	0.198	0.079
0.000	0.050	0.080	0.198	0.079

Standard Deviations	0.0006	0.0012	0.0011	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0008

Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



DESTINEE N ARMSTRONG

Signature and Printed Name

02/04/2025

Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001715 , manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001715</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>OSCEOLA COUNTY SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>02/04/2025</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>09:52</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Destinee
Armstrong

Digitally signed by Destinee
Armstrong
Date: 2025.02.04 13:03:35
-05'00'

02/04/2025

Date

DESTINEE N ARMSTRONG,
Department Inspector

FDLE/ATP Form 69 October 2024

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality