







INSTRUMENT PROCESSING SHEET

Agency Palm Beach CSOS/N 80-001741Florida Department of
Law EnforcementDate In 02/03/2025DI Completion Date 02/14/2025☐ Ship☒ P/U☐ H/D☐ CMI☐ EE

Intake	By TDG	Date	Quality Checks	By TDG	Date	Flow Calibration	By	Date																												
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: <u>Dropped off. No box.</u>		<u>02/03/2025</u>	<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>224</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP101</u> 32 mm <u>0.156</u> (.139 - .169) 36 mm <u>0.175</u> (.156 - .190) 53 mm <u>0.242</u> (.228 - .278) 103 mm <u>0.500</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>26932</u> <input checked="" type="checkbox"/> Stability Checks		<u>02/03/2025</u>	Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																														
			<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6286</td><td>202303K 03/29/2025</td></tr><tr><td>0.080</td><td>MP6287</td><td>202303L 03/29/2025</td></tr><tr><td>0.200</td><td>MP6288</td><td>202304C 04/05/2025</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG429602 10/22/2026</td></tr></tbody></table>	Simulator	Serial #	Lot #/Exp	0.050	MP6286	202303K 03/29/2025	0.080	MP6287	202303L 03/29/2025	0.200	MP6288	202304C 04/05/2025	0.080 DGS	N/A	AG429602 10/22/2026																		
Simulator	Serial #	Lot #/Exp																																		
0.050	MP6286	202303K 03/29/2025																																		
0.080	MP6287	202303L 03/29/2025																																		
0.200	MP6288	202304C 04/05/2025																																		
0.080 DGS	N/A	AG429602 10/22/2026																																		
			<table border="1"><thead><tr><th colspan="2">Maintenance By</th><th>Date</th></tr></thead><tbody><tr><td><input type="checkbox"/> Battery Replacement</td><td></td><td></td></tr><tr><td><input type="checkbox"/> Dry Gas Regulator Replacement</td><td></td><td></td></tr><tr><td><input type="checkbox"/> Breath Tube Replacement</td><td></td><td></td></tr><tr><td><input type="checkbox"/> Other</td><td></td><td></td></tr></tbody></table>						Maintenance By		Date	<input type="checkbox"/> Battery Replacement			<input type="checkbox"/> Dry Gas Regulator Replacement			<input type="checkbox"/> Breath Tube Replacement			<input type="checkbox"/> Other															
Maintenance By		Date																																		
<input type="checkbox"/> Battery Replacement																																				
<input type="checkbox"/> Dry Gas Regulator Replacement																																				
<input type="checkbox"/> Breath Tube Replacement																																				
<input type="checkbox"/> Other																																				
Calibration Adjustment			Department Inspection																																	
By <u>TDG</u>			By <u>TDG</u>																																	
Barometric Pressure Gauge <u>1022</u> ID # <u>28199</u>			Barometric Pressure ID# <u>26932</u>																																	
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td>MP5097</td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td>MP5098</td><td>23400</td><td>10/24/2025</td></tr><tr><td>0.100</td><td>MP5099</td><td>24110</td><td>03/05/2026</td></tr><tr><td>0.200</td><td>MP5100</td><td>24080</td><td>02/13/2026</td></tr><tr><td>0.300</td><td>MP5101</td><td>23410</td><td>11/01/2025</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>06723080A5</td><td>04/05/2025</td></tr></tbody></table>			Simulator	Serial #	Lot #	Expiration	0.000	MP5097	N/A	N/A	0.040	MP5098	23400	10/24/2025	0.100	MP5099	24110	03/05/2026	0.200	MP5100	24080	02/13/2026	0.300	MP5101	23410	11/01/2025	0.080 DGS	N/A	06723080A5	04/05/2025	Gauge <u>1027</u> Instrument <u>1027</u> Mouth Alcohol Solution Lot # <u>2024-A</u> Acetone Stock Solution Lot # <u>2023-B</u>					
Simulator	Serial #	Lot #	Expiration																																	
0.000	MP5097	N/A	N/A																																	
0.040	MP5098	23400	10/24/2025																																	
0.100	MP5099	24110	03/05/2026																																	
0.200	MP5100	24080	02/13/2026																																	
0.300	MP5101	23410	11/01/2025																																	
0.080 DGS	N/A	06723080A5	04/05/2025																																	
<input checked="" type="checkbox"/> Post Calibration Adjustment Stability Checks			<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6284</td></tr><tr><td>Interferent</td><td>MP6285</td></tr><tr><td>0.050</td><td>MP6286</td></tr><tr><td>0.080</td><td>MP6287</td></tr><tr><td>0.200</td><td>MP6288</td></tr></tbody></table>						Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288																
Simulator	Serial Number																																			
0.000	MP6284																																			
Interferent	MP6285																																			
0.050	MP6286																																			
0.080	MP6287																																			
0.200	MP6288																																			
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td>MP6286</td><td>202303K</td><td>03/29/2025</td></tr><tr><td>0.080</td><td>MP6287</td><td>202303L</td><td>03/29/2025</td></tr><tr><td>0.200</td><td>MP6288</td><td>202304C</td><td>04/05/2025</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG429602</td><td>10/22/2026</td></tr></tbody></table>			Simulator	Serial #	Lot #	Expiration	0.050	MP6286	202303K	03/29/2025	0.080	MP6287	202303L	03/29/2025	0.200	MP6288	202304C	04/05/2025	0.080 DGS	N/A	AG429602	10/22/2026	Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Calibration Certificate <input checked="" type="checkbox"/> Calibration Adjustment <input checked="" type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Form 40 <input type="checkbox"/> Other _____													
Simulator	Serial #	Lot #	Expiration																																	
0.050	MP6286	202303K	03/29/2025																																	
0.080	MP6287	202303L	03/29/2025																																	
0.200	MP6288	202304C	04/05/2025																																	
0.080 DGS	N/A	AG429602	10/22/2026																																	
Notes/Suggested Service: Instrument will receive an optical cal adjust. Baro gauge read 1022; instrument read 1012. (TDG 2/3/2025)			<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use																																	
			Destinee <small>Digitally signed by Destinee Armstrong Date: 2025.02.21 11:13:55 -05'00'</small> Phil Nicodemo <small>Digitally signed by Phil Nicodemo Date: 2025.02.25 09:24:12 -05'00'</small>																																	
			Tech Review / Date Admin Review / Date																																	

Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																				
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																				
✓	✓	✓	✓																																																																																																																																				
<p>PALM BEACH COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 02/23/2025 Software: 8100.27</p> <p>SN 80-001741</p> <p>Test Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>10:12</td></tr> <tr><td>Control Test</td><td>0.050</td><td>10:13</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:13</td></tr> <tr><td>Control Test</td><td>0.050</td><td>10:14</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:15</td></tr> <tr><td>Control Test</td><td>0.050</td><td>10:15</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:16</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.050</td><td></td></tr> <tr><td>Std Dev</td><td>0.000</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.000</td><td></td></tr> </table>	Air Blank	0.000	10:12	Control Test	0.050	10:13	Air Blank	0.000	10:13	Control Test	0.050	10:14	Air Blank	0.000	10:15	Control Test	0.050	10:15	Air Blank	0.000	10:16	Control Test Stats			Average	0.050		Std Dev	0.000		Rel. Std Dev(%)	0.000		<p>PALM BEACH COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 02/23/2025 Software: 8100.27</p> <p>SN 80-001741</p> <p>Test Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>09:57</td></tr> <tr><td>Control Test</td><td>0.080</td><td>09:58</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:58</td></tr> <tr><td>Control Test</td><td>0.079</td><td>09:59</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:59</td></tr> <tr><td>Control Test</td><td>0.079</td><td>09:59</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:00</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.079</td><td></td></tr> <tr><td>Std Dev</td><td>0.006</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.7277</td><td></td></tr> </table>	Air Blank	0.000	09:57	Control Test	0.080	09:58	Air Blank	0.000	09:58	Control Test	0.079	09:59	Air Blank	0.000	09:59	Control Test	0.079	09:59	Air Blank	0.000	10:00	Control Test Stats			Average	0.079		Std Dev	0.006		Rel. Std Dev(%)	0.7277		<p>PALM BEACH COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 02/23/2025 Software: 8100.27</p> <p>SN 80-001741</p> <p>Test Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>10:05</td></tr> <tr><td>Control Test</td><td>0.195</td><td>10:06</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:06</td></tr> <tr><td>Control Test</td><td>0.195</td><td>10:07</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:07</td></tr> <tr><td>Control Test</td><td>0.195</td><td>10:08</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:09</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.195</td><td></td></tr> <tr><td>Std Dev</td><td>0.000</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.000</td><td></td></tr> </table>	Air Blank	0.000	10:05	Control Test	0.195	10:06	Air Blank	0.000	10:06	Control Test	0.195	10:07	Air Blank	0.000	10:07	Control Test	0.195	10:08	Air Blank	0.000	10:09	Control Test Stats			Average	0.195		Std Dev	0.000		Rel. Std Dev(%)	0.000		<p>09:08 02/23/2025</p> <p>PCS</p> <p>PALM BEACH COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 02/23/2025 Software: 8100.27</p> <p>SN 80-001741</p> <p>Test Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>09:49</td></tr> <tr><td>Control Test</td><td>0.079</td><td>09:49</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:50</td></tr> <tr><td>Control Test</td><td>0.079</td><td>09:50</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:51</td></tr> <tr><td>Control Test</td><td>0.080</td><td>09:51</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>09:51</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0793</td><td></td></tr> <tr><td>Std Dev</td><td>0.006</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.7277</td><td></td></tr> </table>	Air Blank	0.000	09:49	Control Test	0.079	09:49	Air Blank	0.000	09:50	Control Test	0.079	09:50	Air Blank	0.000	09:51	Control Test	0.080	09:51	Air Blank	0.000	09:51	Control Test Stats			Average	0.0793		Std Dev	0.006		Rel. Std Dev(%)	0.7277	
Air Blank	0.000	10:12																																																																																																																																					
Control Test	0.050	10:13																																																																																																																																					
Air Blank	0.000	10:13																																																																																																																																					
Control Test	0.050	10:14																																																																																																																																					
Air Blank	0.000	10:15																																																																																																																																					
Control Test	0.050	10:15																																																																																																																																					
Air Blank	0.000	10:16																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.050																																																																																																																																						
Std Dev	0.000																																																																																																																																						
Rel. Std Dev(%)	0.000																																																																																																																																						
Air Blank	0.000	09:57																																																																																																																																					
Control Test	0.080	09:58																																																																																																																																					
Air Blank	0.000	09:58																																																																																																																																					
Control Test	0.079	09:59																																																																																																																																					
Air Blank	0.000	09:59																																																																																																																																					
Control Test	0.079	09:59																																																																																																																																					
Air Blank	0.000	10:00																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.079																																																																																																																																						
Std Dev	0.006																																																																																																																																						
Rel. Std Dev(%)	0.7277																																																																																																																																						
Air Blank	0.000	10:05																																																																																																																																					
Control Test	0.195	10:06																																																																																																																																					
Air Blank	0.000	10:06																																																																																																																																					
Control Test	0.195	10:07																																																																																																																																					
Air Blank	0.000	10:07																																																																																																																																					
Control Test	0.195	10:08																																																																																																																																					
Air Blank	0.000	10:09																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.195																																																																																																																																						
Std Dev	0.000																																																																																																																																						
Rel. Std Dev(%)	0.000																																																																																																																																						
Air Blank	0.000	09:49																																																																																																																																					
Control Test	0.079	09:49																																																																																																																																					
Air Blank	0.000	09:50																																																																																																																																					
Control Test	0.079	09:50																																																																																																																																					
Air Blank	0.000	09:51																																																																																																																																					
Control Test	0.080	09:51																																																																																																																																					
Air Blank	0.000	09:51																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0793																																																																																																																																						
Std Dev	0.006																																																																																																																																						
Rel. Std Dev(%)	0.7277																																																																																																																																						
Operator's Signature 	Operator's Signature 	Operator's Signature 	Operator's Signature 																																																																																																																																				

PMU BERCH COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
02/13/2025
SN 80-001741
09:24:16

Auto Calibration
Max Power Res Value = 56
Auto Range Res Value = 37

Auto Calibration
Max Power Res Value = 56
Auto Range Res Value = 37

Auto Calibration
Max Power Res Value = 56
Auto Range Res Value = 37

Auto Calibration
Max Power Res Value = 56
Auto Range Res Value = 37

Auto Calibration
Max Power Res Value = 56
Auto Range Res Value = 37

Auto Calibration
Max Power Res Value = 56
Auto Range Res Value = 37

Auto Calibration
Max Power Res Value = 56
Auto Range Res Value = 37

Auto Calibration
Max Power Res Value = 56
Auto Range Res Value = 37

Auto Calibration
Max Power Res Value = 56
Auto Range Res Value = 37

Auto Calibration
Max Power Res Value = 56
Auto Range Res Value = 37

Auto Calibration
Max Power Res Value = 56
Auto Range Res Value = 37

Auto Calibration
Max Power Res Value = 56
Auto Range Res Value = 37

Auto Calibration
Max Power Res Value = 56
Auto Range Res Value = 37

Auto Calibration
Max Power Res Value = 56
Auto Range Res Value = 37

Auto Calibration
Max Power Res Value = 56
Auto Range Res Value = 37

Auto Calibration
Max Power Res Value = 56
Auto Range Res Value = 37

Auto Calibration
Max Power Res Value = 56
Auto Range Res Value = 37





Auto Calibration
Max Power Res Value = 56
Auto Range Res Value = 37

Auto Calibration
Max Power Res Value = 56
Auto Range Res Value = 37

Auto Calibration
Max Power Res Value = 56
Auto Range Res Value = 37

Optical Calibration
Adjustment
By: TDG

Post-Cal Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																																
<p>  Operator's Signature </p> <p> PALM BEACH COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001741 02/13/2025 Software: 8100.27 </p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:04</td></tr> <tr><td>Control Test</td><td>0.050</td><td>11:05</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:05</td></tr> <tr><td>Control Test</td><td>0.050</td><td>11:06</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:06</td></tr> <tr><td>Control Test</td><td>0.050</td><td>11:07</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:08</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0500</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	11:04	Control Test	0.050	11:05	Air Blank	0.000	11:05	Control Test	0.050	11:06	Air Blank	0.000	11:06	Control Test	0.050	11:07	Air Blank	0.000	11:08	Control Test Stats			Average	0.0500		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>  Operator's Signature </p> <p> PALM BEACH COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001741 02/13/2025 Software: 8100.27 </p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:14</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:14</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:15</td></tr> <tr><td>Control Test</td><td>0.080</td><td>11:15</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:16</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:17</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:17</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0793</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7277</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	11:14	Control Test	0.079	11:14	Air Blank	0.000	11:15	Control Test	0.080	11:15	Air Blank	0.000	11:16	Control Test	0.079	11:17	Air Blank	0.000	11:17	Control Test Stats			Average	0.0793		Std Dev	0.0006		Rel Std Dev(%)	0.7277		<p>  Operator's Signature </p> <p> PALM BEACH COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001741 02/13/2025 Software: 8100.27 </p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:20</td></tr> <tr><td>Control Test</td><td>0.197</td><td>11:21</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:21</td></tr> <tr><td>Control Test</td><td>0.197</td><td>11:22</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:23</td></tr> <tr><td>Control Test</td><td>0.197</td><td>11:23</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:24</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1970</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	11:20	Control Test	0.197	11:21	Air Blank	0.000	11:21	Control Test	0.197	11:22	Air Blank	0.000	11:23	Control Test	0.197	11:23	Air Blank	0.000	11:24	Control Test Stats			Average	0.1970		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>  Operator's Signature </p> <p> PALM BEACH COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001741 02/13/2025 Software: 8100.27 </p> <p style="text-align: center;">DGS</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>10:55</td></tr> <tr><td>Control Test</td><td>0.080</td><td>10:55</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:56</td></tr> <tr><td>Control Test</td><td>0.080</td><td>10:56</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:56</td></tr> <tr><td>Control Test</td><td>0.080</td><td>10:57</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:57</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0800</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	10:55	Control Test	0.080	10:55	Air Blank	0.000	10:56	Control Test	0.080	10:56	Air Blank	0.000	10:56	Control Test	0.080	10:57	Air Blank	0.000	10:57	Control Test Stats			Average	0.0800		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:04																																																																																																																																																	
Control Test	0.050	11:05																																																																																																																																																	
Air Blank	0.000	11:05																																																																																																																																																	
Control Test	0.050	11:06																																																																																																																																																	
Air Blank	0.000	11:06																																																																																																																																																	
Control Test	0.050	11:07																																																																																																																																																	
Air Blank	0.000	11:08																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0500																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:14																																																																																																																																																	
Control Test	0.079	11:14																																																																																																																																																	
Air Blank	0.000	11:15																																																																																																																																																	
Control Test	0.080	11:15																																																																																																																																																	
Air Blank	0.000	11:16																																																																																																																																																	
Control Test	0.079	11:17																																																																																																																																																	
Air Blank	0.000	11:17																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0793																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7277																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:20																																																																																																																																																	
Control Test	0.197	11:21																																																																																																																																																	
Air Blank	0.000	11:21																																																																																																																																																	
Control Test	0.197	11:22																																																																																																																																																	
Air Blank	0.000	11:23																																																																																																																																																	
Control Test	0.197	11:23																																																																																																																																																	
Air Blank	0.000	11:24																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1970																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:55																																																																																																																																																	
Control Test	0.080	10:55																																																																																																																																																	
Air Blank	0.000	10:56																																																																																																																																																	
Control Test	0.080	10:56																																																																																																																																																	
Air Blank	0.000	10:56																																																																																																																																																	
Control Test	0.080	10:57																																																																																																																																																	
Air Blank	0.000	10:57																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0800																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PALM BEACH COUNTY SO
Time of Inspection: 12:30

Date of Inspection: 02/14/2025

Serial Number: 80-001741
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG429602 Exp: 10/22/2026
0.000	0.049	0.079	0.197	0.079
0.000	0.050	0.079	0.197	0.080
0.000	0.051	0.079	0.197	0.080
0.000	0.051	0.079	0.196	0.080
0.000	0.051	0.079	0.197	0.079
0.000	0.052	0.079	0.197	0.080
0.000	0.051	0.079	0.197	0.080
0.000	0.051	0.080	0.197	0.079
0.000	0.051	0.079	0.198	0.079
0.000	0.051	0.079	0.197	0.080


Standard Deviations	0.0007	0.0003	0.0004	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

02/14/2025
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001741, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001741</u>	UNCERTAINTY* \pm
Owning Agency:	<u>PALM BEACH COUNTY SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>02/14/2025</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>12:30</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

02/14/2025

Date

TAYLOR D GUTSCHOW,
Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality