



Agency Hialeah PD

S/N 80-002462

Date In 01/30/2025 DI Completion Date 02/14/2025

☒ Ship ☐ P/U ☐ H/D ☒ CMI ☐ EE

TDG 3/21/25

Calibration Adjustment

By TDG

Notes/Suggested Service: Noticed the preliminary results of the 0.20 Test were bouncing around. Ran additional stabilities after the inspection and decided to perform a discretionary optical cal adjust. The second optical and inspection used the same sims, standards, and gauges as their respective firsts. Additional stabilities were run after the second optical. (TDG)(2/14/25) TDG 3/21/25

Department Inspection

By TDG

Admin Review / Date

FDLE/ATP Form 48 January 2024
Issuing Authority: Alcohol Testing Program

Page 1 of 1

Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																				
<p>0.047 to 0.053</p> <p>✓</p>	<p>0.077 to 0.083</p> <p>✓</p>	<p>0.194 to 0.206</p> <p>✓</p>	<p>0.077 to 0.083</p> <p>✓</p> <p>≤0.003 of Wet</p> <p>✓</p>																																																																																																																																				
<p>HiA/EH PD Intoxilyzer - Alcohol Analyzer Model 8000 02/07/2025 Software: 8100.27</p> <p>SN 80-002462</p> <p>Test g/210L Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>11:49</td></tr> <tr><td>Control Test</td><td>0.048</td><td>11:50</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:50</td></tr> <tr><td>Control Test</td><td>0.049</td><td>11:51</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:51</td></tr> <tr><td>Control Test</td><td>0.049</td><td>11:52</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:52</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0487</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1863</td><td></td></tr> </table> <p>Operator's Signature <i>ML</i></p>	Air Blank	0.000	11:49	Control Test	0.048	11:50	Air Blank	0.000	11:50	Control Test	0.049	11:51	Air Blank	0.000	11:51	Control Test	0.049	11:52	Air Blank	0.000	11:52	Control Test Stats			Average	0.0487		Std Dev	0.0006		Rel Std Dev(%)	1.1863		<p>HiA/EH PD Intoxilyzer - Alcohol Analyzer Model 8000 02/07/2025 Software: 8100.27</p> <p>SN 80-002462</p> <p>Test g/210L Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>11:20</td></tr> <tr><td>Control Test</td><td>0.078</td><td>11:20</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:21</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:21</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:21</td></tr> <tr><td>Control Test</td><td>0.078</td><td>11:22</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:22</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0783</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7370</td><td></td></tr> </table> <p>Operator's Signature <i>ML</i></p>	Air Blank	0.000	11:20	Control Test	0.078	11:20	Air Blank	0.000	11:21	Control Test	0.079	11:21	Air Blank	0.000	11:21	Control Test	0.078	11:22	Air Blank	0.000	11:22	Control Test Stats			Average	0.0783		Std Dev	0.0006		Rel Std Dev(%)	0.7370		<p>HiA/EH PD Intoxilyzer - Alcohol Analyzer Model 8000 02/07/2025 Software: 8100.27</p> <p>SN 80-002462</p> <p>Test g/210L Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>11:39</td></tr> <tr><td>Control Test</td><td>0.196</td><td>11:39</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:40</td></tr> <tr><td>Control Test</td><td>0.196</td><td>11:40</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:41</td></tr> <tr><td>Control Test</td><td>0.196</td><td>11:42</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:42</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1960</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </table> <p>Operator's Signature <i>ML</i></p>	Air Blank	0.000	11:39	Control Test	0.196	11:39	Air Blank	0.000	11:40	Control Test	0.196	11:40	Air Blank	0.000	11:41	Control Test	0.196	11:42	Air Blank	0.000	11:42	Control Test Stats			Average	0.1960		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>HiA/EH PD Intoxilyzer - Alcohol Analyzer Model 8000 02/07/2025 Software: 8100.27</p> <p>SN 80-002462</p> <p>Test g/210L Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>11:30</td></tr> <tr><td>Control Test</td><td>0.078</td><td>11:31</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:32</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:32</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:33</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:34</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:34</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0787</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7339</td><td></td></tr> </table> <p>Operator's Signature <i>ML</i></p>	Air Blank	0.000	11:30	Control Test	0.078	11:31	Air Blank	0.000	11:32	Control Test	0.079	11:32	Air Blank	0.000	11:33	Control Test	0.079	11:34	Air Blank	0.000	11:34	Control Test Stats			Average	0.0787		Std Dev	0.0006		Rel Std Dev(%)	0.7339	
Air Blank	0.000	11:49																																																																																																																																					
Control Test	0.048	11:50																																																																																																																																					
Air Blank	0.000	11:50																																																																																																																																					
Control Test	0.049	11:51																																																																																																																																					
Air Blank	0.000	11:51																																																																																																																																					
Control Test	0.049	11:52																																																																																																																																					
Air Blank	0.000	11:52																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0487																																																																																																																																						
Std Dev	0.0006																																																																																																																																						
Rel Std Dev(%)	1.1863																																																																																																																																						
Air Blank	0.000	11:20																																																																																																																																					
Control Test	0.078	11:20																																																																																																																																					
Air Blank	0.000	11:21																																																																																																																																					
Control Test	0.079	11:21																																																																																																																																					
Air Blank	0.000	11:21																																																																																																																																					
Control Test	0.078	11:22																																																																																																																																					
Air Blank	0.000	11:22																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0783																																																																																																																																						
Std Dev	0.0006																																																																																																																																						
Rel Std Dev(%)	0.7370																																																																																																																																						
Air Blank	0.000	11:39																																																																																																																																					
Control Test	0.196	11:39																																																																																																																																					
Air Blank	0.000	11:40																																																																																																																																					
Control Test	0.196	11:40																																																																																																																																					
Air Blank	0.000	11:41																																																																																																																																					
Control Test	0.196	11:42																																																																																																																																					
Air Blank	0.000	11:42																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.1960																																																																																																																																						
Std Dev	0.0000																																																																																																																																						
Rel Std Dev(%)	0.0000																																																																																																																																						
Air Blank	0.000	11:30																																																																																																																																					
Control Test	0.078	11:31																																																																																																																																					
Air Blank	0.000	11:32																																																																																																																																					
Control Test	0.079	11:32																																																																																																																																					
Air Blank	0.000	11:33																																																																																																																																					
Control Test	0.079	11:34																																																																																																																																					
Air Blank	0.000	11:34																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0787																																																																																																																																						
Std Dev	0.0006																																																																																																																																						
Rel Std Dev(%)	0.7339																																																																																																																																						

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: HIALEAH PD
Time of Inspection: 11:46

Date of Inspection: 02/10/2025

Serial Number: 80-002462
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG429602 Exp: 10/22/2026
0.000	0.049	0.078	0.191	0.078
0.000	0.048	0.077	0.196	0.078
0.000	0.049	0.079	0.196	0.078
0.000	0.048	0.077	0.196	0.079
0.000	0.049	0.079	0.196	0.078
0.000	0.049	0.079	0.196	0.078
0.000	0.049	0.079	0.196	0.078
0.000	0.049	0.079	0.192	0.079
0.000	0.049	0.078	0.196	0.079
0.000	0.049	0.079	0.196	0.079


Standard Deviations	0.0004	0.0008	0.0019	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0009 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 TAYLOR D GUTSCHOW
Signature and Printed Name

02/10/2025
Date

HIALEAH PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-002462
02/13/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:03
Control Test	0.195	12:04
Air Blank	0.000	12:04
Air Blank	0.196	12:05
Control Test	0.195	12:05
Air Blank	0.000	12:07
Control Test	0.195	12:07
Air Blank	0.000	12:08
Control Test	0.192	12:08
Air Blank	0.000	12:09
Control Test	0.196	12:10
Air Blank	0.000	12:10
Control Test	0.192	12:11
Air Blank	0.000	12:11
Control Test	0.195	12:12
Air Blank	0.000	12:13
Control Test	0.195	12:13
Air Blank	0.000	12:14
Control Test	0.195	12:14
Air Blank	0.000	12:15
Control Test	0.195	12:16
Air Blank	0.000	12:16
Control Test	0.195	12:17
Air Blank	0.000	12:17
Control Test	0.196	12:18
Air Blank	0.000	12:19
Control Test	0.195	12:19
Air Blank	0.000	12:20
Control Test	0.195	12:21
Air Blank	0.000	12:21
Control Test	0.193	12:22
Air Blank	0.000	12:22
Control Test	0.192	12:23
Air Blank	0.000	12:23
Control Test	0.190	12:24
Air Blank	0.000	12:25
Control Test	0.195	12:25
Air Blank	0.000	12:26
Control Test	0.194	12:27
Air Blank	0.000	12:27
Control Test Stats		
Average	0.1942	
Std Dev	0.0016	
Rel Std Dev(%)	0.8161	

Extra 0.200 Stabilities

2/10/25 and 2/13/25 ML

<-- Non-visible results are 0.195 and 0.196. Could not obtain better scan. TDG 2/17/2025





HIALEAH PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-002462
02/13/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:19
Control Test	0.195	12:20
Air Blank	0.000	12:20
Control Test	0.194	12:21
Air Blank	0.000	12:22
Control Test	0.195	12:22
Air Blank	0.000	12:23
Control Test	0.198	12:24
Air Blank	0.000	12:24
Control Test	0.193	12:25
Air Blank	0.000	12:25
Control Test	0.198	12:26
Air Blank	0.000	12:27
Control Test	0.193	12:27
Air Blank	0.000	12:28
Control Test	0.193	12:28
Air Blank	0.000	12:29
Control Test	0.198	12:30
Air Blank	0.000	12:30
Control Test	0.194	12:31
Air Blank	0.000	12:31
Control Test	0.194	12:32
Air Blank	0.000	12:33
Control Test	0.198	12:33
Air Blank	0.000	12:34
Control Test	0.194	12:34
Air Blank	0.000	12:35
Control Test	0.198	12:36
Air Blank	0.000	12:36
Control Test	0.195	12:37
Air Blank	0.000	12:37
Control Test	0.195	12:38
Air Blank	0.000	12:39
Control Test	0.195	12:39
Air Blank	0.000	12:40
Control Test	0.194	12:40
Air Blank	0.000	12:41
Control Test	0.193	12:42
Air Blank	0.000	12:42
Control Test	0.194	12:43
Air Blank	0.000	12:43
Control Test Stats		
Average	0.1951	
Std Dev	0.0019	
Rel Std Dev(%)	0.9624	

ML
Operator's Signature

ML
Operator's Signature

Post-Cal Stability Checks #1

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083																																																																																																																																				
<p>HALEPH PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-002462 12/13/2025 Software: 8100.27</p> <p>Test g/210L Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>10:37</td></tr> <tr><td>Control Test</td><td>0.050</td><td>10:38</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:38</td></tr> <tr><td>Control Test</td><td>0.051</td><td>10:39</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:39</td></tr> <tr><td>Control Test</td><td>0.050</td><td>10:40</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:41</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0503</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>1.1471</td><td></td></tr> </table> <p>Operator's Signature </p>	Air Blank	0.000	10:37	Control Test	0.050	10:38	Air Blank	0.000	10:38	Control Test	0.051	10:39	Air Blank	0.000	10:39	Control Test	0.050	10:40	Air Blank	0.000	10:41	Control Test Stats			Average	0.0503		Std Dev	0.0006		Rel. Std Dev(%)	1.1471		<p>HALEPH PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-002462 12/13/2025 Software: 8100.27</p> <p>Test g/210L Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>10:46</td></tr> <tr><td>Control Test</td><td>0.080</td><td>10:46</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:47</td></tr> <tr><td>Control Test</td><td>0.080</td><td>10:48</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:48</td></tr> <tr><td>Control Test</td><td>0.079</td><td>10:49</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:49</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0797</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.7247</td><td></td></tr> </table> <p>Operator's Signature </p>	Air Blank	0.000	10:46	Control Test	0.080	10:46	Air Blank	0.000	10:47	Control Test	0.080	10:48	Air Blank	0.000	10:48	Control Test	0.079	10:49	Air Blank	0.000	10:49	Control Test Stats			Average	0.0797		Std Dev	0.0006		Rel. Std Dev(%)	0.7247		<p>HALEPH PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-002462 12/13/2025 Software: 8100.27</p> <p>Test g/210L Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>10:53</td></tr> <tr><td>Control Test</td><td>0.198</td><td>10:54</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:55</td></tr> <tr><td>Control Test</td><td>0.199</td><td>10:55</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:56</td></tr> <tr><td>Control Test</td><td>0.192</td><td>10:56</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:57</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1963</td><td></td></tr> <tr><td>Std Dev</td><td>0.0038</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>1.9283</td><td></td></tr> </table> <p>Operator's Signature </p>	Air Blank	0.000	10:53	Control Test	0.198	10:54	Air Blank	0.000	10:55	Control Test	0.199	10:55	Air Blank	0.000	10:56	Control Test	0.192	10:56	Air Blank	0.000	10:57	Control Test Stats			Average	0.1963		Std Dev	0.0038		Rel. Std Dev(%)	1.9283		<p>HALEPH PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-002462 12/13/2025 Software: 8100.27</p> <p>Test g/210L Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>10:27</td></tr> <tr><td>Control Test</td><td>0.081</td><td>10:27</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:27</td></tr> <tr><td>Control Test</td><td>0.081</td><td>10:28</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:28</td></tr> <tr><td>Control Test</td><td>0.082</td><td>10:28</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:29</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0813</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.7099</td><td></td></tr> </table> <p>Operator's Signature </p>	Air Blank	0.000	10:27	Control Test	0.081	10:27	Air Blank	0.000	10:27	Control Test	0.081	10:28	Air Blank	0.000	10:28	Control Test	0.082	10:28	Air Blank	0.000	10:29	Control Test Stats			Average	0.0813		Std Dev	0.0006		Rel. Std Dev(%)	0.7099	
Air Blank	0.000	10:37																																																																																																																																					
Control Test	0.050	10:38																																																																																																																																					
Air Blank	0.000	10:38																																																																																																																																					
Control Test	0.051	10:39																																																																																																																																					
Air Blank	0.000	10:39																																																																																																																																					
Control Test	0.050	10:40																																																																																																																																					
Air Blank	0.000	10:41																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0503																																																																																																																																						
Std Dev	0.0006																																																																																																																																						
Rel. Std Dev(%)	1.1471																																																																																																																																						
Air Blank	0.000	10:46																																																																																																																																					
Control Test	0.080	10:46																																																																																																																																					
Air Blank	0.000	10:47																																																																																																																																					
Control Test	0.080	10:48																																																																																																																																					
Air Blank	0.000	10:48																																																																																																																																					
Control Test	0.079	10:49																																																																																																																																					
Air Blank	0.000	10:49																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0797																																																																																																																																						
Std Dev	0.0006																																																																																																																																						
Rel. Std Dev(%)	0.7247																																																																																																																																						
Air Blank	0.000	10:53																																																																																																																																					
Control Test	0.198	10:54																																																																																																																																					
Air Blank	0.000	10:55																																																																																																																																					
Control Test	0.199	10:55																																																																																																																																					
Air Blank	0.000	10:56																																																																																																																																					
Control Test	0.192	10:56																																																																																																																																					
Air Blank	0.000	10:57																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.1963																																																																																																																																						
Std Dev	0.0038																																																																																																																																						
Rel. Std Dev(%)	1.9283																																																																																																																																						
Air Blank	0.000	10:27																																																																																																																																					
Control Test	0.081	10:27																																																																																																																																					
Air Blank	0.000	10:27																																																																																																																																					
Control Test	0.081	10:28																																																																																																																																					
Air Blank	0.000	10:28																																																																																																																																					
Control Test	0.082	10:28																																																																																																																																					
Air Blank	0.000	10:29																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0813																																																																																																																																						
Std Dev	0.0006																																																																																																																																						
Rel. Std Dev(%)	0.7099																																																																																																																																						

H10LEAH 20

Intoxilyzer - Alcohol Analyzer

Model 8000

02/13/2025

SN

60-002462

10:56:47

Auto Calibration

Max Power Res Value = 92

Auto Range Res Value = 64

Sol Value = 0.000 g/210L ***

Fit Value = 0.0000 mg/l %%%

Samples Taken = 4, Discarded = 1

Sum Io = 12804, Sum Io = 13077

**** CHANNEL 1 ****

Sample % Abs (% Abs Ref)

Sample #1 = 0.1050 (-0.0350)

Sample #2 = 0.0710 (-0.0140)

Sample #3 = 0.1030 (-0.0260)

Sample #4 = 0.1800 (-0.0520)

Avg % Abs = 0.0847 (0.0307)

STD DEV = 0.0165 (0.0194)

REL STD DEV = 19.491 (63.343)

**** CHANNEL 2 ****

Sample % Abs (% Abs Ref)

Sample #1 = 0.0850 (-0.0070)

Sample #2 = 0.0890 (-0.0020)

Sample #3 = 0.0900 (-0.0110)

Sample #4 = 0.0820 (-0.0090)

Avg % Abs = 0.0870 (0.0073)

STD DEV = 0.0044 (0.0047)

REL STD DEV = 5.010 (64.443)

Sol Value = 0.040 g/210L ***

Fit Value = 0.1905 mg/l %%%

Samples Taken = 4, Discarded = 1

Sum Io = 12795, Sum Io = 13075

**** CHANNEL 1 ****

Sample % Abs (% Abs Ref)

Sample #1 = 0.8160 (-0.0110)

Sample #2 = 0.8100 (-0.0050)

Sample #3 = 0.8300 (-0.0000)

Sample #4 = 0.8050 (-0.0070)

Avg % Abs = 0.8150 (0.0002)

STD DEV = 0.0132 (0.0060)

REL STD DEV = 1.623 (904.157)

**** CHANNEL 2 ****

Sample % Abs (% Abs Ref)

Sample #1 = 1.4700 (-0.0100)

Sample #2 = 1.4600 (-0.0080)

Sample #3 = 1.4680 (-0.0000)

Sample #4 = 1.4670 (-0.0010)

Avg % Abs = 1.4650 (0.0030)

STD DEV = 0.0044 (0.0044)

REL STD DEV = 0.298 (145.297)

Sol Value = 0.100 g/210L ***

Fit Value = 0.4762 mg/l %%%

Samples Taken = 4, Discarded = 1

Sum Io = 12790, Sum Io = 13071

**** CHANNEL 1 ****

Sample % Abs (% Abs Ref)

Sample #1 = 1.8580 (-0.0290)

Sample #2 = 1.8450 (-0.0270)

Sample #3 = 1.8650 (0.0000)

Sample #4 = 1.8520 (-0.0130)

Avg % Abs = 1.8540 (-0.0133)

STD DEV = 0.0101 (0.0135)

REL STD DEV = 0.547 (101.273)

**** CHANNEL 2 ****

Sample % Abs (% Abs Ref)

Sample #1 = 3.4420 (-0.0260)

Sample #2 = 3.4190 (-0.0150)

Sample #3 = 3.4300 (-0.0110)

Sample #4 = 3.4140 (-0.0100)

Avg % Abs = 3.4210 (-0.0120)

STD DEV = 0.0082 (0.0026)

REL STD DEV = 0.239 (22.048)

Sol Value = 0.200 g/210L ***

Fit Value = 0.9524 mg/l %%%

Samples Taken = 4, Discarded = 1

Sum Io = 12791, Sum Io = 13072

**** CHANNEL 1 ****

Sample % Abs (% Abs Ref)

Sample #1 = 3.5820 (-0.0170)

Sample #2 = 3.5620 (-0.0050)

Sample #3 = 3.5620 (0.0150)

Sample #4 = 3.5700 (-0.0030)

Avg % Abs = 3.5647 (0.0023)

STD DEV = 0.0046 (0.0110)

REL STD DEV = 0.130 (472.077)

Optical Calibration

Adjustment #2

By: TDG

**** AUTO CAL DATA ****

**** CHANNEL 1 ****

Sol Val = 0.000 mg/l or 0.000 g/210L

% Abs = 0.085

Std Dev = 0.02 Rel Std Dev = 19.49

Sol Val = 0.1905 mg/l or 0.040 g/210L

% Abs = 0.815

Std Dev = 0.01 Rel Std Dev = 1.62

Sol Val = 0.4762 mg/l or 0.100 g/210L

% Abs = 1.854

Std Dev = 0.01 Rel Std Dev = 0.55

Sol Val = 0.9524 mg/l or 0.200 g/210L

% Abs = 3.565

Std Dev = 0.00 Rel Std Dev = 0.13

Sol Val = 1.4286 mg/l or 0.300 g/210L

% Abs = 5.269

Std Dev = 0.00 Rel Std Dev = 0.08

Zero Order Coef = -254.08

First Order Coef = 2691.07

Second Order Coef = 15.23

Standard Deviation = 27.164909

**** CHANNEL 2 ****

Sol Val = 0.000 mg/l or 0.000 g/210L

% Abs = 0.087

Std Dev = 0.00 Rel Std Dev = 5.01

Sol Val = 0.1905 mg/l or 0.040 g/210L

% Abs = 1.465

Std Dev = 0.00 Rel Std Dev = 0.30

Sol Val = 0.4762 mg/l or 0.100 g/210L

% Abs = 3.421

Std Dev = 0.01 Rel Std Dev = 0.24

Sol Val = 0.9524 mg/l or 0.200 g/210L

% Abs = 6.567

Std Dev = 0.03 Rel Std Dev = 0.45

Sol Val = 1.4286 mg/l or 0.300 g/210L

% Abs = 9.643

Std Dev = 0.02 Rel Std Dev = 0.19

Zero Order Coef = -145.43

First Order Coef = 1405.03

Second Order Coef = 9.60

Standard Deviation = 24.687298

Solution Stats Quadratic Fit Chan 1

Act Fit Residual

g/210L g/210L g/210L

0.000 -0.001 -0.006

0.040 0.041 -0.008

0.100 0.100 -0.0001

0.200 0.199 0.0006

0.300 0.300 -0.0002

Solution Stats Quadratic Fit Chan 2

Act Fit Residual

g/210L g/210L g/210L

0.000 -0.000 0.005

0.040 0.041 -0.0006

0.100 0.101 -0.0002

0.200 0.199 0.0006

0.300 0.300 -0.0002

Sol Value = 0.080 g/210L ***

Fit Value = 0.3810 mg/l %%%

Samples Taken = 4, Discarded = 1

**** CHANNEL 1 ****

Sample #1 = 3218.00

Sample #2 = 3085.08

Sample #3 = 3221.00

Sample #4 = 3161.00

Average Result = 3155.6667

STD DEV = 68.1567

REL STD DEV = 2.160

**** CHANNEL 2 ****

Sample #1 = 3384.00

Sample #2 = 3382.00

Sample #3 = 3421.00

Sample #4 = 3405.00

Average Result = 3402.6667

STD DEV = 19.6144

REL STD DEV = 0.576

**** CHANNEL 1 ****

Dry Gas H2O Adjust Results *****

Barometric Pressure = 1022

3 um H2O Adjust (mg/l*10,000) = 654

9 um H2O Adjust (mg/l*10,000) = 407

**** AUTO CAL PASS ****

Post-Cal Stability Checks

#2

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
✓	✓	✓	✓
≤0.003 of Wet	✓	✓	✓
<p> 0.05g/210L Intoxilyzer - Alcohol Analyzer Model: 8000 SN: 80-002462 02/13/2025 Software: 8100.27 </p> <p> Test g/210L Time ----- Air Blank 0.000 11:58 Control Test 0.051 11:59 Air Blank 0.000 12:00 Control Test 0.051 12:00 Air Blank 0.000 12:01 Control Test 0.051 12:01 Air Blank 0.000 12:02 Control Test Stats Average 0.0510 Std Dev 0.0000 Rel. Std Dev(%) 0.0000 </p> <p> Operator's Signature _____ </p>	<p> 0.08g/210L Intoxilyzer - Alcohol Analyzer Model: 8000 SN: 80-002462 02/13/2025 Software: 8100.27 </p> <p> Test g/210L Time ----- Air Blank 0.000 12:05 Control Test 0.081 12:06 Air Blank 0.000 12:06 Control Test 0.079 12:07 Air Blank 0.000 12:07 Control Test 0.079 12:08 Air Blank 0.000 12:09 Control Test Stats Average 0.0797 Std Dev 0.0012 Rel. Std Dev(%) 1.4494 </p> <p> Operator's Signature _____ </p>	<p> 0.20g/210L Intoxilyzer - Alcohol Analyzer Model: 8000 SN: 80-002462 02/13/2025 Software: 8100.27 </p> <p> Test g/210L Time ----- Air Blank 0.000 12:14 Control Test 0.195 12:15 Air Blank 0.000 12:15 Control Test 0.195 12:16 Air Blank 0.000 12:16 Control Test 0.198 12:17 Air Blank 0.000 12:18 Control Test Stats Average 0.1960 Std Dev 0.0017 Rel. Std Dev(%) 0.8837 </p> <p> Operator's Signature _____ </p>	<p> DGS 0.08g/210L Intoxilyzer - Alcohol Analyzer Model: 8000 SN: 80-002462 02/13/2025 Software: 8100.27 </p> <p> Test g/210L Time ----- Air Blank 0.000 11:55 Control Test 0.080 11:55 Air Blank 0.000 11:56 Control Test 0.080 11:56 Air Blank 0.000 11:56 Control Test 0.080 11:57 Air Blank 0.000 11:57 Control Test Stats Average 0.0800 Std Dev 0.0000 Rel. Std Dev(%) 0.0000 </p> <p> Operator's Signature _____ </p>

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: HIALEAH PD
Time of Inspection: 11:53

Date of Inspection: 02/14/2025

Serial Number: 80-002462
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG429602 Exp: 10/22/2026
0.000	0.051	0.079	0.200	0.080
0.000	0.051	0.080	0.197	0.080
0.000	0.050	0.080	0.200	0.080
0.000	0.051	0.080	0.196	0.080
0.000	0.051	0.080	0.198	0.080
0.000	0.051	0.080	0.196	0.080
0.000	0.051	0.081	0.197	0.080
0.000	0.051	0.080	0.197	0.080
0.000	0.051	0.080	0.198	0.080
0.000	0.051	0.081	0.198	0.080


Standard Deviations	0.0003	0.0005	0.0014	0.0000
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.


I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

02/14/2025
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-002462, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-002462</u>	UNCERTAINTY* \pm	
Owning Agency:	<u>HIALEAH PD</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>02/14/2025</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>11:53</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

02/14/2025

Date

TAYLOR D GUTSCHOW,
Department Inspector

FDLE/ATP Form 69 December 2021
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Return Material Authorization

Ship to: ☒ CMI, Inc.

☐ Enforcement Electronics

Shipment to repair facility authorized by: Jose Montero on 3/21/2025

Items Returned: Instrument ☒ Supplies ☐ Other ☐ Describe: _____

Instrument Model: Intoxilyzer 8000 Serial Number: 80-002462

Bill To Address:

Hialeah Police Department

Attn: Jose Montero

Ship to Address:

Florida Department of Law Enforcement

Fort Myers Regional Operations Center

Attn: Taylor Gutschow

4700 Terminal Drive, Suite 1

Fort Myers, FL 33907

Reason for Return:

Performed optical calibration adjustment to correct bouncing preliminary results. Instrument
passed the post-cal stability checks, but additional follow-up stabilities did not meet ATP SOP
requirements (3% accuracy). Will be sent to CMI for evaluation.

Please choose one of the following options:

☐ 1. I _____, authorize all repairs.

☐ 2. I _____, authorize repairs up to \$_____.

☒ 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Jose Montero

Phone #: 1-954-445-7077

Email: JMontero@hialeahfl.gov

ATP Contact Name: Taylor Gutschow

ATP Email: TaylorGutschow@fdle.state.fl.us