



INSTRUMENT PROCESSING SHEET

Agency Florida Highway Patrol Troop CS/N 80-003411Florida Department of
Law EnforcementDate In 06/25/2025 DI Completion Date 07/17/2025☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By TDG	Date	Quality Checks	By TDG	Date	Flow Calibration	By TDG	Date																																																
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: <u>Missing back right foot.</u>		<u>07/14/2025</u>	☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>206</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP101</u> 32 mm <u>0.175*</u> (.139 - .169) 36 mm <u>0.195*</u> (.156 - .190) 53 mm <u>0.257</u> (.228 - .278) 103 mm <u>0.519</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28199</u> ☒ Stability Checks		<u>07/15/2025</u>	Flow Column # <u>ATP104</u> ☒ 5L/min - 17mm ☒ 15L/min - 53mm ☒ 30L/min - 103mm ☒ R-Value <u>211</u> ☒ Post Calibration Verification (L/s) Flow Column # <u>ATP101</u> 32 mm <u>0.148</u> (.139 - .169) 36 mm <u>0.167</u> (.156 - .190) 53 mm <u>0.242</u> (.228 - .278) 103 mm <u>0.503</u> (.447 - .547)		<u>07/15/2025</u>																																																
			<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6286</td><td>202406K 06/19/2026</td></tr><tr><td>0.080</td><td>MP6287</td><td>202406L 06/19/2026</td></tr><tr><td>0.200</td><td>MP6288</td><td>202406N 06/20/2026</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG429602 10/22/2026</td></tr></tbody></table>			Simulator	Serial #	Lot #/Exp	0.050	MP6286	202406K 06/19/2026	0.080	MP6287	202406L 06/19/2026	0.200	MP6288	202406N 06/20/2026	0.080 DGS	N/A	AG429602 10/22/2026	<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6286</td><td>202406K 06/19/2026</td></tr><tr><td>0.080</td><td>MP6287</td><td>202406L 06/19/2026</td></tr><tr><td>0.200</td><td>MP6288</td><td>202406N 06/20/2026</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG429602 10/22/2026</td></tr></tbody></table>			Simulator	Serial #	Lot #/Exp	0.050	MP6286	202406K 06/19/2026	0.080	MP6287	202406L 06/19/2026	0.200	MP6288	202406N 06/20/2026	0.080 DGS	N/A	AG429602 10/22/2026	<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6286</td><td>202406K 06/19/2026</td></tr><tr><td>0.080</td><td>MP6287</td><td>202406L 06/19/2026</td></tr><tr><td>0.200</td><td>MP6288</td><td>202406N 06/20/2026</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG429602 10/22/2026</td></tr></tbody></table>			Simulator	Serial #	Lot #/Exp	0.050	MP6286	202406K 06/19/2026	0.080	MP6287	202406L 06/19/2026	0.200	MP6288	202406N 06/20/2026	0.080 DGS	N/A	AG429602 10/22/2026
Simulator	Serial #	Lot #/Exp																																																						
0.050	MP6286	202406K 06/19/2026																																																						
0.080	MP6287	202406L 06/19/2026																																																						
0.200	MP6288	202406N 06/20/2026																																																						
0.080 DGS	N/A	AG429602 10/22/2026																																																						
Simulator	Serial #	Lot #/Exp																																																						
0.050	MP6286	202406K 06/19/2026																																																						
0.080	MP6287	202406L 06/19/2026																																																						
0.200	MP6288	202406N 06/20/2026																																																						
0.080 DGS	N/A	AG429602 10/22/2026																																																						
Simulator	Serial #	Lot #/Exp																																																						
0.050	MP6286	202406K 06/19/2026																																																						
0.080	MP6287	202406L 06/19/2026																																																						
0.200	MP6288	202406N 06/20/2026																																																						
0.080 DGS	N/A	AG429602 10/22/2026																																																						
Calibration Adjustment By <u> </u>			Department Inspection By <u>TDG</u>																																																					
Barometric Pressure Gauge <u> </u> ID # <u> </u>			Barometric Pressure ID# <u>28199</u>																																																					
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>			Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Gauge <u>1019</u> Instrument <u>1014</u>																									
Simulator	Serial #	Lot #	Expiration																																																					
0.000		N/A	N/A																																																					
0.040																																																								
0.100																																																								
0.200																																																								
0.300																																																								
0.080 DGS	N/A																																																							
☐ Post Calibration Adjustment Stability Checks			Mouth Alcohol Solution Lot # <u>2024-A</u>																																																					
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Acetone Stock Solution Lot # <u>2023-B</u>																																	
Simulator	Serial #	Lot #	Expiration																																																					
0.050																																																								
0.080																																																								
0.200																																																								
0.080 DGS	N/A																																																							
			<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6284</td></tr><tr><td>Interferent</td><td>MP6285</td></tr><tr><td>0.050</td><td>MP6286</td></tr><tr><td>0.080</td><td>MP6287</td></tr><tr><td>0.200</td><td>MP6288</td></tr></tbody></table>			Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288																																							
Simulator	Serial Number																																																							
0.000	MP6284																																																							
Interferent	MP6285																																																							
0.050	MP6286																																																							
0.080	MP6287																																																							
0.200	MP6288																																																							
Notes/Suggested Service: <u>*Outside nominal. Performed root cause analysis. No user/equipment error detected. (TDG 7/15/25)</u>			Attachments																																																					
			☒ Form 41 ☐ Post-Stability Checks ☒ Stability Checks ☒ Flow Calibration ☒ Calibration Certificate ☐ Form 40 ☐ Calibration Adjustment ☐ Other <u> </u>																																																					
			☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																																																					
			Digitally signed by Shayla Platt Phil Nicodemus Date: 2025.07.18 12:26:49 -0400																																																					
			Shayla Platt Date: 2025.07.19 12:01:15 -0400																																																					
			Tech Review / Date <u> </u> Admin Review / Date <u> </u>																																																					

Flow Calibration Adjustment(s)

Performed by TDG

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003411
07/15/2025
Software: 8100.27

Flow Rate Calibration*****
1: Rate (Liters/min) = 5
 SQRT(Diff)) = -7.277
2: Rate (Liters/min) = 15
 SQRT(Diff)) = 11.871
3: Rate (Liters/min) = 30
 SQRT(Diff)) = 21.281
Dependent Data Scale Factor = 100000 L/min
Independent Data Scale Factor = 256
Rounded Slope = 686
Rounded Intercept = -700997
Correlation = 0.99672

Stability Checks

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083 ☒ ≤0.003 of Wet ☒																																																																																																																																																
FL HIGHWAY PATROL Intoxilyzer - Alcohol Analyzer Model 8000 07/15/2025 Software: 8100.27 SN 80-003411 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>15:06</td></tr> <tr><td>Control Test</td><td>0.049</td><td>15:07</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:07</td></tr> <tr><td>Control Test</td><td>0.049</td><td>15:08</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:09</td></tr> <tr><td>Control Test</td><td>0.049</td><td>15:09</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:10</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0490</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	15:06	Control Test	0.049	15:07	Air Blank	0.000	15:07	Control Test	0.049	15:08	Air Blank	0.000	15:09	Control Test	0.049	15:09	Air Blank	0.000	15:10	Control Test Stats			Average	0.0490		Std Dev	0.0000		Rel Std Dev(%)	0.0000		FL HIGHWAY PATROL Intoxilyzer - Alcohol Analyzer Model 8000 07/15/2025 Software: 8100.27 SN 80-003411 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>15:14</td></tr> <tr><td>Control Test</td><td>0.080</td><td>15:14</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:15</td></tr> <tr><td>Control Test</td><td>0.079</td><td>15:16</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:16</td></tr> <tr><td>Control Test</td><td>0.079</td><td>15:17</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:17</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0793</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7277</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	15:14	Control Test	0.080	15:14	Air Blank	0.000	15:15	Control Test	0.079	15:16	Air Blank	0.000	15:16	Control Test	0.079	15:17	Air Blank	0.000	15:17	Control Test Stats			Average	0.0793		Std Dev	0.0006		Rel Std Dev(%)	0.7277		FL HIGHWAY PATROL Intoxilyzer - Alcohol Analyzer Model 8000 07/15/2025 Software: 8100.27 SN 80-003411 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>15:00</td></tr> <tr><td>Control Test</td><td>0.198</td><td>15:00</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:01</td></tr> <tr><td>Control Test</td><td>0.198</td><td>15:02</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:02</td></tr> <tr><td>Control Test</td><td>0.198</td><td>15:03</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:03</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1980</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	15:00	Control Test	0.198	15:00	Air Blank	0.000	15:01	Control Test	0.198	15:02	Air Blank	0.000	15:02	Control Test	0.198	15:03	Air Blank	0.000	15:03	Control Test Stats			Average	0.1980		Std Dev	0.0000		Rel Std Dev(%)	0.0000		FL HIGHWAY PATROL Intoxilyzer - Alcohol Analyzer Model 8000 07/15/2025 Software: 8100.27 SN 80-003411 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>15:18</td></tr> <tr><td>Control Test</td><td>0.079</td><td>15:19</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:19</td></tr> <tr><td>Control Test</td><td>0.079</td><td>15:19</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:20</td></tr> <tr><td>Control Test</td><td>0.079</td><td>15:20</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>15:21</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0790</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	15:18	Control Test	0.079	15:19	Air Blank	0.000	15:19	Control Test	0.079	15:19	Air Blank	0.000	15:20	Control Test	0.079	15:20	Air Blank	0.000	15:21	Control Test Stats			Average	0.0790		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	15:06																																																																																																																																																	
Control Test	0.049	15:07																																																																																																																																																	
Air Blank	0.000	15:07																																																																																																																																																	
Control Test	0.049	15:08																																																																																																																																																	
Air Blank	0.000	15:09																																																																																																																																																	
Control Test	0.049	15:09																																																																																																																																																	
Air Blank	0.000	15:10																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0490																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	15:14																																																																																																																																																	
Control Test	0.080	15:14																																																																																																																																																	
Air Blank	0.000	15:15																																																																																																																																																	
Control Test	0.079	15:16																																																																																																																																																	
Air Blank	0.000	15:16																																																																																																																																																	
Control Test	0.079	15:17																																																																																																																																																	
Air Blank	0.000	15:17																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0793																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7277																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	15:00																																																																																																																																																	
Control Test	0.198	15:00																																																																																																																																																	
Air Blank	0.000	15:01																																																																																																																																																	
Control Test	0.198	15:02																																																																																																																																																	
Air Blank	0.000	15:02																																																																																																																																																	
Control Test	0.198	15:03																																																																																																																																																	
Air Blank	0.000	15:03																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1980																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	15:18																																																																																																																																																	
Control Test	0.079	15:19																																																																																																																																																	
Air Blank	0.000	15:19																																																																																																																																																	
Control Test	0.079	15:19																																																																																																																																																	
Air Blank	0.000	15:20																																																																																																																																																	
Control Test	0.079	15:20																																																																																																																																																	
Air Blank	0.000	15:21																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0790																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Operator's Signature 	Operator's Signature 	Operator's Signature 	Operator's Signature 																																																																																																																																																

Florida Department of Law Enforcement

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 14:14

Date of Inspection: 07/17/2025

Serial Number: 80-003411
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG429602 Exp: 10/22/2026
0.000	0.049	0.078	0.197	0.078
0.000	0.048	0.079	0.196	0.078
0.000	0.049	0.078	0.197	0.078
0.000	0.049	0.078	0.197	0.079
0.000	0.048	0.078	0.197	0.078
0.000	0.048	0.078	0.197	0.078
0.000	0.048	0.078	0.197	0.078
0.000	0.048	0.078	0.197	0.078
0.000	0.048	0.078	0.197	0.078
0.000	0.048	0.078	0.197	0.078


Standard Deviations	0.0004	0.0003	0.0003	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


 TAYLOR D GUTSCHOW
 Signature and Printed Name

07/17/2025
 Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-003411, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-003411</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>FL HIGHWAY PATROL</u>	0.050 g/210 L 0.004
Calibration Date:	<u>07/17/2025</u>	0.080 g/210 L 0.004
Calibration Time:	<u>14:14</u>	0.200 g/210 L 0.007
		0.080 g/210 L Dry Gas Control 0.005

All results are reported in g/210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Taylor
Gutschow
Digitally signed by Taylor
Gutschow
Date: 2025.07.17 15:22:49
-04'00'

07/17/2025

Date

TAYLOR D GUTSCHOW,

Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality