



INSTRUMENT PROCESSING SHEET

Agency Gulf Breeze PDS/N 80-004771Florida Department of
Law EnforcementDate In 03/24/2025 DI Completion Date 5/21/25☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake By <u>ALL</u> Date <u>03/24/2025</u> ☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☒ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____	Quality Checks By <u>DA</u> Date <u>03/31/2025</u> ☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>175</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP103</u> 32 mm <u>0.160</u> (.139 - .169) 36 mm <u>0.171</u> (.156 - .190) 53 mm <u>0.242</u> (.228 - .278) 103 mm <u>0.515</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>30793</u> ☒ Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP5088</td><td>202406K 06/19/2026</td></tr><tr><td>0.080</td><td>MP5089</td><td>202406L 06/19/2026</td></tr><tr><td>0.200</td><td>MP5090</td><td>202406N 06/20/2026</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG429602 10/22/2026</td></tr></tbody></table>	Simulator	Serial #	Lot #/Exp	0.050	MP5088	202406K 06/19/2026	0.080	MP5089	202406L 06/19/2026	0.200	MP5090	202406N 06/20/2026	0.080 DGS	N/A	AG429602 10/22/2026	Flow Calibration By _____ Date _____ Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) Maintenance By _____ Date _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ _____ _____ _____ _____ _____																																	
Simulator	Serial #	Lot #/Exp																																																
0.050	MP5088	202406K 06/19/2026																																																
0.080	MP5089	202406L 06/19/2026																																																
0.200	MP5090	202406N 06/20/2026																																																
0.080 DGS	N/A	AG429602 10/22/2026																																																
Calibration Adjustment By <u>SP</u> Barometric Pressure Gauge <u>1009</u> ID # <u>28427</u> <table border="1" style="width:100%; border-collapse: collapse;"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td>MP6294</td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td>MP6295</td><td>24330</td><td>9/10/26</td></tr><tr><td>0.100</td><td>MP6296</td><td>24390</td><td>10/29/26</td></tr><tr><td>0.200</td><td>MP6297</td><td>24080</td><td>2/13/26</td></tr><tr><td>0.300</td><td>MP6298</td><td>24430</td><td>12/10/26</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>28424080A3</td><td>11/5/26</td></tr></tbody></table> ☒ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td>MP5088</td><td>202406K</td><td>06/19/2026</td></tr><tr><td>0.080</td><td>MP5089</td><td>202406L</td><td>06/19/2026</td></tr><tr><td>0.200</td><td>MP5090</td><td>202406N</td><td>06/20/2026</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG429602</td><td>10/22/2026</td></tr></tbody></table> Notes/Suggested Service: <u>I performed root-cause analysis for out-of-range stability checks and did not find any user/equipment error. DA 04/02/2025</u> _____ _____ _____ _____ _____ _____ _____ _____			Simulator	Serial #	Lot #	Expiration	0.000	MP6294	N/A	N/A	0.040	MP6295	24330	9/10/26	0.100	MP6296	24390	10/29/26	0.200	MP6297	24080	2/13/26	0.300	MP6298	24430	12/10/26	0.080 DGS	N/A	28424080A3	11/5/26	Simulator	Serial #	Lot #	Expiration	0.050	MP5088	202406K	06/19/2026	0.080	MP5089	202406L	06/19/2026	0.200	MP5090	202406N	06/20/2026	0.080 DGS	N/A	AG429602	10/22/2026
Simulator	Serial #	Lot #	Expiration																																															
0.000	MP6294	N/A	N/A																																															
0.040	MP6295	24330	9/10/26																																															
0.100	MP6296	24390	10/29/26																																															
0.200	MP6297	24080	2/13/26																																															
0.300	MP6298	24430	12/10/26																																															
0.080 DGS	N/A	28424080A3	11/5/26																																															
Simulator	Serial #	Lot #	Expiration																																															
0.050	MP5088	202406K	06/19/2026																																															
0.080	MP5089	202406L	06/19/2026																																															
0.200	MP5090	202406N	06/20/2026																																															
0.080 DGS	N/A	AG429602	10/22/2026																																															
Department Inspection By <u>SP</u> Barometric Pressure ID# <u>30793</u> Gauge <u>1010</u> Instrument <u>1008</u> Mouth Alcohol Solution Lot # <u>2024-A</u> Acetone Stock Solution Lot # <u>2024-B</u> <table border="1" style="width:100%; border-collapse: collapse;"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP5086</td></tr><tr><td>Interferent</td><td>MP5087</td></tr><tr><td>0.050</td><td>MP5088</td></tr><tr><td>0.080</td><td>MP5089</td></tr><tr><td>0.200</td><td>MP5090</td></tr></tbody></table> Attachments <table border="1" style="width:100%; border-collapse: collapse;"><tbody><tr><td>☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Calibration Adjustment</td><td>☒ Post-Stability Checks ☐ Flow Calibration ☒ Form 40 X2 ☐ Other _____</td></tr></tbody></table> ☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use <table style="width:100%;"><tr><td style="width:50%; text-align: center;">Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2025.05.26 17:22:21 -04'00'</td><td style="width:50%; text-align: center;">Phil Nicodemo Digitally signed by Phil Nicodemo Date: 2025.05.29 09:36:42 -04'00'</td></tr><tr><td style="text-align: center;">Tech Review / Date _____</td><td style="text-align: center;">Admin Review / Date _____</td></tr></table>			Simulator	Serial Number	0.000	MP5086	Interferent	MP5087	0.050	MP5088	0.080	MP5089	0.200	MP5090	☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Calibration Adjustment	☒ Post-Stability Checks ☐ Flow Calibration ☒ Form 40 X2 ☐ Other _____	Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2025.05.26 17:22:21 -04'00'	Phil Nicodemo Digitally signed by Phil Nicodemo Date: 2025.05.29 09:36:42 -04'00'	Tech Review / Date _____	Admin Review / Date _____																														
Simulator	Serial Number																																																	
0.000	MP5086																																																	
Interferent	MP5087																																																	
0.050	MP5088																																																	
0.080	MP5089																																																	
0.200	MP5090																																																	
☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Calibration Adjustment	☒ Post-Stability Checks ☐ Flow Calibration ☒ Form 40 X2 ☐ Other _____																																																	
Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2025.05.26 17:22:21 -04'00'	Phil Nicodemo Digitally signed by Phil Nicodemo Date: 2025.05.29 09:36:42 -04'00'																																																	
Tech Review / Date _____	Admin Review / Date _____																																																	

Stability Checks

80-004771

DA

3/31/25

BULF BREEDER PC
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-004771
03/31/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:32
Control Test	0.048	08:33
Air Blank	0.000	08:33
Control Test	0.047	08:34
Air Blank	0.000	08:34
Control Test	0.047	08:35
Air Blank	0.000	08:35
Control Test Stats		
Average	0.0467	
Std Dev	0.0006	
Rel Std Dev(%)	1.2872	

DA

Operator's Signature

BULF BREEDER PC
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-004771
03/31/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:40
Control Test	0.077	08:40
Air Blank	0.000	08:40
Control Test	0.076	08:40
Air Blank	0.000	08:40
Control Test	0.077	08:47
Air Blank	0.000	08:48
Control Test Stats		
Average	0.0767	
Std Dev	0.0006	
Rel Std Dev(%)	0.7531	

Wet

DA

Operator's Signature

BULF BREEDER PC
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-004771
03/31/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:51
Control Test	0.192	08:51
Air Blank	0.000	08:52
Control Test	0.194	08:53
Air Blank	0.000	08:53
Control Test	0.194	08:54
Air Blank	0.000	08:54
Control Test Stats		
Average	0.1943	
Std Dev	0.0006	
Rel Std Dev(%)	0.2671	

DA

Operator's Signature

BULF BREEDER PC
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-004771
03/31/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:28
Control Test	0.076	08:28
Air Blank	0.000	08:28
Control Test	0.077	08:28
Air Blank	0.000	08:30
Control Test	0.077	08:30
Air Blank	0.000	08:31
Control Test Stats		
Average	0.0773	
Std Dev	0.0006	
Rel Std Dev(%)	0.7486	

DGS

DA

Operator's Signature

SOL 555555 RD

Intelligence = Manual Analysis
Model 6100
10/21/2015

Auto Calibration

%K Power Res Value = 35
Auto Range Res Value = 10

Sol Value = 0.000 g/210.0

Fit Value = 0.0000 g/1.0000

Samples Taken = 4, Discarded = 1

Sum to = 1253, Sum to = 1253

REL STD DEV = 0.0000

REL STD DEV = 0.0000

REL STD DEV = 0.0000

REL STD DEV = 0.0000

REL STD DEV = 0.0000

REL STD DEV = 0.0000

REL STD DEV = 0.0000

REL STD DEV = 0.0000

REL STD DEV = 0.0000

REL STD DEV = 0.0000

REL STD DEV = 0.0000

REL STD DEV = 0.0000

REL STD DEV = 0.0000

REL STD DEV = 0.0000

REL STD DEV = 0.0000

REL STD DEV = 0.0000

REL STD DEV = 0.0000

REL STD DEV = 0.0000

REL STD DEV = 0.0000

REL STD DEV = 0.0000

REL STD DEV = 0.0000

REL STD DEV = 0.0000

REL STD DEV = 0.0000

REL STD DEV = 0.0000

REL STD DEV = 0.0000

REL STD DEV = 0.0000

REL STD DEV = 0.0000

REL STD DEV = 0.0000

REL STD DEV = 0.0000

REL STD DEV = 0.0000

REL STD DEV = 0.0000

REL STD DEV = 0.0000

REL STD DEV = 0.0000

REL STD DEV = 0.0000

REL STD DEV = 0.0000

REL STD DEV = 0.0000

REL STD DEV = 0.0000

REL STD DEV = 0.0000

REL STD DEV = 0.0000

REL STD DEV = 0.0000

Channel 1

Sample % Abs (% Abs Ref)
Sample #1 = 1.5555 (0.0000)
Sample #2 = 1.5555 (0.0000)
Sample #3 = 1.5555 (0.0000)
Sample #4 = 1.5555 (0.0000)
Avg % Abs = 1.5555 (0.0000)
STD DEV = 0.0000
REL STD DEV = 0.0000

Channel 2

Sol Value = 0.100 g/210.0
Fit Value = 0.4762 mg/1.0000
Samples Taken = 4, Discarded = 1
Sum to = 12634, Sum to = 12634
REL STD DEV = 0.0000

Channel 1

Sample % Abs (% Abs Ref)
Sample #1 = 1.7610 (0.0000)
Sample #2 = 1.7610 (0.0000)
Sample #3 = 1.7610 (0.0000)
Sample #4 = 1.7610 (0.0000)
Avg % Abs = 1.7610 (0.0000)
STD DEV = 0.0000
REL STD DEV = 0.0000

Channel 2

Sol Value = 0.200 g/210.0
Fit Value = 0.9524 mg/1.0000
Samples Taken = 4, Discarded = 1
Sum to = 12634, Sum to = 12634
REL STD DEV = 0.0000

Channel 1

Sample % Abs (% Abs Ref)
Sample #1 = 1.5555 (0.0000)
Sample #2 = 1.5555 (0.0000)
Sample #3 = 1.5555 (0.0000)
Sample #4 = 1.5555 (0.0000)
Avg % Abs = 1.5555 (0.0000)
STD DEV = 0.0000
REL STD DEV = 0.0000

Channel 2

Sol Value = 0.300 g/210.0
Fit Value = 1.4266 mg/1.0000
Samples Taken = 4, Discarded = 1
Sum to = 12634, Sum to = 12634
REL STD DEV = 0.0000

Channel 1

Sample % Abs (% Abs Ref)
Sample #1 = 1.5555 (0.0000)
Sample #2 = 1.5555 (0.0000)
Sample #3 = 1.5555 (0.0000)
Sample #4 = 1.5555 (0.0000)
Avg % Abs = 1.5555 (0.0000)
STD DEV = 0.0000
REL STD DEV = 0.0000

Channel 2

Sol Value = 0.400 g/210.0
Fit Value = 1.8528 mg/1.0000
Samples Taken = 4, Discarded = 1
Sum to = 12634, Sum to = 12634
REL STD DEV = 0.0000

Channel 1

Sol Val = 0.500 mg/1.0000 g/210.0
% Abs = 1.5555
Std Dev = 0.0000
Sol Val = 0.5000 mg/1.0000 g/210.0
% Abs = 1.5555
Std Dev = 0.0000
Sol Val = 0.5000 mg/1.0000 g/210.0
% Abs = 1.5555
Std Dev = 0.0000
Sol Val = 0.5000 mg/1.0000 g/210.0
% Abs = 1.5555
Std Dev = 0.0000

Channel 2

Sol Val = 0.600 mg/1.0000 g/210.0
% Abs = 1.5555
Std Dev = 0.0000
Sol Val = 0.6000 mg/1.0000 g/210.0
% Abs = 1.5555
Std Dev = 0.0000
Sol Val = 0.6000 mg/1.0000 g/210.0
% Abs = 1.5555
Std Dev = 0.0000

Channel 1

Sol Val = 0.700 mg/1.0000 g/210.0
% Abs = 1.5555
Std Dev = 0.0000
Sol Val = 0.7000 mg/1.0000 g/210.0
% Abs = 1.5555
Std Dev = 0.0000
Sol Val = 0.7000 mg/1.0000 g/210.0
% Abs = 1.5555
Std Dev = 0.0000

Channel 2

Sol Val = 0.800 mg/1.0000 g/210.0
% Abs = 1.5555
Std Dev = 0.0000
Sol Val = 0.8000 mg/1.0000 g/210.0
% Abs = 1.5555
Std Dev = 0.0000
Sol Val = 0.8000 mg/1.0000 g/210.0
% Abs = 1.5555
Std Dev = 0.0000

Channel 1

Sol Val = 0.900 mg/1.0000 g/210.0
% Abs = 1.5555
Std Dev = 0.0000
Sol Val = 0.9000 mg/1.0000 g/210.0
% Abs = 1.5555
Std Dev = 0.0000
Sol Val = 0.9000 mg/1.0000 g/210.0
% Abs = 1.5555
Std Dev = 0.0000

Channel 2

Sol Val = 1.000 mg/1.0000 g/210.0
% Abs = 1.5555
Std Dev = 0.0000
Sol Val = 1.0000 mg/1.0000 g/210.0
% Abs = 1.5555
Std Dev = 0.0000
Sol Val = 1.0000 mg/1.0000 g/210.0
% Abs = 1.5555
Std Dev = 0.0000

Channel 1

Sol Val = 1.100 mg/1.0000 g/210.0
% Abs = 1.5555
Std Dev = 0.0000
Sol Val = 1.1000 mg/1.0000 g/210.0
% Abs = 1.5555
Std Dev = 0.0000
Sol Val = 1.1000 mg/1.0000 g/210.0
% Abs = 1.5555
Std Dev = 0.0000

Channel 2

Sol Val = 1.200 mg/1.0000 g/210.0
% Abs = 1.5555
Std Dev = 0.0000
Sol Val = 1.2000 mg/1.0000 g/210.0
% Abs = 1.5555
Std Dev = 0.0000
Sol Val = 1.2000 mg/1.0000 g/210.0
% Abs = 1.5555
Std Dev = 0.0000

Dry Gas H2O Adjust Results
Barometric Pressure = 1009
3 um H2O Adjust (mg/1.0000) = 466
9 um H2O Adjust (mg/1.0000) = 507
***** AUTO CAL PASS

CAL ADJUSTMENT
SR

Post Cal Stability Checks

DAS

30LF SREZE PD
Intelligence - Alcorn Analyze
Model 8000
05/21/2025
Software: 8100.27

Test	g/100	mg
Air Blank	0.00	0.00
Control Test	0.079	0.079
Air Blank	0.00	0.00
Control Test	0.079	0.079
Air Blank	0.00	0.00
Control Test	0.079	0.079
Air Blank	0.00	0.00
Control Test	0.079	0.079
Air Blank	0.00	0.00
Control Test	0.079	0.079
Average	0.079	0.079
Std Dev	0.000	0.000
Rel. Std. Dev.	0.7335	0.7335

30LF SREZE PD
Intelligence - Alcorn Analyze
Model 8000
05/21/2025
Software: 8100.27

Test	g/100	mg
Air Blank	0.00	0.00
Control Test	0.080	0.080
Air Blank	0.00	0.00
Control Test	0.080	0.080
Air Blank	0.00	0.00
Control Test	0.080	0.080
Air Blank	0.00	0.00
Control Test	0.080	0.080
Air Blank	0.00	0.00
Control Test	0.080	0.080
Average	0.080	0.080
Std Dev	0.000	0.000
Rel. Std. Dev.	0.2656	0.2656

30LF SREZE PD
Intelligence - Alcorn Analyze
Model 8000
05/21/2025
Software: 8100.27

Test	g/100	mg
Air Blank	0.00	0.00
Control Test	0.079	0.079
Air Blank	0.00	0.00
Control Test	0.079	0.079
Air Blank	0.00	0.00
Control Test	0.079	0.079
Air Blank	0.00	0.00
Control Test	0.079	0.079
Air Blank	0.00	0.00
Control Test	0.079	0.079
Average	0.079	0.079
Std Dev	0.000	0.000
Rel. Std. Dev.	0.7277	0.7277

30LF SREZE PD
Intelligence - Alcorn Analyze
Model 8000
05/21/2025
Software: 8100.27

Test	g/100	mg
Air Blank	0.00	0.00
Control Test	0.048	0.048
Air Blank	0.00	0.00
Control Test	0.048	0.048
Air Blank	0.00	0.00
Control Test	0.048	0.048
Air Blank	0.00	0.00
Control Test	0.048	0.048
Air Blank	0.00	0.00
Control Test	0.048	0.048
Average	0.048	0.048
Std Dev	0.000	0.000
Rel. Std. Dev.	0.663	0.663

SP

SP

SP

SP

Operator's Signature

Operator's Signature

Operator's Signature

Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: GULF BREEZE PD

Time of Inspection: 14:39

Date of Inspection: 05/21/2025

Serial Number: 80-004771

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG429602 Exp: 10/22/2026
0.000	0.050	0.080	0.198	0.078
0.000	0.051	0.080	0.199	0.078
0.000	0.050	0.080	0.199	0.079
0.000	0.049	0.080	0.199	0.078
0.000	0.050	0.080	0.199	0.078
0.000	0.049	0.080	0.199	0.078
0.000	0.050	0.080	0.199	0.078
0.000	0.050	0.080	0.199	0.078
0.000	0.049	0.080	0.199	0.078
0.000	0.050	0.080	0.200	0.078

Standard Deviations	0.0006	0.0000	0.0004	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt

SHAYLA D PLATT

Signature and Printed Name

05/21/2025
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-004771, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-004771</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>GULF BREEZE PD</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>05/21/2025</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>14:39</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Shayla Platt
Digitally signed by Shayla Platt
Date: 2025.05.23 11:26:37 -04'00'

05/21/2025

Date

SHAYLA D PLATT,
Department Inspector

FDLE/ATP Form 69 October 2024

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Florida Department of Law Enforcement

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: GULF BREEZE PD

Serial Number: 80-004771

Time of Inspection: 08:15

Date of Inspection: 03/31/2025

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot #: Exp:

Number of Simulators Used: _____

Remarks:

BYPASS AI TO OPERATE INSTRUMENT COMPLIANCE NOT DETERMINED

N/A DA 3/31/25

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



DESTINEE N ARMSTRONG

Signature and Printed Name

03/31/2025

Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: GULF BREEZE PD
Time of Inspection: 08:19

Date of Inspection: 05/21/2025

Serial Number: 80-004771
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08g/210L Test (g/210L) Lot#:_____ Exp:_____	0.20g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:_____ Exp:_____

Number of Simulators Used: _____

Remarks:

BYPASSED AI TO OPERATE INSTRUMENT

N/A COMPLIANCE NOT DETERMINED.

The above instrument complies (☒) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt

SHAYLA D PLATT

Signature and Printed Name

05/21/2025
Date