





INSTRUMENT PROCESSING SHEET

Agency DAYTONA BEACH PDS/N 80-005537Florida Department of
Law EnforcementDate In 7/2/25DI Completion Date 07/17/2025☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By <u>KTS</u>	Date <u>7/2/25</u>	Quality Checks	By <u>WKP</u>	Date <u>7/10/2025</u>	Flow Calibration	By _____	Date _____																													
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: <u>Black receipt paper cover holder is broken. Came with receipt paper all unraveled. KTS 7/2/25</u>			☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>190</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP-102</u> 32 mm <u>0.152</u> (.139 - .169) 36 mm <u>0.167</u> (.156 - .190) 53 mm <u>0.230</u> (.228 - .278) 103 mm <u>0.503</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28662</u> ☒ Stability Checks			Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																															
			<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP5088</td><td>202406K 06/19/2026</td></tr><tr><td>0.080</td><td>MP5089</td><td>202406L 06/19/2026</td></tr><tr><td>0.200</td><td>MP5090</td><td>202406N 6/20/2026</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG429602 10/22/2026</td></tr></tbody></table>	Simulator	Serial #	Lot #/Exp	0.050	MP5088	202406K 06/19/2026	0.080	MP5089	202406L 06/19/2026	0.200	MP5090	202406N 6/20/2026	0.080 DGS	N/A	AG429602 10/22/2026			<table border="1"><thead><tr><th colspan="2">Maintenance By _____ Date _____</th></tr></thead><tbody><tr><td>☐ Battery Replacement</td><td></td></tr><tr><td>☐ Dry Gas Regulator Replacement</td><td></td></tr><tr><td>☐ Breath Tube Replacement</td><td></td></tr><tr><td>☐ Other _____</td><td></td></tr></tbody></table>	Maintenance By _____ Date _____		☐ Battery Replacement		☐ Dry Gas Regulator Replacement		☐ Breath Tube Replacement		☐ Other _____							
Simulator	Serial #	Lot #/Exp																																			
0.050	MP5088	202406K 06/19/2026																																			
0.080	MP5089	202406L 06/19/2026																																			
0.200	MP5090	202406N 6/20/2026																																			
0.080 DGS	N/A	AG429602 10/22/2026																																			
Maintenance By _____ Date _____																																					
☐ Battery Replacement																																					
☐ Dry Gas Regulator Replacement																																					
☐ Breath Tube Replacement																																					
☐ Other _____																																					
Calibration Adjustment By _____			Department Inspection By <u>SLH</u>																																		
Barometric Pressure Gauge _____ ID # _____			Barometric Pressure ID# <u>28421</u>																																		
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>			Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Gauge <u>1016</u> Instrument <u>1010</u> Mouth Alcohol Solution Lot # <u>2025-A</u> Acetone Stock Solution Lot # <u>2024-B</u>						
Simulator	Serial #	Lot #	Expiration																																		
0.000		N/A	N/A																																		
0.040																																					
0.100																																					
0.200																																					
0.300																																					
0.080 DGS	N/A																																				
☐ Post Calibration Adjustment Stability Checks			<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6289</td></tr><tr><td>Interferent</td><td>MP6290</td></tr><tr><td>0.050</td><td>MP5088</td></tr><tr><td>0.080</td><td>MP5089</td></tr><tr><td>0.200</td><td>MP5090</td></tr></tbody></table>					Simulator	Serial Number	0.000	MP6289	Interferent	MP6290	0.050	MP5088	0.080	MP5089	0.200	MP5090																		
Simulator	Serial Number																																				
0.000	MP6289																																				
Interferent	MP6290																																				
0.050	MP5088																																				
0.080	MP5089																																				
0.200	MP5090																																				
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			<table border="1"><thead><tr><th colspan="2">Attachments</th></tr></thead><tbody><tr><td>☒ Form 41</td><td>☐ Post-Stability Checks</td></tr><tr><td>☒ Stability Checks</td><td>☐ Flow Calibration</td></tr><tr><td>☒ Calibration Certificate</td><td>☐ Form 40</td></tr><tr><td>☐ Calibration Adjustment</td><td>☐ Other _____</td></tr></tbody></table>					Attachments		☒ Form 41	☐ Post-Stability Checks	☒ Stability Checks	☐ Flow Calibration	☒ Calibration Certificate	☐ Form 40	☐ Calibration Adjustment	☐ Other _____
Simulator	Serial #	Lot #	Expiration																																		
0.050																																					
0.080																																					
0.200																																					
0.080 DGS	N/A																																				
Attachments																																					
☒ Form 41	☐ Post-Stability Checks																																				
☒ Stability Checks	☐ Flow Calibration																																				
☒ Calibration Certificate	☐ Form 40																																				
☐ Calibration Adjustment	☐ Other _____																																				
Notes/Suggested Service: _____			☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																																		
			Digitally signed by Shayla Platt Date: 2025.07.29 09:43:09 -04'00'																																		
			Tech Review / Date _____ Admin Review / Date _____																																		

Stability Checks

0.05g/210L		0.08g/210L		0.20g/210L		DGS 0.08g/210L	
0.047 to 0.053		0.077 to 0.083		0.194 to 0.206		0.077 to 0.083	
✓		✓		✓		✓	
Performed Root Case Analysis		Performed Root Case Analysis		Performed Root Case Analysis		Performed Root Case Analysis	
DAYONE BEACH PC Intoxilyzer - Alcohol Analyzer Model 8000 07/10/2025 Software: 8100.27 Test Time Air Blank 0.100 10:46 Control Test 1.049 10:46 Air Blank 0.000 10:47 Control Test 0.000 10:48 Air Blank 0.000 10:48 Control Test 0.049 10:49 Air Blank 0.000 10:49 Control Test 0.000 10:49 Average 0.049 Std Dev 0.000 Rel Std Dev(%) 0.000 Operator's Signature 		DAYONE BEACH PC Intoxilyzer - Alcohol Analyzer Model 8000 07/10/2025 Software: 8100.27 Test Time Air Blank 0.000 11:04 Control Test 0.000 11:05 Air Blank 0.000 11:05 Control Test 0.079 11:06 Air Blank 0.000 11:07 Control Test 0.000 11:07 Air Blank 0.000 11:07 Control Test 0.000 11:08 Average 0.079 Std Dev 0.000 Rel Std Dev(%) 0.7247 Operator's Signature 		DAYONE BEACH PC Intoxilyzer - Alcohol Analyzer Model 8000 07/10/2025 Software: 8100.27 Test Time Air Blank 0.000 11:04 Control Test 0.000 11:05 Air Blank 0.000 11:05 Control Test 0.000 11:06 Air Blank 0.000 11:07 Control Test 0.000 11:07 Air Blank 0.000 11:07 Control Test 0.000 11:08 Average 0.079 Std Dev 0.000 Rel Std Dev(%) 0.7247 Operator's Signature 		DAYONE BEACH PC Intoxilyzer - Alcohol Analyzer Model 8000 07/10/2025 Software: 8100.27 Test Time Air Blank 0.000 11:17 Control Test 0.000 11:17 Air Blank 0.000 11:18 Control Test 0.000 11:18 Air Blank 0.000 11:19 Control Test 0.000 11:19 Air Blank 0.000 11:19 Control Test 0.000 11:19 Average 0.000 Std Dev 0.000 Rel Std Dev(%) 0.000 Operator's Signature 	

DAYTONA BEACH PD
Intoxilyzer - Alcohol Analyzer
Model: 8000 SN 80-005537
07/10/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:59
Control Test	0.000	11:00
Air Blank	0.000	11:00
Control Test	0.000	11:01
Air Blank	0.000	11:01
Control Test	0.000	11:02
Air Blank	0.000	11:03
Control Test Stats		
Average	0.0000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0001	


Operator's Signature

0.08g/210L

wet stability #1

Performed RCA,

repeated

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: DAYTONA BEACH PD
Time of Inspection: 13:27

Date of Inspection: 07/17/2025

Serial Number: 80-005537
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG429602 Exp: 10/22/2026
0.000	0.049	0.079	0.198	0.080
0.000	0.049	0.079	0.199	0.080
0.000	0.049	0.079	0.198	0.080
0.000	0.049	0.079	0.198	0.081
0.000	0.049	0.079	0.198	0.080
0.000	0.050	0.079	0.198	0.081
0.000	0.050	0.079	0.199	0.080
0.000	0.049	0.079	0.199	0.081
0.000	0.050	0.079	0.199	0.081
0.000	0.050	0.079	0.199	0.080

Standard Deviations	0.0005	0.0000	0.0005	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:


The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



LEANDRA HIGGINBOTHAM
Signature and Printed Name

07/17/2025
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-005537, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-005537</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>DAYTONA BEACH PD</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>07/17/2025</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>13:27</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

07/17/2025

Date


LEANDRA HIGGINBOTHAM,

Department Inspector

FDLE/ATP Form 69 October 2024
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality