



INSTRUMENT PROCESSING SHEET

Agency Palm Beach CSOS/N 80-006027Florida Department of
Law EnforcementDate In 01/13/2025 DI Completion Date 01/13/2025 Ship P/U H/D CMI EE

Intake	By TDG	Date <u>01/13/2025</u>	Quality Checks	By TDG	Date <u>01/13/2025</u>	Flow Calibration	By	Date															
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: <u>Hand-delivered by AI.</u>			<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>209</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP101</u> 32 mm <u>0.152</u> (.139 - .169) 36 mm <u>0.171</u> (.156 - .190) 53 mm <u>0.242</u> (.228 - .278) 103 mm <u>0.515</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>26932</u> <input checked="" type="checkbox"/> Stability Checks			Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)			Maintenance By _____ Date _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____														
			<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP6286</td> <td>202303K 03/29/2025</td> </tr> <tr> <td>0.080</td> <td>MP6287</td> <td>202303L 03/29/2025</td> </tr> <tr> <td>0.200</td> <td>MP6288</td> <td>202304C 04/05/2025</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG429602 10/22/2026</td> </tr> </tbody> </table>			Simulator	Serial #	Lot #/Exp	0.050	MP6286	202303K 03/29/2025	0.080	MP6287	202303L 03/29/2025	0.200	MP6288	202304C 04/05/2025	0.080 DGS	N/A	AG429602 10/22/2026			
Simulator	Serial #	Lot #/Exp																					
0.050	MP6286	202303K 03/29/2025																					
0.080	MP6287	202303L 03/29/2025																					
0.200	MP6288	202304C 04/05/2025																					
0.080 DGS	N/A	AG429602 10/22/2026																					

Calibration Adjustment	Department Inspection																																																												
By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			By <u>TDG</u> Barometric Pressure ID# <u>26932</u> Gauge <u>1020</u> Instrument <u>1021</u> Mouth Alcohol Solution Lot # <u>2024-A</u> Acetone Stock Solution Lot # <u>2023-B</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP6284</td> </tr> <tr> <td>Interferent</td> <td>MP6285</td> </tr> <tr> <td>0.050</td> <td>MP6286</td> </tr> <tr> <td>0.080</td> <td>MP6287</td> </tr> <tr> <td>0.200</td> <td>MP6288</td> </tr> </tbody> </table>	Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288
Simulator	Serial #	Lot #	Expiration																																																										
0.000		N/A	N/A																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.300																																																													
0.080 DGS	N/A																																																												
Simulator	Serial #	Lot #	Expiration																																																										
0.050																																																													
0.080																																																													
0.200																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number																																																												
0.000	MP6284																																																												
Interferent	MP6285																																																												
0.050	MP6286																																																												
0.080	MP6287																																																												
0.200	MP6288																																																												
Notes/Suggested Service: _____ _____ _____ _____ _____ _____	<table border="1" style="width:100%; border-collapse: collapse;"> <tr> <td> <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Calibration Adjustment </td> <td> <input type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Form 40 <input type="checkbox"/> Other _____ </td> </tr> <tr> <td colspan="2"> <input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use </td> </tr> <tr> <td> Destinee Armstrong <small>Digitally signed by Destinee Armstrong Date: 2025.01.13 14:01:59 -0500</small> </td> <td> Phil Nicodemo <small>Digitally signed by Phil Nicodemo Date: 2025.01.13 15:28:02 -0500</small> </td> </tr> <tr> <td>Tech Review / Date</td> <td>Admin Review / Date</td> </tr> </table>	<input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Calibration Adjustment	<input type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Form 40 <input type="checkbox"/> Other _____	<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use		Destinee Armstrong <small>Digitally signed by Destinee Armstrong Date: 2025.01.13 14:01:59 -0500</small>	Phil Nicodemo <small>Digitally signed by Phil Nicodemo Date: 2025.01.13 15:28:02 -0500</small>	Tech Review / Date	Admin Review / Date																																																				
<input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Calibration Adjustment	<input type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Form 40 <input type="checkbox"/> Other _____																																																												
<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use																																																													
Destinee Armstrong <small>Digitally signed by Destinee Armstrong Date: 2025.01.13 14:01:59 -0500</small>	Phil Nicodemo <small>Digitally signed by Phil Nicodemo Date: 2025.01.13 15:28:02 -0500</small>																																																												
Tech Review / Date	Admin Review / Date																																																												

Stability Checks

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083 ✓ ≤0.003 of Wet ✓																																																																																																																																																
<p>PALM BEACH COUNTY SO Intoxilyzer - Alcohol Analyzer Model: 8000 SN 80-016027 01/13/2025 Software: 8100.27</p> <table border="0"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>10:26</td></tr> <tr><td>Control Test</td><td>0.049</td><td>10:26</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:27</td></tr> <tr><td>Control Test</td><td>0.050</td><td>10:28</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:28</td></tr> <tr><td>Control Test</td><td>0.049</td><td>10:29</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:29</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0493</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1703</td><td></td></tr> </tbody> </table> <p style="text-align: right;">  ----- Operator's Signature </p>	Test	g/210L	Time	Air Blank	0.000	10:26	Control Test	0.049	10:26	Air Blank	0.000	10:27	Control Test	0.050	10:28	Air Blank	0.000	10:28	Control Test	0.049	10:29	Air Blank	0.000	10:29	Control Test Stats			Average	0.0493		Std Dev	0.0006		Rel Std Dev(%)	1.1703		<p>PALM BEACH COUNTY SO Intoxilyzer - Alcohol Analyzer Model: 8000 SN 80-016027 01/13/2025 Software: 8100.27</p> <table border="0"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>10:33</td></tr> <tr><td>Control Test</td><td>0.079</td><td>10:33</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:34</td></tr> <tr><td>Control Test</td><td>0.080</td><td>10:35</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:35</td></tr> <tr><td>Control Test</td><td>0.078</td><td>10:36</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:36</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0790</td><td></td></tr> <tr><td>Std Dev</td><td>0.0010</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.2658</td><td></td></tr> </tbody> </table> <p style="text-align: right;">  ----- Operator's Signature </p>	Test	g/210L	Time	Air Blank	0.000	10:33	Control Test	0.079	10:33	Air Blank	0.000	10:34	Control Test	0.080	10:35	Air Blank	0.000	10:35	Control Test	0.078	10:36	Air Blank	0.000	10:36	Control Test Stats			Average	0.0790		Std Dev	0.0010		Rel Std Dev(%)	1.2658		<p>PALM BEACH COUNTY SO Intoxilyzer - Alcohol Analyzer Model: 8000 SN 80-016027 01/13/2025 Software: 8100.27</p> <table border="0"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>10:03</td></tr> <tr><td>Control Test</td><td>0.199</td><td>10:03</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:04</td></tr> <tr><td>Control Test</td><td>0.198</td><td>10:05</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:05</td></tr> <tr><td>Control Test</td><td>0.198</td><td>10:06</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:06</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1983</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2911</td><td></td></tr> </tbody> </table> <p style="text-align: right;">  ----- Operator's Signature </p>	Test	g/210L	Time	Air Blank	0.000	10:03	Control Test	0.199	10:03	Air Blank	0.000	10:04	Control Test	0.198	10:05	Air Blank	0.000	10:05	Control Test	0.198	10:06	Air Blank	0.000	10:06	Control Test Stats			Average	0.1983		Std Dev	0.0006		Rel Std Dev(%)	0.2911		<p style="text-align: center; font-size: 2em; color: blue;">DGS</p> <p>PALM BEACH COUNTY SO Intoxilyzer - Alcohol Analyzer Model: 8000 SN 80-006027 01/13/2025 Software: 8100.27</p> <table border="0"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>10:10</td></tr> <tr><td>Control Test</td><td>0.079</td><td>10:10</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:10</td></tr> <tr><td>Control Test</td><td>0.079</td><td>10:11</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:11</td></tr> <tr><td>Control Test</td><td>0.080</td><td>10:12</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:12</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0793</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7277</td><td></td></tr> </tbody> </table> <p style="text-align: right;">  ----- Operator's Signature </p>	Test	g/210L	Time	Air Blank	0.000	10:10	Control Test	0.079	10:10	Air Blank	0.000	10:10	Control Test	0.079	10:11	Air Blank	0.000	10:11	Control Test	0.080	10:12	Air Blank	0.000	10:12	Control Test Stats			Average	0.0793		Std Dev	0.0006		Rel Std Dev(%)	0.7277	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:26																																																																																																																																																	
Control Test	0.049	10:26																																																																																																																																																	
Air Blank	0.000	10:27																																																																																																																																																	
Control Test	0.050	10:28																																																																																																																																																	
Air Blank	0.000	10:28																																																																																																																																																	
Control Test	0.049	10:29																																																																																																																																																	
Air Blank	0.000	10:29																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0493																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1703																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:33																																																																																																																																																	
Control Test	0.079	10:33																																																																																																																																																	
Air Blank	0.000	10:34																																																																																																																																																	
Control Test	0.080	10:35																																																																																																																																																	
Air Blank	0.000	10:35																																																																																																																																																	
Control Test	0.078	10:36																																																																																																																																																	
Air Blank	0.000	10:36																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0790																																																																																																																																																		
Std Dev	0.0010																																																																																																																																																		
Rel Std Dev(%)	1.2658																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:03																																																																																																																																																	
Control Test	0.199	10:03																																																																																																																																																	
Air Blank	0.000	10:04																																																																																																																																																	
Control Test	0.198	10:05																																																																																																																																																	
Air Blank	0.000	10:05																																																																																																																																																	
Control Test	0.198	10:06																																																																																																																																																	
Air Blank	0.000	10:06																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1983																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2911																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:10																																																																																																																																																	
Control Test	0.079	10:10																																																																																																																																																	
Air Blank	0.000	10:10																																																																																																																																																	
Control Test	0.079	10:11																																																																																																																																																	
Air Blank	0.000	10:11																																																																																																																																																	
Control Test	0.080	10:12																																																																																																																																																	
Air Blank	0.000	10:12																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0793																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7277																																																																																																																																																		

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PALM BEACH COUNTY SO
Time of Inspection: 12:28

Date of Inspection: 01/13/2025

Serial Number: 80-006027
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG429602 Exp: 10/22/2026
0.000	0.049	0.079	0.198	0.079
0.000	0.049	0.080	0.199	0.078
0.000	0.049	0.079	0.199	0.078
0.000	0.050	0.079	0.199	0.078
0.000	0.049	0.079	0.199	0.078
0.000	0.049	0.080	0.199	0.078
0.000	0.049	0.079	0.200	0.078
0.000	0.050	0.080	0.199	0.078
0.000	0.050	0.080	0.199	0.078
0.000	0.050	0.080	0.199	0.078
0.000	0.050	0.080	0.199	0.078

Standard Deviations	0.0005	0.0005	0.0004	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

01/13/2025
Date



Florida Department of Law Enforcement
 Alcohol Testing Program
 4700 Terminal Drive, Suite 1
 Ft. Myers, FL 33907

Calibration Certificate

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006027, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006027</u>	UNCERTAINTY* ±
Owning Agency:	<u>PALM BEACH COUNTY SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>01/13/2025</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>12:28</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/210 L.
 Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
 *Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).
 The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/IEC 17025 standards.
 This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.


 TAYLOR D GUTSCHOW,
 Department Inspector

01/13/2025
 Date

FDLE/ATP Form 69 December 2021
 Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality