







INSTRUMENT PROCESSING SHEET

Agency Palm Beach CSOS/N 80-006240Florida Department of
Law EnforcementDate In 01/13/2025 DI Completion Date 01/13/2025☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake By <u>TDG</u> Date <u>01/13/2025</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: <u>Hand-delivered by AI.</u>	Quality Checks By <u>TDG</u> Date <u>01/13/2025</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>216</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP101</u> 32 mm <u>0.152</u> (.139 - .169) 36 mm <u>0.167</u> (.156 - .190) 53 mm <u>0.238</u> (.228 - .278) 103 mm <u>0.507</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>26932</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6286</td><td>202303K 03/29/2025</td></tr><tr><td>0.080</td><td>MP6287</td><td>202303L 03/29/2025</td></tr><tr><td>0.200</td><td>MP6288</td><td>202304C 04/05/2025</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG429602 10/22/2026</td></tr></tbody></table>	Simulator	Serial #	Lot #/Exp	0.050	MP6286	202303K 03/29/2025	0.080	MP6287	202303L 03/29/2025	0.200	MP6288	202304C 04/05/2025	0.080 DGS	N/A	AG429602 10/22/2026	Flow Calibration By _____ Date _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) Maintenance By _____ Date _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____																																																			
Simulator	Serial #	Lot #/Exp																																																																		
0.050	MP6286	202303K 03/29/2025																																																																		
0.080	MP6287	202303L 03/29/2025																																																																		
0.200	MP6288	202304C 04/05/2025																																																																		
0.080 DGS	N/A	AG429602 10/22/2026																																																																		
Calibration Adjustment By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table> Notes/Suggested Service: _____			Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Department Inspection By <u>TDG</u> Barometric Pressure ID# <u>26932</u> Gauge <u>1020</u> Instrument <u>1020</u> Mouth Alcohol Solution Lot # <u>2024-A</u> Acetone Stock Solution Lot # <u>2023-B</u> <table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6284</td></tr><tr><td>Interferent</td><td>MP6285</td></tr><tr><td>0.050</td><td>MP6286</td></tr><tr><td>0.080</td><td>MP6287</td></tr><tr><td>0.200</td><td>MP6288</td></tr></tbody></table> Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Form 40 <input type="checkbox"/> Other _____ <input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use <table style="width:100%"><tr><td style="width:50%">Destinee Armstrong <small>Digitally signed by Destinee Armstrong Date: 2025.01.13 14:00:46 -05'00'</small></td><td style="width:50%">Phil Nicodemo <small>Digitally signed by Phil Nicodemo Date: 2025.01.13 15:28:25 -05'00'</small></td></tr><tr><td>Tech Review / Date</td><td>Admin Review / Date</td></tr></table>		Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288	Destinee Armstrong <small>Digitally signed by Destinee Armstrong Date: 2025.01.13 14:00:46 -05'00'</small>	Phil Nicodemo <small>Digitally signed by Phil Nicodemo Date: 2025.01.13 15:28:25 -05'00'</small>	Tech Review / Date	Admin Review / Date
Simulator	Serial #	Lot #	Expiration																																																																	
0.000		N/A	N/A																																																																	
0.040																																																																				
0.100																																																																				
0.200																																																																				
0.300																																																																				
0.080 DGS	N/A																																																																			
Simulator	Serial #	Lot #	Expiration																																																																	
0.050																																																																				
0.080																																																																				
0.200																																																																				
0.080 DGS	N/A																																																																			
Simulator	Serial Number																																																																			
0.000	MP6284																																																																			
Interferent	MP6285																																																																			
0.050	MP6286																																																																			
0.080	MP6287																																																																			
0.200	MP6288																																																																			
Destinee Armstrong <small>Digitally signed by Destinee Armstrong Date: 2025.01.13 14:00:46 -05'00'</small>	Phil Nicodemo <small>Digitally signed by Phil Nicodemo Date: 2025.01.13 15:28:25 -05'00'</small>																																																																			
Tech Review / Date	Admin Review / Date																																																																			

Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																										
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																										
✓	✓	✓	✓																																																																																																																																										
<p>PALM BEACH COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006240 01/13/2025 Software: 8100.27</p> <p>Test g/210L Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>10:33</td></tr> <tr><td>Control Test</td><td>0.050</td><td>10:33</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:34</td></tr> <tr><td>Control Test</td><td>0.050</td><td>10:34</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:35</td></tr> <tr><td>Control Test</td><td>0.051</td><td>10:35</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:36</td></tr> <tr><td>Control Test</td><td>0.050</td><td>10:36</td></tr> <tr><td>Average</td><td>0.0503</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1471</td><td></td></tr> </table> <p>Operator's Signature </p>	Air Blank	0.000	10:33	Control Test	0.050	10:33	Air Blank	0.000	10:34	Control Test	0.050	10:34	Air Blank	0.000	10:35	Control Test	0.051	10:35	Air Blank	0.000	10:36	Control Test	0.050	10:36	Average	0.0503		Std Dev	0.0006		Rel Std Dev(%)	1.1471		<p>PALM BEACH COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006240 01/13/2025 Software: 8100.27</p> <p>Test g/210L Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>10:09</td></tr> <tr><td>Control Test</td><td>0.080</td><td>10:10</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:10</td></tr> <tr><td>Control Test</td><td>0.080</td><td>10:11</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:12</td></tr> <tr><td>Control Test</td><td>0.080</td><td>10:12</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:13</td></tr> <tr><td>Control Test</td><td>0.080</td><td>10:13</td></tr> <tr><td>Average</td><td>0.0800</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </table> <p>Operator's Signature </p>	Air Blank	0.000	10:09	Control Test	0.080	10:10	Air Blank	0.000	10:10	Control Test	0.080	10:11	Air Blank	0.000	10:12	Control Test	0.080	10:12	Air Blank	0.000	10:13	Control Test	0.080	10:13	Average	0.0800		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>PALM BEACH COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006240 01/13/2025 Software: 8100.27</p> <p>Test g/210L Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>10:17</td></tr> <tr><td>Control Test</td><td>0.198</td><td>10:18</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:18</td></tr> <tr><td>Control Test</td><td>0.198</td><td>10:19</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:20</td></tr> <tr><td>Control Test</td><td>0.198</td><td>10:20</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:21</td></tr> <tr><td>Control Test</td><td>0.1980</td><td>10:21</td></tr> <tr><td>Average</td><td>0.1980</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </table> <p>Operator's Signature </p>	Air Blank	0.000	10:17	Control Test	0.198	10:18	Air Blank	0.000	10:18	Control Test	0.198	10:19	Air Blank	0.000	10:20	Control Test	0.198	10:20	Air Blank	0.000	10:21	Control Test	0.1980	10:21	Average	0.1980		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>PALM BEACH COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006240 01/13/2025 Software: 8100.27</p> <p>Test g/210L Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>10:26</td></tr> <tr><td>Control Test</td><td>0.081</td><td>10:26</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:27</td></tr> <tr><td>Control Test</td><td>0.080</td><td>10:27</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:27</td></tr> <tr><td>Control Test</td><td>0.080</td><td>10:27</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:28</td></tr> <tr><td>Control Test</td><td>0.080</td><td>10:28</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:28</td></tr> <tr><td>Control Test</td><td>0.0803</td><td>10:28</td></tr> <tr><td>Average</td><td>0.0803</td><td></td></tr> <tr><td>Std Dev</td><td>0.0016</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7187</td><td></td></tr> </table> <p>Operator's Signature </p>	Air Blank	0.000	10:26	Control Test	0.081	10:26	Air Blank	0.000	10:27	Control Test	0.080	10:27	Air Blank	0.000	10:27	Control Test	0.080	10:27	Air Blank	0.000	10:28	Control Test	0.080	10:28	Air Blank	0.000	10:28	Control Test	0.0803	10:28	Average	0.0803		Std Dev	0.0016		Rel Std Dev(%)	0.7187	
Air Blank	0.000	10:33																																																																																																																																											
Control Test	0.050	10:33																																																																																																																																											
Air Blank	0.000	10:34																																																																																																																																											
Control Test	0.050	10:34																																																																																																																																											
Air Blank	0.000	10:35																																																																																																																																											
Control Test	0.051	10:35																																																																																																																																											
Air Blank	0.000	10:36																																																																																																																																											
Control Test	0.050	10:36																																																																																																																																											
Average	0.0503																																																																																																																																												
Std Dev	0.0006																																																																																																																																												
Rel Std Dev(%)	1.1471																																																																																																																																												
Air Blank	0.000	10:09																																																																																																																																											
Control Test	0.080	10:10																																																																																																																																											
Air Blank	0.000	10:10																																																																																																																																											
Control Test	0.080	10:11																																																																																																																																											
Air Blank	0.000	10:12																																																																																																																																											
Control Test	0.080	10:12																																																																																																																																											
Air Blank	0.000	10:13																																																																																																																																											
Control Test	0.080	10:13																																																																																																																																											
Average	0.0800																																																																																																																																												
Std Dev	0.0000																																																																																																																																												
Rel Std Dev(%)	0.0000																																																																																																																																												
Air Blank	0.000	10:17																																																																																																																																											
Control Test	0.198	10:18																																																																																																																																											
Air Blank	0.000	10:18																																																																																																																																											
Control Test	0.198	10:19																																																																																																																																											
Air Blank	0.000	10:20																																																																																																																																											
Control Test	0.198	10:20																																																																																																																																											
Air Blank	0.000	10:21																																																																																																																																											
Control Test	0.1980	10:21																																																																																																																																											
Average	0.1980																																																																																																																																												
Std Dev	0.0000																																																																																																																																												
Rel Std Dev(%)	0.0000																																																																																																																																												
Air Blank	0.000	10:26																																																																																																																																											
Control Test	0.081	10:26																																																																																																																																											
Air Blank	0.000	10:27																																																																																																																																											
Control Test	0.080	10:27																																																																																																																																											
Air Blank	0.000	10:27																																																																																																																																											
Control Test	0.080	10:27																																																																																																																																											
Air Blank	0.000	10:28																																																																																																																																											
Control Test	0.080	10:28																																																																																																																																											
Air Blank	0.000	10:28																																																																																																																																											
Control Test	0.0803	10:28																																																																																																																																											
Average	0.0803																																																																																																																																												
Std Dev	0.0016																																																																																																																																												
Rel Std Dev(%)	0.7187																																																																																																																																												

Florida Department of Law Enforcement

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PALM BEACH COUNTY SO
Time of Inspection: 12:54

Date of Inspection: 01/13/2025

Serial Number: 80-006240
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG429602 Exp: 10/22/2026
0.000	0.050	0.080	0.197	0.080
0.000	0.050	0.080	0.197	0.080
0.000	0.050	0.080	0.197	0.080
0.000	0.050	0.079	0.197	0.079
0.000	0.050	0.080	0.197	0.079
0.000	0.050	0.080	0.197	0.079
0.000	0.050	0.080	0.197	0.079
0.000	0.050	0.080	0.197	0.080
0.000	0.050	0.080	0.197	0.080
0.000	0.050	0.080	0.198	0.079


Standard Deviations	0.0000	0.0003	0.0003	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.



I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

01/13/2025
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006240, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006240</u>	UNCERTAINTY* \pm
Owning Agency:	<u>PALM BEACH COUNTY SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>01/13/2025</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>12:54</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

01/13/2025

Date

Taylor D Gutschow

TAYLOR D GUTSCHOW,
Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality