



INSTRUMENT PROCESSING SHEET

Agency PALM BEACH COUNTY SOS/N 80-006478Florida Department of
Law EnforcementDate In 7/16/2025DI Completion Date 07/29/2025☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By <u>KTS</u>	Date <u>7/16/25</u>	Quality Checks	By <u>WKP</u>	Date <u>07/18/2025</u>	Flow Calibration	By _____	Date _____																									
☒ Annual ☐ Registration ☒ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: <u>InstaPak foam packaging</u> <u>in box. KTS 7/16/25</u>			☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>208</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP-103</u> 32 mm <u>0.156</u> (.139 - .169) 36 mm <u>0.171</u> (.156 - .190) 53 mm <u>0.238</u> (.228 - .278) 103 mm <u>0.496</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28421</u> ☒ Stability Checks			Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																											
			<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6291</td><td>202406K 06/19/2026</td></tr><tr><td>0.080</td><td>MP6292</td><td>202406L 06/19/2026</td></tr><tr><td>0.200</td><td>MP6293</td><td>202406N 06/20/2026</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG429602 10/22/2026</td></tr></tbody></table>			Simulator	Serial #	Lot #/Exp	0.050	MP6291	202406K 06/19/2026	0.080	MP6292	202406L 06/19/2026	0.200	MP6293	202406N 06/20/2026	0.080 DGS	N/A	AG429602 10/22/2026	<table border="1"><thead><tr><th colspan="2">Maintenance By _____ Date _____</th></tr><tr><td>☐ Battery Replacement</td><td></td></tr><tr><td>☐ Dry Gas Regulator Replacement</td><td></td></tr><tr><td>☐ Breath Tube Replacement</td><td></td></tr><tr><td>☐ Other _____</td><td></td></tr></thead></table>			Maintenance By _____ Date _____		☐ Battery Replacement		☐ Dry Gas Regulator Replacement		☐ Breath Tube Replacement		☐ Other _____	
Simulator	Serial #	Lot #/Exp																															
0.050	MP6291	202406K 06/19/2026																															
0.080	MP6292	202406L 06/19/2026																															
0.200	MP6293	202406N 06/20/2026																															
0.080 DGS	N/A	AG429602 10/22/2026																															
Maintenance By _____ Date _____																																	
☐ Battery Replacement																																	
☐ Dry Gas Regulator Replacement																																	
☐ Breath Tube Replacement																																	
☐ Other _____																																	
Calibration Adjustment By _____			Department Inspection By <u>SLH</u>																														
Barometric Pressure Gauge _____ ID # _____			Barometric Pressure ID# <u>28421</u>																														
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>			Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Gauge <u>1014</u> Instrument <u>1011</u> Mouth Alcohol Solution Lot # <u>2025-A</u> Acetone Stock Solution Lot # <u>2024-B</u>		
Simulator	Serial #	Lot #	Expiration																														
0.000		N/A	N/A																														
0.040																																	
0.100																																	
0.200																																	
0.300																																	
0.080 DGS	N/A																																
☐ Post Calibration Adjustment Stability Checks			<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6294</td></tr><tr><td>Interferent</td><td>MP5087</td></tr><tr><td>0.050</td><td>MP5088</td></tr><tr><td>0.080</td><td>MP5089</td></tr><tr><td>0.200</td><td>MP5090</td></tr></tbody></table>			Simulator	Serial Number	0.000	MP6294	Interferent	MP5087	0.050	MP5088	0.080	MP5089	0.200	MP5090																
Simulator	Serial Number																																
0.000	MP6294																																
Interferent	MP5087																																
0.050	MP5088																																
0.080	MP5089																																
0.200	MP5090																																
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Attachments ☒ Form 41 ☐ Post-Stability Checks ☒ Stability Checks ☐ Flow Calibration ☒ Calibration Certificate ☐ Form 40 ☐ Calibration Adjustment ☐ Other _____										
Simulator	Serial #	Lot #	Expiration																														
0.050																																	
0.080																																	
0.200																																	
0.080 DGS	N/A																																
Notes/Suggested Service: _____ _____ _____ _____ _____ _____			☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use Shayla Platt Taylor Gutschow 08:08:47 -04'00" Shayla Platt Date: 2025.08.07																														

Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083 ≤ 0.003 of Wet																																																																																																																																																
Performed Root Case Analysis	Performed Root Case Analysis	Performed Root Case Analysis	Performed Root Case Analysis																																																																																																																																																
PALM BEACH CSO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006478 07/18/2025 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:37</td></tr> <tr><td>Control Test</td><td>0.049</td><td>14:38</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:38</td></tr> <tr><td>Control Test</td><td>0.049</td><td>14:39</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:39</td></tr> <tr><td>Control Test</td><td>0.049</td><td>14:40</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:41</td></tr> <tr><td colspan="3">Control Test Stats</td></tr> <tr><td>Average</td><td>0.0490</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature: <i>menpi</i>	Test	g/210L	Time	Air Blank	0.000	14:37	Control Test	0.049	14:38	Air Blank	0.000	14:38	Control Test	0.049	14:39	Air Blank	0.000	14:39	Control Test	0.049	14:40	Air Blank	0.000	14:41	Control Test Stats			Average	0.0490		Std Dev	0.0000		Rel Std Dev(%)	0.0000		PALM BEACH CSO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006478 07/18/2025 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:42</td></tr> <tr><td>Control Test</td><td>0.080</td><td>14:42</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:43</td></tr> <tr><td>Control Test</td><td>0.081</td><td>14:44</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:44</td></tr> <tr><td>Control Test</td><td>0.080</td><td>14:45</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:46</td></tr> <tr><td colspan="3">Control Test Stats</td></tr> <tr><td>Average</td><td>0.0803</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7187</td><td></td></tr> </tbody> </table> Operator's Signature: <i>menpi</i>	Test	g/210L	Time	Air Blank	0.000	14:42	Control Test	0.080	14:42	Air Blank	0.000	14:43	Control Test	0.081	14:44	Air Blank	0.000	14:44	Control Test	0.080	14:45	Air Blank	0.000	14:46	Control Test Stats			Average	0.0803		Std Dev	0.0006		Rel Std Dev(%)	0.7187		PALM BEACH CSO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006478 07/18/2025 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:51</td></tr> <tr><td>Control Test</td><td>0.201</td><td>14:51</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:51</td></tr> <tr><td>Control Test</td><td>0.202</td><td>14:52</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:53</td></tr> <tr><td>Control Test</td><td>0.203</td><td>14:53</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:54</td></tr> <tr><td colspan="3">Control Test Stats</td></tr> <tr><td>Average</td><td>0.2020</td><td></td></tr> <tr><td>Std Dev</td><td>0.0010</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.4950</td><td></td></tr> </tbody> </table> Operator's Signature: <i>menpi</i>	Test	g/210L	Time	Air Blank	0.000	14:51	Control Test	0.201	14:51	Air Blank	0.000	14:51	Control Test	0.202	14:52	Air Blank	0.000	14:53	Control Test	0.203	14:53	Air Blank	0.000	14:54	Control Test Stats			Average	0.2020		Std Dev	0.0010		Rel Std Dev(%)	0.4950		PALM BEACH CSO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006478 07/18/2025 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:56</td></tr> <tr><td>Control Test</td><td>0.078</td><td>14:56</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:56</td></tr> <tr><td>Control Test</td><td>0.079</td><td>14:57</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:57</td></tr> <tr><td>Control Test</td><td>0.079</td><td>14:58</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:58</td></tr> <tr><td colspan="3">Control Test Stats</td></tr> <tr><td>Average</td><td>0.0787</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7339</td><td></td></tr> </tbody> </table> Operator's Signature: <i>menpi</i>	Test	g/210L	Time	Air Blank	0.000	14:56	Control Test	0.078	14:56	Air Blank	0.000	14:56	Control Test	0.079	14:57	Air Blank	0.000	14:57	Control Test	0.079	14:58	Air Blank	0.000	14:58	Control Test Stats			Average	0.0787		Std Dev	0.0006		Rel Std Dev(%)	0.7339	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:37																																																																																																																																																	
Control Test	0.049	14:38																																																																																																																																																	
Air Blank	0.000	14:38																																																																																																																																																	
Control Test	0.049	14:39																																																																																																																																																	
Air Blank	0.000	14:39																																																																																																																																																	
Control Test	0.049	14:40																																																																																																																																																	
Air Blank	0.000	14:41																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0490																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:42																																																																																																																																																	
Control Test	0.080	14:42																																																																																																																																																	
Air Blank	0.000	14:43																																																																																																																																																	
Control Test	0.081	14:44																																																																																																																																																	
Air Blank	0.000	14:44																																																																																																																																																	
Control Test	0.080	14:45																																																																																																																																																	
Air Blank	0.000	14:46																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0803																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7187																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:51																																																																																																																																																	
Control Test	0.201	14:51																																																																																																																																																	
Air Blank	0.000	14:51																																																																																																																																																	
Control Test	0.202	14:52																																																																																																																																																	
Air Blank	0.000	14:53																																																																																																																																																	
Control Test	0.203	14:53																																																																																																																																																	
Air Blank	0.000	14:54																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.2020																																																																																																																																																		
Std Dev	0.0010																																																																																																																																																		
Rel Std Dev(%)	0.4950																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:56																																																																																																																																																	
Control Test	0.078	14:56																																																																																																																																																	
Air Blank	0.000	14:56																																																																																																																																																	
Control Test	0.079	14:57																																																																																																																																																	
Air Blank	0.000	14:57																																																																																																																																																	
Control Test	0.079	14:58																																																																																																																																																	
Air Blank	0.000	14:58																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0787																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7339																																																																																																																																																		

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PALM BEACH CSO

Time of Inspection: 15:06

Date of Inspection: 07/29/2025

Serial Number: 80-006478

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG429602 Exp: 10/22/2026
0.000	0.049	0.078	0.200	0.079
0.000	0.049	0.079	0.199	0.079
0.000	0.048	0.079	0.199	0.078
0.000	0.048	0.079	0.199	0.079
0.000	0.048	0.079	0.199	0.078
0.000	0.048	0.079	0.200	0.079
0.000	0.049	0.079	0.199	0.079
0.000	0.049	0.079	0.200	0.078
0.000	0.049	0.079	0.199	0.078
0.000	0.049	0.079	0.199	0.078


Standard Deviations	0.0005	0.0003	0.0004	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5



Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


 LEANDRA HIGGINBOTHAM
 Signature and Printed Name

07/29/2025
 Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006478, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006478</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>PALM BEACH CSO</u>	0.050 g/210 L	0.004
Calibration Date:	<u>07/29/2025</u>	0.080 g/210 L	0.004
Calibration Time:	<u>15:06</u>	0.200 g/210 L	0.007
		0.080 g/210 L Dry Gas Control	0.005

All results are reported in g/210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or $\pm 1\%$, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

07/29/2025

Date

LEANDRA HIGGINBOTHAM,

Department Inspector

FDLE/ATP Form 69 October 2024

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1