



INSTRUMENT PROCESSING SHEET

Agency Florida Highway PatrolS/N 80-006638Florida Department of
Law EnforcementDate In 6/16/2025DI Completion Date 06/17/2025☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake By <u>KTS</u> Date <u>6/16/25</u>	Quality Checks By <u>SLH</u> Date <u>06/16/2025</u>	Flow Calibration By _____ Date _____																																																																								
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☐ Handle ☒ Keyboard ☐ Dry Gas Shelf ☒ Feet ☐ Breath Tube ☒ Ports ☐ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: <u>DGS attached. KTS</u> <u>6/16/25</u>	☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>210</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP105</u> 32 mm <u>0.152</u> (.139 - .169) 36 mm <u>0.171</u> (.156 - .190) 53 mm <u>0.238</u> (.228 - .278) 103 mm <u>0.503</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28421</u> ☒ Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP5088</td><td>202406K 06/19/2026</td></tr><tr><td>0.080</td><td>MP5089</td><td>202406L 06/19/2026</td></tr><tr><td>0.200</td><td>MP5090</td><td>202406N 06/20/2026</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG429602 10/22/2026</td></tr></tbody></table>	Simulator	Serial #	Lot #/Exp	0.050	MP5088	202406K 06/19/2026	0.080	MP5089	202406L 06/19/2026	0.200	MP5090	202406N 06/20/2026	0.080 DGS	N/A	AG429602 10/22/2026	Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) <table border="1" style="width:100%; border-collapse: collapse;"><thead><tr><th>Maintenance By _____</th><th>Date _____</th></tr></thead><tbody><tr><td colspan="2">☐ Battery Replacement</td></tr><tr><td colspan="2">☐ Dry Gas Regulator Replacement</td></tr><tr><td colspan="2">☐ Breath Tube Replacement</td></tr><tr><td colspan="2">☐ Other _____</td></tr></tbody></table>	Maintenance By _____	Date _____	☐ Battery Replacement		☐ Dry Gas Regulator Replacement		☐ Breath Tube Replacement		☐ Other _____																																																
Simulator	Serial #	Lot #/Exp																																																																								
0.050	MP5088	202406K 06/19/2026																																																																								
0.080	MP5089	202406L 06/19/2026																																																																								
0.200	MP5090	202406N 06/20/2026																																																																								
0.080 DGS	N/A	AG429602 10/22/2026																																																																								
Maintenance By _____	Date _____																																																																									
☐ Battery Replacement																																																																										
☐ Dry Gas Regulator Replacement																																																																										
☐ Breath Tube Replacement																																																																										
☐ Other _____																																																																										
Calibration Adjustment By <u>SLH</u> Barometric Pressure Gauge <u>1013</u> ID # <u>28427</u> <table border="1" style="width:100%; border-collapse: collapse;"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td>MP5086</td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td>MP6295</td><td>24330</td><td>09/10/2026</td></tr><tr><td>0.100</td><td>MP6296</td><td>24390</td><td>10/29/2026</td></tr><tr><td>0.200</td><td>MP6297</td><td>24080</td><td>02/13/2026</td></tr><tr><td>0.300</td><td>MP6298</td><td>24430</td><td>12/10/2026</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>28424080A3</td><td>11/05/2026</td></tr></tbody></table> ☒ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td>MP5088</td><td>202406K</td><td>06/19/2026</td></tr><tr><td>0.080</td><td>MP5089</td><td>202406L</td><td>06/19/2026</td></tr><tr><td>0.200</td><td>MP5090</td><td>202406N</td><td>06/20/2026</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG429602</td><td>10/22/2026</td></tr></tbody></table> Notes/Suggested Service: <u>Barometric pressure on</u> <u>instrument >1% of the barometric pressure gauge.</u> <u>SLH 6/16/25</u> Tech Review: <u>added initials to the Department Inspection box. SLH 6/19/2025</u>		Simulator	Serial #	Lot #	Expiration	0.000	MP5086	N/A	N/A	0.040	MP6295	24330	09/10/2026	0.100	MP6296	24390	10/29/2026	0.200	MP6297	24080	02/13/2026	0.300	MP6298	24430	12/10/2026	0.080 DGS	N/A	28424080A3	11/05/2026	Simulator	Serial #	Lot #	Expiration	0.050	MP5088	202406K	06/19/2026	0.080	MP5089	202406L	06/19/2026	0.200	MP5090	202406N	06/20/2026	0.080 DGS	N/A	AG429602	10/22/2026	Department Inspection By <u>SLH -6/19/25</u> SLH Barometric Pressure ID# <u>28662</u> Gauge <u>1014</u> Instrument <u>1013</u> Mouth Alcohol Solution Lot # <u>2025A</u> Acetone Stock Solution Lot # <u>2024B</u> <table border="1" style="width:100%; border-collapse: collapse;"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6294</td></tr><tr><td>Interferent</td><td>MP5087</td></tr><tr><td>0.050</td><td>MP5088</td></tr><tr><td>0.080</td><td>MP5089</td></tr><tr><td>0.200</td><td>MP5090</td></tr></tbody></table> Attachments <table border="1" style="width:100%; border-collapse: collapse;"><tbody><tr><td>☒ Form 41</td><td>☒ Post-Stability Checks</td></tr><tr><td>☒ Stability Checks</td><td>☐ Flow Calibration</td></tr><tr><td>☒ Calibration Certificate</td><td>☐ Form 40</td></tr><tr><td>☒ Calibration Adjustment</td><td>☐ Other _____</td></tr></tbody></table> ☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use <table style="width:100%;"><tr><td style="width:50%; text-align: center;">Phil Nicodemo Digitally signed by Phil Nicodemo Date: 2025.06.19 11:01:15 -04'00'</td><td style="width:50%; text-align: center;">Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2025.06.19 12:25:00 -04'00'</td></tr><tr><td style="text-align: center;">Tech Review / Date _____</td><td style="text-align: center;">Admin Review / Date _____</td></tr></table>	Simulator	Serial Number	0.000	MP6294	Interferent	MP5087	0.050	MP5088	0.080	MP5089	0.200	MP5090	☒ Form 41	☒ Post-Stability Checks	☒ Stability Checks	☐ Flow Calibration	☒ Calibration Certificate	☐ Form 40	☒ Calibration Adjustment	☐ Other _____	Phil Nicodemo Digitally signed by Phil Nicodemo Date: 2025.06.19 11:01:15 -04'00'	Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2025.06.19 12:25:00 -04'00'	Tech Review / Date _____	Admin Review / Date _____
Simulator	Serial #	Lot #	Expiration																																																																							
0.000	MP5086	N/A	N/A																																																																							
0.040	MP6295	24330	09/10/2026																																																																							
0.100	MP6296	24390	10/29/2026																																																																							
0.200	MP6297	24080	02/13/2026																																																																							
0.300	MP6298	24430	12/10/2026																																																																							
0.080 DGS	N/A	28424080A3	11/05/2026																																																																							
Simulator	Serial #	Lot #	Expiration																																																																							
0.050	MP5088	202406K	06/19/2026																																																																							
0.080	MP5089	202406L	06/19/2026																																																																							
0.200	MP5090	202406N	06/20/2026																																																																							
0.080 DGS	N/A	AG429602	10/22/2026																																																																							
Simulator	Serial Number																																																																									
0.000	MP6294																																																																									
Interferent	MP5087																																																																									
0.050	MP5088																																																																									
0.080	MP5089																																																																									
0.200	MP5090																																																																									
☒ Form 41	☒ Post-Stability Checks																																																																									
☒ Stability Checks	☐ Flow Calibration																																																																									
☒ Calibration Certificate	☐ Form 40																																																																									
☒ Calibration Adjustment	☐ Other _____																																																																									
Phil Nicodemo Digitally signed by Phil Nicodemo Date: 2025.06.19 11:01:15 -04'00'	Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2025.06.19 12:25:00 -04'00'																																																																									
Tech Review / Date _____	Admin Review / Date _____																																																																									


FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006638
06/16/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:13
Control Test	0.050	15:13
Air Blank	0.000	15:14
Control Test	0.050	15:15
Air Blank	0.000	15:15
Control Test	0.050	15:16
Air Blank	0.000	15:16
Control Test Stats		
Average	0.0500	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	


Operator's Signature

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006638
06/16/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:19
Control Test	0.080	15:20
Air Blank	0.000	15:20
Control Test	0.079	15:21
Air Blank	0.000	15:21
Control Test	0.080	15:22
Air Blank	0.000	15:23
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	


Operator's Signature

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006638
06/16/2025
Software: 8100.27


Test	g/210L	Time
Air Blank	0.000	15:26
Control Test	0.197	15:27
Air Blank	0.000	15:27
Control Test	0.197	15:28
Air Blank	0.000	15:29
Control Test	0.196	15:29
Air Blank	0.000	15:30
Control Test Stats		
Average	0.1967	
Std Dev	0.0006	
Rel Std Dev(%)	0.2936	


Operator's Signature

FL HIGHWAY PATROL
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006638
06/16/2025
Software: 8100.27

DGS

Test	g/210L	Time
Air Blank	0.000	15:33
Control Test	0.081	15:33
Air Blank	0.000	15:34
Control Test	0.081	15:34
Air Blank	0.000	15:34
Control Test	0.081	15:35
Air Blank	0.000	15:35
Control Test Stats		
Average	0.0810	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	


Operator's Signature

FL HIGHWAY PATROL
 Intoxilyzer - Alcohol Analyzer
 Model 8000 SN 80-006638
 06/16/2025 16:41:11

Auto Calibration
 Max Power Res Value = 81
 Auto Range Res Value = 74
 Sol Value = 0.000 g/210L ***
 Fit value = 0.0000 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12734, Sum Io = 12638

Channel 1
 Sample % Abs (% Abs Ref)
 Sample #1 = 0.0270 (-0.0040)
 Sample #2 = 0.0370 (0.0100)
 Sample #3 = 0.0360 (0.0280)
 Sample #4 = 0.0100 (0.0520)
 Avg % Abs = 0.0277 (0.0300)
 STD DEV = 0.0153 (0.0211)
 REL STD DEV = 55.330 (70.238)

Channel 2
 Sample % Abs (% Abs Ref)
 Sample #1 = 0.1390 (-0.0050)
 Sample #2 = 0.1570 (-0.0150)
 Sample #3 = 0.1900 (-0.0140)
 Sample #4 = 0.1600 (0.0210)
 Avg % Abs = 0.1690 (-0.0027)
 STD DEV = 0.0182 (0.0205)
 REL STD DEV = 10.798 (768.826)

Sol Value = 0.040 g/210L ***
 Fit value = 0.1905 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12721, Sum Io = 12631

Channel 1
 Sample % Abs (% Abs Ref)
 Sample #1 = 0.7640 (-0.0100)
 Sample #2 = 0.7700 (0.0000)
 Sample #3 = 0.7470 (0.0230)
 Sample #4 = 0.7640 (0.0220)
 Avg % Abs = 0.7603 (0.0150)
 STD DEV = 0.0119 (0.0130)
 REL STD DEV = 1.569 (86.667)

Channel 2
 Sample % Abs (% Abs Ref)
 Sample #1 = 1.5240 (0.0110)
 Sample #2 = 1.5510 (0.0000)
 Sample #3 = 1.5370 (0.0060)
 Sample #4 = 1.5310 (0.0120)
 Avg % Abs = 1.5397 (0.0060)
 STD DEV = 0.0103 (0.0060)
 REL STD DEV = 0.667 (100.000)

Sol Value = 0.100 g/210L ***
 Fit value = 0.4762 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12716, Sum Io = 12629
 Channel 1
 Sample % Abs (% Abs Ref)
 Sample #1 = 1.8390 (-0.0070)
 Sample #2 = 1.8410 (0.0070)
 Sample #3 = 1.8320 (0.0080)
 Sample #4 = 1.8520 (0.0110)
 Avg % Abs = 1.8417 (0.0087)
 STD DEV = 0.0100 (0.0021)
 REL STD DEV = 0.544 (24.019)

Channel 2
 Sample % Abs (% Abs Ref)
 Sample #1 = 3.5440 (0.0130)
 Sample #2 = 3.5450 (0.0230)
 Sample #3 = 3.5360 (0.0190)
 Sample #4 = 3.5400 (0.0150)
 Avg % Abs = 3.5403 (0.0190)
 STD DEV = 0.0045 (0.0040)
 REL STD DEV = 0.127 (21.053)

Sol Value = 0.200 g/210L ***
 Fit value = 0.9524 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12713, Sum Io = 12626

Channel 1
 Sample % Abs (% Abs Ref)
 Sample #1 = 3.6210 (-0.0110)
 Sample #2 = 3.5910 (0.0130)
 Sample #3 = 3.5930 (0.0170)
 Sample #4 = 3.6080 (0.0180)
 Avg % Abs = 3.5973 (0.0160)
 STD DEV = 0.0093 (0.0026)
 REL STD DEV = 0.258 (16.536)

Channel 2
 Sample % Abs (% Abs Ref)
 Sample #1 = 6.7690 (-0.0140)
 Sample #2 = 6.7230 (0.0090)
 Sample #3 = 6.7340 (0.0060)
 Sample #4 = 6.7210 (0.0110)
 Avg % Abs = 6.7260 (0.0087)
 STD DEV = 0.0070 (0.0025)
 REL STD DEV = 0.104 (29.038)

Sol Value = 0.300 g/210L ***
 Fit value = 1.4286 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12709, Sum Io = 12625

Channel 1
 Sample % Abs (% Abs Ref)
 Sample #1 = 5.3100 (-0.0070)
 Sample #2 = 5.3000 (0.0030)
 Sample #3 = 5.3020 (0.0010)
 Sample #4 = 5.3080 (0.0030)
 Avg % Abs = 5.3033 (0.0023)
 STD DEV = 0.0042 (0.0012)
 REL STD DEV = 0.079 (49.487)

Channel 2
 Sample % Abs (% Abs Ref)
 Sample #1 = 9.7820 (0.0000)
 Sample #2 = 9.7620 (0.0120)
 Sample #3 = 9.7420 (0.0260)
 Sample #4 = 9.7520 (0.0240)
 Avg % Abs = 9.7520 (0.0207)
 STD DEV = 0.0100 (0.0076)
 REL STD DEV = 0.103 (36.638)

***** AUTO CAL DATA *****
 Channel 1
 Sol Val = 0.0000 mg/l or 0.000 g/210L
 % Abs = 0.028
 Std Dev = 0.02 Rel Std Dev = 55.33
 Sol Val = 0.1905 mg/l or 0.040 g/210L
 % Abs = 0.760
 Std Dev = 0.01 Rel Std Dev = 1.57
 Sol Val = 0.4762 mg/l or 0.100 g/210L
 % Abs = 1.842
 Std Dev = 0.01 Rel Std Dev = 0.54
 Sol Val = 0.9524 mg/l or 0.200 g/210L
 % Abs = 3.597
 Std Dev = 0.01 Rel Std Dev = 0.26
 Sol Val = 1.4286 mg/l or 0.300 g/210L
 % Abs = 5.303
 Std Dev = 0.00 Rel Std Dev = 0.08
 Zero Order Coef = -71.97
 First Order Coef = 2581.95
 Second Order Coef = 23.65
 Standard Deviation = 1.121778

Channel 2
 Sol Val = 0.0000 mg/l or 0.000 g/210L
 % Abs = 0.169
 Std Dev = 0.02 Rel Std Dev = 10.80
 Sol Val = 0.1905 mg/l or 0.040 g/210L
 % Abs = 1.540
 Std Dev = 0.01 Rel Std Dev = 0.67
 Sol Val = 0.4762 mg/l or 0.100 g/210L
 % Abs = 3.540
 Std Dev = 0.00 Rel Std Dev = 0.13
 Sol Val = 0.9524 mg/l or 0.200 g/210L
 % Abs = 6.726
 Std Dev = 0.01 Rel Std Dev = 0.10
 Sol Val = 1.4286 mg/l or 0.300 g/210L
 % Abs = 9.752
 Std Dev = 0.01 Rel Std Dev = 0.10
 Zero Order Coef = -229.69
 First Order Coef = 1365.39
 Second Order Coef = 12.61
 Standard Deviation = 1.444573

80-006638
Sut 6/16/2025 pg 2 of 2

Solution Stats Quadratic Fit Chan 1		
Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	-0.000	0.0000
0.040	0.040	-0.0000
0.100	0.100	-0.0000
0.200	0.200	0.0000
0.300	0.300	-0.0000

Solution Stats Quadratic Fit Chan 2		
Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.000	-0.0000
0.040	0.040	0.0000
0.100	0.100	-0.0000
0.200	0.200	-0.0000
0.300	0.300	0.0000

Sol Value = 0.080 g/210L ***
Fit value = 0.3810 mg/l %%%
Samples Taken = 4, Discarded = 1

***** CHANNEL 1

Sample #1 = 3561.00
Sample #2 = 3537.00
Sample #3 = 3519.00
Sample #4 = 3519.00
Average Result = 3525.0000
STD DEV = 10.3923
REL STD DEV = 0.295

***** CHANNEL 2

Sample #1 = 3291.00
Sample #2 = 3308.00
Sample #3 = 3306.00
Sample #4 = 3281.00
Average Result = 3298.3333
STD DEV = 15.0444
REL STD DEV = 0.456

Dry Gas H2O Adjust Results *****

Barometric Pressure = 1013

3 um H2O Adjust (mg/l*10,000) = 284

9 um H2O Adjust (mg/l*10,000) = 511

**** AUTO CAL PASS

Post-Calibration Adjustment Stability Checks

80-006638
6/16/2025 sent

0.050 g/210L	0.080 g/210L	0.200 g/210L	DGS 0.080 g/210L																																																																																																																																																
0.047 to 0.053 g/210L ☒	0.077 to 0.083 g/210L ☒	0.194 to 0.206 g/210L ☒	0.077 to 0.083 g/210L ☒ ≤0.003 g/210L of Wet ☒																																																																																																																																																
Performed Root Case Analysis	Performed Root Case Analysis	Performed Root Case Analysis	Performed Root Case Analysis																																																																																																																																																
FL HIGHWAY PATROL Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006638 06/16/2025 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>17:31</td></tr> <tr><td>Control Test</td><td>0.049</td><td>17:31</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>17:32</td></tr> <tr><td>Control Test</td><td>0.049</td><td>17:32</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>17:33</td></tr> <tr><td>Control Test</td><td>0.049</td><td>17:34</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>17:34</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0490</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	17:31	Control Test	0.049	17:31	Air Blank	0.000	17:32	Control Test	0.049	17:32	Air Blank	0.000	17:33	Control Test	0.049	17:34	Air Blank	0.000	17:34	Control Test Stats			Average	0.0490		Std Dev	0.0000		Rel Std Dev(%)	0.0000		FL HIGHWAY PATROL Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006638 06/16/2025 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>17:36</td></tr> <tr><td>Control Test</td><td>0.079</td><td>17:37</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>17:38</td></tr> <tr><td>Control Test</td><td>0.078</td><td>17:38</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>17:39</td></tr> <tr><td>Control Test</td><td>0.078</td><td>17:40</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>17:40</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0783</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7370</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	17:36	Control Test	0.079	17:37	Air Blank	0.000	17:38	Control Test	0.078	17:38	Air Blank	0.000	17:39	Control Test	0.078	17:40	Air Blank	0.000	17:40	Control Test Stats			Average	0.0783		Std Dev	0.0006		Rel Std Dev(%)	0.7370		FL HIGHWAY PATROL Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006638 06/16/2025 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>17:46</td></tr> <tr><td>Control Test</td><td>0.199</td><td>17:46</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>17:47</td></tr> <tr><td>Control Test</td><td>0.199</td><td>17:48</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>17:48</td></tr> <tr><td>Control Test</td><td>0.199</td><td>17:49</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>17:49</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1990</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	17:46	Control Test	0.199	17:46	Air Blank	0.000	17:47	Control Test	0.199	17:48	Air Blank	0.000	17:48	Control Test	0.199	17:49	Air Blank	0.000	17:49	Control Test Stats			Average	0.1990		Std Dev	0.0000		Rel Std Dev(%)	0.0000		FL HIGHWAY PATROL Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006638 06/16/2025 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>17:51</td></tr> <tr><td>Control Test</td><td>0.075</td><td>17:51</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>17:51</td></tr> <tr><td>Control Test</td><td>0.079</td><td>17:52</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>17:52</td></tr> <tr><td>Control Test</td><td>0.080</td><td>17:53</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>17:53</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0793</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7277</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	17:51	Control Test	0.075	17:51	Air Blank	0.000	17:51	Control Test	0.079	17:52	Air Blank	0.000	17:52	Control Test	0.080	17:53	Air Blank	0.000	17:53	Control Test Stats			Average	0.0793		Std Dev	0.0006		Rel Std Dev(%)	0.7277	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	17:31																																																																																																																																																	
Control Test	0.049	17:31																																																																																																																																																	
Air Blank	0.000	17:32																																																																																																																																																	
Control Test	0.049	17:32																																																																																																																																																	
Air Blank	0.000	17:33																																																																																																																																																	
Control Test	0.049	17:34																																																																																																																																																	
Air Blank	0.000	17:34																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0490																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	17:36																																																																																																																																																	
Control Test	0.079	17:37																																																																																																																																																	
Air Blank	0.000	17:38																																																																																																																																																	
Control Test	0.078	17:38																																																																																																																																																	
Air Blank	0.000	17:39																																																																																																																																																	
Control Test	0.078	17:40																																																																																																																																																	
Air Blank	0.000	17:40																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0783																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7370																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	17:46																																																																																																																																																	
Control Test	0.199	17:46																																																																																																																																																	
Air Blank	0.000	17:47																																																																																																																																																	
Control Test	0.199	17:48																																																																																																																																																	
Air Blank	0.000	17:48																																																																																																																																																	
Control Test	0.199	17:49																																																																																																																																																	
Air Blank	0.000	17:49																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1990																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	17:51																																																																																																																																																	
Control Test	0.075	17:51																																																																																																																																																	
Air Blank	0.000	17:51																																																																																																																																																	
Control Test	0.079	17:52																																																																																																																																																	
Air Blank	0.000	17:52																																																																																																																																																	
Control Test	0.080	17:53																																																																																																																																																	
Air Blank	0.000	17:53																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0793																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7277																																																																																																																																																		

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 15:35

Date of Inspection: 06/17/2025

Serial Number: 80-006638
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG429602 Exp: 10/22/2026
0.000	0.048	0.078	0.198	0.079
0.000	0.049	0.078	0.199	0.079
0.000	0.049	0.079	0.199	0.078
0.000	0.049	0.079	0.199	0.079
0.000	0.049	0.078	0.199	0.078
0.000	0.049	0.079	0.199	0.078
0.000	0.049	0.079	0.199	0.078
0.000	0.049	0.079	0.200	0.078
0.000	0.049	0.079	0.200	0.078
0.000	0.049	0.079	0.199	0.078


Standard Deviations	0.0003	0.0004	0.0005	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5


Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


LEANDRA HIGGINBOTHAM
Signature and Printed Name

06/17/2025
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006638, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006638</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>FL HIGHWAY PATROL</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>06/17/2025</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>15:35</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.


Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

06/17/2025

Date


LEANDRA HIGGINBOTHAM,
Department Inspector

FDLE/ATP Form 69 October 2024

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1