



INSTRUMENT PROCESSING SHEET

Agency Monroe County Sheriff's OfficeS/N 80-006693Florida Department of
Law EnforcementDate In 07/01/2025 DI Completion Date 07/18/2025☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By	TDG	Date	07/18/2025	Quality Checks	By	TDG	Date	07/18/2025	Flow Calibration	By	Date																																																																	
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☒ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____					☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>229</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>0.152</u> (.139 - .169) 36 mm <u>0.171</u> (.156 - .190) 53 mm <u>0.242</u> (.228 - .278) 103 mm <u>0.507</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28199</u> ☒ Stability Checks					Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																																																			
					<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6286</td><td>202406K 06/19/2026</td></tr><tr><td>0.080</td><td>MP6287</td><td>202406L 06/19/2026</td></tr><tr><td>0.200</td><td>MP6288</td><td>202406N 06/20/2026</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG429602 10/22/2026</td></tr></tbody></table>	Simulator	Serial #	Lot #/Exp	0.050	MP6286	202406K 06/19/2026	0.080	MP6287	202406L 06/19/2026	0.200	MP6288	202406N 06/20/2026	0.080 DGS	N/A	AG429602 10/22/2026						<table border="1"><thead><tr><th colspan="2">Maintenance By _____ Date _____</th></tr><tr><td>☐ Battery Replacement</td><td></td></tr><tr><td>☐ Dry Gas Regulator Replacement</td><td></td></tr><tr><td>☐ Breath Tube Replacement</td><td></td></tr><tr><td>☐ Other _____</td><td></td></tr><tr><td>_____</td><td></td></tr><tr><td>_____</td><td></td></tr><tr><td>_____</td><td></td></tr><tr><td>_____</td><td></td></tr></thead></table>	Maintenance By _____ Date _____		☐ Battery Replacement		☐ Dry Gas Regulator Replacement		☐ Breath Tube Replacement		☐ Other _____		_____		_____		_____		_____																																		
Simulator	Serial #	Lot #/Exp																																																																											
0.050	MP6286	202406K 06/19/2026																																																																											
0.080	MP6287	202406L 06/19/2026																																																																											
0.200	MP6288	202406N 06/20/2026																																																																											
0.080 DGS	N/A	AG429602 10/22/2026																																																																											
Maintenance By _____ Date _____																																																																													
☐ Battery Replacement																																																																													
☐ Dry Gas Regulator Replacement																																																																													
☐ Breath Tube Replacement																																																																													
☐ Other _____																																																																													

Calibration Adjustment By _____					Department Inspection By <u>TDG</u>																																																																								
Barometric Pressure Gauge _____ ID # _____ <table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table> ☐ Post Calibration Adjustment Stability Checks <table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>					Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Barometric Pressure ID# <u>28199</u> Gauge <u>1022</u> Instrument <u>1020</u> Mouth Alcohol Solution Lot # <u>2024-A</u> Acetone Stock Solution Lot # <u>2023-B</u> <table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6284</td></tr><tr><td>Interferent</td><td>MP6285</td></tr><tr><td>0.050</td><td>MP6286</td></tr><tr><td>0.080</td><td>MP6287</td></tr><tr><td>0.200</td><td>MP6288</td></tr></tbody></table> Attachments <table border="1"><tbody><tr><td>☒ Form 41</td><td>☐ Post-Stability Checks</td></tr><tr><td>☒ Stability Checks</td><td>☐ Flow Calibration</td></tr><tr><td>☒ Calibration Certificate</td><td>☐ Form 40</td></tr><tr><td>☐ Calibration Adjustment</td><td>☐ Other _____</td></tr></tbody></table> ☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use					Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288	☒ Form 41	☐ Post-Stability Checks	☒ Stability Checks	☐ Flow Calibration	☒ Calibration Certificate	☐ Form 40	☐ Calibration Adjustment	☐ Other _____
Simulator	Serial #	Lot #	Expiration																																																																										
0.000		N/A	N/A																																																																										
0.040																																																																													
0.100																																																																													
0.200																																																																													
0.300																																																																													
0.080 DGS	N/A																																																																												
Simulator	Serial #	Lot #	Expiration																																																																										
0.050																																																																													
0.080																																																																													
0.200																																																																													
0.080 DGS	N/A																																																																												
Simulator	Serial Number																																																																												
0.000	MP6284																																																																												
Interferent	MP6285																																																																												
0.050	MP6286																																																																												
0.080	MP6287																																																																												
0.200	MP6288																																																																												
☒ Form 41	☐ Post-Stability Checks																																																																												
☒ Stability Checks	☐ Flow Calibration																																																																												
☒ Calibration Certificate	☐ Form 40																																																																												
☐ Calibration Adjustment	☐ Other _____																																																																												
Notes/Suggested Service: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____					Phil Nicodemo Digitally signed by Phil Nicodemo Date: 2025.08.04 10:35:30 -04'00' Tech Review / Date _____					Shayla Platt Digitally signed by Shayla Platt Date: 2025.08.07 09:04:20 -04'00' Admin Review / Date _____																																																																			

Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																						
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																																						
MONROE COUNTY SO Intoxilizer - Alcohol Analyzer Model 8000 SN 80-006693 07/18/2025 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:18</td></tr> <tr><td>Control Test</td><td>0.048</td><td>12:19</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:19</td></tr> <tr><td>Control Test</td><td>0.048</td><td>12:20</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:20</td></tr> <tr><td>Control Test</td><td>0.048</td><td>12:21</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:22</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0480</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	12:18	Control Test	0.048	12:19	Air Blank	0.000	12:19	Control Test	0.048	12:20	Air Blank	0.000	12:20	Control Test	0.048	12:21	Air Blank	0.000	12:22	Control Test Stats			Average	0.0480		Std Dev	0.0000		Rel Std Dev(%)	0.0000		MONROE COUNTY SO Intoxilizer - Alcohol Analyzer Model 8000 SN 80-006693 07/18/2025 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:26</td></tr> <tr><td>Control Test</td><td>0.078</td><td>12:26</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:27</td></tr> <tr><td>Control Test</td><td>0.078</td><td>12:27</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:28</td></tr> <tr><td>Control Test</td><td>0.078</td><td>12:28</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:29</td></tr> <tr><td>Control Test</td><td>0.078</td><td>12:29</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:29</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0780</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	12:26	Control Test	0.078	12:26	Air Blank	0.000	12:27	Control Test	0.078	12:27	Air Blank	0.000	12:28	Control Test	0.078	12:28	Air Blank	0.000	12:29	Control Test	0.078	12:29	Air Blank	0.000	12:29	Control Test Stats			Average	0.0780		Std Dev	0.0000		Rel Std Dev(%)	0.0000		MONROE COUNTY SO Intoxilizer - Alcohol Analyzer Model 8000 SN 80-006693 07/18/2025 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:33</td></tr> <tr><td>Control Test</td><td>0.197</td><td>12:33</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:34</td></tr> <tr><td>Control Test</td><td>0.197</td><td>12:34</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:35</td></tr> <tr><td>Control Test</td><td>0.196</td><td>12:35</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:35</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1967</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2936</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	12:33	Control Test	0.197	12:33	Air Blank	0.000	12:34	Control Test	0.197	12:34	Air Blank	0.000	12:35	Control Test	0.196	12:35	Air Blank	0.000	12:35	Control Test Stats			Average	0.1967		Std Dev	0.0006		Rel Std Dev(%)	0.2936		MONROE COUNTY SO Intoxilizer - Alcohol Analyzer Model 8000 SN 80-006693 07/18/2025 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:40</td></tr> <tr><td>Control Test</td><td>0.077</td><td>12:40</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:41</td></tr> <tr><td>Control Test</td><td>0.077</td><td>12:41</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:42</td></tr> <tr><td>Control Test</td><td>0.077</td><td>12:42</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:43</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0770</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	12:40	Control Test	0.077	12:40	Air Blank	0.000	12:41	Control Test	0.077	12:41	Air Blank	0.000	12:42	Control Test	0.077	12:42	Air Blank	0.000	12:43	Control Test Stats			Average	0.0770		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																							
Air Blank	0.000	12:18																																																																																																																																																							
Control Test	0.048	12:19																																																																																																																																																							
Air Blank	0.000	12:19																																																																																																																																																							
Control Test	0.048	12:20																																																																																																																																																							
Air Blank	0.000	12:20																																																																																																																																																							
Control Test	0.048	12:21																																																																																																																																																							
Air Blank	0.000	12:22																																																																																																																																																							
Control Test Stats																																																																																																																																																									
Average	0.0480																																																																																																																																																								
Std Dev	0.0000																																																																																																																																																								
Rel Std Dev(%)	0.0000																																																																																																																																																								
Test	g/210L	Time																																																																																																																																																							
Air Blank	0.000	12:26																																																																																																																																																							
Control Test	0.078	12:26																																																																																																																																																							
Air Blank	0.000	12:27																																																																																																																																																							
Control Test	0.078	12:27																																																																																																																																																							
Air Blank	0.000	12:28																																																																																																																																																							
Control Test	0.078	12:28																																																																																																																																																							
Air Blank	0.000	12:29																																																																																																																																																							
Control Test	0.078	12:29																																																																																																																																																							
Air Blank	0.000	12:29																																																																																																																																																							
Control Test Stats																																																																																																																																																									
Average	0.0780																																																																																																																																																								
Std Dev	0.0000																																																																																																																																																								
Rel Std Dev(%)	0.0000																																																																																																																																																								
Test	g/210L	Time																																																																																																																																																							
Air Blank	0.000	12:33																																																																																																																																																							
Control Test	0.197	12:33																																																																																																																																																							
Air Blank	0.000	12:34																																																																																																																																																							
Control Test	0.197	12:34																																																																																																																																																							
Air Blank	0.000	12:35																																																																																																																																																							
Control Test	0.196	12:35																																																																																																																																																							
Air Blank	0.000	12:35																																																																																																																																																							
Control Test Stats																																																																																																																																																									
Average	0.1967																																																																																																																																																								
Std Dev	0.0006																																																																																																																																																								
Rel Std Dev(%)	0.2936																																																																																																																																																								
Test	g/210L	Time																																																																																																																																																							
Air Blank	0.000	12:40																																																																																																																																																							
Control Test	0.077	12:40																																																																																																																																																							
Air Blank	0.000	12:41																																																																																																																																																							
Control Test	0.077	12:41																																																																																																																																																							
Air Blank	0.000	12:42																																																																																																																																																							
Control Test	0.077	12:42																																																																																																																																																							
Air Blank	0.000	12:43																																																																																																																																																							
Control Test Stats																																																																																																																																																									
Average	0.0770																																																																																																																																																								
Std Dev	0.0000																																																																																																																																																								
Rel Std Dev(%)	0.0000																																																																																																																																																								

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MONROE COUNTY SO
Time of Inspection: 15:22

Date of Inspection: 07/18/2025

Serial Number: 80-006693
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG429602 Exp: 10/22/2026
0.000	0.047	0.078	0.196	0.077
0.000	0.048	0.078	0.196	0.077
0.000	0.048	0.078	0.196	0.077
0.000	0.048	0.078	0.196	0.077
0.000	0.048	0.077	0.196	0.077
0.000	0.048	0.078	0.196	0.077
0.000	0.048	0.078	0.196	0.077
0.000	0.048	0.078	0.196	0.077
0.000	0.048	0.078	0.196	0.077
0.000	0.048	0.078	0.196	0.077
0.000	0.048	0.078	0.196	0.077


Standard Deviations	0.0003	0.0003	0.0000	0.0000
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0001 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

07/18/2025
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006693, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006693</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>MONROE COUNTY SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>07/18/2025</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>15:22</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Taylor
Gutschow
Digitally signed by Taylor
Gutschow
Date: 2025.07.18 15:39:06
-04'00'

07/18/2025

Date

TAYLOR D GUTSCHOW,
Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality