



INSTRUMENT PROCESSING SHEET

Agency Broward CSOS/N 80-006925Florida Department of
Law EnforcementDate In 01/23/2025 DI Completion Date N/A☐ Ship ☐ P/U ☐ H/D ☒ CMI ☐ EE

Intake By TDG _____ Date <u>02/04/2025</u> ☒ Annual ☐ Registration ☒ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☒ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____	Quality Checks By TDG _____ Date <u>02/05/2025</u> ☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>114</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP101</u> 32 mm <u>0.171*</u> (.139 - .169) 36 mm <u>0.191*</u> (.156 - .190) 53 mm <u>0.273</u> (.228 - .278) 103 mm <u>0.539</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>26932</u> ☒ Stability Checks <table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6286</td><td>202303K 03/29/2025</td></tr><tr><td>0.080</td><td>MP6287</td><td>202303L 03/29/2025</td></tr><tr><td>0.200</td><td>MP6288</td><td>202304C 04/05/2025</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG429602 10/22/2026</td></tr></tbody></table>	Simulator	Serial #	Lot #/Exp	0.050	MP6286	202303K 03/29/2025	0.080	MP6287	202303L 03/29/2025	0.200	MP6288	202304C 04/05/2025	0.080 DGS	N/A	AG429602 10/22/2026	Flow Calibration By TDG _____ Date <u>02/05/2025</u> Flow Column # <u>ATP106</u> ☒ 5L/min – 17mm ☒ 15L/min – 53mm ☒ 30L/min – 103mm ☒ R-Value <u>115</u> ☒ Post Calibration Verification (L/s) Flow Column # <u>ATP101</u> 32 mm <u>0.156</u> (.139 - .169) 36 mm <u>0.175</u> (.156 - .190) 53 mm <u>0.250</u> (.228 - .278) 103 mm <u>0.515</u> (.447 - .547) Maintenance By _____ Date _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ _____ _____ _____ _____																																	
Simulator	Serial #	Lot #/Exp																																																
0.050	MP6286	202303K 03/29/2025																																																
0.080	MP6287	202303L 03/29/2025																																																
0.200	MP6288	202304C 04/05/2025																																																
0.080 DGS	N/A	AG429602 10/22/2026																																																
Calibration Adjustment By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table> ☐ Post Calibration Adjustment Stability Checks <table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table> Notes/Suggested Service: <u>*Outside nominal. (TDG 2/5/25)</u> <u>Instrument reads 0's on simulator solutions. Returning to CMI for evaluation. (TDG 2/5/25)</u> _____ _____ _____ _____			Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A		
Simulator	Serial #	Lot #	Expiration																																															
0.000		N/A	N/A																																															
0.040																																																		
0.100																																																		
0.200																																																		
0.300																																																		
0.080 DGS	N/A																																																	
Simulator	Serial #	Lot #	Expiration																																															
0.050																																																		
0.080																																																		
0.200																																																		
0.080 DGS	N/A																																																	
Department Inspection By _____ Barometric Pressure ID# _____ Gauge _____ Instrument _____ Mouth Alcohol Solution Lot # _____ Acetone Stock Solution Lot # _____ <table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td></td></tr><tr><td>Interferent</td><td></td></tr><tr><td>0.050</td><td></td></tr><tr><td>0.080</td><td></td></tr><tr><td>0.200</td><td></td></tr></tbody></table> Attachments ☐ Form 41 ☐ Post-Stability Checks ☒ Stability Checks ☒ Flow Calibration ☐ Calibration Certificate ☐ Form 40 ☐ Calibration Adjustment ☒ Other <u>Form 51</u> ☐ Instrument Complies with Chapter 11D-8, FAC ☒ Instrument Does Not Comply with Chapter 11D-8, FAC ☐ Return to/Place into Evidentiary Use ☒ Remain Out of Evidentiary Use ☐ Conduct an Agency Inspection Before Evidentiary Use Phil Nicodemo Digitally signed by Phil Nicodemo Date: 2025.02.12 11:24:05 -05'00' Shayla Platt by Shayla Platt Date: 2025.02.12 11:52:38 -05'00' Tech Review / Date _____ Admin Review / Date _____			Simulator	Serial Number	0.000		Interferent		0.050		0.080		0.200																																					
Simulator	Serial Number																																																	
0.000																																																		
Interferent																																																		
0.050																																																		
0.080																																																		
0.200																																																		

Flow Calibration Adjustment(s)

Performed by TDG

BROWARD COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006925
02/05/2025
Software: 8100.27

Flow Rate Calibration*****

1: Rate (Liters/min) = 5
SQRT(Diff) = 5.383
2: Rate (Liters/min) = 15
SQRT(Diff) = 10.953
3: Rate (Liters/min) = 30
SQRT(Diff) = 20.270
Dependent Data Scale Factor = 100000 L/min
Independent Data Scale Factor = 256
Rounded Slope = 653
Rounded Intercept = -373752
Correlation = 0.99957

Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																				
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																				
BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 02/05/2025 Software: 8100.27 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>11:21</td></tr> <tr><td>Control Test</td><td>0.000</td><td>11:22</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:22</td></tr> <tr><td>Control Test</td><td>0.000</td><td>11:23</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:24</td></tr> <tr><td>Control Test</td><td>0.000</td><td>11:24</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:25</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0000</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </table>	Air Blank	0.000	11:21	Control Test	0.000	11:22	Air Blank	0.000	11:22	Control Test	0.000	11:23	Air Blank	0.000	11:24	Control Test	0.000	11:24	Air Blank	0.000	11:25	Control Test Stats			Average	0.0000		Std Dev	0.0000		Rel Std Dev(%)	0.0000		BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 02/05/2025 Software: 8100.27 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>11:28</td></tr> <tr><td>Control Test</td><td>0.000</td><td>11:28</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:29</td></tr> <tr><td>Control Test</td><td>0.000</td><td>11:29</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:30</td></tr> <tr><td>Control Test</td><td>0.000</td><td>11:31</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:31</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0000</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </table>	Air Blank	0.000	11:28	Control Test	0.000	11:28	Air Blank	0.000	11:29	Control Test	0.000	11:29	Air Blank	0.000	11:30	Control Test	0.000	11:31	Air Blank	0.000	11:31	Control Test Stats			Average	0.0000		Std Dev	0.0000		Rel Std Dev(%)	0.0000		BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 02/05/2025 Software: 8100.27 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>11:15</td></tr> <tr><td>Control Test</td><td>0.000</td><td>11:15</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:16</td></tr> <tr><td>Control Test</td><td>0.000</td><td>11:16</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:17</td></tr> <tr><td>Control Test</td><td>0.000</td><td>11:18</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:18</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0000</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </table>	Air Blank	0.000	11:15	Control Test	0.000	11:15	Air Blank	0.000	11:16	Control Test	0.000	11:16	Air Blank	0.000	11:17	Control Test	0.000	11:18	Air Blank	0.000	11:18	Control Test Stats			Average	0.0000		Std Dev	0.0000		Rel Std Dev(%)	0.0000		BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 02/05/2025 Software: 8100.27 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>11:03</td></tr> <tr><td>Control Test</td><td>0.076</td><td>11:04</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:04</td></tr> <tr><td>Control Test</td><td>0.076</td><td>11:04</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:05</td></tr> <tr><td>Control Test</td><td>0.076</td><td>11:05</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:06</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0760</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </table>	Air Blank	0.000	11:03	Control Test	0.076	11:04	Air Blank	0.000	11:04	Control Test	0.076	11:04	Air Blank	0.000	11:05	Control Test	0.076	11:05	Air Blank	0.000	11:06	Control Test Stats			Average	0.0760		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Air Blank	0.000	11:21																																																																																																																																					
Control Test	0.000	11:22																																																																																																																																					
Air Blank	0.000	11:22																																																																																																																																					
Control Test	0.000	11:23																																																																																																																																					
Air Blank	0.000	11:24																																																																																																																																					
Control Test	0.000	11:24																																																																																																																																					
Air Blank	0.000	11:25																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0000																																																																																																																																						
Std Dev	0.0000																																																																																																																																						
Rel Std Dev(%)	0.0000																																																																																																																																						
Air Blank	0.000	11:28																																																																																																																																					
Control Test	0.000	11:28																																																																																																																																					
Air Blank	0.000	11:29																																																																																																																																					
Control Test	0.000	11:29																																																																																																																																					
Air Blank	0.000	11:30																																																																																																																																					
Control Test	0.000	11:31																																																																																																																																					
Air Blank	0.000	11:31																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0000																																																																																																																																						
Std Dev	0.0000																																																																																																																																						
Rel Std Dev(%)	0.0000																																																																																																																																						
Air Blank	0.000	11:15																																																																																																																																					
Control Test	0.000	11:15																																																																																																																																					
Air Blank	0.000	11:16																																																																																																																																					
Control Test	0.000	11:16																																																																																																																																					
Air Blank	0.000	11:17																																																																																																																																					
Control Test	0.000	11:18																																																																																																																																					
Air Blank	0.000	11:18																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0000																																																																																																																																						
Std Dev	0.0000																																																																																																																																						
Rel Std Dev(%)	0.0000																																																																																																																																						
Air Blank	0.000	11:03																																																																																																																																					
Control Test	0.076	11:04																																																																																																																																					
Air Blank	0.000	11:04																																																																																																																																					
Control Test	0.076	11:04																																																																																																																																					
Air Blank	0.000	11:05																																																																																																																																					
Control Test	0.076	11:05																																																																																																																																					
Air Blank	0.000	11:06																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0760																																																																																																																																						
Std Dev	0.0000																																																																																																																																						
Rel Std Dev(%)	0.0000																																																																																																																																						
Operator's Signature	Operator's Signature	Operator's Signature	Operator's Signature																																																																																																																																				

Return Material Authorization

Ship to: ☒ CMI, Inc.

☐ Enforcement Electronics

Shipment to repair facility authorized by: Anaya Frazier on 2/5/2025

Items Returned: Instrument ☒ Supplies ☐ Other ☐ Describe: _____

Instrument Model: Intoxilyzer 8000 Serial Number: 80-006925

Bill To Address:

Broward County Sheriff's Office

Attn: Anaya Frazier

Ship to Address:

Florida Department of Law Enforcement

Fort Myers Regional Operations Center

Attn: Taylor Gutschow

4700 Terminal Drive, Suite 1

Fort Myers, FL 33907

Reason for Return:

The instrument just returned from repair (see work order 409021). It reads simulator solutions as 0 (there are no bubbles in the simulator when a sample is collected). The R-value is close to 100. Instrument also required a flow calibration adjustment, which was performed by FDLE.

Please choose one of the following options:

☐ 1. I _____, authorize all repairs.

☐ 2. I _____, authorize repairs up to \$_____.

☒ 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Anaya Frazier

Phone #: 1-305-218-6752

Email: Anaya_Frazier@sheriff.org

ATP Contact Name: Taylor Gutschow

ATP Email: TaylorGutschow@fdle.state.fl.us