

Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083 ≤ 0.003 of Wet																																																																																																																																																
BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007008 01/16/2025 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:14</td></tr> <tr><td>Control Test</td><td>0.049</td><td>12:15</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:15</td></tr> <tr><td>Control Test</td><td>0.050</td><td>12:16</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:16</td></tr> <tr><td>Control Test</td><td>0.049</td><td>12:17</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:18</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0493</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1703</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	12:14	Control Test	0.049	12:15	Air Blank	0.000	12:15	Control Test	0.050	12:16	Air Blank	0.000	12:16	Control Test	0.049	12:17	Air Blank	0.000	12:18	Control Test Stats			Average	0.0493		Std Dev	0.0006		Rel Std Dev(%)	1.1703		BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007008 01/16/2025 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:21</td></tr> <tr><td>Control Test</td><td>0.080</td><td>12:21</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:22</td></tr> <tr><td>Control Test</td><td>0.080</td><td>12:23</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:23</td></tr> <tr><td>Control Test</td><td>0.080</td><td>12:24</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:24</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0800</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	12:21	Control Test	0.080	12:21	Air Blank	0.000	12:22	Control Test	0.080	12:23	Air Blank	0.000	12:23	Control Test	0.080	12:24	Air Blank	0.000	12:24	Control Test Stats			Average	0.0800		Std Dev	0.0000		Rel Std Dev(%)	0.0000		BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007008 01/16/2025 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:07</td></tr> <tr><td>Control Test</td><td>0.198</td><td>12:07</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:08</td></tr> <tr><td>Control Test</td><td>0.198</td><td>12:08</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:09</td></tr> <tr><td>Control Test</td><td>0.198</td><td>12:10</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:10</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1980</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	12:07	Control Test	0.198	12:07	Air Blank	0.000	12:08	Control Test	0.198	12:08	Air Blank	0.000	12:09	Control Test	0.198	12:10	Air Blank	0.000	12:10	Control Test Stats			Average	0.1980		Std Dev	0.0000		Rel Std Dev(%)	0.0000		ALD #1 BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007008 01/16/2025 Software: 8100.27 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:02</td></tr> <tr><td>Control Test</td><td>0.081</td><td>12:03</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:03</td></tr> <tr><td>Control Test</td><td>0.081</td><td>12:03</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:04</td></tr> <tr><td>Control Test</td><td>INT*</td><td>12:04</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:05</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0807</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7157</td><td></td></tr> </tbody> </table> *Interferent Detect Operator's Signature 	Test	g/210L	Time	Air Blank	0.000	12:02	Control Test	0.081	12:03	Air Blank	0.000	12:03	Control Test	0.081	12:03	Air Blank	0.000	12:04	Control Test	INT*	12:04	Air Blank	0.000	12:05	Control Test Stats			Average	0.0807		Std Dev	0.0006		Rel Std Dev(%)	0.7157	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:14																																																																																																																																																	
Control Test	0.049	12:15																																																																																																																																																	
Air Blank	0.000	12:15																																																																																																																																																	
Control Test	0.050	12:16																																																																																																																																																	
Air Blank	0.000	12:16																																																																																																																																																	
Control Test	0.049	12:17																																																																																																																																																	
Air Blank	0.000	12:18																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0493																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1703																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:21																																																																																																																																																	
Control Test	0.080	12:21																																																																																																																																																	
Air Blank	0.000	12:22																																																																																																																																																	
Control Test	0.080	12:23																																																																																																																																																	
Air Blank	0.000	12:23																																																																																																																																																	
Control Test	0.080	12:24																																																																																																																																																	
Air Blank	0.000	12:24																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0800																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:07																																																																																																																																																	
Control Test	0.198	12:07																																																																																																																																																	
Air Blank	0.000	12:08																																																																																																																																																	
Control Test	0.198	12:08																																																																																																																																																	
Air Blank	0.000	12:09																																																																																																																																																	
Control Test	0.198	12:10																																																																																																																																																	
Air Blank	0.000	12:10																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1980																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:02																																																																																																																																																	
Control Test	0.081	12:03																																																																																																																																																	
Air Blank	0.000	12:03																																																																																																																																																	
Control Test	0.081	12:03																																																																																																																																																	
Air Blank	0.000	12:04																																																																																																																																																	
Control Test	INT*	12:04																																																																																																																																																	
Air Blank	0.000	12:05																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0807																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7157																																																																																																																																																		

Stability Checks

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083 ✓ ≤0.003 of Wet ✓																								
<i>015</i> <i>#2</i> BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 01/16/2025 Software: 8100.27 Test <table border="1"> <thead> <tr> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000 12:26</td></tr> <tr><td>Control Test</td><td>0.080 12:26</td></tr> <tr><td>Air Blank</td><td>0.000 12:27</td></tr> <tr><td>Control Test</td><td>0.080 12:27</td></tr> <tr><td>Air Blank</td><td>0.000 12:28</td></tr> <tr><td>Control Test</td><td>0.081 12:28</td></tr> <tr><td>Air Blank</td><td>0.000 12:28</td></tr> <tr><td>Control Test Stats</td><td></td></tr> <tr><td>Average</td><td>0.0803</td></tr> <tr><td>Std Dev</td><td>0.0006</td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7187</td></tr> </tbody> </table> <i>MK</i> Operator's Signature				g/210L	Time	Air Blank	0.000 12:26	Control Test	0.080 12:26	Air Blank	0.000 12:27	Control Test	0.080 12:27	Air Blank	0.000 12:28	Control Test	0.081 12:28	Air Blank	0.000 12:28	Control Test Stats		Average	0.0803	Std Dev	0.0006	Rel Std Dev(%)	0.7187
g/210L	Time																										
Air Blank	0.000 12:26																										
Control Test	0.080 12:26																										
Air Blank	0.000 12:27																										
Control Test	0.080 12:27																										
Air Blank	0.000 12:28																										
Control Test	0.081 12:28																										
Air Blank	0.000 12:28																										
Control Test Stats																											
Average	0.0803																										
Std Dev	0.0006																										
Rel Std Dev(%)	0.7187																										

DLS

Extra

DLS

Stabilities
Stabilities

DLS

BROWARD COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007008
01/16/2025
Software: 8100.27

BROWARD COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007008
01/17/2025
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:30
Control Test	0.081	12:30
Air Blank	0.000	12:30
Control Test	0.081	12:31
Air Blank	0.000	12:31
Control Test	0.081	12:32
Air Blank	0.000	12:32
Control Test	0.081	12:32
Air Blank	0.000	12:33
Control Test	0.081	12:33
Air Blank	0.000	12:34
Control Test	0.081	12:34
Air Blank	0.000	12:35
Control Test	0.082	12:35
Air Blank	0.000	12:35
Control Test	0.081	12:36
Air Blank	0.000	12:36
Control Test	0.081	12:37
Air Blank	0.000	12:37
Control Test	0.081	12:37
Air Blank	0.000	12:38
Control Test	0.081	12:38
Air Blank	0.000	12:39
Control Test	0.081	12:39
Air Blank	0.000	12:40
Control Test	0.081	12:40
Air Blank	0.000	12:40
Control Test	0.082	12:41
Air Blank	0.000	12:41
Control Test	0.081	12:42
Air Blank	0.000	12:42
Control Test	0.081	12:42
Air Blank	0.000	12:43
Control Test	0.082	12:43
Air Blank	0.000	12:44
Control Test	0.081	12:44
Air Blank	0.000	12:45
Control Test	0.081	12:45
Air Blank	0.000	12:45
Control Test	0.081	12:46
Air Blank	0.000	12:46
Control Test Stats		
Average	0.0811	
Std Dev	0.0004	
Rel Std Dev(%)	0.4514	

Ran extra stabilities on two
different days.

ML
11/17/25

Test	g/210L	Time
Air Blank	0.000	09:31
Control Test	0.081	09:31
Air Blank	0.000	09:32
Control Test	0.081	09:32
Air Blank	0.000	09:33
Control Test	0.082	09:33
Air Blank	0.000	09:33
Control Test	0.081	09:34
Air Blank	0.000	09:34
Control Test	0.082	09:35
Air Blank	0.000	09:35
Control Test	0.081	09:35
Air Blank	0.000	09:36
Control Test	0.082	09:36
Air Blank	0.000	09:37
Control Test	0.082	09:37
Air Blank	0.000	09:38
Control Test	0.082	09:38
Air Blank	0.000	09:39
Control Test	0.082	09:39
Air Blank	0.000	09:39
Control Test	0.081	09:40
Air Blank	0.000	09:40
Control Test	0.082	09:41
Air Blank	0.000	09:41
Control Test	0.081	09:41
Air Blank	0.000	09:42
Control Test	0.082	09:42
Air Blank	0.000	09:43
Control Test	0.082	09:43
Air Blank	0.000	09:44
Control Test	0.082	09:44
Air Blank	0.000	09:44
Control Test	0.082	09:45
Air Blank	0.000	09:45
Control Test	0.082	09:46
Air Blank	0.000	09:46
Control Test	0.082	09:46
Air Blank	0.000	09:47
Control Test	0.082	09:47
Air Blank	0.000	09:48
Control Test Stats		
Average	0.0817	
Std Dev	0.0005	
Rel Std Dev(%)	0.5755	

Operator's Signature

Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: BROWARD COUNTY SO
Time of Inspection: 13:52

Date of Inspection: 01/17/2025

Serial Number: 80-007008
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#: 202303L 202303L Exp: 03/29/2025 706-112415	0.20g/210L Test (g/210L) Lot#: 202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: AG429602 Exp: 10/22/2026
0.000	0.049	0.079	0.199	0.080
0.000	0.050	0.079	0.198	0.080
0.000	0.049	0.079	0.198	0.081
0.000	0.050	0.079	0.197	0.081
0.000	0.050	0.079	0.198	0.080
0.000	0.050	0.079	0.198	0.080
0.000	0.050	0.078	0.198	0.080
0.000	0.050	0.079	0.198	0.080
0.000	0.050	0.079	0.198	0.080
0.000	0.050	0.079	0.198	0.080


Standard Deviations	0.0004	0.0003	0.0004	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

01/17/2025
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007008, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-007008</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>BROWARD COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>01/17/2025</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>13:52</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

01/17/2025

Date

TAYLOR D GUTSCHOW,
Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality