



INSTRUMENT PROCESSING SHEET

Agency FFWCCS/N 80-007168Florida Department of
Law EnforcementDate In 06/02/2025DI Completion Date 06/30/2025☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake By <u>TDG</u> Date <u>06/24/2025</u>		Quality Checks By <u>TDG</u> Date <u>06/30/2025</u>		Flow Calibration By <u>TDG</u> Date <u>06/30/2025</u>																																									
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☒ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: <u>Oscar Saavedra dropped off with DGS in pelican case.</u>		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>232</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP106</u> 32 mm <u>0.144</u> (.139 - .169) 36 mm <u>0.160</u> (.156 - .190) 53 mm <u>0.226*</u> (.228 - .278) 103 mm <u>0.472</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28199</u> ☒ Stability Checks		☒ Flow Column # <u>ATP101</u> ☒ 5L/min – 17mm ☒ 15L/min – 53mm ☒ 30L/min – 103mm ☒ R-Value <u>235</u> ☒ Post Calibration Verification (L/s) Flow Column # <u>ATP106</u> 32 mm <u>0.144</u> (.139 - .169) 36 mm <u>0.160</u> (.156 - .190) 53 mm <u>0.230</u> (.228 - .278) 103 mm <u>0.500</u> (.447 - .547)																																									
		<table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6286</td><td>202406K 06/19/2026</td></tr><tr><td>0.080</td><td>MP6287</td><td>202406L 06/19/2026</td></tr><tr><td>0.200</td><td>MP6288</td><td>202406N 06/20/2026</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG429602 10/22/2026</td></tr></tbody></table>		Simulator	Serial #	Lot #/Exp	0.050	MP6286	202406K 06/19/2026	0.080	MP6287	202406L 06/19/2026	0.200	MP6288	202406N 06/20/2026	0.080 DGS	N/A	AG429602 10/22/2026	Maintenance By _____ Date _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____																										
Simulator	Serial #	Lot #/Exp																																											
0.050	MP6286	202406K 06/19/2026																																											
0.080	MP6287	202406L 06/19/2026																																											
0.200	MP6288	202406N 06/20/2026																																											
0.080 DGS	N/A	AG429602 10/22/2026																																											
Calibration Adjustment By _____		Department Inspection By <u>TDG</u>																																											
Barometric Pressure Gauge _____ ID # _____		Barometric Pressure ID# <u>28199</u> Gauge <u>1021</u> Instrument <u>1018</u> Mouth Alcohol Solution Lot # <u>2024-A</u> Acetone Stock Solution Lot # <u>2023-B</u>																																											
<table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			<table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6284</td></tr><tr><td>Interferent</td><td>MP6285</td></tr><tr><td>0.050</td><td>MP6286</td></tr><tr><td>0.080</td><td>MP6287</td></tr><tr><td>0.200</td><td>MP6288</td></tr></tbody></table>				Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288
Simulator	Serial #	Lot #	Expiration																																										
0.000		N/A	N/A																																										
0.040																																													
0.100																																													
0.200																																													
0.300																																													
0.080 DGS	N/A																																												
Simulator	Serial Number																																												
0.000	MP6284																																												
Interferent	MP6285																																												
0.050	MP6286																																												
0.080	MP6287																																												
0.200	MP6288																																												
☐ Post Calibration Adjustment Stability Checks		Attachments ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment ☐ Post-Stability Checks ☒ Flow Calibration ☐ Form 40 ☐ Other _____																																											
<table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																							
Simulator	Serial #	Lot #	Expiration																																										
0.050																																													
0.080																																													
0.200																																													
0.080 DGS	N/A																																												
Notes/Suggested Service: <u>*Outside nominal (TDG 6/30/25)</u>		<table border="1" style="width:100%"><tr><td>Phil Nicodemo Digitally signed by Phil Nicodemo Date: 2025.07.01 12:34:14 +04'00'</td><td>LeAndra Higginbotham Digitally signed by LeAndra Higginbotham Date: 2025.07.08 08:17:10 -04'00'</td></tr><tr><td>Tech Review / Date</td><td>Admin Review / Date</td></tr></table>				Phil Nicodemo Digitally signed by Phil Nicodemo Date: 2025.07.01 12:34:14 +04'00'	LeAndra Higginbotham Digitally signed by LeAndra Higginbotham Date: 2025.07.08 08:17:10 -04'00'	Tech Review / Date	Admin Review / Date																																				
Phil Nicodemo Digitally signed by Phil Nicodemo Date: 2025.07.01 12:34:14 +04'00'	LeAndra Higginbotham Digitally signed by LeAndra Higginbotham Date: 2025.07.08 08:17:10 -04'00'																																												
Tech Review / Date	Admin Review / Date																																												

Flow Calibration Adjustment(s)

Performed by TDG

Performed root cause analysis on failed flow verification prior to cal adjustment. No user / equipment error identified.

ML
6/30/25

FFWCC
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007168
06/30/2025
Software: 8100.27

Flow Rate Calibration*****
1: Rate (Liters/min) = 5
SQRT(Diff)) = 6.926
2: Rate (Liters/min) = 15
SQRT(Diff)) = 11.574
3: Rate (Liters/min) = 30
SQRT(Diff)) = 19.973
Dependent Data Scale Factor = 100000 L/min
Independent Data Scale Factor = 256
Rounded Slope = 742
Rounded Intercept = -770636
Correlation = 0.99878

Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																																
✓	✓	✓	✓																																																																																																																																																
0.003 of Wet	✓	✓	✓																																																																																																																																																
FFWCC Intoxilyzer - Alcohol Analyzer Model 8000 06/30/2025 Software: 8100.27 SN 80-007168 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:16</td></tr> <tr><td>Control Test</td><td>0.049</td><td>11:17</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:17</td></tr> <tr><td>Control Test</td><td>0.048</td><td>11:18</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:18</td></tr> <tr><td>Control Test</td><td>0.048</td><td>11:19</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:20</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0483</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1945</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	11:16	Control Test	0.049	11:17	Air Blank	0.000	11:17	Control Test	0.048	11:18	Air Blank	0.000	11:18	Control Test	0.048	11:19	Air Blank	0.000	11:20	Control Test Stats			Average	0.0483		Std Dev	0.0006		Rel Std Dev(%)	1.1945		FFWCC Intoxilyzer - Alcohol Analyzer Model 8000 06/30/2025 Software: 8100.27 SN 80-007168 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:26</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:26</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:27</td></tr> <tr><td>Control Test</td><td>0.078</td><td>11:28</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:28</td></tr> <tr><td>Control Test</td><td>0.078</td><td>11:29</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:29</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0783</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7370</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	11:26	Control Test	0.079	11:26	Air Blank	0.000	11:27	Control Test	0.078	11:28	Air Blank	0.000	11:28	Control Test	0.078	11:29	Air Blank	0.000	11:29	Control Test Stats			Average	0.0783		Std Dev	0.0006		Rel Std Dev(%)	0.7370		FFWCC Intoxilyzer - Alcohol Analyzer Model 8000 06/30/2025 Software: 8100.27 SN 80-007168 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>10:59</td></tr> <tr><td>Control Test</td><td>0.196</td><td>11:00</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:00</td></tr> <tr><td>Control Test</td><td>0.196</td><td>11:01</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:02</td></tr> <tr><td>Control Test</td><td>0.195</td><td>11:02</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:03</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1957</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2951</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	10:59	Control Test	0.196	11:00	Air Blank	0.000	11:00	Control Test	0.196	11:01	Air Blank	0.000	11:02	Control Test	0.195	11:02	Air Blank	0.000	11:03	Control Test Stats			Average	0.1957		Std Dev	0.0006		Rel Std Dev(%)	0.2951		FFWCC Intoxilyzer - Alcohol Analyzer Model 8000 06/30/2025 Software: 8100.27 SN 80-007168 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:05</td></tr> <tr><td>Control Test</td><td>0.077</td><td>11:05</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:06</td></tr> <tr><td>Control Test</td><td>0.077</td><td>11:06</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:07</td></tr> <tr><td>Control Test</td><td>0.078</td><td>11:07</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:08</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0773</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7466</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	11:05	Control Test	0.077	11:05	Air Blank	0.000	11:06	Control Test	0.077	11:06	Air Blank	0.000	11:07	Control Test	0.078	11:07	Air Blank	0.000	11:08	Control Test Stats			Average	0.0773		Std Dev	0.0006		Rel Std Dev(%)	0.7466	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:16																																																																																																																																																	
Control Test	0.049	11:17																																																																																																																																																	
Air Blank	0.000	11:17																																																																																																																																																	
Control Test	0.048	11:18																																																																																																																																																	
Air Blank	0.000	11:18																																																																																																																																																	
Control Test	0.048	11:19																																																																																																																																																	
Air Blank	0.000	11:20																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0483																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1945																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:26																																																																																																																																																	
Control Test	0.079	11:26																																																																																																																																																	
Air Blank	0.000	11:27																																																																																																																																																	
Control Test	0.078	11:28																																																																																																																																																	
Air Blank	0.000	11:28																																																																																																																																																	
Control Test	0.078	11:29																																																																																																																																																	
Air Blank	0.000	11:29																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0783																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7370																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:59																																																																																																																																																	
Control Test	0.196	11:00																																																																																																																																																	
Air Blank	0.000	11:00																																																																																																																																																	
Control Test	0.196	11:01																																																																																																																																																	
Air Blank	0.000	11:02																																																																																																																																																	
Control Test	0.195	11:02																																																																																																																																																	
Air Blank	0.000	11:03																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1957																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2951																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:05																																																																																																																																																	
Control Test	0.077	11:05																																																																																																																																																	
Air Blank	0.000	11:06																																																																																																																																																	
Control Test	0.077	11:06																																																																																																																																																	
Air Blank	0.000	11:07																																																																																																																																																	
Control Test	0.078	11:07																																																																																																																																																	
Air Blank	0.000	11:08																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0773																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7466																																																																																																																																																		
Operator's Signature <i>ML</i>	Operator's Signature <i>ML</i>	Operator's Signature <i>ML</i>	Operator's Signature <i>ML</i>																																																																																																																																																

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FFWCC

Serial Number: 80-007168

Time of Inspection: 13:38

Date of Inspection: 06/30/2025

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test#			Diagnostic Check		
Interferent Detect	Yes		(Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202406K Exp: 06/19/2026	0.08g/210L Test (g/210L) Lot#:202406L Exp: 06/19/2026	0.20g/210L Test (g/210L) Lot#:202406N Exp: 06/20/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG429602 Exp: 10/22/2026
0.000	0.048	0.079	0.196	0.078
0.000	0.048	0.078	0.196	0.078
0.000	0.048	0.078	0.195	0.078
0.000	0.048	0.078	0.195	0.078
0.000	0.048	0.078	0.195	0.078
0.000	0.048	0.078	0.195	0.078
0.000	0.049	0.078	0.195	0.078
0.000	0.048	0.079	0.195	0.078
0.000	0.048	0.078	0.195	0.077
0.000	0.049	0.079	0.195	0.077


Standard Deviations	0.0004	0.0004	0.0004	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.


I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

06/30/2025
Date



Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

Calibration Certificate

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007168 , manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-007168</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>FFWCC</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>06/30/2025</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>13:38</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Taylor
Gutschow
Digitally signed by Taylor
Gutschow
Date: 2025.07.01 10:41:38
-04'00'

06/30/2025

Date

TAYLOR D GUTSCHOW,
Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality